

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cobalt	1920	2000	96	90 - 110	P	10/31/2025	11:34	LB137727

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cobalt	30.1	30.0	100	80 - 120	P	10/31/2025	11:45	LB137727

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Yannuzzi Group, Inc.

SDG No.: Q3391

Contract: YANN01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cobalt	2470	2500	99	90 - 110	P	10/31/2025	12:15	LB137727
CCV02	Cobalt	2500	2500	100	90 - 110	P	10/31/2025	13:17	LB137727
CCV03	Cobalt	2440	2500	98	90 - 110	P	10/31/2025	14:09	LB137727
CCV04	Cobalt	2370	2500	95	90 - 110	P	10/31/2025	15:08	LB137727
CCV05	Cobalt	2470	2500	99	90 - 110	P	10/31/2025	15:58	LB137727
CCV06	Cobalt	2480	2500	99	90 - 110	P	10/31/2025	16:48	LB137727
CCV07	Cobalt	2440	2500	98	90 - 110	P	10/31/2025	17:39	LB137727
CCV08	Cobalt	2400	2500	96	90 - 110	P	10/31/2025	18:28	LB137727
CCV09	Cobalt	2410	2500	96	90 - 110	P	10/31/2025	18:45	LB137727