

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3393

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137593

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
TOC		mg/L	968	1000	97	90-110	07/22/2025
Sample ID:	CCV1						
TOC		mg/L	984	1000	98	90-110	10/21/2025
Sample ID:	CCV2						
TOC		mg/L	1070	1000	107	90-110	10/21/2025
Sample ID:	CCV3						
TOC		mg/L	1080	1000	108	90-110	10/21/2025
Sample ID:	CCV4						
TOC		mg/L	1070	1000	107	90-110	10/22/2025
Sample ID:	CCV5						
TOC		mg/L	985	1000	98	90-110	10/22/2025

Initial and Continuing Calibration Verification

Client: AECOM

SDG No.: Q3393

Project: AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206

RunNo.: LB137625

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.489	0.5	98	90-110	10/23/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.494	0.5	99	90-110	10/23/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.489	0.5	98	90-110	10/23/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.494	0.5	99	90-110	10/23/2025
Sample ID: CCV4 Hexavalent Chromium	mg/L	0.495	0.5	99	90-110	10/23/2025