

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3393

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 17:50

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.29       | 5.29      | 5.19      | 5.39 | 0.00       |
| Aroclor-1254-2 (2)   | 5.51       | 5.51      | 5.41      | 5.61 | 0.00       |
| Aroclor-1254-3 (3)   | 5.85       | 5.86      | 5.76      | 5.96 | 0.01       |
| Aroclor-1254-4 (4)   | 6.14       | 6.14      | 6.04      | 6.24 | 0.00       |
| Aroclor-1254-5 (5)   | 6.54       | 6.54      | 6.44      | 6.64 | 0.00       |
| Tetrachloro-m-xylene | 3.40       | 3.40      | 3.30      | 3.50 | 0.00       |
| Decachlorobiphenyl   | 8.50       | 8.50      | 8.40      | 8.60 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3393

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/10/2025

10/10/2025

Continuing Calib Time: 17:50

Initial Calibration Time(s): 08:57

16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 4.53       | 4.53      | 4.43      | 4.63 | 0.00       |
| Aroclor-1254-2 (2)   | 4.68       | 4.68      | 4.58      | 4.78 | 0.00       |
| Aroclor-1254-3 (3)   | 5.06       | 5.06      | 4.96      | 5.16 | 0.00       |
| Aroclor-1254-4 (4)   | 5.28       | 5.29      | 5.19      | 5.39 | 0.01       |
| Aroclor-1254-5 (5)   | 5.69       | 5.69      | 5.59      | 5.79 | 0.00       |
| Tetrachloro-m-xylene | 2.78       | 2.78      | 2.68      | 2.88 | 0.00       |
| Decachlorobiphenyl   | 7.53       | 7.53      | 7.43      | 7.63 | 0.00       |