

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3393
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114557.D **Time Analyzed:** 20:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.494	5.392	5.592	522.940	500.000	4.6
Aroclor-1016-2	5.515	5.414	5.614	499.430	500.000	-0.1
Aroclor-1016-3	5.577	5.476	5.676	499.400	500.000	-0.1
Aroclor-1016-4	5.673	5.571	5.771	532.610	500.000	6.5
Aroclor-1016-5	5.964	5.863	6.063	514.000	500.000	2.8
Aroclor-1260-1	7.078	6.975	7.175	506.160	500.000	1.2
Aroclor-1260-2	7.332	7.229	7.429	500.870	500.000	0.2
Aroclor-1260-3	7.689	7.585	7.785	452.440	500.000	-9.5
Aroclor-1260-4	7.910	7.807	8.007	497.590	500.000	-0.5
Aroclor-1260-5	8.217	8.113	8.313	486.950	500.000	-2.6
Decachlorobiphenyl	9.942	9.834	10.034	49.450	50.000	-1.1
Tetrachloro-m-xylene	4.355	4.255	4.455	53.780	50.000	7.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3393
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114557.D **Time Analyzed:** 20:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.737	4.643	4.843	519.520	500.000	3.9
Aroclor-1016-2	4.757	4.662	4.862	512.130	500.000	2.4
Aroclor-1016-3	4.931	4.836	5.036	542.990	500.000	8.6
Aroclor-1016-4	4.973	4.879	5.079	544.110	500.000	8.8
Aroclor-1016-5	5.186	5.092	5.292	530.130	500.000	6.0
Aroclor-1260-1	6.218	6.125	6.325	558.310	500.000	11.7
Aroclor-1260-2	6.405	6.313	6.513	595.730	500.000	19.1
Aroclor-1260-3	6.559	6.466	6.666	530.560	500.000	6.1
Aroclor-1260-4	7.030	6.939	7.139	584.700	500.000	16.9
Aroclor-1260-5	7.271	7.179	7.379	559.000	500.000	11.8
Decachlorobiphenyl	8.658	8.571	8.771	48.330	50.000	-3.3
Tetrachloro-m-xylene	3.657	3.560	3.760	51.890	50.000	3.8