

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3393
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114565.D **Time Analyzed:** 23:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.493	5.392	5.592	550.940	500.000	10.2
Aroclor-1016-2	5.515	5.414	5.614	525.580	500.000	5.1
Aroclor-1016-3	5.577	5.476	5.676	527.690	500.000	5.5
Aroclor-1016-4	5.672	5.571	5.771	557.630	500.000	11.5
Aroclor-1016-5	5.964	5.863	6.063	543.990	500.000	8.8
Aroclor-1260-1	7.077	6.975	7.175	536.150	500.000	7.2
Aroclor-1260-2	7.331	7.229	7.429	525.470	500.000	5.1
Aroclor-1260-3	7.689	7.585	7.785	465.140	500.000	-7.0
Aroclor-1260-4	7.910	7.807	8.007	515.720	500.000	3.1
Aroclor-1260-5	8.217	8.113	8.313	502.000	500.000	0.4
Decachlorobiphenyl	9.941	9.834	10.034	50.980	50.000	2.0
Tetrachloro-m-xylene	4.354	4.255	4.455	55.600	50.000	11.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3393
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114565.D **Time Analyzed:** 23:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.737	4.643	4.843	529.290	500.000	5.9
Aroclor-1016-2	4.756	4.662	4.862	519.570	500.000	3.9
Aroclor-1016-3	4.931	4.836	5.036	555.390	500.000	11.1
Aroclor-1016-4	4.973	4.879	5.079	551.540	500.000	10.3
Aroclor-1016-5	5.185	5.092	5.292	556.390	500.000	11.3
Aroclor-1260-1	6.216	6.125	6.325	585.790	500.000	17.2
Aroclor-1260-2	6.404	6.313	6.513	627.550	500.000	25.5
Aroclor-1260-3	6.558	6.466	6.666	554.840	500.000	11.0
Aroclor-1260-4	7.029	6.939	7.139	604.360	500.000	20.9
Aroclor-1260-5	7.269	7.179	7.379	583.300	500.000	16.7
Decachlorobiphenyl	8.658	8.571	8.771	50.050	50.000	0.1
Tetrachloro-m-xylene	3.656	3.560	3.760	52.850	50.000	5.7