

Surrogate Summary

SDG No.: Q3393

Client: AECOM

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO114312.D	PIBLK-PO114312.D	Tetrachloro-m-xyl	1	20	18.6	93		60	140
		Decachlorobiphen	1	20	19.2	96		60	140
		Tetrachloro-m-xyl	2	20	18.9	94		60	140
		Decachlorobiphen	2	20	19.0	95		60	140
I.BLK-PO114546.D	PIBLK-PO114546.D	Tetrachloro-m-xyl	1	20	21.7	109		60	140
		Decachlorobiphen	1	20	21.3	106		60	140
		Tetrachloro-m-xyl	2	20	21.9	110		60	140
		Decachlorobiphen	2	20	21.5	107		60	140
Q3393-01	PR132-S10-000022-20251017	Tetrachloro-m-xyl	1	20	17.3	86		32	144
		Decachlorobiphen	1	20	16.2	81		32	175
		Tetrachloro-m-xyl	2	20	25.3	127		32	144
		Decachlorobiphen	2	20	16.4	82		32	175
Q3393-02	PR132-S10-022047-20251017	Tetrachloro-m-xyl	1	20	23.0	115		32	144
		Decachlorobiphen	1	20	20.0	100		32	175
		Tetrachloro-m-xyl	2	20	23.2	116		32	144
		Decachlorobiphen	2	20	20.1	101		32	175
I.BLK-PO114561.D	PIBLK-PO114561.D	Tetrachloro-m-xyl	1	20	22.3	112		60	140
		Decachlorobiphen	1	20	21.7	108		60	140
		Tetrachloro-m-xyl	2	20	22.1	110		60	140
		Decachlorobiphen	2	20	21.9	110		60	140
Q3393-03	PR132-S08-013038-20251017	Tetrachloro-m-xyl	1	20	18.3	92		32	144
		Decachlorobiphen	1	20	16.0	80		32	175
		Tetrachloro-m-xyl	2	20	22.9	114		32	144
		Decachlorobiphen	2	20	15.9	80		32	175
Q3393-04	PR132-S06-000021-20251017	Tetrachloro-m-xyl	1	20	25.5	128		32	144
		Decachlorobiphen	1	20	20.9	104		32	175
		Tetrachloro-m-xyl	2	20	27.4	137		32	144
		Decachlorobiphen	2	20	21.0	105		32	175
Q3393-05	PR132-S08-000013-20251017	Tetrachloro-m-xyl	1	20	20.9	105		32	144
		Decachlorobiphen	1	20	20.2	101		32	175
		Tetrachloro-m-xyl	2	20	30.2	151	*	32	144
		Decachlorobiphen	2	20	20.1	100		32	175
I.BLK-PO114569.D	PIBLK-PO114569.D	Tetrachloro-m-xyl	1	20	23.4	117		60	140
		Decachlorobiphen	1	20	22.4	112		60	140
		Tetrachloro-m-xyl	2	20	22.6	113		60	140
		Decachlorobiphen	2	20	22.3	112		60	140
I.BLK-PQ070972.D	PIBLK-PQ070972.D	Tetrachloro-m-xyl	1	20	18.3	91		60	140
		Decachlorobiphen	1	20	19.0	95		60	140
		Tetrachloro-m-xyl	2	20	19.1	96		60	140
		Decachlorobiphen	2	20	19.4	97		60	140
I.BLK-PQ071033.D	PIBLK-PQ071033.D	Tetrachloro-m-xyl	1	20	21.8	109		60	140

Surrogate Summary

SDG No.: Q3393

Client: AECOM

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PQ071033.D	PIBLK-PQ071033.D	Decachlorobiphen	1	20	24.8	124		60	140
		Tetrachloro-m-xyl	2	20	20.7	104		60	140
PB170178BL	PB170178BL	Decachlorobiphen	2	20	24.5	123		60	140
		Tetrachloro-m-xyl	1	20	26.1	130		32	144
		Decachlorobiphen	1	20	29.7	148		32	175
		Tetrachloro-m-xyl	2	20	23.8	119		32	144
		Decachlorobiphen	2	20	28.1	141		32	175
Q3402-01MS	41-LAMS	Tetrachloro-m-xyl	1	20	23.0	115		32	144
		Decachlorobiphen	1	20	14.8	74		32	175
		Tetrachloro-m-xyl	2	20	20.9	104		32	144
		Decachlorobiphen	2	20	14.0	70		32	175
Q3402-01MSD	41-LAMSD	Tetrachloro-m-xyl	1	20	23.0	115		32	144
		Decachlorobiphen	1	20	14.4	72		32	175
		Tetrachloro-m-xyl	2	20	20.5	103		32	144
		Decachlorobiphen	2	20	13.7	68		32	175
I.BLK-PQ071048.D	PIBLK-PQ071048.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	22.3	112		60	140
		Tetrachloro-m-xyl	2	20	18.9	95		60	140
		Decachlorobiphen	2	20	22.5	113		60	140
I.BLK-PQ071138.D	PIBLK-PQ071138.D	Tetrachloro-m-xyl	1	20	21.4	107		60	140
		Decachlorobiphen	1	20	24.7	124		60	140
		Tetrachloro-m-xyl	2	20	19.5	98		60	140
		Decachlorobiphen	2	20	23.6	118		60	140
PB170178BS	PB170178BS	Tetrachloro-m-xyl	1	20	22.4	112		32	144
		Decachlorobiphen	1	20	25.7	129		32	175
		Tetrachloro-m-xyl	2	20	19.0	95		32	144
		Decachlorobiphen	2	20	24.3	122		32	175
I.BLK-PQ071147.D	PIBLK-PQ071147.D	Tetrachloro-m-xyl	1	20	20.9	105		60	140
		Decachlorobiphen	1	20	24.7	123		60	140
		Tetrachloro-m-xyl	2	20	19.4	97		60	140
		Decachlorobiphen	2	20	23.3	117		60	140