

Surrogate Summary

SDG No.: **Q3393**

Client: **AECOM**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		57	171
		Tetrachloro-m-xyl	1	20	19.4	97		61	148
		Decachlorobiphen	2	20	20.3	102		57	171
		Tetrachloro-m-xyl	2	20	20.0	100		61	148
I.BLK-PD090722.D	PIBLK-PD090722.D	Decachlorobiphen	1	20	21.2	106		57	171
		Tetrachloro-m-xyl	1	20	21.3	106		61	148
		Decachlorobiphen	2	20	22.1	111		57	171
		Tetrachloro-m-xyl	2	20	22.1	110		61	148
PB170179BL	PB170179BL	Decachlorobiphen	1	20	21.6	108		20	144
		Tetrachloro-m-xyl	1	20	20.3	101		19	148
		Decachlorobiphen	2	20	23.1	115		20	144
		Tetrachloro-m-xyl	2	20	21.2	106		19	148
PB170179BS	PB170179BS	Decachlorobiphen	1	20	20.3	101		20	144
		Tetrachloro-m-xyl	1	20	16.8	84		19	148
		Decachlorobiphen	2	20	21.9	110		20	144
		Tetrachloro-m-xyl	2	20	17.1	86		19	148
Q3402-07MS	41-LBMS	Decachlorobiphen	1	20	7.89	39		20	144
		Tetrachloro-m-xyl	1	20	13.4	67		19	148
		Decachlorobiphen	2	20	8.23	41		20	144
		Tetrachloro-m-xyl	2	20	14.0	70		19	148
Q3402-07MSD	41-LBMSD	Decachlorobiphen	1	20	8.21	41		20	144
		Tetrachloro-m-xyl	1	20	13.1	65		19	148
		Decachlorobiphen	2	20	8.28	41		20	144
		Tetrachloro-m-xyl	2	20	14.0	70		19	148
I.BLK-PD090732.D	PIBLK-PD090732.D	Decachlorobiphen	1	20	20.8	104		57	171
		Tetrachloro-m-xyl	1	20	20.4	102		61	148
		Decachlorobiphen	2	20	22.7	114		57	171
		Tetrachloro-m-xyl	2	20	21.1	106		61	148
Q3393-01	PR132-S10-000022-20251017	Decachlorobiphen	1	20	13.5	68		20	144
		Tetrachloro-m-xyl	1	20	14.4	72		19	148
		Decachlorobiphen	2	20	13.1	66		20	144
		Tetrachloro-m-xyl	2	20	15.8	79		19	148
Q3393-02	PR132-S10-022047-20251017	Decachlorobiphen	1	20	16.2	81		20	144
		Tetrachloro-m-xyl	1	20	16.2	81		19	148
		Decachlorobiphen	2	20	17.4	87		20	144
		Tetrachloro-m-xyl	2	20	17.2	86		19	148
I.BLK-PD090744.D	PIBLK-PD090744.D	Decachlorobiphen	1	20	21.0	105		57	171
		Tetrachloro-m-xyl	1	20	20.1	101		61	148
		Decachlorobiphen	2	20	22.9	115		57	171
		Tetrachloro-m-xyl	2	20	20.8	104		61	148
Q3393-03	PR132-S08-013038-20251017	Decachlorobiphen	1	20	13.8	69		20	144

Surrogate Summary

SDG No.: Q3393

Client: AECOM

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3393-03	PR132-S08-013038-20251017	Tetrachloro-m-xyl	1	20	14.0	70		19	148
		Decachlorobiphen	2	20	13.0	65		20	144
		Tetrachloro-m-xyl	2	20	15.8	79		19	148
Q3393-04	PR132-S06-000021-20251017	Decachlorobiphen	1	20	17.4	87		20	144
		Tetrachloro-m-xyl	1	20	17.5	87		19	148
		Decachlorobiphen	2	20	18.0	90		20	144
		Tetrachloro-m-xyl	2	20	19.2	96		19	148
Q3393-05	PR132-S08-000013-20251017	Decachlorobiphen	1	20	16.4	82		20	144
		Tetrachloro-m-xyl	1	20	15.4	77		19	148
		Decachlorobiphen	2	20	17.2	86		20	144
		Tetrachloro-m-xyl	2	20	17.5	87		19	148
I.BLK-PD090750.D	PIBLK-PD090750.D	Decachlorobiphen	1	20	22.0	110		57	171
		Tetrachloro-m-xyl	1	20	20.7	103		61	148
		Decachlorobiphen	2	20	24.4	122		57	171
		Tetrachloro-m-xyl	2	20	21.5	108		61	148