

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF100625

Review By	Rahul	Review On	10/6/2025 4:17:39 PM
Supervise By	Jagrut	Supervise On	10/7/2025 11:32:28 AM
SubDirectory	BF100625	HP Acquire Method	BNA_F
		HP Processing Method	bf100625

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13177,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143874.D	06 Oct 2025 09:30		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF143875.D	06 Oct 2025 09:59		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF143876.D	06 Oct 2025 10:28		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF143877.D	06 Oct 2025 10:57		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF143878.D	06 Oct 2025 11:56		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BF143879.D	06 Oct 2025 12:54	compound #54 kept on LR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF143880.D	06 Oct 2025 13:23		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF143881.D	06 Oct 2025 13:53		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF143882.D	06 Oct 2025 14:22	Compound #45 & #77 removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBF100625	BF143883.D	06 Oct 2025 16:08		RC/JU	Ok,M
11	PB169953BL	PB169953BL	BF143884.D	06 Oct 2025 16:38	Surrogate Fail	RC/JU	Not Ok
12	PB169953BS	PB169953BS	BF143885.D	06 Oct 2025 17:07	Surrogate Fail	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF102125

Review By	rahul	Review On	10/22/2025 2:35:54 PM
Supervise By	Jagrut	Supervise On	10/22/2025 2:50:46 PM
SubDirectory	BF102125	HP Acquire Method	BNA_F
		HP Processing Method	BF100625

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13178,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144008.D	21 Oct 2025 12:43		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF144009.D	21 Oct 2025 13:29		RC/JU	Ok
3	PB170166BL	PB170166BL	BF144010.D	21 Oct 2025 13:58		RC/JU	Ok
4	PB170166BS	PB170166BS	BF144011.D	21 Oct 2025 14:26		RC/JU	Ok,M
5	PB170166BSD	PB170166BSD	BF144012.D	21 Oct 2025 14:59		RC/JU	Ok,M
6	Q3386-03	3443	BF144013.D	21 Oct 2025 15:33	Internal standard fail	RC/JU	Ok,M
7	Q3386-04	3444	BF144014.D	21 Oct 2025 16:03		RC/JU	Ok,M
8	Q3375-01	PR132-S02-074099-20	BF144015.D	21 Oct 2025 16:32		RC/JU	Ok
9	Q3375-02	PR132-S04-000013-20	BF144016.D	21 Oct 2025 17:02		RC/JU	Ok
10	Q3375-03	PR132-S05-000057-20	BF144017.D	21 Oct 2025 17:31		RC/JU	Ok
11	Q3375-04	PR132-S02-010074-20	BF144018.D	21 Oct 2025 18:01		RC/JU	Ok
12	Q3375-05	PR132-DP02-20251016	BF144019.D	21 Oct 2025 18:30		RC/JU	Ok
13	Q3402-07	41-LB	BF144020.D	21 Oct 2025 18:59		RC/JU	Ok
14	Q3402-07MS	41-LBMS	BF144021.D	21 Oct 2025 19:28		RC/JU	Ok
15	Q3402-07MSD	41-LBMSD	BF144022.D	21 Oct 2025 19:58		RC/JU	Ok
16	Q3402-10	41-RB	BF144023.D	21 Oct 2025 20:27		RC/JU	Ok,M
17	Q3403-01	MH-18	BF144024.D	21 Oct 2025 20:56		RC/JU	Ok,M
18	Q3397-01	TP-7-WC	BF144025.D	21 Oct 2025 21:25		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF102125

Review By	rahul	Review On	10/22/2025 2:35:54 PM		
Supervise By	Jagrut	Supervise On	10/22/2025 2:50:46 PM		
SubDirectory	BF102125	HP Acquire Method	BNA_F	HP Processing Method	BF100625
STD. NAME	STD REF.#				
Tune/Reschk	SP6875				
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864				
CCC	SP6860				
Internal Standard/PEM	S13178,10ul/1000ul sample				
ICV/I.BLK	SP6872				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q3402-04	41-RA	BF144026.D	21 Oct 2025 21:55		RC/JU	Ok
20	Q3402-01	41-LA	BF144027.D	21 Oct 2025 22:24		RC/JU	Ok,M
21	Q3376-01	PR132-S05-057082-20	BF144028.D	21 Oct 2025 22:53		RC/JU	Ok,M
22	Q3376-02	PR132-DP01-20251016	BF144029.D	21 Oct 2025 23:22		RC/JU	Ok,M
23	Q3393-02	PR132-S10-022047-20	BF144030.D	21 Oct 2025 23:51		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF102225

Review By	rahul	Review On	10/28/2025 9:35:43 AM
Supervise By	mohammad	Supervise On	10/28/2025 9:38:00 AM
SubDirectory	BF102225	HP Acquire Method	BNA_F
		HP Processing Method	BF100625

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13179,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF144031.D	22 Oct 2025 09:26		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF144032.D	22 Oct 2025 10:25	CCC fail high side for com. #92	RC/JU	Ok
3	PB170180BL	PB170180BL	BF144033.D	22 Oct 2025 10:54		RC/JU	Ok
4	PB170180BS	PB170180BS	BF144034.D	22 Oct 2025 11:23		RC/JU	Ok,M
5	Q3393-05	PR132-S08-000013-20	BF144035.D	22 Oct 2025 11:57		RC/JU	Ok
6	Q3393-01	PR132-S10-000022-20	BF144036.D	22 Oct 2025 12:26		RC/JU	Ok
7	Q3393-04	PR132-S06-000021-20	BF144037.D	22 Oct 2025 12:55		RC/JU	Ok
8	Q3393-03	PR132-S08-013038-20	BF144038.D	22 Oct 2025 13:24		RC/JU	Ok
9	Q3376-03	PR132-COMP2-202510	BF144039.D	22 Oct 2025 13:53		RC/JU	Ok
10	Q3376-04	PR132-S12-047072-20	BF144040.D	22 Oct 2025 14:22		RC/JU	Ok
11	Q3376-05	PR132-S12-020047-20	BF144041.D	22 Oct 2025 14:51	Internal standard fail	RC/JU	ReRun
12	Q3392-08RE	MH-19RE	BF144042.D	22 Oct 2025 15:20	Internal Standard Fail	RC/JU	Confirms
13	Q3399-01	COMP-ROLLOFF	BF144043.D	22 Oct 2025 15:49	Internal Standard Fail- confirmed in MS & MSD.	RC/JU	Ok
14	Q3404-01	MOO-25-0285-0288	BF144044.D	22 Oct 2025 16:18	Internal standard fail	RC/JU	ReRun
15	Q3399-01MS	COMP-ROLLOFFMS	BF144045.D	22 Oct 2025 16:48	Internal standard fail	RC/JU	Ok
16	Q3399-01MSD	COMP-ROLLOFFMSD	BF144046.D	22 Oct 2025 17:17	Internal standard fail	RC/JU	Ok
17	Q3416-01	NB-08-10212025	BF144047.D	22 Oct 2025 17:46	Internal standard fail	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF102225

Review By	rahul	Review On	10/28/2025 9:35:43 AM
Supervise By	mohammad	Supervise On	10/28/2025 9:38:00 AM
SubDirectory	BF102225	HP Acquire Method	BNA_F
		HP Processing Method	BF100625

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13179,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q3403-04	MH-18	BF144048.D	22 Oct 2025 18:15	Internal standard fail	RC/JU	Ok
19	Q3412-04	TP-6	BF144049.D	22 Oct 2025 18:45	Internal Standard Fail	RC/JU	ReRun
20	Q3404-04	MOO-25-0292	BF144050.D	22 Oct 2025 19:14	Internal Standard Fail	RC/JU	Ok

M : Manual Integration