

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3393

Client ID : SSTDCCC040

Date Analyzed: 10/21/2025

Lab File ID: BF144009.D

Time Analyzed: 13:29

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	126907	6.881	483648	8.16	263342	9.92
UPPER LIMIT	253814	7.381	967296	8.663	526684	10.422
LOWER LIMIT	63453.5	6.381	241824	7.663	131671	9.422
EPA SAMPLE NO.						
01 PR132-S10-022047-20251017	96300	6.88	359740	8.16	172359	9.92
02 41-LBMS	97883	6.88	362327	8.16	171519	9.92
03 41-LBMSD	105784	6.88	388491	8.16	194134	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.