

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3393

Lab File ID: BF144008.D

DFTPP Injection Date: 10/21/2025

Instrument ID: BNA_F

DFTPP Injection Time: 12:43

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.4 (1.6) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	85.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.7 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF144009.D	10/21/2025	13:29
41-LBMS	Q3402-07MS	BF144021.D	10/21/2025	19:28
41-LBMSD	Q3402-07MSD	BF144022.D	10/21/2025	19:58
PR132-S10-022047-20251017	Q3393-02	BF144030.D	10/21/2025	23:51