

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102025\
 Data File : VY023495.D
 Acq On : 20 Oct 2025 09:10
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 21 03:53:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.360	0.313	13.1	90	0.00
3 P	Chloromethane	0.490	0.411	16.1	85	0.00
4 C	Vinyl Chloride	0.639	0.566	11.4#	87	0.00
5 T	Bromomethane	0.536	0.473	11.8	93	0.00
6 T	Chloroethane	0.435	0.409	6.0	91	0.00
7 T	Trichlorofluoromethane	0.887	0.849	4.3	93	0.00
8 T	Diethyl Ether	0.232	0.218	6.0	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.460	0.0	97	0.00
10 T	Methyl Iodide	0.512	0.568	-10.9	97	0.00
11 T	Tert butyl alcohol	0.031	0.030	3.2	110	0.00
12 CM	1,1-Dichloroethene	0.446	0.432	3.1#	93	0.00
13 T	Acrolein	0.044	0.015	65.9#	34#	0.01
14 T	Allyl chloride	0.544	0.496	8.8	87	0.00
15 T	Acrylonitrile	0.089	0.083	6.7	94	0.00
16 T	Acetone	0.100	0.098	2.0	108	0.00
17 T	Carbon Disulfide	1.277	1.195	6.4	90	0.00
18 T	Methyl Acetate	0.261	0.266	-1.9	100	0.00
19 T	Methyl tert-butyl Ether	1.072	1.041	2.9	94	0.00
20 T	Methylene Chloride	0.538	0.486	9.7	98	0.00
21 T	trans-1,2-Dichloroethene	0.491	0.495	-0.8	97	0.00
22 T	Diisopropyl ether	1.202	1.160	3.5	92	0.00
23 T	Vinyl Acetate	0.651	0.592	9.1	88	0.00
24 P	1,1-Dichloroethane	0.796	0.781	1.9	95	0.00
25 T	2-Butanone	0.122	0.109	10.7	96	0.00
26 T	2,2-Dichloropropane	0.731	0.745	-1.9	98	0.00
27 T	cis-1,2-Dichloroethene	0.567	0.587	-3.5	100	0.00
28 T	Bromochloromethane	0.297	0.251	15.5	87	0.00
29 T	Tetrahydrofuran	0.069	0.061	11.6	88	0.00
30 C	Chloroform	0.875	0.897	-2.5#	99	0.00
31 T	Cyclohexane	0.696	0.625	10.2	90	0.00
32 T	1,1,1-Trichloroethane	0.795	0.812	-2.1	98	0.00
33 S	1,2-Dichloroethane-d4	0.418	0.377	9.8	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.307	0.316	-2.9	93	0.00
36 T	1,1-Dichloropropene	0.407	0.436	-7.1	96	0.00
37 T	Ethyl Acetate	0.161	0.156	3.1	91	0.00
38 T	Carbon Tetrachloride	0.493	0.541	-9.7	98	0.00
39 T	Methylcyclohexane	0.530	0.553	-4.3	92	0.00
40 TM	Benzene	1.288	1.374	-6.7	95	0.00
41 T	Methacrylonitrile	0.087	0.086	1.1	94	0.00
42 TM	1,2-Dichloroethane	0.332	0.341	-2.7	95	0.00
43 T	Isopropyl Acetate	0.321	0.303	5.6	88	0.00
44 TM	Trichloroethene	0.377	0.407	-8.0	97	0.00
45 C	1,2-Dichloropropane	0.285	0.306	-7.4#	98	0.00
46 T	Dibromomethane	0.180	0.192	-6.7	97	0.00
47 T	Bromodichloromethane	0.453	0.495	-9.3	99	0.00
48 T	Methyl methacrylate	0.151	0.145	4.0	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102025\
 Data File : VY023495.D
 Acq On : 20 Oct 2025 09:10
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 21 03:53:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	93	0.00
50 S	Toluene-d8	1.144	1.163	-1.7	90	0.00
51 T	4-Methyl-2-Pentanone	0.166	0.166	0.0	91	0.00
52 CM	Toluene	0.836	0.928	-11.0#	98	0.00
53 T	t-1,3-Dichloropropene	0.394	0.415	-5.3	94	0.00
54 T	cis-1,3-Dichloropropene	0.465	0.498	-7.1	96	0.00
55 T	1,1,2-Trichloroethane	0.242	0.265	-9.5	100	0.00
56 T	Ethyl methacrylate	0.279	0.297	-6.5	93	0.00
57 T	1,3-Dichloropropane	0.383	0.409	-6.8	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.127	-1.6	86	0.00
59 T	2-Hexanone	0.119	0.120	-0.8	94	0.00
60 T	Dibromochloromethane	0.339	0.380	-12.1	102	0.00
61 T	1,2-Dibromoethane	0.231	0.247	-6.9	97	0.00
62 S	4-Bromofluorobenzene	0.378	0.379	-0.3	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.516	0.560	-8.5	96	0.00
65 PM	Chlorobenzene	1.087	1.159	-6.6	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.424	-9.3	102	0.00
67 C	Ethyl Benzene	1.760	1.910	-8.5#	96	0.00
68 T	m/p-Xylenes	0.706	0.781	-10.6	97	0.00
69 T	o-Xylene	0.662	0.731	-10.4	98	0.00
70 T	Styrene	1.089	1.231	-13.0	99	0.00
71 P	Bromoform	0.240	0.262	-9.2	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.295	3.555	-7.9	97	0.00
74 T	N-amyl acetate	0.592	0.540	8.8	84	0.00
75 P	1,1,2,2-Tetrachloroethane	0.539	0.552	-2.4	100	0.00
76 T	1,2,3-Trichloropropane	0.441	0.428	2.9	90	0.00
77 T	Bromobenzene	0.876	0.921	-5.1	99	0.00
78 T	n-propylbenzene	3.813	4.166	-9.3	97	0.00
79 T	2-Chlorotoluene	2.233	2.362	-5.8	96	0.00
80 T	1,3,5-Trimethylbenzene	2.663	2.929	-10.0	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.180	0.182	-1.1	98	0.00
82 T	4-Chlorotoluene	2.300	2.463	-7.1	98	0.00
83 T	tert-Butylbenzene	2.458	2.672	-8.7	96	0.00
84 T	1,2,4-Trimethylbenzene	2.657	2.910	-9.5	98	0.00
85 T	sec-Butylbenzene	3.527	3.879	-10.0	98	0.00
86 T	p-Isopropyltoluene	3.003	3.312	-10.3	97	0.00
87 T	1,3-Dichlorobenzene	1.706	1.790	-4.9	98	0.00
88 T	1,4-Dichlorobenzene	1.685	1.748	-3.7	98	0.00
89 T	n-Butylbenzene	2.616	2.831	-8.2	96	0.00
90 T	Hexachloroethane	0.646	0.688	-6.5	100	0.00
91 T	1,2-Dichlorobenzene	1.496	1.577	-5.4	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.088	0.085	3.4	93	0.00
93 T	1,2,4-Trichlorobenzene	0.932	0.949	-1.8	94	0.00
94 T	Hexachlorobutadiene	0.581	0.615	-5.9	97	0.00
95 T	Naphthalene	1.485	1.429	3.8	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102025\
Data File : VY023495.D
Acq On : 20 Oct 2025 09:10
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 21 03:53:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.808	0.813	-0.6	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6