



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102125\
Data File : VY023520.D
Acq On : 21 Oct 2025 08:29
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Title : SW846 8260
Last Update : Wed Oct 08 05:12:50 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1743

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	31051	PASS
75	95	30	60	48.5	91288	PASS
95	95	100	100	100.0	188203	PASS
96	95	5	9	6.7	12595	PASS
173	174	0.00	2	1.4	2489	PASS
174	95	50	100	94.7	178283	PASS
175	174	5	9	7.1	12726	PASS
176	174	95	101	97.7	174101	PASS
177	176	5	9	6.7	11711	PASS