

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3393
 Lab File ID: VY023521.D Date Analyzed: 10/21/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:05
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	499961	7.71	698196	8.62	642915	11.42
UPPER LIMIT	999922	8.213	1396390	9.116	1285830	11.92
LOWER LIMIT	249981	7.213	349098	8.116	321458	10.92
EPA SAMPLE NO.						
PR132-S08-013038-20251017	730515	7.71	1210975	8.62	1010042	11.41
VY1021SBL01	767531	7.71	1271980	8.62	1000258	11.42
VY1021SBS01	494016	7.71	712868	8.62	635785	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.