

Laboratory Control Sample Summary

Client:	AECOM	SDG No.:	Q3395
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Run No.:	LB137593

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137593BSS							
TOC	mg/Kg	1000	992		99	1	90-110	10/21/2025

Laboratory Control Sample Summary

Client:	AECOM	SDG No.:	Q3395
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Run No.:	LB137593

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137593BSS2							
TOC	mg/Kg	1000	1020		102	1	90-110	10/22/2025

Laboratory Control Sample Summary

Client:	AECOM	SDG No.:	Q3395
Project:	AE1-CTY 3.2.Z-PR-DOCK-Des-ODC - 60693795-1719206	Run No.:	LB137625

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170212BS							
Hexavalent Chromium	mg/Kg	20	19.7		98	1	84-110	10/23/2025