

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3395

Continuing Calib Date: 10/23/2025

Initial Calibration Date(s): 10/10/2025

10/10/2025

Continuing Calib Time: 10:27

Initial Calibration Time(s): 08:57

16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.29	5.29	5.19	5.39	0.00
Aroclor-1254-2 (2)	5.51	5.51	5.41	5.61	0.00
Aroclor-1254-3 (3)	5.86	5.86	5.76	5.96	0.00
Aroclor-1254-4 (4)	6.14	6.14	6.04	6.24	0.00
Aroclor-1254-5 (5)	6.54	6.54	6.44	6.64	0.00
Tetrachloro-m-xylene	3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl	8.50	8.50	8.40	8.60	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3395

Continuing Calib Date: 10/23/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 10:27

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	4.53	4.53	4.43	4.63	0.00
Aroclor-1254-2	(2)	4.68	4.68	4.58	4.78	0.00
Aroclor-1254-3	(3)	5.06	5.06	4.96	5.16	0.00
Aroclor-1254-4	(4)	5.29	5.29	5.19	5.39	0.01
Aroclor-1254-5	(5)	5.69	5.69	5.59	5.79	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.53	7.53	7.43	7.63	0.00