

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3395

Continuing Calib Date: 10/23/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 14:06

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.48       | 4.48      | 4.38      | 4.58 | 0.00       |
| Aroclor-1248-2 (2)   | 4.74       | 4.74      | 4.64      | 4.84 | 0.00       |
| Aroclor-1248-3 (3)   | 4.93       | 4.93      | 4.83      | 5.03 | 0.00       |
| Aroclor-1248-4 (4)   | 5.32       | 5.32      | 5.22      | 5.42 | 0.00       |
| Aroclor-1248-5 (5)   | 5.35       | 5.35      | 5.25      | 5.45 | 0.00       |
| Tetrachloro-m-xylene | 3.40       | 3.40      | 3.30      | 3.50 | 0.00       |
| Decachlorobiphenyl   | 8.50       | 8.50      | 8.40      | 8.60 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q3395

Continuing Calib Date: 10/23/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 14:06

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 3.78       | 3.78      | 3.68      | 3.88 | 0.00       |
| Aroclor-1248-2 (2)   | 4.00       | 4.00      | 3.90      | 4.10 | 0.00       |
| Aroclor-1248-3 (3)   | 4.04       | 4.04      | 3.94      | 4.14 | 0.00       |
| Aroclor-1248-4 (4)   | 4.20       | 4.20      | 4.10      | 4.30 | 0.00       |
| Aroclor-1248-5 (5)   | 4.57       | 4.57      | 4.47      | 4.67 | 0.00       |
| Tetrachloro-m-xylene | 2.78       | 2.78      | 2.68      | 2.88 | 0.00       |
| Decachlorobiphenyl   | 7.53       | 7.53      | 7.43      | 7.63 | 0.00       |