

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3395
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071124.D **Time Analyzed:** 10:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.480	4.380	4.580	535.180	500.000	7.0
Aroclor-1016-2	4.499	4.399	4.599	521.020	500.000	4.2
Aroclor-1016-3	4.556	4.456	4.656	530.020	500.000	6.0
Aroclor-1016-4	4.646	4.546	4.746	529.540	500.000	5.9
Aroclor-1016-5	4.930	4.831	5.031	526.950	500.000	5.4
Aroclor-1260-1	6.020	5.919	6.119	515.670	500.000	3.1
Aroclor-1260-2	6.276	6.175	6.375	516.240	500.000	3.2
Aroclor-1260-3	6.627	6.526	6.726	541.410	500.000	8.3
Aroclor-1260-4	6.840	6.740	6.940	553.630	500.000	10.7
Aroclor-1260-5	7.147	7.045	7.245	544.110	500.000	8.8
Decachlorobiphenyl	8.497	8.396	8.596	54.590	50.000	9.2
Tetrachloro-m-xylene	3.400	3.300	3.500	52.250	50.000	4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** AECO02
Lab Code: ACE **SDG NO.:** Q3395
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071124.D **Time Analyzed:** 10:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	427.210	500.000	-14.6
Aroclor-1016-2	3.792	3.693	3.893	424.110	500.000	-15.2
Aroclor-1016-3	3.951	3.853	4.053	434.870	500.000	-13.0
Aroclor-1016-4	4.000	3.902	4.102	441.890	500.000	-11.6
Aroclor-1016-5	4.195	4.097	4.297	439.520	500.000	-12.1
Aroclor-1260-1	5.187	5.089	5.289	435.740	500.000	-12.9
Aroclor-1260-2	5.376	5.278	5.478	433.990	500.000	-13.2
Aroclor-1260-3	5.517	5.419	5.619	438.480	500.000	-12.3
Aroclor-1260-4	5.978	5.879	6.079	442.200	500.000	-11.6
Aroclor-1260-5	6.222	6.124	6.324	441.170	500.000	-11.8
Decachlorobiphenyl	7.531	7.433	7.633	51.410	50.000	2.8
Tetrachloro-m-xylene	2.776	2.678	2.878	46.210	50.000	-7.6