

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM
SubDirectory	PQ101025	HP Acquire Method	HP Processing Method PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ070971.D	10 Oct 2025 08:27		YP\AJ	Ok
2	I.BLK	I.BLK	PQ070972.D	10 Oct 2025 08:42		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PQ070973.D	10 Oct 2025 08:57		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PQ070974.D	10 Oct 2025 09:11		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PQ070975.D	10 Oct 2025 09:26		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PQ070976.D	10 Oct 2025 09:40		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PQ070977.D	10 Oct 2025 09:55		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PQ070978.D	10 Oct 2025 10:09		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PQ070979.D	10 Oct 2025 10:24		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PQ070980.D	10 Oct 2025 10:38		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PQ070981.D	10 Oct 2025 10:53		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PQ070982.D	10 Oct 2025 11:08		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PQ070983.D	10 Oct 2025 11:22		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PQ070984.D	10 Oct 2025 11:37		YP\AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PQ070985.D	10 Oct 2025 11:51		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PQ070986.D	10 Oct 2025 12:06		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PQ070987.D	10 Oct 2025 12:20		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PQ070988.D	10 Oct 2025 12:35		YP\AJ	Ok,M

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ101025

Review By	yogesh	Review On	10/10/2025 10:11:25 AM
Supervise By	mohammad	Supervise On	10/14/2025 2:48:24 AM
SubDirectory	PQ101025	HP Acquire Method	HP Processing Method PQ101025
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PQ070989.D	10 Oct 2025 12:50		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PQ070990.D	10 Oct 2025 13:04		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PQ070991.D	10 Oct 2025 13:19		YPIAJ	Ok
22	AR1254ICC500	AR1254ICC500	PQ070992.D	10 Oct 2025 13:33		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PQ070993.D	10 Oct 2025 13:48		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PQ070994.D	10 Oct 2025 15:15		YPIAJ	Ok
25	AR1262ICC500	AR1262ICC500	PQ070995.D	10 Oct 2025 15:30		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PQ070996.D	10 Oct 2025 15:44		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PQ070997.D	10 Oct 2025 15:59		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PQ070998.D	10 Oct 2025 16:13		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PQ070999.D	10 Oct 2025 16:28		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PQ071000.D	10 Oct 2025 16:43		YPIAJ	Ok
31	PQ101025ICV	ICVPQ101025	PQ071001.D	10 Oct 2025 16:57		YPIAJ	Ok,M
32	AR1242ICV500	ICVPQ101025AR1242	PQ071002.D	10 Oct 2025 17:41		YPIAJ	Ok,M
33	AR1248ICV500	ICVPQ101025AR1248	PQ071003.D	10 Oct 2025 18:25		YPIAJ	Ok,M
34	AR1254ICV500	ICVPQ101025AR1254	PQ071004.D	10 Oct 2025 19:08		YPIAJ	Ok
35	AR1268ICV500	ICVPQ101025AR1268	PQ071005.D	10 Oct 2025 19:52		YPIAJ	Ok
36	DDT ANALOG	DDT ANALOG	PQ071006.D	10 Oct 2025 20:36		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ102325

Review By	yogesh	Review On	10/23/2025 10:13:01 AM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:08 AM
SubDirectory	PQ102325	HP Acquire Method	HP Processing Method PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ071085.D	23 Oct 2025 08:45		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PQ071086.D	23 Oct 2025 08:59		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PQ071087.D	23 Oct 2025 09:58		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PQ071088.D	23 Oct 2025 10:13		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PQ071089.D	23 Oct 2025 10:27		YP\AJ	Ok
6	I.BLK	I.BLK	PQ071090.D	23 Oct 2025 10:42		YP\AJ	Ok
7	PB170211BL	PB170211BL	PQ071091.D	23 Oct 2025 10:56		YP\AJ	Ok
8	PB170211BS	PB170211BS	PQ071092.D	23 Oct 2025 11:11	AR1016&1260 recovery fail	YP\AJ	Not Ok
9	PB170211BSD	PB170211BSD	PQ071093.D	23 Oct 2025 11:25	AR1016&1260 recovery fail	YP\AJ	Not Ok
10	Q3398-01	A748	PQ071094.D	23 Oct 2025 11:40		YP\AJ	Ok
11	Q3400-01	FRAC-TANK-0760450	PQ071095.D	23 Oct 2025 11:55		YP\AJ	Ok
12	Q3417-01	101425	PQ071096.D	23 Oct 2025 12:09	All surrogate fail	YP\AJ	ReRun
13	Q3417-01RE	101425RE	PQ071097.D	23 Oct 2025 12:38	All surrogate fail	YP\AJ	Confirms
14	AR1660CCC500	AR1660CCC500	PQ071098.D	23 Oct 2025 13:37		YP\AJ	Ok
15	AR1242CCC500	AR1242CCC500	PQ071099.D	23 Oct 2025 13:51		YP\AJ	Ok,M
16	AR1248CCC500	AR1248CCC500	PQ071100.D	23 Oct 2025 14:06		YP\AJ	Ok
17	AR1254CCC500	AR1254CCC500	PQ071101.D	23 Oct 2025 14:20		YP\AJ	Ok
18	I.BLK	I.BLK	PQ071102.D	23 Oct 2025 14:35		YP\AJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ102325

Review By	yogesh	Review On	10/23/2025 10:13:01 AM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:08 AM
SubDirectory	PQ102325	HP Acquire Method	HP Processing Method PQ101025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PB170227BL	PB170227BL	PQ071103.D	23 Oct 2025 14:50		YP\AJ	Ok
20	PB170227BS	PB170227BS	PQ071104.D	23 Oct 2025 15:04	AR1016 recovery fail	YP\AJ	Not Ok
21	Q3395-01	PR132-S09-000051-20	PQ071105.D	23 Oct 2025 15:19	AR1242+1254 Hit	YP\AJ	Ok
22	Q3395-02	PR132-S07-000102-20	PQ071106.D	23 Oct 2025 15:33	AR1254 Hit	YP\AJ	Ok,M
23	Q3395-03MS	PR132-S07-000102-20	PQ071107.D	23 Oct 2025 15:48		YP\AJ	Ok,M
24	Q3395-04MSD	PR132-S07-000102-20	PQ071108.D	23 Oct 2025 16:03		YP\AJ	Ok,M
25	Q3395-05	PR132-S07-102116-20	PQ071109.D	23 Oct 2025 16:17		YP\AJ	Ok
26	Q3395-06	PR132-COMP03-2025	PQ071110.D	23 Oct 2025 16:32		YP\AJ	Ok
27	Q3405-01	PAD-10-20-25-A	PQ071111.D	23 Oct 2025 16:46	DCB high in 1st column	YP\AJ	Ok
28	Q3405-02	PAD-10-20-25-B	PQ071112.D	23 Oct 2025 17:01	DCB high in 1st column	YP\AJ	Ok
29	AR1660CCC500	AR1660CCC500	PQ071113.D	23 Oct 2025 17:59		YP\AJ	Ok
30	AR1242CCC500	AR1242CCC500	PQ071114.D	23 Oct 2025 19:12		YP\AJ	Ok,M
31	AR1248CCC500	AR1248CCC500	PQ071115.D	23 Oct 2025 19:27		YP\AJ	Ok
32	AR1254CCC500	AR1254CCC500	PQ071116.D	23 Oct 2025 19:41		YP\AJ	Ok
33	I.BLK	I.BLK	PQ071117.D	23 Oct 2025 19:56		YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ102425

Review By	yogesh	Review On	10/24/2025 11:07:53 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:20 AM
SubDirectory	PQ102425	HP Acquire Method	HP Processing Method PQ101025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ071118.D	24 Oct 2025 08:37		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PQ071119.D	24 Oct 2025 08:52		YPIAJ	Ok
3	I.BLK	I.BLK	PQ071120.D	24 Oct 2025 09:06		YPIAJ	Ok
4	PB170211BS	PB170211BS	PQ071121.D	24 Oct 2025 09:34		YPIAJ	Ok,M
5	PB170211BSD	PB170211BSD	PQ071122.D	24 Oct 2025 09:51		YPIAJ	Ok,M
6	PB170227BS	PB170227BS	PQ071123.D	24 Oct 2025 10:07		YPIAJ	Ok
7	AR1660CCC500	AR1660CCC500	PQ071124.D	24 Oct 2025 10:36		YPIAJ	Ok
8	AR1242CCC500	AR1242CCC500	PQ071125.D	24 Oct 2025 10:57		YPIAJ	Ok,M
9	AR1248CCC500	AR1248CCC500	PQ071126.D	24 Oct 2025 11:13		YPIAJ	Ok
10	AR1254CCC500	AR1254CCC500	PQ071127.D	24 Oct 2025 11:28		YPIAJ	Ok
11	I.BLK	I.BLK	PQ071128.D	24 Oct 2025 11:46		YPIAJ	Ok
12	PB170228BS	PB170228BS	PQ071130.D	24 Oct 2025 12:15		YPIAJ	Ok
13	Q3429-01	40309	PQ071131.D	24 Oct 2025 12:29	AR1254 hits, DCB high in 1st column	YPIAJ	Dilution
14	Q3429-02	40310	PQ071132.D	24 Oct 2025 12:44	AR1254 hits	YPIAJ	Dilution
15	Q3430-01	BUR-25-0095	PQ071133.D	24 Oct 2025 12:58	DCB high in both column & TCMX high in 1st column	YPIAJ	ReRun
16	AR1660CCC500	AR1660CCC500	PQ071134.D	24 Oct 2025 13:57		YPIAJ	Ok
17	AR1242CCC500	AR1242CCC500	PQ071135.D	24 Oct 2025 14:12		YPIAJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ102425

Review By	yogesh	Review On	10/24/2025 11:07:53 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:20 AM
SubDirectory	PQ102425	HP Acquire Method	HP Processing Method PQ101025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1248CCC500	AR1248CCC500	PQ071136.D	24 Oct 2025 14:26		YPIAJ	Ok
19	AR1254CCC500	AR1254CCC500	PQ071137.D	24 Oct 2025 14:41		YPIAJ	Ok
20	I.BLK	I.BLK	PQ071138.D	24 Oct 2025 14:55		YPIAJ	Ok
21	Q3430-01RE	BUR-25-0095RE	PQ071139.D	24 Oct 2025 15:10	DCB high in both column & TCMX high in 1st column	YPIAJ	Confirms
22	PB170178BS	PB170178BS	PQ071140.D	24 Oct 2025 15:24		YPIAJ	Ok,M
23	Q3429-01DL	40309DL	PQ071141.D	24 Oct 2025 15:39	AR1254 hits	YPIAJ	Dilution
24	Q3429-01DL2	40309DL2	PQ071142.D	24 Oct 2025 15:54	AR1254 hits	YPIAJ	Ok
25	Q3429-02DL	40310DL	PQ071143.D	24 Oct 2025 16:08	AR1254 hits, DCB High in 2nd column & TCMX high in 1st column	YPIAJ	Dilution
26	Q3429-02DL2	40310DL2	PQ071144.D	24 Oct 2025 16:23	AR1254 hits	YPIAJ	Ok
27	AR1660CCC500	AR1660CCC500	PQ071145.D	24 Oct 2025 17:21		YPIAJ	Ok,M
28	AR1254CCC500	AR1254CCC500	PQ071146.D	24 Oct 2025 17:50		YPIAJ	Ok
29	I.BLK	I.BLK	PQ071147.D	24 Oct 2025 18:05		YPIAJ	Ok

M : Manual Integration