

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q3395
 Lab File ID: VW032378.D Date Analyzed: 10/20/2025
 Instrument ID: MSVOA_W Time Analyzed: 10:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	234766	13.556				
UPPER LIMIT	469532	14.056				
LOWER LIMIT	117383	13.056				
EPA SAMPLE NO.						
PR132-S09-000051-20251017	278056	13.56				
PR132-S07-000102-20251017	270145	13.56				
PR132-S07-000102-20251017MS	54387 *	13.56				
PR132-S07-000102-20251017MSD	56507 *	13.56				
PR132-S07-102116-20251017	242674	13.55				
PR132-S08-038048-20251017	246462	13.56				
VW1020SBL01	188560	13.56				
VW1020SBS01	240746	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.