



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3399

MATRIX: Solid

METHOD: 8015D/3541

| | NA | NO | YES |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----|----|-----|
| 1. Chromatograms Labeled/Compounds Identified. | | | ✓ |
| 2. Standard Summary Submitted. | | | ✓ |
| 3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD. | | | ✓ |
| 4. Blank Contamination - If yes, list compounds and concentrations in each blank: | | ✓ | |
| 5. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. | | | ✓ |
| 6. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The Blank Spike met requirements for all compounds. The Blank Spike Duplicate met requirements for all compounds. The RPD were met for all analysis. | | | ✓ |
| 7. Retention Time Shift Meet Criteria (if applicable) Comments: | | | ✓ |
| 8. Extraction Holding Time Met If not met, list number of days exceeded for each sample: | | ✓ | |
| 9. Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range. The Holding Times were met for all analysis. | | ✓ | |

ADDITIONAL COMMENTS:

The soil samples results are based on a dry weight basis.

QA REVIEW

Date



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES