

Laboratory Control Sample Summary

Client:	ICE Service Group, Inc.	SDG No.:	Q3399
Project:	NWIRP Northrup Grumman Site – Bethpage, NY	Run No.:	LB137601

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170189BS							
Hexavalent Chromium	mg/Kg	20	19.8		99	1	84-110	10/21/2025