

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV85	Mercury	4.26	4.0	106	90 - 110	CV	10/23/2025	13:51	LB137626

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV79	Mercury	4.94	5.0	99	90 - 110	CV	10/23/2025	14:01	LB137626
CCV80	Mercury	4.97	5.0	99	90 - 110	CV	10/23/2025	14:35	LB137626
CCV81	Mercury	4.86	5.0	97	90 - 110	CV	10/23/2025	15:09	LB137626
CCV82	Mercury	5.04	5.0	101	90 - 110	CV	10/23/2025	15:29	LB137626
CCV83	Mercury	4.68	5.0	94	90 - 110	CV	10/23/2025	15:59	LB137626
CCV84	Mercury	4.71	5.0	94	90 - 110	CV	10/23/2025	16:14	LB137626

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7550	8000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Antimony	3880	4000	97	90 - 110	P	10/21/2025	12:40	LB137613
	Arsenic	3800	4000	95	90 - 110	P	10/21/2025	12:40	LB137613
	Barium	7500	8000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Beryllium	194	200	97	90 - 110	P	10/21/2025	12:40	LB137613
	Cadmium	1890	2000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Calcium	18800	20000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Chromium	784	800	98	90 - 110	P	10/21/2025	12:40	LB137613
	Cobalt	1900	2000	95	90 - 110	P	10/21/2025	12:40	LB137613
	Copper	976	1000	98	90 - 110	P	10/21/2025	12:40	LB137613
	Iron	3800	4000	95	90 - 110	P	10/21/2025	12:40	LB137613
	Lead	3700	4000	92	90 - 110	P	10/21/2025	12:40	LB137613
	Magnesium	18800	20000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Manganese	1860	2000	93	90 - 110	P	10/21/2025	12:40	LB137613
	Nickel	1910	2000	95	90 - 110	P	10/21/2025	12:40	LB137613
	Potassium	18900	20000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Selenium	3850	4000	96	90 - 110	P	10/21/2025	12:40	LB137613
	Silver	956	1000	96	90 - 110	P	10/21/2025	12:40	LB137613
	Sodium	18300	20000	92	90 - 110	P	10/21/2025	12:40	LB137613
	Thallium	3840	4000	96	90 - 110	P	10/21/2025	12:40	LB137613
	Vanadium	1880	2000	94	90 - 110	P	10/21/2025	12:40	LB137613
	Zinc	1920	2000	96	90 - 110	P	10/21/2025	12:40	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	110	100	110	80 - 120	P	10/21/2025	12:44	LB137613
	Antimony	52.1	50.0	104	80 - 120	P	10/21/2025	12:44	LB137613
	Arsenic	20.1	20.0	100	80 - 120	P	10/21/2025	12:44	LB137613
	Barium	95.0	100	95	80 - 120	P	10/21/2025	12:44	LB137613
	Beryllium	6.21	6.0	104	80 - 120	P	10/21/2025	12:44	LB137613
	Cadmium	5.68	6.0	95	80 - 120	P	10/21/2025	12:44	LB137613
	Calcium	1990	2000	100	80 - 120	P	10/21/2025	12:44	LB137613
	Chromium	9.98	10.0	100	80 - 120	P	10/21/2025	12:44	LB137613
	Cobalt	28.8	30.0	96	80 - 120	P	10/21/2025	12:44	LB137613
	Copper	21.5	20.0	107	80 - 120	P	10/21/2025	12:44	LB137613
	Iron	109	100	109	80 - 120	P	10/21/2025	12:44	LB137613
	Lead	11.0	12.0	92	80 - 120	P	10/21/2025	12:44	LB137613
	Magnesium	2060	2000	103	80 - 120	P	10/21/2025	12:44	LB137613
	Manganese	19.7	20.0	98	80 - 120	P	10/21/2025	12:44	LB137613
	Nickel	38.9	40.0	97	80 - 120	P	10/21/2025	12:44	LB137613
	Potassium	1980	2000	99	80 - 120	P	10/21/2025	12:44	LB137613
	Selenium	23.1	20.0	115	80 - 120	P	10/21/2025	12:44	LB137613
	Silver	10.3	10.0	103	80 - 120	P	10/21/2025	12:44	LB137613
	Sodium	1790	2000	90	80 - 120	P	10/21/2025	12:44	LB137613
	Thallium	41.7	40.0	104	80 - 120	P	10/21/2025	12:44	LB137613
	Vanadium	34.8	40.0	87	80 - 120	P	10/21/2025	12:44	LB137613
	Zinc	41.7	40.0	104	80 - 120	P	10/21/2025	12:44	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9780	10000	98	90 - 110	P	10/21/2025	13:45	LB137613
	Antimony	4830	5000	97	90 - 110	P	10/21/2025	13:45	LB137613
	Arsenic	4820	5000	96	90 - 110	P	10/21/2025	13:45	LB137613
	Barium	9660	10000	97	90 - 110	P	10/21/2025	13:45	LB137613
	Beryllium	242	250	97	90 - 110	P	10/21/2025	13:45	LB137613
	Cadmium	2420	2500	97	90 - 110	P	10/21/2025	13:45	LB137613
	Calcium	24200	25000	97	90 - 110	P	10/21/2025	13:45	LB137613
	Chromium	976	1000	98	90 - 110	P	10/21/2025	13:45	LB137613
	Cobalt	2410	2500	96	90 - 110	P	10/21/2025	13:45	LB137613
	Copper	1220	1250	98	90 - 110	P	10/21/2025	13:45	LB137613
	Iron	4750	5000	95	90 - 110	P	10/21/2025	13:45	LB137613
	Lead	4800	5000	96	90 - 110	P	10/21/2025	13:45	LB137613
	Magnesium	24300	25000	97	90 - 110	P	10/21/2025	13:45	LB137613
	Manganese	2410	2500	96	90 - 110	P	10/21/2025	13:45	LB137613
	Nickel	2410	2500	97	90 - 110	P	10/21/2025	13:45	LB137613
	Potassium	23800	25000	95	90 - 110	P	10/21/2025	13:45	LB137613
	Selenium	4820	5000	96	90 - 110	P	10/21/2025	13:45	LB137613
	Silver	1200	1250	96	90 - 110	P	10/21/2025	13:45	LB137613
	Sodium	23900	25000	96	90 - 110	P	10/21/2025	13:45	LB137613
	Thallium	5230	5000	104	90 - 110	P	10/21/2025	13:45	LB137613
	Vanadium	2430	2500	97	90 - 110	P	10/21/2025	13:45	LB137613
	Zinc	2430	2500	97	90 - 110	P	10/21/2025	13:45	LB137613
CCV02	Aluminum	9530	10000	95	90 - 110	P	10/21/2025	15:03	LB137613
	Antimony	4980	5000	100	90 - 110	P	10/21/2025	15:03	LB137613
	Arsenic	4830	5000	97	90 - 110	P	10/21/2025	15:03	LB137613
	Barium	9800	10000	98	90 - 110	P	10/21/2025	15:03	LB137613
	Beryllium	227	250	91	90 - 110	P	10/21/2025	15:03	LB137613
	Cadmium	2310	2500	92	90 - 110	P	10/21/2025	15:03	LB137613
	Calcium	23700	25000	95	90 - 110	P	10/21/2025	15:03	LB137613
	Chromium	934	1000	93	90 - 110	P	10/21/2025	15:03	LB137613
	Cobalt	2340	2500	94	90 - 110	P	10/21/2025	15:03	LB137613
	Copper	1210	1250	96	90 - 110	P	10/21/2025	15:03	LB137613
	Iron	4910	5000	98	90 - 110	P	10/21/2025	15:03	LB137613
	Lead	4660	5000	93	90 - 110	P	10/21/2025	15:03	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	23200	25000	93	90 - 110	P	10/21/2025	15:03	LB137613
	Manganese	2380	2500	95	90 - 110	P	10/21/2025	15:03	LB137613
	Nickel	2330	2500	93	90 - 110	P	10/21/2025	15:03	LB137613
	Potassium	24700	25000	99	90 - 110	P	10/21/2025	15:03	LB137613
	Selenium	4920	5000	98	90 - 110	P	10/21/2025	15:03	LB137613
	Silver	1200	1250	96	90 - 110	P	10/21/2025	15:03	LB137613
	Sodium	24800	25000	99	90 - 110	P	10/21/2025	15:03	LB137613
	Thallium	4950	5000	99	90 - 110	P	10/21/2025	15:03	LB137613
	Vanadium	2390	2500	96	90 - 110	P	10/21/2025	15:03	LB137613
	Zinc	2370	2500	95	90 - 110	P	10/21/2025	15:03	LB137613
CCV03	Aluminum	9580	10000	96	90 - 110	P	10/21/2025	15:56	LB137613
	Antimony	4900	5000	98	90 - 110	P	10/21/2025	15:56	LB137613
	Arsenic	4820	5000	96	90 - 110	P	10/21/2025	15:56	LB137613
	Barium	9710	10000	97	90 - 110	P	10/21/2025	15:56	LB137613
	Beryllium	233	250	93	90 - 110	P	10/21/2025	15:56	LB137613
	Cadmium	2330	2500	93	90 - 110	P	10/21/2025	15:56	LB137613
	Calcium	23800	25000	95	90 - 110	P	10/21/2025	15:56	LB137613
	Chromium	941	1000	94	90 - 110	P	10/21/2025	15:56	LB137613
	Cobalt	2360	2500	94	90 - 110	P	10/21/2025	15:56	LB137613
	Copper	1210	1250	96	90 - 110	P	10/21/2025	15:56	LB137613
	Iron	4850	5000	97	90 - 110	P	10/21/2025	15:56	LB137613
	Lead	4680	5000	94	90 - 110	P	10/21/2025	15:56	LB137613
	Magnesium	23400	25000	94	90 - 110	P	10/21/2025	15:56	LB137613
	Manganese	2380	2500	95	90 - 110	P	10/21/2025	15:56	LB137613
	Nickel	2350	2500	94	90 - 110	P	10/21/2025	15:56	LB137613
	Potassium	25200	25000	101	90 - 110	P	10/21/2025	15:56	LB137613
	Selenium	4870	5000	97	90 - 110	P	10/21/2025	15:56	LB137613
	Silver	1200	1250	96	90 - 110	P	10/21/2025	15:56	LB137613
	Sodium	25300	25000	101	90 - 110	P	10/21/2025	15:56	LB137613
	Thallium	5060	5000	101	90 - 110	P	10/21/2025	15:56	LB137613
	Vanadium	2400	2500	96	90 - 110	P	10/21/2025	15:56	LB137613
	Zinc	2390	2500	96	90 - 110	P	10/21/2025	15:56	LB137613
CCV04	Aluminum	9280	10000	93	90 - 110	P	10/21/2025	16:48	LB137613
	Antimony	4870	5000	98	90 - 110	P	10/21/2025	16:48	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4750	5000	95	90 - 110	P	10/21/2025	16:48	LB137613
	Barium	9490	10000	95	90 - 110	P	10/21/2025	16:48	LB137613
	Beryllium	232	250	93	90 - 110	P	10/21/2025	16:48	LB137613
	Cadmium	2300	2500	92	90 - 110	P	10/21/2025	16:48	LB137613
	Calcium	23200	25000	93	90 - 110	P	10/21/2025	16:48	LB137613
	Chromium	934	1000	93	90 - 110	P	10/21/2025	16:48	LB137613
	Cobalt	2330	2500	93	90 - 110	P	10/21/2025	16:48	LB137613
	Copper	1190	1250	96	90 - 110	P	10/21/2025	16:48	LB137613
	Iron	4840	5000	97	90 - 110	P	10/21/2025	16:48	LB137613
	Lead	4640	5000	93	90 - 110	P	10/21/2025	16:48	LB137613
	Magnesium	22900	25000	92	90 - 110	P	10/21/2025	16:48	LB137613
	Manganese	2350	2500	94	90 - 110	P	10/21/2025	16:48	LB137613
	Nickel	2330	2500	93	90 - 110	P	10/21/2025	16:48	LB137613
	Potassium	24100	25000	96	90 - 110	P	10/21/2025	16:48	LB137613
	Selenium	4830	5000	97	90 - 110	P	10/21/2025	16:48	LB137613
	Silver	1210	1250	97	90 - 110	P	10/21/2025	16:48	LB137613
	Sodium	24000	25000	96	90 - 110	P	10/21/2025	16:48	LB137613
	Thallium	4900	5000	98	90 - 110	P	10/21/2025	16:48	LB137613
	Vanadium	2330	2500	93	90 - 110	P	10/21/2025	16:48	LB137613
	Zinc	2400	2500	96	90 - 110	P	10/21/2025	16:48	LB137613
CCV05	Aluminum	9540	10000	95	90 - 110	P	10/21/2025	17:37	LB137613
	Antimony	5030	5000	101	90 - 110	P	10/21/2025	17:37	LB137613
	Arsenic	4900	5000	98	90 - 110	P	10/21/2025	17:37	LB137613
	Barium	9470	10000	95	90 - 110	P	10/21/2025	17:37	LB137613
	Beryllium	234	250	94	90 - 110	P	10/21/2025	17:37	LB137613
	Cadmium	2370	2500	95	90 - 110	P	10/21/2025	17:37	LB137613
	Calcium	23800	25000	95	90 - 110	P	10/21/2025	17:37	LB137613
	Chromium	964	1000	96	90 - 110	P	10/21/2025	17:37	LB137613
	Cobalt	2390	2500	96	90 - 110	P	10/21/2025	17:37	LB137613
	Copper	1220	1250	98	90 - 110	P	10/21/2025	17:37	LB137613
	Iron	4840	5000	97	90 - 110	P	10/21/2025	17:37	LB137613
	Lead	4780	5000	96	90 - 110	P	10/21/2025	17:37	LB137613
	Magnesium	23600	25000	94	90 - 110	P	10/21/2025	17:37	LB137613
	Manganese	2360	2500	94	90 - 110	P	10/21/2025	17:37	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2390	2500	96	90 - 110	P	10/21/2025	17:37	LB137613
	Potassium	23700	25000	95	90 - 110	P	10/21/2025	17:37	LB137613
	Selenium	5030	5000	100	90 - 110	P	10/21/2025	17:37	LB137613
	Silver	1210	1250	97	90 - 110	P	10/21/2025	17:37	LB137613
	Sodium	23100	25000	92	90 - 110	P	10/21/2025	17:37	LB137613
	Thallium	5030	5000	100	90 - 110	P	10/21/2025	17:37	LB137613
	Vanadium	2390	2500	96	90 - 110	P	10/21/2025	17:37	LB137613
	Zinc	2430	2500	97	90 - 110	P	10/21/2025	17:37	LB137613
CCV06	Aluminum	9220	10000	92	90 - 110	P	10/21/2025	18:27	LB137613
	Antimony	4900	5000	98	90 - 110	P	10/21/2025	18:27	LB137613
	Arsenic	4750	5000	95	90 - 110	P	10/21/2025	18:27	LB137613
	Barium	9370	10000	94	90 - 110	P	10/21/2025	18:27	LB137613
	Beryllium	226	250	90	90 - 110	P	10/21/2025	18:27	LB137613
	Cadmium	2270	2500	91	90 - 110	P	10/21/2025	18:27	LB137613
	Calcium	23000	25000	92	90 - 110	P	10/21/2025	18:27	LB137613
	Chromium	905	1000	90	90 - 110	P	10/21/2025	18:27	LB137613
	Cobalt	2310	2500	92	90 - 110	P	10/21/2025	18:27	LB137613
	Copper	1180	1250	95	90 - 110	P	10/21/2025	18:27	LB137613
	Iron	4730	5000	95	90 - 110	P	10/21/2025	18:27	LB137613
	Lead	4610	5000	92	90 - 110	P	10/21/2025	18:27	LB137613
	Magnesium	22800	25000	91	90 - 110	P	10/21/2025	18:27	LB137613
	Manganese	2330	2500	93	90 - 110	P	10/21/2025	18:27	LB137613
	Nickel	2300	2500	92	90 - 110	P	10/21/2025	18:27	LB137613
	Potassium	24100	25000	96	90 - 110	P	10/21/2025	18:27	LB137613
	Selenium	4870	5000	97	90 - 110	P	10/21/2025	18:27	LB137613
	Silver	1170	1250	94	90 - 110	P	10/21/2025	18:27	LB137613
	Sodium	24400	25000	98	90 - 110	P	10/21/2025	18:27	LB137613
	Thallium	4850	5000	97	90 - 110	P	10/21/2025	18:27	LB137613
	Vanadium	2320	2500	93	90 - 110	P	10/21/2025	18:27	LB137613
	Zinc	2320	2500	93	90 - 110	P	10/21/2025	18:27	LB137613
CCV07	Aluminum	9650	10000	96	90 - 110	P	10/21/2025	19:18	LB137613
	Antimony	4900	5000	98	90 - 110	P	10/21/2025	19:18	LB137613
	Arsenic	4770	5000	95	90 - 110	P	10/21/2025	19:18	LB137613
	Barium	9640	10000	96	90 - 110	P	10/21/2025	19:18	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	236	250	94	90 - 110	P	10/21/2025	19:18	LB137613
	Cadmium	2320	2500	93	90 - 110	P	10/21/2025	19:18	LB137613
	Calcium	24100	25000	96	90 - 110	P	10/21/2025	19:18	LB137613
	Chromium	945	1000	94	90 - 110	P	10/21/2025	19:18	LB137613
	Cobalt	2350	2500	94	90 - 110	P	10/21/2025	19:18	LB137613
	Copper	1200	1250	96	90 - 110	P	10/21/2025	19:18	LB137613
	Iron	4820	5000	96	90 - 110	P	10/21/2025	19:18	LB137613
	Lead	4680	5000	94	90 - 110	P	10/21/2025	19:18	LB137613
	Magnesium	23800	25000	95	90 - 110	P	10/21/2025	19:18	LB137613
	Manganese	2410	2500	96	90 - 110	P	10/21/2025	19:18	LB137613
	Nickel	2350	2500	94	90 - 110	P	10/21/2025	19:18	LB137613
	Potassium	23500	25000	94	90 - 110	P	10/21/2025	19:18	LB137613
	Selenium	4850	5000	97	90 - 110	P	10/21/2025	19:18	LB137613
	Silver	1190	1250	95	90 - 110	P	10/21/2025	19:18	LB137613
	Sodium	22700	25000	91	90 - 110	P	10/21/2025	19:18	LB137613
	Thallium	4820	5000	96	90 - 110	P	10/21/2025	19:18	LB137613
	Vanadium	2430	2500	97	90 - 110	P	10/21/2025	19:18	LB137613
	Zinc	2370	2500	95	90 - 110	P	10/21/2025	19:18	LB137613
CCV08	Aluminum	9520	10000	95	90 - 110	P	10/21/2025	20:03	LB137613
	Antimony	5000	5000	100	90 - 110	P	10/21/2025	20:03	LB137613
	Arsenic	4850	5000	97	90 - 110	P	10/21/2025	20:03	LB137613
	Barium	9610	10000	96	90 - 110	P	10/21/2025	20:03	LB137613
	Beryllium	234	250	94	90 - 110	P	10/21/2025	20:03	LB137613
	Cadmium	2330	2500	93	90 - 110	P	10/21/2025	20:03	LB137613
	Calcium	23700	25000	95	90 - 110	P	10/21/2025	20:03	LB137613
	Chromium	915	1000	92	90 - 110	P	10/21/2025	20:03	LB137613
	Cobalt	2360	2500	94	90 - 110	P	10/21/2025	20:03	LB137613
	Copper	1210	1250	97	90 - 110	P	10/21/2025	20:03	LB137613
	Iron	4830	5000	97	90 - 110	P	10/21/2025	20:03	LB137613
	Lead	4730	5000	94	90 - 110	P	10/21/2025	20:03	LB137613
	Magnesium	23600	25000	94	90 - 110	P	10/21/2025	20:03	LB137613
	Manganese	2410	2500	96	90 - 110	P	10/21/2025	20:03	LB137613
	Nickel	2350	2500	94	90 - 110	P	10/21/2025	20:03	LB137613
	Potassium	23100	25000	93	90 - 110	P	10/21/2025	20:03	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4970	5000	100	90 - 110	P	10/21/2025	20:03	LB137613
	Silver	1190	1250	95	90 - 110	P	10/21/2025	20:03	LB137613
	Sodium	23300	25000	93	90 - 110	P	10/21/2025	20:03	LB137613
	Thallium	5050	5000	101	90 - 110	P	10/21/2025	20:03	LB137613
	Vanadium	2400	2500	96	90 - 110	P	10/21/2025	20:03	LB137613
	Zinc	2340	2500	94	90 - 110	P	10/21/2025	20:03	LB137613

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7590	8000	95	90 - 110	P	10/23/2025	13:05	LB137630
	Antimony	3990	4000	100	90 - 110	P	10/23/2025	13:05	LB137630
	Arsenic	3870	4000	97	90 - 110	P	10/23/2025	13:05	LB137630
	Barium	7690	8000	96	90 - 110	P	10/23/2025	13:05	LB137630
	Beryllium	195	200	98	90 - 110	P	10/23/2025	13:05	LB137630
	Cadmium	2050	2000	103	90 - 110	P	10/23/2025	13:05	LB137630
	Calcium	19200	20000	96	90 - 110	P	10/23/2025	13:05	LB137630
	Chromium	800	800	100	90 - 110	P	10/23/2025	13:05	LB137630
	Cobalt	1940	2000	97	90 - 110	P	10/23/2025	13:05	LB137630
	Copper	995	1000	100	90 - 110	P	10/23/2025	13:05	LB137630
	Iron	3980	4000	100	90 - 110	P	10/23/2025	13:05	LB137630
	Lead	3900	4000	98	90 - 110	P	10/23/2025	13:05	LB137630
	Magnesium	19100	20000	96	90 - 110	P	10/23/2025	13:05	LB137630
	Manganese	1920	2000	96	90 - 110	P	10/23/2025	13:05	LB137630
	Nickel	1940	2000	97	90 - 110	P	10/23/2025	13:05	LB137630
	Potassium	19500	20000	97	90 - 110	P	10/23/2025	13:05	LB137630
	Selenium	3910	4000	98	90 - 110	P	10/23/2025	13:05	LB137630
	Silver	1010	1000	101	90 - 110	P	10/23/2025	13:05	LB137630
	Sodium	18900	20000	94	90 - 110	P	10/23/2025	13:05	LB137630
	Thallium	3800	4000	95	90 - 110	P	10/23/2025	13:05	LB137630
	Vanadium	1920	2000	96	90 - 110	P	10/23/2025	13:05	LB137630
	Zinc	1970	2000	99	90 - 110	P	10/23/2025	13:05	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	93.2	100	93	80 - 120	P	10/23/2025	13:14	LB137630
	Antimony	51.7	50.0	103	80 - 120	P	10/23/2025	13:14	LB137630
	Arsenic	22.1	20.0	110	80 - 120	P	10/23/2025	13:14	LB137630
	Barium	98.7	100	99	80 - 120	P	10/23/2025	13:14	LB137630
	Beryllium	6.48	6.0	108	80 - 120	P	10/23/2025	13:14	LB137630
	Cadmium	5.95	6.0	99	80 - 120	P	10/23/2025	13:14	LB137630
	Calcium	2070	2000	103	80 - 120	P	10/23/2025	13:14	LB137630
	Chromium	11.0	10.0	110	80 - 120	P	10/23/2025	13:14	LB137630
	Cobalt	29.9	30.0	100	80 - 120	P	10/23/2025	13:14	LB137630
	Copper	22.5	20.0	113	80 - 120	P	10/23/2025	13:14	LB137630
	Iron	108	100	108	80 - 120	P	10/23/2025	13:14	LB137630
	Lead	13.2	12.0	110	80 - 120	P	10/23/2025	13:14	LB137630
	Magnesium	2120	2000	106	80 - 120	P	10/23/2025	13:14	LB137630
	Manganese	22.1	20.0	110	80 - 120	P	10/23/2025	13:14	LB137630
	Nickel	39.8	40.0	100	80 - 120	P	10/23/2025	13:14	LB137630
	Potassium	2040	2000	102	80 - 120	P	10/23/2025	13:14	LB137630
	Selenium	23.6	20.0	118	80 - 120	P	10/23/2025	13:14	LB137630
	Silver	10.7	10.0	106	80 - 120	P	10/23/2025	13:14	LB137630
	Sodium	1890	2000	95	80 - 120	P	10/23/2025	13:14	LB137630
	Thallium	40.5	40.0	101	80 - 120	P	10/23/2025	13:14	LB137630
	Vanadium	36.3	40.0	91	80 - 120	P	10/23/2025	13:14	LB137630
	Zinc	43.5	40.0	109	80 - 120	P	10/23/2025	13:14	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10100	10000	101	90 - 110	P	10/23/2025	13:52	LB137630
	Antimony	5270	5000	105	90 - 110	P	10/23/2025	13:52	LB137630
	Arsenic	5250	5000	105	90 - 110	P	10/23/2025	13:52	LB137630
	Barium	9900	10000	99	90 - 110	P	10/23/2025	13:52	LB137630
	Beryllium	260	250	104	90 - 110	P	10/23/2025	13:52	LB137630
	Cadmium	2580	2500	103	90 - 110	P	10/23/2025	13:52	LB137630
	Calcium	25300	25000	101	90 - 110	P	10/23/2025	13:52	LB137630
	Chromium	1060	1000	106	90 - 110	P	10/23/2025	13:52	LB137630
	Cobalt	2570	2500	103	90 - 110	P	10/23/2025	13:52	LB137630
	Copper	1300	1250	104	90 - 110	P	10/23/2025	13:52	LB137630
	Iron	5110	5000	102	90 - 110	P	10/23/2025	13:52	LB137630
	Lead	5160	5000	103	90 - 110	P	10/23/2025	13:52	LB137630
	Magnesium	25300	25000	101	90 - 110	P	10/23/2025	13:52	LB137630
	Manganese	2480	2500	99	90 - 110	P	10/23/2025	13:52	LB137630
	Nickel	2580	2500	103	90 - 110	P	10/23/2025	13:52	LB137630
	Potassium	25200	25000	101	90 - 110	P	10/23/2025	13:52	LB137630
	Selenium	5330	5000	107	90 - 110	P	10/23/2025	13:52	LB137630
	Silver	1300	1250	104	90 - 110	P	10/23/2025	13:52	LB137630
	Sodium	24300	25000	97	90 - 110	P	10/23/2025	13:52	LB137630
	Thallium	5300	5000	106	90 - 110	P	10/23/2025	13:52	LB137630
	Vanadium	2530	2500	101	90 - 110	P	10/23/2025	13:52	LB137630
	Zinc	2620	2500	105	90 - 110	P	10/23/2025	13:52	LB137630
CCV02	Aluminum	10400	10000	104	90 - 110	P	10/23/2025	15:04	LB137630
	Antimony	5160	5000	103	90 - 110	P	10/23/2025	15:04	LB137630
	Arsenic	5240	5000	105	90 - 110	P	10/23/2025	15:04	LB137630
	Barium	10600	10000	106	90 - 110	P	10/23/2025	15:04	LB137630
	Beryllium	274	250	110	90 - 110	P	10/23/2025	15:04	LB137630
	Cadmium	2630	2500	105	90 - 110	P	10/23/2025	15:04	LB137630
	Calcium	26200	25000	105	90 - 110	P	10/23/2025	15:04	LB137630
	Chromium	1050	1000	105	90 - 110	P	10/23/2025	15:04	LB137630
	Cobalt	2600	2500	104	90 - 110	P	10/23/2025	15:04	LB137630
	Copper	1310	1250	105	90 - 110	P	10/23/2025	15:04	LB137630
	Iron	5060	5000	101	90 - 110	P	10/23/2025	15:04	LB137630
	Lead	5230	5000	105	90 - 110	P	10/23/2025	15:04	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	26100	25000	104	90 - 110	P	10/23/2025	15:04	LB137630
	Manganese	2650	2500	106	90 - 110	P	10/23/2025	15:04	LB137630
	Nickel	2610	2500	104	90 - 110	P	10/23/2025	15:04	LB137630
	Potassium	25500	25000	102	90 - 110	P	10/23/2025	15:04	LB137630
	Selenium	5210	5000	104	90 - 110	P	10/23/2025	15:04	LB137630
	Silver	1300	1250	104	90 - 110	P	10/23/2025	15:04	LB137630
	Sodium	25400	25000	102	90 - 110	P	10/23/2025	15:04	LB137630
	Thallium	5320	5000	106	90 - 110	P	10/23/2025	15:04	LB137630
	Vanadium	2600	2500	104	90 - 110	P	10/23/2025	15:04	LB137630
	Zinc	2610	2500	104	90 - 110	P	10/23/2025	15:04	LB137630
CCV03	Aluminum	9770	10000	98	90 - 110	P	10/23/2025	15:56	LB137630
	Antimony	4840	5000	97	90 - 110	P	10/23/2025	15:56	LB137630
	Arsenic	4960	5000	99	90 - 110	P	10/23/2025	15:56	LB137630
	Barium	9590	10000	96	90 - 110	P	10/23/2025	15:56	LB137630
	Beryllium	265	250	106	90 - 110	P	10/23/2025	15:56	LB137630
	Cadmium	2520	2500	101	90 - 110	P	10/23/2025	15:56	LB137630
	Calcium	24500	25000	98	90 - 110	P	10/23/2025	15:56	LB137630
	Chromium	1030	1000	103	90 - 110	P	10/23/2025	15:56	LB137630
	Cobalt	2470	2500	99	90 - 110	P	10/23/2025	15:56	LB137630
	Copper	1240	1250	99	90 - 110	P	10/23/2025	15:56	LB137630
	Iron	4680	5000	94	90 - 110	P	10/23/2025	15:56	LB137630
	Lead	5020	5000	100	90 - 110	P	10/23/2025	15:56	LB137630
	Magnesium	24800	25000	99	90 - 110	P	10/23/2025	15:56	LB137630
	Manganese	2420	2500	97	90 - 110	P	10/23/2025	15:56	LB137630
	Nickel	2490	2500	100	90 - 110	P	10/23/2025	15:56	LB137630
	Potassium	24800	25000	99	90 - 110	P	10/23/2025	15:56	LB137630
	Selenium	4960	5000	99	90 - 110	P	10/23/2025	15:56	LB137630
	Silver	1230	1250	99	90 - 110	P	10/23/2025	15:56	LB137630
	Sodium	25100	25000	100	90 - 110	P	10/23/2025	15:56	LB137630
	Thallium	4720	5000	94	90 - 110	P	10/23/2025	15:56	LB137630
	Vanadium	2430	2500	97	90 - 110	P	10/23/2025	15:56	LB137630
	Zinc	2480	2500	99	90 - 110	P	10/23/2025	15:56	LB137630
CCV04	Aluminum	9680	10000	97	90 - 110	P	10/23/2025	16:43	LB137630
	Antimony	4830	5000	97	90 - 110	P	10/23/2025	16:43	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4960	5000	99	90 - 110	P	10/23/2025	16:43	LB137630
	Barium	9850	10000	98	90 - 110	P	10/23/2025	16:43	LB137630
	Beryllium	266	250	106	90 - 110	P	10/23/2025	16:43	LB137630
	Cadmium	2520	2500	101	90 - 110	P	10/23/2025	16:43	LB137630
	Calcium	24600	25000	98	90 - 110	P	10/23/2025	16:43	LB137630
	Chromium	1000	1000	100	90 - 110	P	10/23/2025	16:43	LB137630
	Cobalt	2480	2500	99	90 - 110	P	10/23/2025	16:43	LB137630
	Copper	1240	1250	99	90 - 110	P	10/23/2025	16:43	LB137630
	Iron	4670	5000	93	90 - 110	P	10/23/2025	16:43	LB137630
	Lead	5020	5000	100	90 - 110	P	10/23/2025	16:43	LB137630
	Magnesium	24600	25000	98	90 - 110	P	10/23/2025	16:43	LB137630
	Manganese	2470	2500	99	90 - 110	P	10/23/2025	16:43	LB137630
	Nickel	2490	2500	100	90 - 110	P	10/23/2025	16:43	LB137630
	Potassium	23500	25000	94	90 - 110	P	10/23/2025	16:43	LB137630
	Selenium	4920	5000	98	90 - 110	P	10/23/2025	16:43	LB137630
	Silver	1220	1250	98	90 - 110	P	10/23/2025	16:43	LB137630
	Sodium	22800	25000	91	90 - 110	P	10/23/2025	16:43	LB137630
	Thallium	4760	5000	95	90 - 110	P	10/23/2025	16:43	LB137630
	Vanadium	2430	2500	97	90 - 110	P	10/23/2025	16:43	LB137630
	Zinc	2450	2500	98	90 - 110	P	10/23/2025	16:43	LB137630
CCV05	Aluminum	10200	10000	102	90 - 110	P	10/23/2025	17:21	LB137630
	Antimony	5070	5000	101	90 - 110	P	10/23/2025	17:21	LB137630
	Arsenic	5210	5000	104	90 - 110	P	10/23/2025	17:21	LB137630
	Barium	10300	10000	103	90 - 110	P	10/23/2025	17:21	LB137630
	Beryllium	247	250	99	90 - 110	P	10/23/2025	17:21	LB137630
	Cadmium	2640	2500	105	90 - 110	P	10/23/2025	17:21	LB137630
	Calcium	25800	25000	103	90 - 110	P	10/23/2025	17:21	LB137630
	Chromium	967	1000	97	90 - 110	P	10/23/2025	17:21	LB137630
	Cobalt	2590	2500	104	90 - 110	P	10/23/2025	17:21	LB137630
	Copper	1300	1250	104	90 - 110	P	10/23/2025	17:21	LB137630
	Iron	4920	5000	98	90 - 110	P	10/23/2025	17:21	LB137630
	Lead	5250	5000	105	90 - 110	P	10/23/2025	17:21	LB137630
	Magnesium	25900	25000	104	90 - 110	P	10/23/2025	17:21	LB137630
	Manganese	2580	2500	103	90 - 110	P	10/23/2025	17:21	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2610	2500	104	90 - 110	P	10/23/2025	17:21	LB137630
	Potassium	25500	25000	102	90 - 110	P	10/23/2025	17:21	LB137630
	Selenium	5180	5000	104	90 - 110	P	10/23/2025	17:21	LB137630
	Silver	1290	1250	103	90 - 110	P	10/23/2025	17:21	LB137630
	Sodium	25100	25000	101	90 - 110	P	10/23/2025	17:21	LB137630
	Thallium	4740	5000	95	90 - 110	P	10/23/2025	17:21	LB137630
	Vanadium	2560	2500	102	90 - 110	P	10/23/2025	17:21	LB137630
	Zinc	2600	2500	104	90 - 110	P	10/23/2025	17:21	LB137630
CCV06	Aluminum	10200	10000	102	90 - 110	P	10/23/2025	18:11	LB137630
	Antimony	5070	5000	101	90 - 110	P	10/23/2025	18:11	LB137630
	Arsenic	5210	5000	104	90 - 110	P	10/23/2025	18:11	LB137630
	Barium	10000	10000	100	90 - 110	P	10/23/2025	18:11	LB137630
	Beryllium	253	250	101	90 - 110	P	10/23/2025	18:11	LB137630
	Cadmium	2670	2500	107	90 - 110	P	10/23/2025	18:11	LB137630
	Calcium	25700	25000	103	90 - 110	P	10/23/2025	18:11	LB137630
	Chromium	987	1000	99	90 - 110	P	10/23/2025	18:11	LB137630
	Cobalt	2610	2500	104	90 - 110	P	10/23/2025	18:11	LB137630
	Copper	1310	1250	105	90 - 110	P	10/23/2025	18:11	LB137630
	Iron	4840	5000	97	90 - 110	P	10/23/2025	18:11	LB137630
	Lead	5320	5000	106	90 - 110	P	10/23/2025	18:11	LB137630
	Magnesium	26100	25000	104	90 - 110	P	10/23/2025	18:11	LB137630
	Manganese	2530	2500	101	90 - 110	P	10/23/2025	18:11	LB137630
	Nickel	2630	2500	105	90 - 110	P	10/23/2025	18:11	LB137630
	Potassium	25300	25000	101	90 - 110	P	10/23/2025	18:11	LB137630
	Selenium	5190	5000	104	90 - 110	P	10/23/2025	18:11	LB137630
	Silver	1320	1250	105	90 - 110	P	10/23/2025	18:11	LB137630
	Sodium	23800	25000	95	90 - 110	P	10/23/2025	18:11	LB137630
	Thallium	4660	5000	93	90 - 110	P	10/23/2025	18:11	LB137630
	Vanadium	2560	2500	102	90 - 110	P	10/23/2025	18:11	LB137630
	Zinc	2630	2500	105	90 - 110	P	10/23/2025	18:11	LB137630
CCV07	Aluminum	9840	10000	98	90 - 110	P	10/23/2025	18:59	LB137630
	Antimony	4900	5000	98	90 - 110	P	10/23/2025	18:59	LB137630
	Arsenic	5030	5000	101	90 - 110	P	10/23/2025	18:59	LB137630
	Barium	9590	10000	96	90 - 110	P	10/23/2025	18:59	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	244	250	98	90 - 110	P	10/23/2025	18:59	LB137630
	Cadmium	2530	2500	101	90 - 110	P	10/23/2025	18:59	LB137630
	Calcium	24600	25000	98	90 - 110	P	10/23/2025	18:59	LB137630
	Chromium	952	1000	95	90 - 110	P	10/23/2025	18:59	LB137630
	Cobalt	2470	2500	99	90 - 110	P	10/23/2025	18:59	LB137630
	Copper	1250	1250	100	90 - 110	P	10/23/2025	18:59	LB137630
	Iron	4710	5000	94	90 - 110	P	10/23/2025	18:59	LB137630
	Lead	5060	5000	101	90 - 110	P	10/23/2025	18:59	LB137630
	Magnesium	25100	25000	100	90 - 110	P	10/23/2025	18:59	LB137630
	Manganese	2430	2500	97	90 - 110	P	10/23/2025	18:59	LB137630
	Nickel	2490	2500	100	90 - 110	P	10/23/2025	18:59	LB137630
	Potassium	23800	25000	95	90 - 110	P	10/23/2025	18:59	LB137630
	Selenium	5090	5000	102	90 - 110	P	10/23/2025	18:59	LB137630
	Silver	1280	1250	102	90 - 110	P	10/23/2025	18:59	LB137630
	Sodium	22700	25000	91	90 - 110	P	10/23/2025	18:59	LB137630
	Thallium	4530	5000	90	90 - 110	P	10/23/2025	18:59	LB137630
	Vanadium	2460	2500	98	90 - 110	P	10/23/2025	18:59	LB137630
	Zinc	2550	2500	102	90 - 110	P	10/23/2025	18:59	LB137630
CCV08	Aluminum	10300	10000	103	90 - 110	P	10/23/2025	19:48	LB137630
	Antimony	5040	5000	101	90 - 110	P	10/23/2025	19:48	LB137630
	Arsenic	5200	5000	104	90 - 110	P	10/23/2025	19:48	LB137630
	Barium	9920	10000	99	90 - 110	P	10/23/2025	19:48	LB137630
	Beryllium	254	250	102	90 - 110	P	10/23/2025	19:48	LB137630
	Cadmium	2670	2500	107	90 - 110	P	10/23/2025	19:48	LB137630
	Calcium	25600	25000	102	90 - 110	P	10/23/2025	19:48	LB137630
	Chromium	999	1000	100	90 - 110	P	10/23/2025	19:48	LB137630
	Cobalt	2600	2500	104	90 - 110	P	10/23/2025	19:48	LB137630
	Copper	1300	1250	104	90 - 110	P	10/23/2025	19:48	LB137630
	Iron	4770	5000	96	90 - 110	P	10/23/2025	19:48	LB137630
	Lead	5320	5000	106	90 - 110	P	10/23/2025	19:48	LB137630
	Magnesium	26100	25000	104	90 - 110	P	10/23/2025	19:48	LB137630
	Manganese	2520	2500	101	90 - 110	P	10/23/2025	19:48	LB137630
	Nickel	2620	2500	105	90 - 110	P	10/23/2025	19:48	LB137630
	Potassium	24000	25000	96	90 - 110	P	10/23/2025	19:48	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5190	5000	104	90 - 110	P	10/23/2025	19:48	LB137630
	Silver	1300	1250	104	90 - 110	P	10/23/2025	19:48	LB137630
	Sodium	23100	25000	92	90 - 110	P	10/23/2025	19:48	LB137630
	Thallium	4630	5000	93	90 - 110	P	10/23/2025	19:48	LB137630
	Vanadium	2560	2500	102	90 - 110	P	10/23/2025	19:48	LB137630
	Zinc	2610	2500	104	90 - 110	P	10/23/2025	19:48	LB137630
CCV09	Aluminum	10300	10000	103	90 - 110	P	10/23/2025	20:39	LB137630
	Antimony	5140	5000	103	90 - 110	P	10/23/2025	20:39	LB137630
	Arsenic	5320	5000	106	90 - 110	P	10/23/2025	20:39	LB137630
	Barium	9990	10000	100	90 - 110	P	10/23/2025	20:39	LB137630
	Beryllium	260	250	104	90 - 110	P	10/23/2025	20:39	LB137630
	Cadmium	2690	2500	108	90 - 110	P	10/23/2025	20:39	LB137630
	Calcium	25700	25000	103	90 - 110	P	10/23/2025	20:39	LB137630
	Chromium	1000	1000	100	90 - 110	P	10/23/2025	20:39	LB137630
	Cobalt	2630	2500	105	90 - 110	P	10/23/2025	20:39	LB137630
	Copper	1320	1250	105	90 - 110	P	10/23/2025	20:39	LB137630
	Iron	4850	5000	97	90 - 110	P	10/23/2025	20:39	LB137630
	Lead	5380	5000	108	90 - 110	P	10/23/2025	20:39	LB137630
	Magnesium	26300	25000	105	90 - 110	P	10/23/2025	20:39	LB137630
	Manganese	2540	2500	102	90 - 110	P	10/23/2025	20:39	LB137630
	Nickel	2650	2500	106	90 - 110	P	10/23/2025	20:39	LB137630
	Potassium	25600	25000	103	90 - 110	P	10/23/2025	20:39	LB137630
	Selenium	5360	5000	107	90 - 110	P	10/23/2025	20:39	LB137630
	Silver	1320	1250	106	90 - 110	P	10/23/2025	20:39	LB137630
	Sodium	25700	25000	103	90 - 110	P	10/23/2025	20:39	LB137630
	Thallium	4940	5000	99	90 - 110	P	10/23/2025	20:39	LB137630
	Vanadium	2570	2500	103	90 - 110	P	10/23/2025	20:39	LB137630
	Zinc	2650	2500	106	90 - 110	P	10/23/2025	20:39	LB137630
CCV10	Aluminum	9720	10000	97	90 - 110	P	10/23/2025	21:30	LB137630
	Antimony	4820	5000	96	90 - 110	P	10/23/2025	21:30	LB137630
	Arsenic	5010	5000	100	90 - 110	P	10/23/2025	21:30	LB137630
	Barium	9490	10000	95	90 - 110	P	10/23/2025	21:30	LB137630
	Beryllium	254	250	102	90 - 110	P	10/23/2025	21:30	LB137630
	Cadmium	2570	2500	103	90 - 110	P	10/23/2025	21:30	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Calcium	24700	25000	99	90 - 110	P	10/23/2025	21:30	LB137630
	Chromium	962	1000	96	90 - 110	P	10/23/2025	21:30	LB137630
	Cobalt	2490	2500	100	90 - 110	P	10/23/2025	21:30	LB137630
	Copper	1250	1250	100	90 - 110	P	10/23/2025	21:30	LB137630
	Iron	4660	5000	93	90 - 110	P	10/23/2025	21:30	LB137630
	Lead	5110	5000	102	90 - 110	P	10/23/2025	21:30	LB137630
	Magnesium	25400	25000	102	90 - 110	P	10/23/2025	21:30	LB137630
	Manganese	2430	2500	97	90 - 110	P	10/23/2025	21:30	LB137630
	Nickel	2510	2500	100	90 - 110	P	10/23/2025	21:30	LB137630
	Potassium	24400	25000	98	90 - 110	P	10/23/2025	21:30	LB137630
	Selenium	5040	5000	101	90 - 110	P	10/23/2025	21:30	LB137630
	Silver	1280	1250	102	90 - 110	P	10/23/2025	21:30	LB137630
	Sodium	23700	25000	95	90 - 110	P	10/23/2025	21:30	LB137630
	Thallium	4830	5000	96	90 - 110	P	10/23/2025	21:30	LB137630
	Vanadium	2460	2500	99	90 - 110	P	10/23/2025	21:30	LB137630
	Zinc	2530	2500	101	90 - 110	P	10/23/2025	21:30	LB137630
CCV11	Aluminum	9500	10000	95	90 - 110	P	10/23/2025	22:20	LB137630
	Antimony	4720	5000	94	90 - 110	P	10/23/2025	22:20	LB137630
	Arsenic	4980	5000	100	90 - 110	P	10/23/2025	22:20	LB137630
	Barium	9350	10000	94	90 - 110	P	10/23/2025	22:20	LB137630
	Beryllium	268	250	107	90 - 110	P	10/23/2025	22:20	LB137630
	Cadmium	2590	2500	104	90 - 110	P	10/23/2025	22:20	LB137630
	Calcium	24300	25000	97	90 - 110	P	10/23/2025	22:20	LB137630
	Chromium	951	1000	95	90 - 110	P	10/23/2025	22:20	LB137630
	Cobalt	2490	2500	100	90 - 110	P	10/23/2025	22:20	LB137630
	Copper	1250	1250	100	90 - 110	P	10/23/2025	22:20	LB137630
	Iron	4390	5000	88	90 - 110	P	10/23/2025	22:20	LB137630
	Lead	5160	5000	103	90 - 110	P	10/23/2025	22:20	LB137630
	Magnesium	25100	25000	100	90 - 110	P	10/23/2025	22:20	LB137630
	Manganese	2420	2500	97	90 - 110	P	10/23/2025	22:20	LB137630
	Nickel	2530	2500	101	90 - 110	P	10/23/2025	22:20	LB137630
	Potassium	28300	25000	113	90 - 110	P	10/23/2025	22:20	LB137630
	Selenium	4940	5000	99	90 - 110	P	10/23/2025	22:20	LB137630
	Silver	1230	1250	99	90 - 110	P	10/23/2025	22:20	LB137630

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV11	Sodium	32800	25000	131	90 - 110	P	10/23/2025	22:20	LB137630
	Thallium	4440	5000	89	90 - 110	P	10/23/2025	22:20	LB137630
	Vanadium	2410	2500	96	90 - 110	P	10/23/2025	22:20	LB137630
	Zinc	2450	2500	98	90 - 110	P	10/23/2025	22:20	LB137630
CCV12	Aluminum	9290	10000	93	90 - 110	P	10/23/2025	22:33	LB137630
	Antimony	4660	5000	93	90 - 110	P	10/23/2025	22:33	LB137630
	Arsenic	4920	5000	98	90 - 110	P	10/23/2025	22:33	LB137630
	Barium	9070	10000	91	90 - 110	P	10/23/2025	22:33	LB137630
	Beryllium	258	250	103	90 - 110	P	10/23/2025	22:33	LB137630
	Cadmium	2540	2500	102	90 - 110	P	10/23/2025	22:33	LB137630
	Calcium	23700	25000	95	90 - 110	P	10/23/2025	22:33	LB137630
	Chromium	932	1000	93	90 - 110	P	10/23/2025	22:33	LB137630
	Cobalt	2440	2500	98	90 - 110	P	10/23/2025	22:33	LB137630
	Copper	1230	1250	98	90 - 110	P	10/23/2025	22:33	LB137630
	Iron	4220	5000	84	90 - 110	P	10/23/2025	22:33	LB137630
	Lead	5070	5000	101	90 - 110	P	10/23/2025	22:33	LB137630
	Magnesium	24500	25000	98	90 - 110	P	10/23/2025	22:33	LB137630
	Manganese	2350	2500	94	90 - 110	P	10/23/2025	22:33	LB137630
	Nickel	2470	2500	99	90 - 110	P	10/23/2025	22:33	LB137630
	Potassium	27000	25000	108	90 - 110	P	10/23/2025	22:33	LB137630
	Selenium	4930	5000	99	90 - 110	P	10/23/2025	22:33	LB137630
	Silver	1220	1250	98	90 - 110	P	10/23/2025	22:33	LB137630
	Sodium	30000	25000	120	90 - 110	P	10/23/2025	22:33	LB137630
	Thallium	4550	5000	91	90 - 110	P	10/23/2025	22:33	LB137630
	Vanadium	2360	2500	94	90 - 110	P	10/23/2025	22:33	LB137630
	Zinc	2410	2500	96	90 - 110	P	10/23/2025	22:33	LB137630