



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB85	Mercury	0.10	+/-0.2	J	0.16	0.20	CV	10/23/2025	13:57	LB137626

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB79	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	10/23/2025	14:05	LB137626
CCB80	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	10/23/2025	14:38	LB137626
CCB81	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	10/23/2025	15:11	LB137626
CCB82	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	10/23/2025	15:32	LB137626
CCB83	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	10/23/2025	16:02	LB137626
CCB84	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	10/23/2025	16:16	LB137626

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	80.0	100	P	10/21/2025	12:50	LB137613
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/21/2025	12:50	LB137613
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/21/2025	12:50	LB137613
	Barium	14.6	+/-50	U	25.0	100	P	10/21/2025	12:50	LB137613
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/21/2025	12:50	LB137613
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/21/2025	12:50	LB137613
	Calcium	234	+/-1000	U	500	2000	P	10/21/2025	12:50	LB137613
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/21/2025	12:50	LB137613
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/21/2025	12:50	LB137613
	Copper	4.60	+/-10	U	16.0	20.0	P	10/21/2025	12:50	LB137613
	Iron	23.4	+/-50	U	80.0	100	P	10/21/2025	12:50	LB137613
	Lead	2.30	+/-6	U	9.60	12.0	P	10/21/2025	12:50	LB137613
	Magnesium	244	+/-1000	U	500	2000	P	10/21/2025	12:50	LB137613
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/21/2025	12:50	LB137613
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/21/2025	12:50	LB137613
	Potassium	918	+/-1000	U	1600	2000	P	10/21/2025	12:50	LB137613
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/21/2025	12:50	LB137613
	Silver	1.62	+/-5	U	5.00	10.0	P	10/21/2025	12:50	LB137613
	Sodium	868	+/-1000	U	1000	2000	P	10/21/2025	12:50	LB137613
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/21/2025	12:50	LB137613
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/21/2025	12:50	LB137613
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/21/2025	12:50	LB137613

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	80.0	100	P	10/21/2025	13:54	LB137613
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/21/2025	13:54	LB137613
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/21/2025	13:54	LB137613
	Barium	14.6	+/-50	U	25.0	100	P	10/21/2025	13:54	LB137613
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/21/2025	13:54	LB137613
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/21/2025	13:54	LB137613
	Calcium	234	+/-1000	U	500	2000	P	10/21/2025	13:54	LB137613
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/21/2025	13:54	LB137613
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/21/2025	13:54	LB137613
	Copper	4.60	+/-10	U	16.0	20.0	P	10/21/2025	13:54	LB137613
	Iron	23.4	+/-50	U	80.0	100	P	10/21/2025	13:54	LB137613
	Lead	2.30	+/-6	U	9.60	12.0	P	10/21/2025	13:54	LB137613
	Magnesium	244	+/-1000	U	500	2000	P	10/21/2025	13:54	LB137613
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/21/2025	13:54	LB137613
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/21/2025	13:54	LB137613
	Potassium	918	+/-1000	U	1600	2000	P	10/21/2025	13:54	LB137613
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/21/2025	13:54	LB137613
	Silver	1.62	+/-5	U	5.00	10.0	P	10/21/2025	13:54	LB137613
	Sodium	868	+/-1000	U	1000	2000	P	10/21/2025	13:54	LB137613
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/21/2025	13:54	LB137613
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/21/2025	13:54	LB137613
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/21/2025	13:54	LB137613
CCB02	Aluminum	11.3	+/-50	U	80.0	100	P	10/21/2025	15:07	LB137613
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/21/2025	15:07	LB137613
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/21/2025	15:07	LB137613
	Barium	14.6	+/-50	U	25.0	100	P	10/21/2025	15:07	LB137613
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/21/2025	15:07	LB137613
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/21/2025	15:07	LB137613
	Calcium	234	+/-1000	U	500	2000	P	10/21/2025	15:07	LB137613
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/21/2025	15:07	LB137613
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/21/2025	15:07	LB137613
	Copper	4.60	+/-10	U	16.0	20.0	P	10/21/2025	15:07	LB137613
	Iron	36.0	+/-50	J	80.0	100	P	10/21/2025	15:07	LB137613
	Lead	2.30	+/-6	U	9.60	12.0	P	10/21/2025	15:07	LB137613
	Magnesium	244	+/-1000	U	500	2000	P	10/21/2025	15:07	LB137613
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/21/2025	15:07	LB137613
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/21/2025	15:07	LB137613
	Potassium	918	+/-1000	U	1600	2000	P	10/21/2025	15:07	LB137613
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/21/2025	15:07	LB137613

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	5.00	10.0	P	10/21/2025	15:07	LB137613
	Sodium	868	+/-1000	U	1000	2000	P	10/21/2025	15:07	LB137613
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/21/2025	15:07	LB137613
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/21/2025	15:07	LB137613
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/21/2025	15:07	LB137613
CCB03	Aluminum	41.4	+/-50	J	80.0	100	P	10/21/2025	16:00	LB137613
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/21/2025	16:00	LB137613
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/21/2025	16:00	LB137613
	Barium	14.6	+/-50	U	25.0	100	P	10/21/2025	16:00	LB137613
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/21/2025	16:00	LB137613
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/21/2025	16:00	LB137613
	Calcium	234	+/-1000	U	500	2000	P	10/21/2025	16:00	LB137613
	Chromium	2.83	+/-5	J	5.00	10.0	P	10/21/2025	16:00	LB137613
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/21/2025	16:00	LB137613
	Copper	4.60	+/-10	U	16.0	20.0	P	10/21/2025	16:00	LB137613
	Iron	43.3	+/-50	J	80.0	100	P	10/21/2025	16:00	LB137613
	Lead	2.30	+/-6	U	9.60	12.0	P	10/21/2025	16:00	LB137613
	Magnesium	244	+/-1000	U	500	2000	P	10/21/2025	16:00	LB137613
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/21/2025	16:00	LB137613
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/21/2025	16:00	LB137613
	Potassium	918	+/-1000	U	1600	2000	P	10/21/2025	16:00	LB137613
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/21/2025	16:00	LB137613
	Silver	1.62	+/-5	U	5.00	10.0	P	10/21/2025	16:00	LB137613
	Sodium	868	+/-1000	U	1000	2000	P	10/21/2025	16:00	LB137613
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/21/2025	16:00	LB137613
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/21/2025	16:00	LB137613
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/21/2025	16:00	LB137613
CCB04	Aluminum	11.3	+/-50	U	80.0	100	P	10/21/2025	16:52	LB137613
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/21/2025	16:52	LB137613
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/21/2025	16:52	LB137613
	Barium	14.6	+/-50	U	25.0	100	P	10/21/2025	16:52	LB137613
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/21/2025	16:52	LB137613
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/21/2025	16:52	LB137613
	Calcium	234	+/-1000	U	500	2000	P	10/21/2025	16:52	LB137613
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/21/2025	16:52	LB137613
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/21/2025	16:52	LB137613
	Copper	4.60	+/-10	U	16.0	20.0	P	10/21/2025	16:52	LB137613
	Iron	23.4	+/-50	U	80.0	100	P	10/21/2025	16:52	LB137613
	Lead	2.30	+/-6	U	9.60	12.0	P	10/21/2025	16:52	LB137613

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	500	2000	P	10/21/2025	16:52	LB137613
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/21/2025	16:52	LB137613
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/21/2025	16:52	LB137613
	Potassium	918	+/-1000	U	1600	2000	P	10/21/2025	16:52	LB137613
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/21/2025	16:52	LB137613
	Silver	1.62	+/-5	U	5.00	10.0	P	10/21/2025	16:52	LB137613
	Sodium	868	+/-1000	U	1000	2000	P	10/21/2025	16:52	LB137613
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/21/2025	16:52	LB137613
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/21/2025	16:52	LB137613
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/21/2025	16:52	LB137613
CCB05	Aluminum	16.1	+/-50	J	80.0	100	P	10/21/2025	17:41	LB137613
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/21/2025	17:41	LB137613
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/21/2025	17:41	LB137613
	Barium	14.6	+/-50	U	25.0	100	P	10/21/2025	17:41	LB137613
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/21/2025	17:41	LB137613
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/21/2025	17:41	LB137613
	Calcium	234	+/-1000	U	500	2000	P	10/21/2025	17:41	LB137613
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/21/2025	17:41	LB137613
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/21/2025	17:41	LB137613
	Copper	4.60	+/-10	U	16.0	20.0	P	10/21/2025	17:41	LB137613
	Iron	23.4	+/-50	U	80.0	100	P	10/21/2025	17:41	LB137613
	Lead	2.30	+/-6	U	9.60	12.0	P	10/21/2025	17:41	LB137613
	Magnesium	244	+/-1000	U	500	2000	P	10/21/2025	17:41	LB137613
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/21/2025	17:41	LB137613
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/21/2025	17:41	LB137613
	Potassium	918	+/-1000	U	1600	2000	P	10/21/2025	17:41	LB137613
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/21/2025	17:41	LB137613
	Silver	1.62	+/-5	U	5.00	10.0	P	10/21/2025	17:41	LB137613
	Sodium	868	+/-1000	U	1000	2000	P	10/21/2025	17:41	LB137613
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/21/2025	17:41	LB137613
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/21/2025	17:41	LB137613
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/21/2025	17:41	LB137613
CCB06	Aluminum	11.4	+/-50	J	80.0	100	P	10/21/2025	18:31	LB137613
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/21/2025	18:31	LB137613
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/21/2025	18:31	LB137613
	Barium	14.6	+/-50	U	25.0	100	P	10/21/2025	18:31	LB137613
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/21/2025	18:31	LB137613
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/21/2025	18:31	LB137613
	Calcium	234	+/-1000	U	500	2000	P	10/21/2025	18:31	LB137613

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	5.00	10.0	P	10/21/2025	18:31	LB137613
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/21/2025	18:31	LB137613
	Copper	4.60	+/-10	U	16.0	20.0	P	10/21/2025	18:31	LB137613
	Iron	23.4	+/-50	U	80.0	100	P	10/21/2025	18:31	LB137613
	Lead	2.30	+/-6	U	9.60	12.0	P	10/21/2025	18:31	LB137613
	Magnesium	244	+/-1000	U	500	2000	P	10/21/2025	18:31	LB137613
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/21/2025	18:31	LB137613
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/21/2025	18:31	LB137613
	Potassium	956	+/-1000	J	1600	2000	P	10/21/2025	18:31	LB137613
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/21/2025	18:31	LB137613
	Silver	1.62	+/-5	U	5.00	10.0	P	10/21/2025	18:31	LB137613
	Sodium	1000	+/-1000	J	1000	2000	P	10/21/2025	18:31	LB137613
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/21/2025	18:31	LB137613
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/21/2025	18:31	LB137613
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/21/2025	18:31	LB137613
CCB07	Aluminum	11.3	+/-50	U	80.0	100	P	10/21/2025	19:22	LB137613
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/21/2025	19:22	LB137613
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/21/2025	19:22	LB137613
	Barium	14.6	+/-50	U	25.0	100	P	10/21/2025	19:22	LB137613
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/21/2025	19:22	LB137613
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/21/2025	19:22	LB137613
	Calcium	234	+/-1000	U	500	2000	P	10/21/2025	19:22	LB137613
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/21/2025	19:22	LB137613
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/21/2025	19:22	LB137613
	Copper	4.60	+/-10	U	16.0	20.0	P	10/21/2025	19:22	LB137613
	Iron	23.4	+/-50	U	80.0	100	P	10/21/2025	19:22	LB137613
	Lead	2.30	+/-6	U	9.60	12.0	P	10/21/2025	19:22	LB137613
	Magnesium	244	+/-1000	U	500	2000	P	10/21/2025	19:22	LB137613
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/21/2025	19:22	LB137613
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/21/2025	19:22	LB137613
	Potassium	918	+/-1000	U	1600	2000	P	10/21/2025	19:22	LB137613
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/21/2025	19:22	LB137613
	Silver	1.62	+/-5	U	5.00	10.0	P	10/21/2025	19:22	LB137613
	Sodium	868	+/-1000	U	1000	2000	P	10/21/2025	19:22	LB137613
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/21/2025	19:22	LB137613
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/21/2025	19:22	LB137613
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/21/2025	19:22	LB137613
CCB08	Aluminum	11.3	+/-50	U	80.0	100	P	10/21/2025	20:07	LB137613
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/21/2025	20:07	LB137613

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/21/2025	20:07	LB137613
	Barium	14.6	+/-50	U	25.0	100	P	10/21/2025	20:07	LB137613
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/21/2025	20:07	LB137613
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/21/2025	20:07	LB137613
	Calcium	234	+/-1000	U	500	2000	P	10/21/2025	20:07	LB137613
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/21/2025	20:07	LB137613
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/21/2025	20:07	LB137613
	Copper	4.60	+/-10	U	16.0	20.0	P	10/21/2025	20:07	LB137613
	Iron	23.4	+/-50	U	80.0	100	P	10/21/2025	20:07	LB137613
	Lead	2.30	+/-6	U	9.60	12.0	P	10/21/2025	20:07	LB137613
	Magnesium	244	+/-1000	U	500	2000	P	10/21/2025	20:07	LB137613
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/21/2025	20:07	LB137613
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/21/2025	20:07	LB137613
	Potassium	918	+/-1000	U	1600	2000	P	10/21/2025	20:07	LB137613
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/21/2025	20:07	LB137613
	Silver	1.62	+/-5	U	5.00	10.0	P	10/21/2025	20:07	LB137613
	Sodium	868	+/-1000	U	1000	2000	P	10/21/2025	20:07	LB137613
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/21/2025	20:07	LB137613
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/21/2025	20:07	LB137613
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/21/2025	20:07	LB137613

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.9	+/-50	J	80.0	100	P	10/23/2025	13:20	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	13:20	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	13:20	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	13:20	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	13:20	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	13:20	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	13:20	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	13:20	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	13:20	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	13:20	LB137630
	Iron	25.5	+/-50	J	80.0	100	P	10/23/2025	13:20	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	13:20	LB137630
	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	13:20	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	13:20	LB137630
	Nickel	6.19	+/-20	J	10.0	40.0	P	10/23/2025	13:20	LB137630
	Potassium	918	+/-1000	U	1600	2000	P	10/23/2025	13:20	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	13:20	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	13:20	LB137630
	Sodium	868	+/-1000	U	1000	2000	P	10/23/2025	13:20	LB137630
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/23/2025	13:20	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	13:20	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	13:20	LB137630

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	80.0	100	P	10/23/2025	13:57	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	13:57	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	13:57	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	13:57	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	13:57	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	13:57	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	13:57	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	13:57	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	13:57	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	13:57	LB137630
	Iron	23.4	+/-50	U	80.0	100	P	10/23/2025	13:57	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	13:57	LB137630
	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	13:57	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	13:57	LB137630
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/23/2025	13:57	LB137630
	Potassium	918	+/-1000	U	1600	2000	P	10/23/2025	13:57	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	13:57	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	13:57	LB137630
	Sodium	868	+/-1000	U	1000	2000	P	10/23/2025	13:57	LB137630
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/23/2025	13:57	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	13:57	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	13:57	LB137630
CCB02	Aluminum	22.6	+/-50	J	80.0	100	P	10/23/2025	15:08	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	15:08	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	15:08	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	15:08	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	15:08	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	15:08	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	15:08	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	15:08	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	15:08	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	15:08	LB137630
	Iron	34.2	+/-50	J	80.0	100	P	10/23/2025	15:08	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	15:08	LB137630
	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	15:08	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	15:08	LB137630
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/23/2025	15:08	LB137630
	Potassium	918	+/-1000	U	1600	2000	P	10/23/2025	15:08	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	15:08	LB137630

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	15:08	LB137630
	Sodium	868	+/-1000	U	1000	2000	P	10/23/2025	15:08	LB137630
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/23/2025	15:08	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	15:08	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	15:08	LB137630
CCB03	Aluminum	11.3	+/-50	U	80.0	100	P	10/23/2025	16:11	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	16:11	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	16:11	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	16:11	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	16:11	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	16:11	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	16:11	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	16:11	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	16:11	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	16:11	LB137630
	Iron	23.4	+/-50	U	80.0	100	P	10/23/2025	16:11	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	16:11	LB137630
	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	16:11	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	16:11	LB137630
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/23/2025	16:11	LB137630
	Potassium	918	+/-1000	U	1600	2000	P	10/23/2025	16:11	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	16:11	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	16:11	LB137630
	Sodium	868	+/-1000	U	1000	2000	P	10/23/2025	16:11	LB137630
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/23/2025	16:11	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	16:11	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	16:11	LB137630
CCB04	Aluminum	11.3	+/-50	U	80.0	100	P	10/23/2025	16:47	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	16:47	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	16:47	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	16:47	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	16:47	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	16:47	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	16:47	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	16:47	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	16:47	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	16:47	LB137630
	Iron	23.4	+/-50	U	80.0	100	P	10/23/2025	16:47	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	16:47	LB137630

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	16:47	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	16:47	LB137630
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/23/2025	16:47	LB137630
	Potassium	918	+/-1000	U	1600	2000	P	10/23/2025	16:47	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	16:47	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	16:47	LB137630
	Sodium	868	+/-1000	U	1000	2000	P	10/23/2025	16:47	LB137630
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/23/2025	16:47	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	16:47	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	16:47	LB137630
CCB05	Aluminum	11.3	+/-50	U	80.0	100	P	10/23/2025	17:25	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	17:25	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	17:25	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	17:25	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	17:25	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	17:25	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	17:25	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	17:25	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	17:25	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	17:25	LB137630
	Iron	23.4	+/-50	U	80.0	100	P	10/23/2025	17:25	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	17:25	LB137630
	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	17:25	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	17:25	LB137630
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/23/2025	17:25	LB137630
	Potassium	918	+/-1000	U	1600	2000	P	10/23/2025	17:25	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	17:25	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	17:25	LB137630
	Sodium	868	+/-1000	U	1000	2000	P	10/23/2025	17:25	LB137630
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/23/2025	17:25	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	17:25	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	17:25	LB137630
CCB06	Aluminum	11.3	+/-50	U	80.0	100	P	10/23/2025	18:15	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	18:15	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	18:15	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	18:15	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	18:15	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	18:15	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	18:15	LB137630

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	18:15	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	18:15	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	18:15	LB137630
	Iron	23.4	+/-50	U	80.0	100	P	10/23/2025	18:15	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	18:15	LB137630
	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	18:15	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	18:15	LB137630
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/23/2025	18:15	LB137630
	Potassium	918	+/-1000	U	1600	2000	P	10/23/2025	18:15	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	18:15	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	18:15	LB137630
	Sodium	868	+/-1000	U	1000	2000	P	10/23/2025	18:15	LB137630
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/23/2025	18:15	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	18:15	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	18:15	LB137630
CCB07	Aluminum	11.3	+/-50	U	80.0	100	P	10/23/2025	19:04	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	19:04	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	19:04	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	19:04	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	19:04	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	19:04	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	19:04	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	19:04	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	19:04	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	19:04	LB137630
	Iron	23.4	+/-50	U	80.0	100	P	10/23/2025	19:04	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	19:04	LB137630
	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	19:04	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	19:04	LB137630
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/23/2025	19:04	LB137630
	Potassium	918	+/-1000	U	1600	2000	P	10/23/2025	19:04	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	19:04	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	19:04	LB137630
	Sodium	868	+/-1000	U	1000	2000	P	10/23/2025	19:04	LB137630
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/23/2025	19:04	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	19:04	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	19:04	LB137630
CCB08	Aluminum	11.3	+/-50	U	80.0	100	P	10/23/2025	19:52	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	19:52	LB137630

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	19:52	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	19:52	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	19:52	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	19:52	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	19:52	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	19:52	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	19:52	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	19:52	LB137630
	Iron	23.4	+/-50	U	80.0	100	P	10/23/2025	19:52	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	19:52	LB137630
	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	19:52	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	19:52	LB137630
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/23/2025	19:52	LB137630
	Potassium	918	+/-1000	U	1600	2000	P	10/23/2025	19:52	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	19:52	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	19:52	LB137630
	Sodium	868	+/-1000	U	1000	2000	P	10/23/2025	19:52	LB137630
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/23/2025	19:52	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	19:52	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	19:52	LB137630
CCB09	Aluminum	11.3	+/-50	U	80.0	100	P	10/23/2025	20:43	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	20:43	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	20:43	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	20:43	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	20:43	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	20:43	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	20:43	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	20:43	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	20:43	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	20:43	LB137630
	Iron	23.4	+/-50	U	80.0	100	P	10/23/2025	20:43	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	20:43	LB137630
	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	20:43	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	20:43	LB137630
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/23/2025	20:43	LB137630
	Potassium	918	+/-1000	U	1600	2000	P	10/23/2025	20:43	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	20:43	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	20:43	LB137630
	Sodium	868	+/-1000	U	1000	2000	P	10/23/2025	20:43	LB137630

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	4.38	+/-20	U	20.0	40.0	P	10/23/2025	20:43	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	20:43	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	20:43	LB137630
CCB10	Aluminum	11.3	+/-50	U	80.0	100	P	10/23/2025	21:34	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	21:34	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	21:34	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	21:34	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	21:34	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	21:34	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	21:34	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	21:34	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	21:34	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	21:34	LB137630
	Iron	23.4	+/-50	U	80.0	100	P	10/23/2025	21:34	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	21:34	LB137630
	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	21:34	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	21:34	LB137630
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/23/2025	21:34	LB137630
	Potassium	1010	+/-1000	J*	1600	2000	P	10/23/2025	21:34	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	21:34	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	21:34	LB137630
	Sodium	1020	+/-1000	J*	1000	2000	P	10/23/2025	21:34	LB137630
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/23/2025	21:34	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	21:34	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	21:34	LB137630
	Aluminum	11.3	+/-50	U	80.0	100	P	10/23/2025	22:24	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	22:24	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	22:24	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	22:24	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	22:24	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	22:24	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	22:24	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	22:24	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	22:24	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	22:24	LB137630
	Iron	23.4	+/-50	U	80.0	100	P	10/23/2025	22:24	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	22:24	LB137630
	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	22:24	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	22:24	LB137630

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Nickel	3.06	+/-20	U	10.0	40.0	P	10/23/2025	22:24	LB137630
	Potassium	6460	+/-1000	*	1600	2000	P	10/23/2025	22:24	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	22:24	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	22:24	LB137630
	Sodium	10400	+/-1000	*	1000	2000	P	10/23/2025	22:24	LB137630
	Thallium	6.40	+/-20	J	20.0	40.0	P	10/23/2025	22:24	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	22:24	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	22:24	LB137630
CCB12	Aluminum	11.3	+/-50	U	80.0	100	P	10/23/2025	22:37	LB137630
	Antimony	6.76	+/-25	U	12.5	50.0	P	10/23/2025	22:37	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	22:37	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	22:37	LB137630
	Beryllium	0.56	+/-3	U	1.50	6.00	P	10/23/2025	22:37	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	22:37	LB137630
	Calcium	234	+/-1000	U	500	2000	P	10/23/2025	22:37	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	22:37	LB137630
	Cobalt	2.26	+/-15	U	7.50	30.0	P	10/23/2025	22:37	LB137630
	Copper	4.60	+/-10	U	16.0	20.0	P	10/23/2025	22:37	LB137630
	Iron	23.4	+/-50	U	80.0	100	P	10/23/2025	22:37	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	22:37	LB137630
	Magnesium	244	+/-1000	U	500	2000	P	10/23/2025	22:37	LB137630
	Manganese	5.94	+/-10	U	15.0	20.0	P	10/23/2025	22:37	LB137630
	Nickel	3.06	+/-20	U	10.0	40.0	P	10/23/2025	22:37	LB137630
	Potassium	4640	+/-1000	*	1600	2000	P	10/23/2025	22:37	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	22:37	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	22:37	LB137630
	Sodium	7480	+/-1000	*	1000	2000	P	10/23/2025	22:37	LB137630
	Thallium	4.38	+/-20	U	20.0	40.0	P	10/23/2025	22:37	LB137630
	Vanadium	6.26	+/-20	U	20.0	40.0	P	10/23/2025	22:37	LB137630
	Zinc	16.7	+/-20	U	15.0	40.0	P	10/23/2025	22:37	LB137630