

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137620

Review By	jaswal	Review On	10/24/2025 6:52:34 PM
Supervise By	mohan	Supervise On	10/27/2025 8:25:17 AM

STD. NAME	STD REF.#
ICAL Standard	MP87686,MP87687,MP87688,MP87689,MP87690,MP87691
ICV Standard	MP87692
CCV Standard	MP87694
ICSA Standard	
CRI Standard	MP87696
LCS Standard	
Chk Standard	MP87693,MP87695,MP87697,MP87701

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/23/25 08:00		Sagar	OK
2	S0.2	S0.2	CAL2	10/23/25 08:03		Sagar	OK
3	S2.5	S2.5	CAL3	10/23/25 08:05		Sagar	OK
4	S5	S5	CAL4	10/23/25 08:07		Sagar	OK
5	S7.5	S7.5	CAL5	10/23/25 08:10		Sagar	OK
6	S10	S10	CAL6	10/23/25 08:12		Sagar	OK
7	ICV84	ICV84	ICV	10/23/25 08:15		Sagar	OK
8	ICB84	ICB84	ICB	10/23/25 08:18		Sagar	OK
9	CCV75	CCV75	CCV	10/23/25 08:22		Sagar	OK
10	CCB75	CCB75	CCB	10/23/25 08:24		Sagar	OK
11	CRA	CRA	CRDL	10/23/25 08:27		Sagar	OK
12	HighStd	HighStd	HIGH STD	10/23/25 08:29		Sagar	OK
13	ChkStd	ChkStd	SAM	10/23/25 08:31		Sagar	OK
14	PB170219BL	PB170219BL	MB	10/23/25 08:34		Sagar	OK
15	PB170219BS	PB170219BS	LCS	10/23/25 08:36		Sagar	OK
16	Q3368-06	OUTSIDE-DUMPSTE	SAM	10/23/25 08:38		Sagar	OK
17	Q3368-07	PRESS-SLUDGE	SAM	10/23/25 08:41		Sagar	OK
18	Q3397-04	TP-7-WC	SAM	10/23/25 08:43		Sagar	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137620

Review By	jaswal	Review On	10/24/2025 6:52:34 PM
Supervise By	mohan	Supervise On	10/27/2025 8:25:17 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87686,MP87687,MP87688,MP87689,MP87690,MP87691		
ICV Standard	MP87692		
CCV Standard	MP87694		
ICSA Standard			
CRI Standard	MP87696		
LCS Standard			
Chk Standard	MP87693,MP87695,MP87697,MP87701		

19	Q3399-02	COMP-ROLLOFF	SAM	10/23/25 08:45		Sagar	OK
20	Q3402-03	41-LA	SAM	10/23/25 08:48		Sagar	OK
21	CCV76	CCV76	CCV	10/23/25 08:50		Sagar	OK
22	CCB76	CCB76	CCB	10/23/25 08:52		Sagar	OK
23	Q3402-06	41-RA	SAM	10/23/25 08:54		Sagar	OK
24	Q3402-09	41-LB	SAM	10/23/25 08:57		Sagar	OK
25	Q3402-12	41-RB	SAM	10/23/25 08:59		Sagar	OK
26	Q3403-04	MH-18	SAM	10/23/25 09:01		Sagar	OK
27	Q3404-02	MOO-25-0285-0288	SAM	10/23/25 09:04		Sagar	OK
28	Q3404-06	MOO-25-0292	SAM	10/23/25 09:06		Sagar	OK
29	Q3406-02	ARS20-0037	SAM	10/23/25 09:08		Sagar	OK
30	Q3412-04	TP-6	SAM	10/23/25 09:10		Sagar	OK
31	Q3412-04DUP	TP-6DUP	DUP	10/23/25 09:13		Sagar	OK
32	Q3412-04MS	TP-6MS	MS	10/23/25 09:16		Sagar	OK
33	CCV77	CCV77	CCV	10/23/25 09:18		Sagar	OK
34	CCB77	CCB77	CCB	10/23/25 09:20		Sagar	OK
35	Q3412-04MSD	TP-6MSD	MSD	10/23/25 09:22		Sagar	OK
36	PB170172TB	PB170172TB	MB	10/23/25 09:25		Sagar	OK
37	Q3412-04L	TP-6L	SD	10/23/25 09:27		Sagar	OK
38	Q3412-04A	TP-6A	PS	10/23/25 09:33		Sagar	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137620

Review By	jaswal	Review On	10/24/2025 6:52:34 PM
Supervise By	mohan	Supervise On	10/27/2025 8:25:17 AM

STD. NAME	STD REF.#
ICAL Standard	MP87686,MP87687,MP87688,MP87689,MP87690,MP87691
ICV Standard	MP87692
CCV Standard	MP87694
ICSA Standard	
CRI Standard	MP87696
LCS Standard	
Chk Standard	MP87693,MP87695,MP87697,MP87701

39	CCV78	CCV78	CCV	10/23/25 09:36		Sagar	OK
40	CCB78	CCB78	CCB	10/23/25 09:38		Sagar	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137630

Review By	Janvi	Review On	10/28/2025 12:32:41 PM
Supervise By	jaswal	Supervise On	10/28/2025 3:14:16 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/23/25 12:40		Jaswal	OK
2	S1	S1	CAL2	10/23/25 12:45		Jaswal	OK
3	S2	S2	CAL3	10/23/25 12:49		Jaswal	OK
4	S3	S3	CAL4	10/23/25 12:53		Jaswal	OK
5	S4	S4	CAL5	10/23/25 12:57		Jaswal	OK
6	S5	S5	CAL6	10/23/25 13:01		Jaswal	OK
7	ICV01	ICV01	ICV	10/23/25 13:05		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/23/25 13:14		Jaswal	OK
9	ICB01	ICB01	ICB	10/23/25 13:20		Jaswal	OK
10	CRI01	CRI01	CRDL	10/23/25 13:24		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/23/25 13:28		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/23/25 13:39		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/23/25 13:43		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/23/25 13:47		Jaswal	OK
15	CCV01	CCV01	CCV	10/23/25 13:52		Jaswal	OK
16	CCB01	CCB01	CCB	10/23/25 13:57		Jaswal	OK
17	LR-1	LR-1	HIGH STD	10/23/25 14:01		Jaswal	OK
18	LR-2	LR-2	HIGH STD	10/23/25 14:06		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137630

Review By	Janvi	Review On	10/28/2025 12:32:41 PM
Supervise By	jaswal	Supervise On	10/28/2025 3:14:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

19	LR-3	LR-3	HIGH STD	10/23/25 14:10		Jaswal	OK
20	Q3398-01	A748	SAM	10/23/25 14:15	Bad Injection	Jaswal	Not Ok
21	PB170190BS	PB170190BS	LCS	10/23/25 14:23		Jaswal	OK
22	Q3365-01DL	MW25DL	SAM	10/23/25 14:28	5x For Na	Jaswal	Confirms
23	Q3365-02DL	MW2DDL	SAM	10/23/25 14:32	5x For Na	Jaswal	Confirms
24	PB170190BL	PB170190BL	MB	10/23/25 14:36		Jaswal	OK
25	Q3398-01DL	A748DL	SAM	10/23/25 14:41	5X Mg,V oversaturated	Jaswal	Dilution
26	Q3398-01DL2	A748DL2	SAM	10/23/25 14:49	10x Mg still oversaturated	Jaswal	Dilution
27	CCV02	CCV02	CCV	10/23/25 15:04		Jaswal	OK
28	CCB02	CCB02	CCB	10/23/25 15:08		Jaswal	OK
29	Q3398-01DL3	A748DL	SAM	10/23/25 15:12	25x for Mg	Jaswal	Confirms
30	Q3368-06	OUTSIDE-DUMPSTE	SAM	10/23/25 15:17		Jaswal	OK
31	Q3368-07	PRESS-SLUDGE	SAM	10/23/25 15:21		Jaswal	OK
32	Q3397-04	TP-7-WC	SAM	10/23/25 15:25		Jaswal	OK
33	Q3399-02	COMP-ROLLOFF	SAM	10/23/25 15:30		Jaswal	OK
34	Q3402-03	41-LA	SAM	10/23/25 15:34		Jaswal	OK
35	Q3402-06	41-RA	SAM	10/23/25 15:38		Jaswal	OK
36	Q3402-09	41-LB	SAM	10/23/25 15:43		Jaswal	OK
37	Q3402-12	41-RB	SAM	10/23/25 15:47		Jaswal	OK
38	Q3403-04	MH-18	SAM	10/23/25 15:51		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137630

Review By	Janvi	Review On	10/28/2025 12:32:41 PM
Supervise By	jaswal	Supervise On	10/28/2025 3:14:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

39	CCV03	CCV03	CCV	10/23/25 15:56		Jaswal	OK
40	CCB03	CCB03	CCB	10/23/25 16:11		Jaswal	OK
41	Q3404-02	MOO-25-0285-0288	SAM	10/23/25 16:15		Jaswal	OK
42	Q3404-06	MOO-25-0292	SAM	10/23/25 16:19		Jaswal	OK
43	Q3416-01	NB-08-10212025	SAM	10/23/25 16:24		Jaswal	OK
44	CCV04	CCV04	CCV	10/23/25 16:43		Jaswal	OK
45	CCB04	CCB04	CCB	10/23/25 16:47		Jaswal	OK
46	Q3412-04	TP-6	SAM	10/23/25 16:51		Jaswal	OK
47	Q3412-04DUP	TP-6DUP	DUP	10/23/25 16:55		Jaswal	OK
48	Q3412-04L	TP-6L	SD	10/23/25 17:00		Jaswal	OK
49	Q3412-04MS	TP-6MS	MS	10/23/25 17:04		Jaswal	OK
50	Q3412-04MSD	TP-6MSD	MSD	10/23/25 17:08		Jaswal	OK
51	Q3412-04A	TP-6A	PS	10/23/25 17:12	0.1 ML OF M6008 AND M6017 IN 10ML	Jaswal	OK
52	Q3406-02	ARS20-0037	SAM	10/23/25 17:16		Jaswal	OK
53	CCV05	CCV05	CCV	10/23/25 17:21		Jaswal	OK
54	CCB05	CCB05	CCB	10/23/25 17:25		Jaswal	OK
55	PB170185BL	PB170185BL	MB	10/23/25 17:30		Jaswal	OK
56	PB170185BS	PB170185BS	LCS	10/23/25 17:34		Jaswal	OK
57	Q3385-04	EFF-WW	SAM	10/23/25 17:38		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137630

Review By	Janvi	Review On	10/28/2025 12:32:41 PM
Supervise By	jaswal	Supervise On	10/28/2025 3:14:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

58	Q3385-04DUP	EFF-WWDUP	DUP	10/23/25 17:42		Jaswal	OK
59	Q3385-04L	EFF-WWL	SD	10/23/25 17:46		Jaswal	OK
60	Q3385-05	EFF-WWMS	MS	10/23/25 17:51		Jaswal	OK
61	Q3385-06	EFF-WWMSD	MSD	10/23/25 17:55		Jaswal	OK
62	Q3385-04A	EFF-WWA	PS	10/23/25 17:59	0.1 ML OF M6008 AND M6017 IN 10ML	Jaswal	OK
63	Q3412-01	TP-6	SAM	10/23/25 18:03		Jaswal	OK
64	Q3412-01DUP	TP-6DUP	DUP	10/23/25 18:07		Jaswal	OK
65	CCV06	CCV06	CCV	10/23/25 18:11		Jaswal	OK
66	CCB06	CCB06	CCB	10/23/25 18:15		Jaswal	OK
67	Q3412-01L	TP-6L	SD	10/23/25 18:19		Jaswal	OK
68	Q3412-01MS	TP-6MS	MS	10/23/25 18:23		Jaswal	OK
69	Q3412-01MSD	TP-6MSD	MSD	10/23/25 18:27		Jaswal	OK
70	Q3412-01A	TP-6A	PS	10/23/25 18:31	0.1 ML OF M6008 AND M6017 IN 10ML	Jaswal	OK
71	PB170224BL	PB170224BL	MB	10/23/25 18:35		Jaswal	OK
72	PB170224BS	PB170224BS	LCS	10/23/25 18:39		Jaswal	OK
73	Q3427-01	TP-8	SAM	10/23/25 18:43		Jaswal	OK
74	Q3431-01	EO-02-10222025	SAM	10/23/25 18:47		Jaswal	OK
75	Q3431-01DUP	EO-02-10222025DUP	DUP	10/23/25 18:51		Jaswal	OK
76	Q3431-01L	EO-02-10222025L	SD	10/23/25 18:55		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137630

Review By	Janvi	Review On	10/28/2025 12:32:41 PM
Supervise By	jaswal	Supervise On	10/28/2025 3:14:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

77	CCV07	CCV07	CCV	10/23/25 18:59		Jaswal	OK
78	CCB07	CCB07	CCB	10/23/25 19:04		Jaswal	OK
79	Q3431-01MS	EO-02-10222025MS	MS	10/23/25 19:08		Jaswal	OK
80	Q3431-01MSD	EO-02-10222025MSD	MSD	10/23/25 19:12		Jaswal	OK
81	Q3431-01A	EO-02-10222025A	PS	10/23/25 19:16	0.1 ML OF M6008 AND M6017 IN 10ML	Jaswal	OK
82	Q3439-01	TP-9	SAM	10/23/25 19:19		Jaswal	OK
83	Q3439-05	TP-10	SAM	10/23/25 19:23		Jaswal	OK
84	PB170214BL	PB170214BL	MB	10/23/25 19:27		Jaswal	OK
85	PB170214BS	PB170214BS	LCS	10/23/25 19:32		Jaswal	OK
86	PB170214TB	PB170214TB	MB	10/23/25 19:36		Jaswal	OK
87	PB170223BL	PB170223BL	MB	10/23/25 19:40		Jaswal	OK
88	PB170223BS	PB170223BS	LCS	10/23/25 19:44		Jaswal	OK
89	CCV08	CCV08	CCV	10/23/25 19:48		Jaswal	OK
90	CCB08	CCB08	CCB	10/23/25 19:52		Jaswal	OK
91	Q3414-01	SB-1	SAM	10/23/25 20:01		Jaswal	OK
92	Q3414-02	SB-2	SAM	10/23/25 20:05		Jaswal	OK
93	Q3414-03	SB-3	SAM	10/23/25 20:10		Jaswal	OK
94	Q3414-04	SB-4	SAM	10/23/25 20:14		Jaswal	OK
95	Q3414-05	SB-5	SAM	10/23/25 20:18		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137630

Review By	Janvi	Review On	10/28/2025 12:32:41 PM
Supervise By	jaswal	Supervise On	10/28/2025 3:14:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

96	Q3414-06	SB-6	SAM	10/23/25 20:22		Jaswal	OK
97	Q3414-07	SB-7	SAM	10/23/25 20:26		Jaswal	OK
98	Q3414-08	SB-8	SAM	10/23/25 20:30		Jaswal	OK
99	Q3414-09	SB-9	SAM	10/23/25 20:35		Jaswal	OK
100	CCV09	CCV09	CCV	10/23/25 20:39		Jaswal	OK
101	CCB09	CCB09	CCB	10/23/25 20:43		Jaswal	OK
102	Q3414-10	SB-10	SAM	10/23/25 20:47		Jaswal	OK
103	Q3414-11	SB-11	SAM	10/23/25 20:52		Jaswal	OK
104	Q3414-12	SB-12	SAM	10/23/25 20:56		Jaswal	OK
105	Q3414-13	SB-13	SAM	10/23/25 21:00		Jaswal	OK
106	Q3427-04	TP-8	SAM	10/23/25 21:04		Jaswal	OK
107	Q3427-04DUP	TP-8DUP	DUP	10/23/25 21:08		Jaswal	OK
108	Q3427-04L	TP-8L	SD	10/23/25 21:13		Jaswal	OK
109	Q3427-04MS	TP-8MS	MS	10/23/25 21:17		Jaswal	OK
110	Q3427-04MSD	TP-8MSD	MSD	10/23/25 21:21		Jaswal	OK
111	Q3427-04A	TP-8A	PS	10/23/25 21:25	0.1 ML OF M6008 AND M6017 IN 10ML	Jaswal	OK
112	CCV10	CCV10	CCV	10/23/25 21:30		Jaswal	OK
113	CCB10	CCB10	CCB	10/23/25 21:34		Jaswal	OK
114	PB170225BL	PB170225BL	MB	10/23/25 21:38	NOT USE	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137630

Review By	Janvi	Review On	10/28/2025 12:32:41 PM
Supervise By	jaswal	Supervise On	10/28/2025 3:14:16 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

115	PB170225BS	PB170225BS	LCS	10/23/25 21:42	NOT USE	Jaswal	Not Ok
116	Q3417-01	101425	SAM	10/23/25 21:46	NOT USE	Jaswal	Not Ok
117	Q3420-05	COMPOSITE	SAM	10/23/25 21:50	NOT USE	Jaswal	Not Ok
118	Q3420-05DUP	COMPOSITEDUP	DUP	10/23/25 21:55	NOT USE	Jaswal	Not Ok
119	Q3420-05L	COMPOSITEL	SD	10/23/25 21:59	NOT USE	Jaswal	Not Ok
120	Q3420-05MS	COMPOSITEMS	MS	10/23/25 22:03	NOT USE	Jaswal	Not Ok
121	Q3420-05MSD	COMPOSITEMSD	MSD	10/23/25 22:08	NOT USE	Jaswal	Not Ok
122	Q3420-05A	COMPOSITEA	PS	10/23/25 22:12	NOT USE	Jaswal	Not Ok
123	Q3422-02	PR132-FB2-2025102	SAM	10/23/25 22:16	NOT USE	Jaswal	Not Ok
124	CCV11	CCV11	CCV	10/23/25 22:20	CCV fail for Fe,K,Na,Tl	Jaswal	OK
125	CCB11	CCB11	CCB	10/23/25 22:24	CCB fail for K,Na	Jaswal	OK
126	Q3434-02	PR132-WC2-2025102	SAM	10/23/25 22:29	NOT USE	Jaswal	Not Ok
127	CCV12	CCV12	CCV	10/23/25 22:33	CCV fail for Fe,Na	Jaswal	OK
128	CCB12	CCB12	CCB	10/23/25 22:37	CCB fail for K,Na	Jaswal	OK