



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB84	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	10/23/2025	08:18	LB137620

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB75	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	10/23/2025	08:24	LB137620
CCB76	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	10/23/2025	08:52	LB137620
CCB77	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	10/23/2025	09:20	LB137620
CCB78	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	10/23/2025	09:38	LB137620

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	13:20	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	13:20	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	13:20	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	13:20	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	13:20	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	13:20	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	13:20	LB137630

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	13:57	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	13:57	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	13:57	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	13:57	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	13:57	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	13:57	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	13:57	LB137630
CCB02	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	15:08	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	15:08	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	15:08	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	15:08	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	15:08	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	15:08	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	15:08	LB137630
CCB03	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	16:11	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	16:11	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	16:11	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	16:11	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	16:11	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	16:11	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	16:11	LB137630
CCB04	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	16:47	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	16:47	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	16:47	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	16:47	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	16:47	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	16:47	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	16:47	LB137630
CCB05	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	17:25	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	17:25	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	17:25	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	17:25	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	17:25	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	17:25	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	17:25	LB137630
CCB06	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	18:15	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	18:15	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	18:15	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	18:15	LB137630

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	18:15	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	18:15	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	18:15	LB137630
CCB07	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	19:04	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	19:04	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	19:04	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	19:04	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	19:04	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	19:04	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	19:04	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	19:52	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	19:52	LB137630
CCB08	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	19:52	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	19:52	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	19:52	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	19:52	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	19:52	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	20:43	LB137630
CCB09	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	20:43	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	20:43	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	20:43	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	20:43	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	20:43	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	20:43	LB137630
CCB10	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	21:34	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	21:34	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	21:34	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	21:34	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	21:34	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	21:34	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	21:34	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	22:24	LB137630
	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	22:24	LB137630
CCB11	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	22:24	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	22:24	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	22:24	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	22:24	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	22:24	LB137630
	Arsenic	5.12	+/-10	U	15.0	20.0	P	10/23/2025	22:37	LB137630
CCB12		5.12	+/-10	U	15.0	20.0	P	10/23/2025		

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Contract: ICES02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB12	Barium	14.6	+/-50	U	25.0	100	P	10/23/2025	22:37	LB137630
	Cadmium	0.50	+/-3	U	1.50	6.00	P	10/23/2025	22:37	LB137630
	Chromium	2.12	+/-5	U	5.00	10.0	P	10/23/2025	22:37	LB137630
	Lead	2.30	+/-6	U	9.60	12.0	P	10/23/2025	22:37	LB137630
	Selenium	9.64	+/-10	U	16.0	20.0	P	10/23/2025	22:37	LB137630
	Silver	1.62	+/-5	U	5.00	10.0	P	10/23/2025	22:37	LB137630