

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170172TB		WATER		Batch Number:		PB170219		Prep Date:	10/22/2025	
	Mercury	0.76	<2	U	1.60	2.00	CV	10/23/2025	09:25	LB137620
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170219BL		WATER		Batch Number:		PB170219		Prep Date:	10/22/2025	
	Mercury	0.076	<0.2	U	0.16	0.20	CV	10/23/2025	08:34	LB137620

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: ICE Service Group, Inc.

SDG No.: Q3399

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170214BL		WATER		Batch Number:		PB170214		Prep Date:	10/22/2025	
	Arsenic	25.6	<50	U	75.0	100	P	10/23/2025	19:27	LB137630
	Barium	72.8	<250	U	125	500	P	10/23/2025	19:27	LB137630
	Cadmium	2.50	<15	U	7.50	30.0	P	10/23/2025	19:27	LB137630
	Chromium	10.6	<25	U	25.0	50.0	P	10/23/2025	19:27	LB137630
	Lead	11.5	<30	U	48.0	60.0	P	10/23/2025	19:27	LB137630
	Selenium	48.2	<50	U	80.0	100	P	10/23/2025	19:27	LB137630
	Silver	8.10	<25	U	25.0	50.0	P	10/23/2025	19:27	LB137630
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170172TB		WATER		Batch Number:		PB170214		Prep Date:	10/22/2025	
	Arsenic	25.6	<50	U	75.0	100	P	10/23/2025	19:36	LB137630
	Barium	72.8	<250	U	125	500	P	10/23/2025	19:36	LB137630
	Cadmium	2.50	<15	U	7.50	30.0	P	10/23/2025	19:36	LB137630
	Chromium	10.6	<25	U	25.0	50.0	P	10/23/2025	19:36	LB137630
	Lead	11.5	<30	U	48.0	60.0	P	10/23/2025	19:36	LB137630
	Selenium	48.2	<50	U	80.0	100	P	10/23/2025	19:36	LB137630
	Silver	8.10	<25	U	25.0	50.0	P	10/23/2025	19:36	LB137630