

**Metals**  
**- 4 -**  
**INTERFERENCE CHECK SAMPLE**

|                                        |                          |
|----------------------------------------|--------------------------|
| <b>Client:</b> ICE Service Group, Inc. | <b>SDG No.:</b> Q3399    |
| <b>Contract:</b> ICES02                | <b>Lab Code:</b> ACE     |
| <b>ICS Source:</b> EPA                 | <b>Instrument ID:</b> P4 |

| Sample ID | Analyte  | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Low<br>Limit<br>(ug/L) | High<br>Limit<br>(ug/L) | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|--------------------|---------------|------------------------|-------------------------|------------------|------------------|---------------|
| ICSA01    | Arsenic  | 8.83           |                    |               | -20                    | 20                      | 10/23/2025       | 13:28            | LB137630      |
|           | Barium   | 2.45           | 6.0                | 41            | -94                    | 106                     | 10/23/2025       | 13:28            | LB137630      |
|           | Cadmium  | -0.071         | 1.0                | 7             | -5                     | 7                       | 10/23/2025       | 13:28            | LB137630      |
|           | Chromium | 48.5           | 52.0               | 93            | 42                     | 62                      | 10/23/2025       | 13:28            | LB137630      |
|           | Lead     | -2.09          |                    |               | -12                    | 12                      | 10/23/2025       | 13:28            | LB137630      |
|           | Selenium | 4.51           |                    |               | -20                    | 20                      | 10/23/2025       | 13:28            | LB137630      |
|           | Silver   | -2.45          |                    |               | -10                    | 10                      | 10/23/2025       | 13:28            | LB137630      |
| ICSAB01   | Arsenic  | 107            | 104                | 103           | 88.4                   | 120                     | 10/23/2025       | 13:39            | LB137630      |
|           | Barium   | 483            | 537                | 90            | 437                    | 637                     | 10/23/2025       | 13:39            | LB137630      |
|           | Cadmium  | 994            | 972                | 102           | 826                    | 1120                    | 10/23/2025       | 13:39            | LB137630      |
|           | Chromium | 550            | 542                | 102           | 460                    | 624                     | 10/23/2025       | 13:39            | LB137630      |
|           | Lead     | 46.0           | 49.0               | 94            | 37                     | 61                      | 10/23/2025       | 13:39            | LB137630      |
|           | Selenium | 55.7           | 46.0               | 121           | 26                     | 66                      | 10/23/2025       | 13:39            | LB137630      |
|           | Silver   | 206            | 201                | 102           | 170                    | 232                     | 10/23/2025       | 13:39            | LB137630      |