

rteres

Instrument :

FID_F

LabSampleId :

10 TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 10/03/2025

Supervised By :mohammad ahmed 10/04/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF100225
Data File : FF016512.D
Signal(s) : FID2B.ch
Acq On : 02 Oct 2025 12:54
Sample : 10 TRPH STD
Misc :
ALS Vial : 34 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100225.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.048	1.945	2.110	BB	103891	1231207	87.96%	5.627%
2	4.583	4.547	4.639	BB	120532	1248028	89.16%	5.704%
3	6.762	6.730	6.822	BB	125135	1268209	90.60%	5.796%
4	8.595	8.560	8.675	BB	119309	1279939	91.44%	5.849%
5	10.204	10.170	10.287	BB	115546	1309042	93.52%	5.982%
6	11.650	11.607	11.720	BB	116749	1366007	97.59%	6.243%
7	12.963	12.925	13.039	BB	115010	1399742	100.00%	6.397%
8	14.164	14.124	14.224	BB	111215	1372675	98.07%	6.273%
9	15.065	15.025	15.098	BV	89557	1196583	85.49%	5.469%
10	15.270	15.232	15.325	BB	108723	1385923	99.01%	6.334%
11	16.294	16.249	16.350	BB	107479	1386979	99.09%	6.339%
12	17.248	17.200	17.302	BB	100840	1372048	98.02%	6.270%
13	18.139	18.079	18.185	BB	101339	1393151	99.53%	6.367%
14	18.975	18.940	19.040	BB	92292	1361594	97.27%	6.223%
15	19.762	19.720	19.803	BV	76202	1111280	79.39%	5.079%
16	20.504	20.460	20.535	BV	59839	884066	63.16%	4.040%
17	21.263	21.200	21.320	BB	40494	731150	52.23%	3.341%
18	22.211	22.140	22.257	BV	24454	583651	41.70%	2.667%
Sum of corrected areas:						21881275		

FF100225.M Fri Oct 03 03:49:27 2025