

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Contract: ICES02
 ProjectID: NWIRP Northrup Grumman Site – Bethpage, NY
 Lab Code: ACE SDG No.: Q3399

Calibration Sequence : FF100225		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	112385153	112385	FF016509.D
500	63771192	127542	FF016510.D
200	27823338	139117	FF016511.D
100	13388592	133886	FF016512.D
50	6367675	127354	FF016513.D
AVG RF : 128057		% RSD : 7.835	AVG RT : 15.0682