

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB170221BS

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: PB170221BS

Date(s) Analyzed: 10/23/2025 10/23/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
DICHLORPROP	1	8.17	8.12	8.22	158	2.6
	2	8.66	8.61	8.71	154	
2,4-D	1	8.41	8.36	8.46	164	8.3
	2	9.01	8.96	9.06	151	
2,4,5-TP(Silvex)	1	9.28	9.23	9.33	161	1.9
	2	9.90	9.85	9.95	158	
2,4,5-T	1	9.58	9.53	9.63	164	4.4
	2	10.33	10.28	10.38	157	
2,4-DB	1	10.16	10.11	10.21	164	13
	2	10.90	10.85	10.95	144	
Dinoseb	1	11.35	11.30	11.40	153	2.6
	2	11.27	11.22	11.32	149	
DICAMBA	1	7.46	7.41	7.51	167	7.5
	2	7.94	7.89	7.99	155	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TP-6MS

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: Q3412-01MS

Date(s) Analyzed: 10/24/2025 10/24/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4,5-T	1	9.58	9.53	9.63	21.3	18.3
	2	10.33	10.28	10.38	25.6	
2,4,5-TP(Silvex)	1	9.28	9.23	9.33	18.0	84.8
	2	9.88	9.83	9.93	44.5	
2,4-D	1	8.39	8.34	8.44	200	157.8
	2	9.01	8.96	9.06	23.6	
DICHLORPROP	1	8.17	8.12	8.22	25.5	35
	2	8.66	8.61	8.71	17.9	
DICAMBA	1	7.46	7.41	7.51	27.5	20
	2	7.93	7.88	7.98	33.6	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

TP-6MSD

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: Q3412-01MSD

Date(s) Analyzed: 10/24/2025 10/24/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4-D	1	8.38	8.33	8.43	200	153.8
	2	9.01	8.96	9.06	26.1	
2,4,5-TP(Silvex)	1	9.28	9.23	9.33	18.0	86.4
	2	9.88	9.83	9.93	45.4	
2,4,5-T	1	9.58	9.53	9.63	20.9	22.9
	2	10.33	10.28	10.38	26.3	
DICHLORPROP	1	8.17	8.12	8.22	27.2	30.5
	2	8.66	8.61	8.71	20.0	
DICAMBA	1	7.46	7.41	7.51	27.4	17
	2	7.94	7.89	7.99	32.5	