

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/23/2025

Initial Calibration Date(s): 10/13/2025 10/13/2025

Continuing Calib Time: 10:04

Initial Calibration Time(s): 10:53 12:29

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.46	7.46	7.36	7.56	0.00
2,4-DCAA	7.28	7.28	7.18	7.38	0.00
DICHLORPROP	8.17	8.17	8.07	8.27	0.00
2,4-D	8.41	8.41	8.31	8.51	0.00
2,4,5-TP(Silvex)	9.28	9.28	9.18	9.38	0.00
2,4,5-T	9.58	9.58	9.48	9.68	0.00
2,4-DB	10.15	10.15	10.05	10.25	0.00
Dinoseb	11.34	11.35	11.25	11.45	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/23/2025

Initial Calibration Date(s): 10/13/2025 10/13/2025

Continuing Calib Time: 10:04

Initial Calibration Time(s): 10:53 12:29

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.94	7.94	7.84	8.04	0.00
2,4-DCAA	7.74	7.74	7.64	7.84	0.00
DICHLORPROP	8.66	8.66	8.56	8.76	0.00
2,4-D	9.00	9.00	8.90	9.10	0.00
2,4,5-TP(Silvex)	9.89	9.89	9.79	9.99	0.00
2,4,5-T	10.32	10.32	10.22	10.42	0.00
2,4-DB	10.89	10.89	10.79	10.99	0.00
Dinoseb	11.26	11.26	11.16	11.36	0.00