

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032132.D **Time Analyzed:** 10:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.576	9.478	9.678	799.240	712.500	12.2
2,4,5-TP(Silvex)	9.276	9.180	9.380	768.020	712.500	7.8
2,4-D	8.409	8.312	8.512	751.270	705.000	6.6
2,4-DB	10.152	10.054	10.254	722.460	712.500	1.4
2,4-DCAA	7.280	7.182	7.382	810.260	750.000	8.0
DICAMBA	7.462	7.364	7.564	779.910	705.000	10.6
DICHLORPROP	8.170	8.073	8.273	739.890	705.000	4.9
Dinoseb	11.343	11.245	11.445	747.900	705.000	6.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032132.D **Time Analyzed:** 10:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.321	10.221	10.421	795.800	712.500	11.7
2,4,5-TP(Silvex)	9.892	9.792	9.992	773.160	712.500	8.5
2,4-D	9.001	8.902	9.102	774.470	705.000	9.9
2,4-DB	10.889	10.789	10.989	805.500	712.500	13.1
2,4-DCAA	7.744	7.644	7.844	877.970	750.000	17.1
DICAMBA	7.939	7.838	8.038	766.180	705.000	8.7
DICHLORPROP	8.658	8.558	8.758	756.080	705.000	7.2
Dinoseb	11.263	11.163	11.363	777.120	705.000	10.2