

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032144.D **Time Analyzed:** 15:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.575	9.478	9.678	795.180	712.500	11.6
2,4,5-TP(Silvex)	9.277	9.180	9.380	756.490	712.500	6.2
2,4-D	8.407	8.312	8.512	764.020	705.000	8.4
2,4-DB	10.150	10.054	10.254	759.070	712.500	6.5
2,4-DCAA	7.279	7.182	7.382	794.670	750.000	6.0
DICAMBA	7.462	7.364	7.564	759.800	705.000	7.8
DICHLORPROP	8.170	8.073	8.273	729.090	705.000	3.4
Dinoseb	11.343	11.245	11.445	706.040	705.000	0.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032144.D **Time Analyzed:** 15:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.321	10.221	10.421	793.440	712.500	11.4
2,4,5-TP(Silvex)	9.892	9.792	9.992	756.750	712.500	6.2
2,4-D	9.000	8.902	9.102	778.940	705.000	10.5
2,4-DB	10.889	10.789	10.989	795.650	712.500	11.7
2,4-DCAA	7.743	7.644	7.844	878.420	750.000	17.1
DICAMBA	7.939	7.838	8.038	753.230	705.000	6.8
DICHLORPROP	8.658	8.558	8.758	744.020	705.000	5.5
Dinoseb	11.264	11.163	11.363	738.560	705.000	4.8