

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ICES02  
**Lab Code:** ACE **SDG NO.:** Q3399  
**GC Column:** RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 10/24/2025  
**Lab Sample No.:** HSTDCCC750 **Data File :** PS032154.D **Time Analyzed:** 14:28

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|------------------|--------|-----------|--------|--------------------|-------------------|-----|
|                  |        | FROM      | TO     |                    |                   |     |
| 2,4,5-T          | 9.574  | 9.478     | 9.678  | 776.340            | 712.500           | 9.0 |
| 2,4,5-TP(Silvex) | 9.275  | 9.180     | 9.380  | 738.430            | 712.500           | 3.6 |
| 2,4-D            | 8.407  | 8.312     | 8.512  | 733.220            | 705.000           | 4.0 |
| 2,4-DB           | 10.149 | 10.054    | 10.254 | 718.400            | 712.500           | 0.8 |
| 2,4-DCAA         | 7.278  | 7.182     | 7.382  | 790.230            | 750.000           | 5.4 |
| DICAMBA          | 7.460  | 7.364     | 7.564  | 716.720            | 705.000           | 1.7 |
| DICHLORPROP      | 8.169  | 8.073     | 8.273  | 710.070            | 705.000           | 0.7 |
| Dinoseb          | 11.341 | 11.245    | 11.445 | 713.910            | 705.000           | 1.3 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ICES02  
**Lab Code:** ACE **SDG NO.:** Q3399  
**GC Column:** RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 10/24/2025  
**Lab Sample No.:** HSTDCCC750 **Data File :** PS032154.D **Time Analyzed:** 14:28

| COMPOUND         | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|----------------------|--------|--------------------|-------------------|------|
| 2,4,5-T          | 10.320 | 10.221               | 10.421 | 783.510            | 712.500           | 10.0 |
| 2,4,5-TP(Silvex) | 9.891  | 9.792                | 9.992  | 750.010            | 712.500           | 5.3  |
| 2,4-D            | 9.000  | 8.902                | 9.102  | 761.860            | 705.000           | 8.1  |
| 2,4-DB           | 10.889 | 10.789               | 10.989 | 752.770            | 712.500           | 5.7  |
| 2,4-DCAA         | 7.743  | 7.644                | 7.844  | 856.540            | 750.000           | 14.2 |
| DICAMBA          | 7.938  | 7.838                | 8.038  | 739.720            | 705.000           | 4.9  |
| DICHLORPROP      | 8.658  | 8.558                | 8.758  | 728.780            | 705.000           | 3.4  |
| Dinoseb          | 11.263 | 11.163               | 11.363 | 747.230            | 705.000           | 6.0  |