

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032160.D **Time Analyzed:** 17:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.571	9.478	9.678	747.470	712.500	4.9
2,4,5-TP(Silvex)	9.274	9.180	9.380	736.630	712.500	3.4
2,4-D	8.405	8.312	8.512	746.330	705.000	5.9
2,4-DB	10.147	10.054	10.254	728.700	712.500	2.3
2,4-DCAA	7.277	7.182	7.382	793.030	750.000	5.7
DICAMBA	7.460	7.364	7.564	714.540	705.000	1.4
DICHLORPROP	8.167	8.073	8.273	708.050	705.000	0.4
Dinoseb	11.341	11.245	11.445	699.200	705.000	-0.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032160.D **Time Analyzed:** 17:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.319	10.221	10.421	785.300	712.500	10.2
2,4,5-TP(Silvex)	9.890	9.792	9.992	747.340	712.500	4.9
2,4-D	8.998	8.902	9.102	768.550	705.000	9.0
2,4-DB	10.887	10.789	10.989	762.850	712.500	7.1
2,4-DCAA	7.742	7.644	7.844	858.900	750.000	14.5
DICAMBA	7.937	7.838	8.038	745.430	705.000	5.7
DICHLORPROP	8.656	8.558	8.758	733.520	705.000	4.0
Dinoseb	11.262	11.163	11.363	743.490	705.000	5.5