

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Report of Analysis

Client: ICE Service Group, Inc.

Project: NWIRP Northrup Grumman Site – Bethpage, NY

Client Sample ID: TP-6MSD Lab Sample ID: Q3412-01MSD

Analytical Method: 8151A

Sample Wt/Vol: 30.04 g Final Vol: 10000 uL Prep Method: 8151A Prep Date 10/22/25 Date Collected: 10/20/25
Date Received: 10/21/25
SDG No.: Q3399
Matrix: SOIL
% Solid: 89.7

Test: Herbicide

CAS Number	· Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
1918-00-9	DICAMBA	32.5	J	1	8.60	36.7	74.6	ug/Kg	10/24/25 16:14	PB170221
120-36-5	DICHLORPROP	27.2	J	1	14.3	36.7	74.6	ug/Kg	10/24/25 16:14	PB170221
94-75-7	2,4-D	200	P	1	10.1	36.7	74.6	ug/Kg	10/24/25 16:14	PB170221
93-72-1	2,4,5-TP (Silvex)	45.4	JP	1	10.1	36.7	74.6	ug/Kg	10/24/25 16:14	PB170221
93-76-5	2,4,5-T	26.3	J	1	9.70	36.7	74.6	ug/Kg	10/24/25 16:14	PB170221
94-82-6	2,4-DB	36.7	U	1	26.9	36.7	74.6	ug/Kg	10/24/25 16:14	PB170221
88-85-7	DINOSEB	36.7	U	1	12.0	36.7	74.6	ug/Kg	10/24/25 16:14	PB170221
SURROGATES										
19719-28-9	2,4-DCAA	196			27 - 122		39%	SPK: 500		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit