

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS101325

Review By	Abdul	Review On	10/14/2025 8:45:14 AM
Supervise By	mohammad	Supervise On	10/15/2025 1:15:14 AM
SubDirectory	PS101325	HP Acquire Method	HP Processing Method ps101325 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24923
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031981.D	13 Oct 2025 09:35		AR\AJ	Ok
2	I.BLK	I.BLK	PS031982.D	13 Oct 2025 10:29		AR\AJ	Ok,M
3	HSTDICC200	HSTDICC200	PS031983.D	13 Oct 2025 10:53		AR\AJ	Ok,M
4	HSTDICC500	HSTDICC500	PS031984.D	13 Oct 2025 11:17		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS031985.D	13 Oct 2025 11:41	Method fail for comp@10	AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS031986.D	13 Oct 2025 12:05		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS031987.D	13 Oct 2025 12:29		AR\AJ	Ok,M
8	HSTDICV750	ICVPS101325	PS031988.D	13 Oct 2025 12:52		AR\AJ	Ok
9	I.BLK	I.BLK	PS031989.D	13 Oct 2025 13:16		AR\AJ	Ok,M
10	HSTDCCC750	HSTDCCC750	PS031990.D	13 Oct 2025 13:41		AR\AJ	Ok
11	Q3297-01	TP1	PS031991.D	13 Oct 2025 14:05		AR\AJ	Ok,M
12	Q3297-11	TP2	PS031992.D	13 Oct 2025 14:29		AR\AJ	Ok
13	PB170051BL	PB170051BL	PS031993.D	13 Oct 2025 14:53		AR\AJ	Ok
14	PB170051BS	PB170051BS	PS031994.D	13 Oct 2025 15:17	Comp#3,9,10,15 recovery fail	AR\AJ	Not Ok
15	Q3297-01MS	TP1MS	PS031995.D	13 Oct 2025 15:41	some compound recovery fail	AR\AJ	Ok,M
16	Q3297-01MSD	TP1MSD	PS031996.D	13 Oct 2025 16:06	some compound recovery fail	AR\AJ	Ok,M
17	Q3305-01	AU-06-100725	PS031997.D	13 Oct 2025 16:30		AR\AJ	Ok
18	I.BLK	I.BLK	PS031998.D	13 Oct 2025 16:54		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS101325

Review By	Abdul	Review On	10/14/2025 8:45:14 AM
Supervise By	mohammad	Supervise On	10/15/2025 1:15:14 AM
SubDirectory	PS101325	HP Acquire Method	HP Processing Method ps101325 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921
CCC Internal Standard/PEM	PP24559
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24923

19	HSTDCCC750	HSTDCCC750	PS031999.D	13 Oct 2025 17:18		ARIAJ	Ok,M
----	------------	------------	------------	-------------------	--	-------	------

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102325

Review By	Review On
Supervise By	Supervise On
SubDirectory	PS102325
HP Acquire Method	HP Processing Method
	ps101325 8151
STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24923
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS032130.D	23 Oct 2025 09:16		AR\AJ	Ok
2	I.BLK	I.BLK	PS032131.D	23 Oct 2025 09:40		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS032132.D	23 Oct 2025 10:04		AR\AJ	Ok
4	PB170220BL	PB170220BL	PS032133.D	23 Oct 2025 10:31		AR\AJ	Ok
5	PB170220BS	PB170220BS	PS032134.D	23 Oct 2025 10:55		AR\AJ	Ok,M
6	Q3399-01	COMP-ROLLOFF	PS032135.D	23 Oct 2025 11:20		AR\AJ	Ok,M
7	Q3399-02	COMP-ROLLOFF	PS032136.D	23 Oct 2025 11:44		AR\AJ	Ok,M
8	PB170172TB	PB170172TB	PS032137.D	23 Oct 2025 12:09	Surrogate low in both column	AR\AJ	Not Ok
9	PB170221BL	PB170221BL	PS032138.D	23 Oct 2025 12:33		AR\AJ	Ok,M
10	PB170221BS	PB170221BS	PS032139.D	23 Oct 2025 12:58		AR\AJ	Ok,M
11	Q3412-01	TP-6	PS032140.D	23 Oct 2025 13:22		AR\AJ	Ok,M
12	Q3412-01MS	TP-6MS	PS032141.D	23 Oct 2025 13:47	recovery fail for some compounds, comp # 1 & 14 having F flag in 1st column	AR\AJ	Not Ok
13	Q3412-01MSD	TP-6MSD	PS032142.D	23 Oct 2025 14:11	recovery fail for some compounds, comp # 1 & 14 having F flag in 1st column	AR\AJ	Not Ok
14	I.BLK	I.BLK	PS032143.D	23 Oct 2025 14:36		AR\AJ	Ok
15	HSTDCCC750	HSTDCCC750	PS032144.D	23 Oct 2025 15:00		AR\AJ	Ok,M
16	Q3399-02MS	COMP-ROLLOFFMS	PS032145.D	23 Oct 2025 15:25	recovery fail for some compounds	AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102325

Review By	Review On
Supervise By	Supervise On
SubDirectory	PS102325
HP Acquire Method	HP Processing Method
	ps101325 8151
STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24923
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

17	Q3399-02MSD	COMP-ROLLOFFMSD	PS032146.D	23 Oct 2025 15:49	recovery fail for some compounds	AR\AJ	Ok,M
18	I.BLK	I.BLK	PS032147.D	23 Oct 2025 16:49		AR\AJ	Ok
19	HSTDCCC750	HSTDCCC750	PS032148.D	23 Oct 2025 17:13		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS102425

Review By	Abdul	Review On	10/27/2025 9:16:47 AM
Supervise By	mohammad	Supervise On	10/28/2025 6:10:55 AM
SubDirectory	PS102425	HP Acquire Method	HP Processing Method ps101325 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24556,PP24557,PP24558,PP24559,PP24560,PP24921
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24923
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS032149.D	24 Oct 2025 09:46		AR\AJ	Ok
2	I.BLK	I.BLK	PS032150.D	24 Oct 2025 10:10		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS032151.D	24 Oct 2025 10:34		AR\AJ	Ok,M
4	PB170172TB	PB170172TB	PS032152.D	24 Oct 2025 12:28		AR\AJ	Ok
5	I.BLK	I.BLK	PS032153.D	24 Oct 2025 14:04		AR\AJ	Ok
6	HSTDCCC750	HSTDCCC750	PS032154.D	24 Oct 2025 14:28		AR\AJ	Ok,M
7	Q3412-01MS	TP-6MS	PS032155.D	24 Oct 2025 15:50	F Flag in comp#1,3,14. some compound recovery fail	AR\AJ	Ok,M
8	Q3412-01MSD	TP-6MSD	PS032156.D	24 Oct 2025 16:14	F Flag in comp#1,3,14. some compound recovery fail , RPD Fail	AR\AJ	Ok,M
9	Q3412-01MSRE	TP-6MSRE	PS032157.D	24 Oct 2025 16:39	F Flag in comp#1,3,14. some compound recovery fail ,Not used	AR\AJ	Not Ok
10	Q3412-01MSDRE	TP-6MSDRE	PS032158.D	24 Oct 2025 17:04	F Flag in comp#1,3,14. some compound recovery fail , RPD Fail,Not used	AR\AJ	Not Ok
11	I.BLK	I.BLK	PS032159.D	24 Oct 2025 17:28		AR\AJ	Ok
12	HSTDCCC750	HSTDCCC750	PS032160.D	24 Oct 2025 17:53		AR\AJ	Ok

M : Manual Integration