

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075936.D **Time Analyzed:** 14:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.799	5.695	5.895	470.840	500.000	-5.8
Aroclor-1016-2	5.821	5.717	5.917	474.670	500.000	-5.1
Aroclor-1016-3	5.883	5.780	5.980	473.780	500.000	-5.2
Aroclor-1016-4	5.980	5.877	6.077	476.450	500.000	-4.7
Aroclor-1016-5	6.273	6.169	6.369	466.280	500.000	-6.7
Aroclor-1260-1	7.391	7.287	7.487	500.580	500.000	0.1
Aroclor-1260-2	7.644	7.540	7.740	474.300	500.000	-5.1
Aroclor-1260-3	8.002	7.899	8.099	487.640	500.000	-2.5
Aroclor-1260-4	8.231	8.127	8.327	528.640	500.000	5.7
Aroclor-1260-5	8.556	8.452	8.652	487.680	500.000	-2.5
Decachlorobiphenyl	10.425	10.319	10.519	48.820	50.000	-2.4
Tetrachloro-m-xylene	4.646	4.542	4.742	46.990	50.000	-6.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075936.D **Time Analyzed:** 14:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.879	4.778	4.978	528.020	500.000	5.6
Aroclor-1016-2	4.937	4.836	5.036	553.110	500.000	10.6
Aroclor-1016-3	5.057	4.954	5.154	499.640	500.000	-0.1
Aroclor-1016-4	5.098	4.996	5.196	547.630	500.000	9.5
Aroclor-1016-5	5.312	5.210	5.410	513.760	500.000	2.8
Aroclor-1260-1	6.339	6.238	6.438	507.530	500.000	1.5
Aroclor-1260-2	6.527	6.425	6.625	510.050	500.000	2.0
Aroclor-1260-3	6.680	6.578	6.778	527.650	500.000	5.5
Aroclor-1260-4	7.149	7.047	7.247	521.350	500.000	4.3
Aroclor-1260-5	7.390	7.287	7.487	505.960	500.000	1.2
Decachlorobiphenyl	8.785	8.682	8.882	46.960	50.000	-6.1
Tetrachloro-m-xylene	3.784	3.681	3.881	49.710	50.000	-0.6