

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075950.D **Time Analyzed:** 19:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.800	5.695	5.895	488.030	500.000	-2.4
Aroclor-1016-2	5.821	5.717	5.917	491.100	500.000	-1.8
Aroclor-1016-3	5.884	5.780	5.980	491.820	500.000	-1.6
Aroclor-1016-4	5.981	5.877	6.077	496.440	500.000	-0.7
Aroclor-1016-5	6.274	6.169	6.369	487.670	500.000	-2.5
Aroclor-1260-1	7.392	7.287	7.487	525.000	500.000	5.0
Aroclor-1260-2	7.644	7.540	7.740	492.510	500.000	-1.5
Aroclor-1260-3	8.002	7.899	8.099	511.510	500.000	2.3
Aroclor-1260-4	8.231	8.127	8.327	553.360	500.000	10.7
Aroclor-1260-5	8.556	8.452	8.652	504.440	500.000	0.9
Decachlorobiphenyl	10.425	10.319	10.519	50.330	50.000	0.7
Tetrachloro-m-xylene	4.647	4.542	4.742	48.150	50.000	-3.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075950.D **Time Analyzed:** 19:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.880	4.778	4.978	541.730	500.000	8.3
Aroclor-1016-2	4.939	4.836	5.036	554.500	500.000	10.9
Aroclor-1016-3	5.058	4.954	5.154	528.400	500.000	5.7
Aroclor-1016-4	5.099	4.996	5.196	555.970	500.000	11.2
Aroclor-1016-5	5.312	5.210	5.410	511.680	500.000	2.3
Aroclor-1260-1	6.340	6.238	6.438	521.230	500.000	4.2
Aroclor-1260-2	6.527	6.425	6.625	540.440	500.000	8.1
Aroclor-1260-3	6.680	6.578	6.778	546.110	500.000	9.2
Aroclor-1260-4	7.150	7.047	7.247	577.530	500.000	15.5
Aroclor-1260-5	7.389	7.287	7.487	556.220	500.000	11.2
Decachlorobiphenyl	8.784	8.682	8.882	48.760	50.000	-2.5
Tetrachloro-m-xylene	3.784	3.681	3.881	51.410	50.000	2.8