

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075960.D **Time Analyzed:** 23:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.797	5.695	5.895	479.930	500.000	-4.0
Aroclor-1016-2	5.818	5.717	5.917	473.690	500.000	-5.3
Aroclor-1016-3	5.881	5.780	5.980	479.690	500.000	-4.1
Aroclor-1016-4	5.978	5.877	6.077	484.760	500.000	-3.0
Aroclor-1016-5	6.270	6.169	6.369	460.940	500.000	-7.8
Aroclor-1260-1	7.388	7.287	7.487	500.800	500.000	0.2
Aroclor-1260-2	7.641	7.540	7.740	467.710	500.000	-6.5
Aroclor-1260-3	8.000	7.899	8.099	485.880	500.000	-2.8
Aroclor-1260-4	8.227	8.127	8.327	517.750	500.000	3.6
Aroclor-1260-5	8.553	8.452	8.652	475.440	500.000	-4.9
Decachlorobiphenyl	10.418	10.319	10.519	48.370	50.000	-3.3
Tetrachloro-m-xylene	4.644	4.542	4.742	47.950	50.000	-4.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/21/2025 10/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075960.D **Time Analyzed:** 23:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.879	4.778	4.978	548.720	500.000	9.7
Aroclor-1016-2	4.936	4.836	5.036	627.760	500.000	25.6
Aroclor-1016-3	5.056	4.954	5.154	557.340	500.000	11.5
Aroclor-1016-4	5.097	4.996	5.196	578.510	500.000	15.7
Aroclor-1016-5	5.311	5.210	5.410	499.120	500.000	-0.2
Aroclor-1260-1	6.338	6.238	6.438	435.450	500.000	-12.9
Aroclor-1260-2	6.525	6.425	6.625	459.310	500.000	-8.1
Aroclor-1260-3	6.678	6.578	6.778	488.450	500.000	-2.3
Aroclor-1260-4	7.147	7.047	7.247	477.190	500.000	-4.6
Aroclor-1260-5	7.387	7.287	7.487	451.150	500.000	-9.8
Decachlorobiphenyl	8.782	8.682	8.882	42.880	50.000	-14.2
Tetrachloro-m-xylene	3.783	3.681	3.881	54.340	50.000	8.7