

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP075886.D	21 Oct 2025 08:33	YP\AJ	Ok
2	I.BLK	PP075887.D	21 Oct 2025 08:49	YP\AJ	Ok,M
3	AR1660ICC1000	PP075888.D	21 Oct 2025 09:06	YP\AJ	Ok
4	AR1660ICC750	PP075889.D	21 Oct 2025 09:22	YP\AJ	Ok
5	AR1660ICC500	PP075890.D	21 Oct 2025 09:38	YP\AJ	Ok
6	AR1660ICC250	PP075891.D	21 Oct 2025 09:54	YP\AJ	Ok
7	AR1660ICC050	PP075892.D	21 Oct 2025 10:11	YP\AJ	Ok,M
8	AR1221ICC500	PP075893.D	21 Oct 2025 10:27	YP\AJ	Ok
9	AR1232ICC500	PP075894.D	21 Oct 2025 10:43	YP\AJ	Ok,M
10	AR1242ICC1000	PP075895.D	21 Oct 2025 10:59	YP\AJ	Ok,M
11	AR1242ICC750	PP075896.D	21 Oct 2025 11:16	YP\AJ	Ok,M
12	AR1242ICC500	PP075897.D	21 Oct 2025 11:48	YP\AJ	Ok,M
13	AR1242ICC250	PP075898.D	21 Oct 2025 12:04	YP\AJ	Ok,M
14	AR1242ICC050	PP075899.D	21 Oct 2025 12:21	YP\AJ	Ok,M
15	AR1248ICC1000	PP075900.D	21 Oct 2025 12:41	YP\AJ	Ok,M
16	AR1248ICC750	PP075901.D	21 Oct 2025 12:57	YP\AJ	Ok,M
17	AR1248ICC500	PP075902.D	21 Oct 2025 13:13	YP\AJ	Ok,M
18	AR1248ICC250	PP075903.D	21 Oct 2025 13:30	YP\AJ	Ok,M
19	AR1248ICC050	PP075904.D	21 Oct 2025 13:46	YP\AJ	Ok,M
20	AR1254ICC1000	PP075905.D	21 Oct 2025 14:02	YP\AJ	Ok,M
21	AR1254ICC750	PP075906.D	21 Oct 2025 14:18	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102125

Review By	yogesh	Review On	10/21/2025 12:01:15 PM
Supervise By	mohammad	Supervise On	10/25/2025 7:28:37 AM
SubDirectory	PP102125	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP075907.D	21 Oct 2025 14:35	YP\AJ	Ok,M
23	AR1254ICC250	PP075908.D	21 Oct 2025 14:51	YP\AJ	Ok,M
24	AR1254ICC050	PP075909.D	21 Oct 2025 15:07	YP\AJ	Ok,M
25	AR1262ICC500	PP075910.D	21 Oct 2025 15:40	YP\AJ	Ok
26	AR1268ICC1000	PP075911.D	21 Oct 2025 15:56	YP\AJ	Ok
27	AR1268ICC750	PP075912.D	21 Oct 2025 16:12	YP\AJ	Ok
28	AR1268ICC500	PP075913.D	21 Oct 2025 16:29	YP\AJ	Ok
29	AR1268ICC250	PP075914.D	21 Oct 2025 16:45	YP\AJ	Ok
30	AR1268ICC050	PP075915.D	21 Oct 2025 17:01	YP\AJ	Ok
31	PP102125ICV500	PP075916.D	21 Oct 2025 17:34	YP\AJ	Ok
32	AR1242ICV500	PP075917.D	21 Oct 2025 18:06	YP\AJ	Ok,M
33	AR1248ICV500	PP075918.D	21 Oct 2025 18:39	YP\AJ	Ok,M
34	AR1254ICV500	PP075919.D	22 Oct 2025 08:49	YP\AJ	Ok,M
35	AR1268ICV500	PP075920.D	22 Oct 2025 09:07	YP\AJ	Ok
36	DDT ANALOG	PP075921.D	22 Oct 2025 09:25	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102225

Review By	yogesh	Review On	10/22/2025 11:56:05 AM
Supervise By	mohammad	Supervise On	10/27/2025 7:09:09 AM
SubDirectory	PP102225	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP075922.D	22 Oct 2025 10:15	YP\AJ	Ok
2	AR1660CCC500	PP075923.D	22 Oct 2025 10:32	YP\AJ	Ok
3	AR1242CCC500	PP075924.D	22 Oct 2025 10:52	YP\AJ	Ok,M
4	AR1248CCC500	PP075925.D	22 Oct 2025 11:08	YP\AJ	Ok,M
5	AR1254CCC500	PP075926.D	22 Oct 2025 11:24	YP\AJ	Ok,M
6	I.BLK	PP075927.D	22 Oct 2025 11:40	YP\AJ	Ok
7	Q3407-13	PP075928.D	22 Oct 2025 11:57	YP\AJ	Ok,M
8	Q3407-14	PP075929.D	22 Oct 2025 12:13	YP\AJ	Ok,M
9	Q3407-15	PP075930.D	22 Oct 2025 12:29	YP\AJ	Ok
10	Q3407-16	PP075931.D	22 Oct 2025 12:45	YP\AJ	Ok,M
11	Q3407-17	PP075932.D	22 Oct 2025 13:02	YP\AJ	Ok,M
12	Q3407-18	PP075933.D	22 Oct 2025 13:18	YP\AJ	Ok,M
13	Q3407-19	PP075934.D	22 Oct 2025 13:34	YP\AJ	Ok
14	Q3407-20	PP075935.D	22 Oct 2025 13:51	YP\AJ	Ok
15	AR1660CCC500	PP075936.D	22 Oct 2025 14:56	YP\AJ	Ok
16	AR1242CCC500	PP075937.D	22 Oct 2025 15:12	YP\AJ	Ok,M
17	AR1248CCC500	PP075938.D	22 Oct 2025 15:28	YP\AJ	Ok,M
18	AR1254CCC500	PP075939.D	22 Oct 2025 15:45	YP\AJ	Ok
19	I.BLK	PP075940.D	22 Oct 2025 16:01	YP\AJ	Ok
20	PB170199BL	PP075941.D	22 Oct 2025 16:17	YP\AJ	Ok
21	PB170199BS	PP075942.D	22 Oct 2025 16:34	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP102225

Review By	yogesh	Review On	10/22/2025 11:56:05 AM
Supervise By	mohammad	Supervise On	10/27/2025 7:09:09 AM
SubDirectory	PP102225	HP Acquire Method	HP Processing Method PP102125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3415-01	PP075943.D	22 Oct 2025 16:50	YP\AJ	Dilution
23	Q3415-02	PP075944.D	22 Oct 2025 17:06	YP\AJ	Dilution
24	Q3415-03	PP075945.D	22 Oct 2025 17:22	YP\AJ	Dilution
25	Q3418-01	PP075946.D	22 Oct 2025 17:39	YP\AJ	Ok,M
26	Q3418-02	PP075947.D	22 Oct 2025 17:55	YP\AJ	Ok
27	Q3418-03	PP075948.D	22 Oct 2025 18:11	YP\AJ	Ok,M
28	Q3399-01	PP075949.D	22 Oct 2025 18:27	YP\AJ	Ok
29	AR1660CCC500	PP075950.D	22 Oct 2025 19:33	YP\AJ	Ok
30	AR1242CCC500	PP075951.D	22 Oct 2025 20:21	YP\AJ	Ok,M
31	AR1248CCC500	PP075952.D	22 Oct 2025 20:38	YP\AJ	Ok,M
32	AR1254CCC500	PP075953.D	22 Oct 2025 20:54	YP\AJ	Ok,M
33	I.BLK	PP075954.D	22 Oct 2025 21:10	YP\AJ	Ok
34	PB170204BL	PP075955.D	22 Oct 2025 21:26	YP\AJ	Ok
35	PB170204BS	PP075956.D	22 Oct 2025 21:43	YP\AJ	Ok
36	Q3416-01	PP075957.D	22 Oct 2025 21:59	YP\AJ	Ok
37	Q3416-01MS	PP075958.D	22 Oct 2025 22:15	YP\AJ	Ok
38	Q3416-01MSD	PP075959.D	22 Oct 2025 22:31	YP\AJ	Ok
39	AR1660CCC500	PP075960.D	22 Oct 2025 23:37	YP\AJ	Ok,NR
40	AR1242CCC500	PP075961.D	23 Oct 2025 00:25	YP\AJ	Ok,M
41	AR1248CCC500	PP075962.D	23 Oct 2025 00:41	YP\AJ	Ok,M
42	AR1254CCC500	PP075963.D	23 Oct 2025 00:58	YP\AJ	Ok,M
43	I.BLK	PP075964.D	23 Oct 2025 01:14	YP\AJ	Ok

M : Manual Integration