

Surrogate Summary

SDG No.: Q3399

Client: ICE Service Group, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP075887.D	PIBLK-PP075887.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	20.4	102		60	140
		Tetrachloro-m-xyl	2	20	18.8	94		60	140
		Decachlorobiphen	2	20	19.3	96		60	140
I.BLK-PP075940.D	PIBLK-PP075940.D	Tetrachloro-m-xyl	1	20	19.8	99		60	140
		Decachlorobiphen	1	20	20.6	103		60	140
		Tetrachloro-m-xyl	2	20	17.2	86		60	140
		Decachlorobiphen	2	20	18.6	93		60	140
Q3399-01	COMP-ROLLOFF	Tetrachloro-m-xyl	1	20	22.4	112		44	130
		Decachlorobiphen	1	20	18.0	90		60	125
		Tetrachloro-m-xyl	2	20	22.5	113		44	130
		Decachlorobiphen	2	20	16.3	81		60	125
I.BLK-PP075954.D	PIBLK-PP075954.D	Tetrachloro-m-xyl	1	20	19.3	96		60	140
		Decachlorobiphen	1	20	21.3	107		60	140
		Tetrachloro-m-xyl	2	20	17.3	87		60	140
		Decachlorobiphen	2	20	19.5	98		60	140
PB170204BL	PB170204BL	Tetrachloro-m-xyl	1	20	23.3	116		44	130
		Decachlorobiphen	1	20	24.9	125		60	125
		Tetrachloro-m-xyl	2	20	21.1	105		44	130
		Decachlorobiphen	2	20	23.0	115		60	125
PB170204BS	PB170204BS	Tetrachloro-m-xyl	1	20	24.0	120		44	130
		Decachlorobiphen	1	20	24.8	124		60	125
		Tetrachloro-m-xyl	2	20	21.8	109		44	130
		Decachlorobiphen	2	20	23.1	116		60	125
Q3416-01MS	NB-08-10212025MS	Tetrachloro-m-xyl	1	20	20.2	101		44	130
		Decachlorobiphen	1	20	24.1	120		60	125
		Tetrachloro-m-xyl	2	20	22.9	115		44	130
		Decachlorobiphen	2	20	18.1	91		60	125
Q3416-01MSD	NB-08-10212025MSD	Tetrachloro-m-xyl	1	20	19.6	98		44	130
		Decachlorobiphen	1	20	22.6	113		60	125
		Tetrachloro-m-xyl	2	20	21.3	106		44	130
		Decachlorobiphen	2	20	16.7	84		60	125
I.BLK-PP075964.D	PIBLK-PP075964.D	Tetrachloro-m-xyl	1	20	19.5	97		60	140
		Decachlorobiphen	1	20	20.8	104		60	140
		Tetrachloro-m-xyl	2	20	17.9	90		60	140
		Decachlorobiphen	2	20	16.6	83		60	140