

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090755.D **Time Analyzed:** 14:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.702	6.600	6.800	50.540	50.000	1.1
4,4'-DDE	6.192	6.090	6.290	47.660	50.000	-4.7
4,4'-DDT	7.018	6.915	7.115	50.860	50.000	1.7
Aldrin	5.265	5.164	5.364	48.440	50.000	-3.1
alpha-BHC	3.994	3.894	4.094	48.140	50.000	-3.7
alpha-Chlordane	6.022	5.921	6.121	48.240	50.000	-3.5
beta-BHC	4.513	4.412	4.612	47.070	50.000	-5.9
Decachlorobiphenyl	9.068	8.964	9.164	46.320	50.000	-7.4
delta-BHC	4.761	4.660	4.860	48.380	50.000	-3.2
Dieldrin	6.342	6.240	6.440	52.600	50.000	5.2
Endosulfan I	6.069	5.968	6.168	48.290	50.000	-3.4
Endosulfan II	6.783	6.681	6.881	48.960	50.000	-2.1
Endosulfan sulfate	7.146	7.045	7.245	48.270	50.000	-3.5
Endrin	6.570	6.468	6.668	47.990	50.000	-4.0
Endrin aldehyde	6.912	6.810	7.010	49.240	50.000	-1.5
Endrin ketone	7.627	7.525	7.725	49.560	50.000	-0.9
gamma-BHC (Lindane)	4.326	4.225	4.425	47.990	50.000	-4.0
gamma-Chlordane	5.941	5.840	6.040	48.010	50.000	-4.0
Heptachlor	4.924	4.823	5.023	57.070	50.000	14.1
Heptachlor epoxide	5.685	5.584	5.784	49.030	50.000	-1.9
Methoxychlor	7.490	7.388	7.588	49.890	50.000	-0.2
Tetrachloro-m-xylene	3.545	3.445	3.645	46.510	50.000	-7.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090755.D **Time Analyzed:** 14:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.920	5.821	6.021	50.670	50.000	1.3
4,4'-DDE	5.366	5.266	5.466	49.060	50.000	-1.9
4,4'-DDT	6.175	6.074	6.274	52.370	50.000	4.7
Aldrin	4.360	4.261	4.461	48.160	50.000	-3.7
alpha-BHC	3.386	3.287	3.487	47.870	50.000	-4.3
alpha-Chlordane	5.181	5.081	5.281	48.660	50.000	-2.7
beta-BHC	4.018	3.919	4.119	47.530	50.000	-4.9
Decachlorobiphenyl	8.062	7.962	8.162	54.910	50.000	9.8
delta-BHC	4.254	4.155	4.355	48.500	50.000	-3.0
Dieldrin	5.504	5.404	5.604	48.850	50.000	-2.3
Endosulfan I	5.238	5.138	5.338	48.350	50.000	-3.3
Endosulfan II	6.072	5.972	6.172	50.380	50.000	0.8
Endosulfan sulfate	6.474	6.374	6.574	50.840	50.000	1.7
Endrin	5.781	5.681	5.881	48.490	50.000	-3.0
Endrin aldehyde	6.251	6.150	6.350	51.380	50.000	2.8
Endrin ketone	6.983	6.883	7.083	53.380	50.000	6.8
gamma-BHC (Lindane)	3.722	3.623	3.823	49.500	50.000	-1.0
gamma-Chlordane	5.117	5.017	5.217	48.880	50.000	-2.2
Heptachlor	4.074	3.975	4.175	49.470	50.000	-1.1
Heptachlor epoxide	4.864	4.764	4.964	47.790	50.000	-4.4
Methoxychlor	6.745	6.645	6.845	52.980	50.000	6.0
Tetrachloro-m-xylene	2.874	2.775	2.975	47.350	50.000	-5.3