

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090764.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.701	6.600	6.800	48.540	50.000	-2.9
4,4'-DDE	6.191	6.090	6.290	45.890	50.000	-8.2
4,4'-DDT	7.016	6.915	7.115	47.220	50.000	-5.6
Aldrin	5.265	5.164	5.364	47.430	50.000	-5.1
alpha-BHC	3.994	3.894	4.094	47.810	50.000	-4.4
alpha-Chlordane	6.022	5.921	6.121	45.360	50.000	-9.3
beta-BHC	4.513	4.412	4.612	46.500	50.000	-7.0
Decachlorobiphenyl	9.067	8.964	9.164	44.170	50.000	-11.7
delta-BHC	4.761	4.660	4.860	47.540	50.000	-4.9
Dieldrin	6.342	6.240	6.440	50.320	50.000	0.6
Endosulfan I	6.069	5.968	6.168	46.400	50.000	-7.2
Endosulfan II	6.782	6.681	6.881	46.850	50.000	-6.3
Endosulfan sulfate	7.146	7.045	7.245	45.950	50.000	-8.1
Endrin	6.569	6.468	6.668	46.250	50.000	-7.5
Endrin aldehyde	6.911	6.810	7.010	46.610	50.000	-6.8
Endrin ketone	7.627	7.525	7.725	46.940	50.000	-6.1
gamma-BHC (Lindane)	4.324	4.225	4.425	47.530	50.000	-4.9
gamma-Chlordane	5.941	5.840	6.040	46.490	50.000	-7.0
Heptachlor	4.923	4.823	5.023	55.620	50.000	11.2
Heptachlor epoxide	5.685	5.584	5.784	46.870	50.000	-6.3
Methoxychlor	7.489	7.388	7.588	46.180	50.000	-7.6
Tetrachloro-m-xylene	3.544	3.445	3.645	46.320	50.000	-7.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090764.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.920	5.821	6.021	48.200	50.000	-3.6
4,4'-DDE	5.366	5.266	5.466	47.300	50.000	-5.4
4,4'-DDT	6.174	6.074	6.274	48.220	50.000	-3.6
Aldrin	4.360	4.261	4.461	47.330	50.000	-5.3
alpha-BHC	3.386	3.287	3.487	47.470	50.000	-5.1
alpha-Chlordane	5.181	5.081	5.281	47.290	50.000	-5.4
beta-BHC	4.018	3.919	4.119	47.020	50.000	-6.0
Decachlorobiphenyl	8.061	7.962	8.162	48.710	50.000	-2.6
delta-BHC	4.254	4.155	4.355	47.810	50.000	-4.4
Dieldrin	5.503	5.404	5.604	47.130	50.000	-5.7
Endosulfan I	5.238	5.138	5.338	46.640	50.000	-6.7
Endosulfan II	6.072	5.972	6.172	47.490	50.000	-5.0
Endosulfan sulfate	6.474	6.374	6.574	47.640	50.000	-4.7
Endrin	5.780	5.681	5.881	46.650	50.000	-6.7
Endrin aldehyde	6.250	6.150	6.350	47.580	50.000	-4.8
Endrin ketone	6.983	6.883	7.083	48.300	50.000	-3.4
gamma-BHC (Lindane)	3.722	3.623	3.823	48.880	50.000	-2.2
gamma-Chlordane	5.116	5.017	5.217	47.570	50.000	-4.9
Heptachlor	4.074	3.975	4.175	48.710	50.000	-2.6
Heptachlor epoxide	4.864	4.764	4.964	46.510	50.000	-7.0
Methoxychlor	6.745	6.645	6.845	48.070	50.000	-3.9
Tetrachloro-m-xylene	2.874	2.775	2.975	47.070	50.000	-5.9