

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090754.D Date Analyzed: 10/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 14:45

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.070	8.970	9.170	22.040	20.000	10.2
Tetrachloro-m-xylene	3.544	3.490	3.590	20.920	20.000	4.6
alpha-BHC	3.994	3.940	4.040	8.790	10.000	-12.1
beta-BHC	4.513	4.460	4.560	10.210	10.000	2.1
gamma-BHC (Lindane)	4.325	4.270	4.380	9.200	10.000	-8.0
Endrin	6.570	6.500	6.640	48.180	50.000	-3.6
4,4'-DDT	7.018	6.950	7.090	100.910	100.000	0.9
Methoxychlor	7.491	7.420	7.560	235.430	250.000	-5.8

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090754.D Date Analyzed: 10/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 14:45

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.062	7.960	8.160	25.380	20.000	26.9
Tetrachloro-m-xylene	2.873	2.820	2.920	21.110	20.000	5.6
alpha-BHC	3.385	3.330	3.440	10.320	10.000	3.2
beta-BHC	4.017	3.970	4.070	10.570	10.000	5.7
gamma-BHC (Lindane)	3.721	3.670	3.770	10.640	10.000	6.4
Endrin	5.779	5.710	5.850	49.190	50.000	-1.6
4,4'-DDT	6.173	6.100	6.240	103.470	100.000	3.5
Methoxychlor	6.745	6.670	6.820	217.530	250.000	-13.0