

DATA PACKAGE
GC SEMI-VOLATILES

PROJECT NAME : NWIRP NORTHRUP GRUMMAN SITE – BETHPAGE, NY

ICE SERVICE GROUP, INC.
191 Central Avenue Suite 200C

Newark, NJ - 07103
Phone No: 4129165710

ORDER ID : Q3399
ATTENTION : Dennis D. Morgan, II



Laboratory Certification ID # 20012



1) PESTICIDE-TCL Data	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	7
5) Conformance/Non Conformance	8
6) QA Checklist	10
7) Chronicle	11
8) Hit Summary	12
9) QC Data Summary For Pesticide-TCL	13
9.1) Deuterated Monitoring Compound Summary	14
9.2) MS/MSD Summary	15
9.3) LCS/LCSD Summary	19
9.4) Method Blank Summary	20
10) Sample Data	21
10.1) COMP-ROLLOFF	22
11) Calibration Data Summary	29
11.1) Initial Calibration Data	30
11.1.1) PD101725	30
11.2) Continued Calibration Data	144
11.2.1) PD090755.D	144
11.2.2) PD090764.D	162
11.2.3) PD090773.D	180
11.2.4) PEM Files	198
11.3) RESCHK Data	217
11.4) Analytical Seq	225
12) Compound Detection Summary	227
13) QC Sample Data	234
13.1) Method Blank Data	235
13.2) PIBLK Data	239
13.3) LCS Data	255
13.4) MS Data	270
13.5) MSD Data	285
14) Manual Integration	300
15) Analytical Runlogs	302
16) Percent Solid	311

Table Of Contents for Q3399

17) Extraction Logs	324
17.1) PB170205.pdf	324
17.2) PB170205IC.pdf	326
18) Standard Prep Logs	327
19) Shipping Document	380
19.1) Chain Of Custody	381
19.2) Lab Certificate	383
19.3) Internal COC	384
20) Not Reviewed Data	385

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20

Cover Page

Order ID : Q3399

Project ID : NWIRP Northrup Grumman Site – Bethpage, NY

Client : ICE Service Group, Inc.

Lab Sample Number

Q3399-01
Q3399-02
Q3399-03
Q3399-04

Client Sample Number

COMP-ROLLOFF
COMP-ROLLOFF
COMP-ROLLOFF
COMP-ROLLOFF

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 12/17/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

ICE Service Group, Inc.

Project Name: NWIRP Northrup Grumman Site – Bethpage, NY

Project # N/A

Order ID # Q3399

Test Name: Pesticide-TCL

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 10/20/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Pesticide-TCL. This data package contains results for Pesticide-TCL.

C. Analytical Techniques:

The analysis was performed on instrument ECD_D. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df.; Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of Pesticide-TCLs was based on method 8081B and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the requirements for all compounds.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

E. Additional Comments:

Sample COMP-ROLLOFF was reported with J flag on form 1 for compound Dieldrin based on reporting criteria of high concentration from both column. Now for other column compound detection is below MDL therefore it is not detecting on form 10.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19
- 20

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3399

MATRIX: Solid

METHOD: 8081B/3541

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified.			✓
2. Standard Summary Submitted.			✓
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD.			✓
4. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
6. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. The MS recoveries met the requirements for all compounds. The MSD recoveries met the requirements for all compounds. The Blank Spike met requirements for all compounds. The RPD were met for all analysis.		✓	
7. Retention Time Shift Meet Criteria (if applicable) Comments:			✓
8. Extraction Holding Time Met If not met, list number of days exceeded for each sample:			✓
9. Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.			✓

ADDITIONAL COMMENTS:

Sample COMP-ROLLOFF was reported with J flag on form 1 for compound Dieldrin based on reporting criteria of high concentration from both column. Now for other column compound detection is below MDL therefore it is not detecting on form 10.



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3399

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 12/17/2025

LAB CHRONICLE

OrderID:	Q3399	OrderDate:	10/20/2025 1:23:08 PM
Client:	ICE Service Group, Inc.	Project:	NWIRP Northrup Grumman Site – Bethpage, NY
Contact:	Dennis D. Morgan, II	Location:	D41,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3399-01	COMP-ROLLOFF	SOIL			10/20/25			10/20/25
			Diesel Range Organics	8015D		10/22/25	10/22/25	
			Gasoline Range Organics	8015D			10/21/25	
			Herbicide	8151A		10/22/25	10/23/25	
			PCB	8082A		10/22/25	10/22/25	
			Pesticide-TCL	8081B		10/22/25	10/22/25	
Q3399-02	COMP-ROLLOFF	TCLP			10/20/25			10/20/25
			TCLP Herbicide	8151A		10/22/25	10/23/25	
			TCLP Pesticide	8081B		10/22/25	10/22/25	

Hit Summary Sheet
SW-846

SDG No.: Q3399

Order ID: Q3399

Client: ICE Service Group, Inc.

Project ID: NWIRP Northrup Grumman Site – Be

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : COMP-ROLLOFF									
Q3399-01	COMP-ROLLOFF	SOIL	Dieldrin	0.25	J	0.16	0.37	1.90	ug/kg
Q3399-01	COMP-ROLLOFF	SOIL	4,4-DDE	0.21	J	0.16	0.37	1.90	ug/kg
Q3399-01	COMP-ROLLOFF	SOIL	Endrin	0.38	JP	0.16	0.37	1.90	ug/kg
Q3399-01	COMP-ROLLOFF	SOIL	4,4-DDT	0.75	JP	0.16	0.37	1.90	ug/kg
Q3399-01	COMP-ROLLOFF	SOIL	alpha-Chlordane	0.37	J	0.13	0.37	1.90	ug/kg
Total Concentration:				1.960					



QC SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19
- 20

Surrogate Summary

SDG No.: Q3399

Client: ICE Service Group, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		30	135
		Tetrachloro-m-xyl	1	20	19.4	97		44	124
		Decachlorobiphen	2	20	20.3	102		30	135
		Tetrachloro-m-xyl	2	20	20.0	100		44	124
I.BLK-PD090753.D	PIBLK-PD090753.D	Decachlorobiphen	1	20	22.1	111		30	135
		Tetrachloro-m-xyl	1	20	19.9	99		44	124
		Decachlorobiphen	2	20	25.1	126		30	135
		Tetrachloro-m-xyl	2	20	20.7	104		44	124
PB170205BL	PB170205BL	Decachlorobiphen	1	20	21.5	107		55	130
		Tetrachloro-m-xyl	1	20	19.9	99		42	129
		Decachlorobiphen	2	20	24.3	121		55	130
		Tetrachloro-m-xyl	2	20	20.4	102		42	129
PB170205BS	PB170205BS	Decachlorobiphen	1	20	21.6	108		55	130
		Tetrachloro-m-xyl	1	20	16.9	84		42	129
		Decachlorobiphen	2	20	24.5	123		55	130
		Tetrachloro-m-xyl	2	20	17.3	86		42	129
Q3399-01	COMP-ROLLOFF	Decachlorobiphen	1	20	14.8	74		55	130
		Tetrachloro-m-xyl	1	20	18.3	91		42	129
		Decachlorobiphen	2	20	16.0	80		55	130
		Tetrachloro-m-xyl	2	20	19.4	97		42	129
I.BLK-PD090763.D	PIBLK-PD090763.D	Decachlorobiphen	1	20	21.0	105		30	135
		Tetrachloro-m-xyl	1	20	19.9	100		44	124
		Decachlorobiphen	2	20	21.0	105		30	135
		Tetrachloro-m-xyl	2	20	20.6	103		44	124
Q3399-01MS	COMP-ROLLOFFMS	Decachlorobiphen	1	20	14.1	70		55	130
		Tetrachloro-m-xyl	1	20	17.7	89		42	129
		Decachlorobiphen	2	20	15.3	77		55	130
		Tetrachloro-m-xyl	2	20	19.1	96		42	129
Q3399-01MSD	COMP-ROLLOFFMSD	Decachlorobiphen	1	20	14.3	72		55	130
		Tetrachloro-m-xyl	1	20	17.5	88		42	129
		Decachlorobiphen	2	20	15.7	79		55	130
		Tetrachloro-m-xyl	2	20	19.4	97		42	129
I.BLK-PD090772.D	PIBLK-PD090772.D	Decachlorobiphen	1	20	21.4	107		30	135
		Tetrachloro-m-xyl	1	20	19.9	100		44	124
		Decachlorobiphen	2	20	24.0	120		30	135
		Tetrachloro-m-xyl	2	20	20.5	103		44	124

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ICE Service Group, Inc.</u>	DataFile :	<u>PD090768.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits		RPD
			Result								High		
Lab Sample ID:	Q3399-01MS (Column 1)	Client Sample ID:		COMP-ROLLOFFMS									
	alpha-BHC	18.48	0	18.2	ug/kg	98				45	137		
	beta-BHC	18.48	0	17.5	ug/kg	95				50	136		
	delta-BHC	18.48	0	18.1	ug/kg	98				47	139		
	gamma-BHC (Lindane)	18.48	0	17.7	ug/kg	96				49	135		
	Heptachlor	18.48	0	20.9	ug/kg	113				47	136		
	Aldrin	18.48	0	17.8	ug/kg	96				45	136		
	Heptachlor epoxide	18.48	0	18.1	ug/kg	98				52	136		
	Endosulfan I	18.48	0	17.7	ug/kg	96				53	132		
	Dieldrin	18.48	0	18.0	ug/kg	97				56	136		
	4,4'-DDE	18.48	0.21	17.6	ug/kg	94				56	134		
	Endrin	18.48	0.19	17.9	ug/kg	96				57	140		
	Endosulfan II	18.48	0	16.8	ug/kg	91				53	134		
	4,4'-DDD	18.48	0	18.4	ug/kg	100				56	139		
	Endosulfan sulfate	18.48	0	17.4	ug/kg	94				55	136		
	4,4'-DDT	18.48	0.22	18.5	ug/kg	99				50	141		
	Methoxychlor	18.48	0	18.0	ug/kg	97				52	143		
	Endrin ketone	18.48	0	18.0	ug/kg	97				55	136		
	Endrin aldehyde	18.48	0	17.6	ug/kg	95				35	137		
	alpha-Chlordane	18.48	0.37	17.7	ug/kg	94				54	133		
	gamma-Chlordane	18.48	0	17.6	ug/kg	95				53	135		
Lab Sample ID:	Q3399-01MS (Column 2)	Client Sample ID:		COMP-ROLLOFFMS									
	alpha-BHC	18.48	0	17.6	ug/kg	95				45	137		
	beta-BHC	18.48	0	17.5	ug/kg	95				50	136		
	delta-BHC	18.48	0	17.7	ug/kg	96				47	139		
	gamma-BHC (Lindane)	18.48	0	17.6	ug/kg	95				49	135		
	Heptachlor	18.48	0	18.1	ug/kg	98				47	136		
	Aldrin	18.48	0	17.8	ug/kg	96				45	136		
	Heptachlor epoxide	18.48	0	17.6	ug/kg	95				52	136		
	Endosulfan I	18.48	0	17.7	ug/kg	96				53	132		
	Dieldrin	18.48	0.25	17.9	ug/kg	96				56	136		
	4,4'-DDE	18.48	0.18	18.0	ug/kg	96				56	134		
	Endrin	18.48	0.38	17.9	ug/kg	95				57	140		
	Endosulfan II	18.48	0	18.2	ug/kg	98				53	134		
	4,4'-DDD	18.48	0	17.4	ug/kg	94				56	139		
	Endosulfan sulfate	18.48	0	18.5	ug/kg	100				55	136		

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ICE Service Group, Inc.</u>	DataFile :	<u>PD090768.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits		RPD
		Result	Result							High		
4,4'-DDT	18.48	0.75	19.2	ug/kg	100				50	141		
Methoxychlor	18.48	0	18.9	ug/kg	102				52	143		
Endrin ketone	18.48	0	18.9	ug/kg	102				55	136		
Endrin aldehyde	18.48	0	18.3	ug/kg	99				35	137		
alpha-Chlordane	18.48	0.27	17.9	ug/kg	95				54	133		
gamma-Chlordane	18.48	0	17.9	ug/kg	97				53	135		

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3399 Analytical Method: 8081B
Client: ICE Service Group, Inc. DataFile : PD090769.D

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD		Low	Limits		RPD
			Result					RPD	Qual		High		
Lab Sample ID:	Q3399-01MSD (Column 1)	Client Sample ID:	COMP-ROLLOFFMSD										
	alpha-BHC	18.49	0	18.5	ug/kg	100		2		45	137		20
	beta-BHC	18.49	0	17.7	ug/kg	96		1		50	136		20
	delta-BHC	18.49	0	18.5	ug/kg	100		2		47	139		20
	gamma-BHC (Lindane)	18.49	0	18.0	ug/kg	97		1		49	135		20
	Heptachlor	18.49	0	21.3	ug/kg	115		2		47	136		20
	Aldrin	18.49	0	18.1	ug/kg	98		2		45	136		20
	Heptachlor epoxide	18.49	0	18.3	ug/kg	99		1		52	136		20
	Endosulfan I	18.49	0	18.0	ug/kg	97		1		53	132		20
	Dieldrin	18.49	0	18.3	ug/kg	99		2		56	136		20
	4,4'-DDE	18.49	0.21	18.1	ug/kg	97		3		56	134		20
	Endrin	18.49	0.19	18.3	ug/kg	98		2		57	140		20
	Endosulfan II	18.49	0	17.1	ug/kg	92		1		53	134		20
	4,4'-DDD	18.49	0	18.7	ug/kg	101		1		56	139		20
	Endosulfan sulfate	18.49	0	17.7	ug/kg	96		2		55	136		20
	4,4'-DDT	18.49	0.22	19.0	ug/kg	102		3		50	141		20
	Methoxychlor	18.49	0	18.3	ug/kg	99		2		52	143		20
	Endrin ketone	18.49	0	18.4	ug/kg	100		3		55	136		20
	Endrin aldehyde	18.49	0	18.0	ug/kg	97		2		35	137		20
	alpha-Chlordane	18.49	0.37	17.7	ug/kg	94		0		54	133		20
	gamma-Chlordane	18.49	0	17.9	ug/kg	97		2		53	135		20
Lab Sample ID:	Q3399-01MSD (Column 2)	Client Sample ID:	COMP-ROLLOFFMSD										
	alpha-BHC	18.49	0	17.9	ug/kg	97		2		45	137		20
	beta-BHC	18.49	0	17.7	ug/kg	96		1		50	136		20
	delta-BHC	18.49	0	18.0	ug/kg	97		1		47	139		20
	gamma-BHC (Lindane)	18.49	0	17.9	ug/kg	97		2		49	135		20
	Heptachlor	18.49	0	18.3	ug/kg	99		1		47	136		20
	Aldrin	18.49	0	18.0	ug/kg	97		1		45	136		20
	Heptachlor epoxide	18.49	0	17.8	ug/kg	96		1		52	136		20
	Endosulfan I	18.49	0	17.9	ug/kg	97		1		53	132		20
	Dieldrin	18.49	0.25	18.2	ug/kg	97		1		56	136		20
	4,4'-DDE	18.49	0.18	18.3	ug/kg	98		2		56	134		20
	Endrin	18.49	0.38	18.0	ug/kg	95		0		57	140		20
	Endosulfan II	18.49	0	18.6	ug/kg	101		3		53	134		20
	4,4'-DDD	18.49	0	17.7	ug/kg	96		2		56	139		20
	Endosulfan sulfate	18.49	0	18.8	ug/kg	102		2		55	136		20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ICE Service Group, Inc.</u>	DataFile :	<u>PD090769.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD		Low	Limits	
		Result	Result				RPD	Qual		High	RPD
4,4'-DDT	18.49	0.75	19.7	ug/kg	102		2		50	141	20
Methoxychlor	18.49	0	19.3	ug/kg	104		2		52	143	20
Endrin ketone	18.49	0	19.2	ug/kg	104		2		55	136	20
Endrin aldehyde	18.49	0	18.6	ug/kg	101		2		35	137	20
alpha-Chlordane	18.49	0.27	18.2	ug/kg	97		2		54	133	20
gamma-Chlordane	18.49	0	18.2	ug/kg	98		1		53	135	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ICE Service Group, Inc.</u>	Datafile :	<u>PD090757.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170205BS (Column 1)	alpha-BHC	16.65	15.0	ug/kg	90				45	137	
	beta-BHC	16.65	14.7	ug/kg	88				50	136	
	delta-BHC	16.65	15.2	ug/kg	91				47	139	
	gamma-BHC (Lindane)	16.65	14.9	ug/kg	89				49	135	
	Heptachlor	16.65	17.9	ug/kg	108				47	136	
	Aldrin	16.65	15.2	ug/kg	91				45	136	
	Heptachlor epoxide	16.65	15.4	ug/kg	92				52	136	
	Endosulfan I	16.65	15.2	ug/kg	91				53	132	
	Dieldrin	16.65	15.5	ug/kg	93				56	136	
	4,4'-DDE	16.65	15.0	ug/kg	90				56	134	
	Endrin	16.65	15.4	ug/kg	92				57	140	
	Endosulfan II	16.65	15.6	ug/kg	94				53	134	
	4,4'-DDD	16.65	16.1	ug/kg	97				56	139	
	Endosulfan sulfate	16.65	15.4	ug/kg	92				55	136	
	4,4'-DDT	16.65	16.2	ug/kg	97				50	141	
	Methoxychlor	16.65	16.0	ug/kg	96				52	143	
	Endrin ketone	16.65	15.8	ug/kg	95				55	136	
	Endrin aldehyde	16.65	15.8	ug/kg	95				35	137	
	alpha-Chlordane	16.65	14.8	ug/kg	89				54	133	
	gamma-Chlordane	16.65	15.1	ug/kg	91				53	135	
PB170205BS (Column 2)	alpha-BHC	16.65	14.8	ug/kg	89				45	137	
	beta-BHC	16.65	14.9	ug/kg	89				50	136	
	delta-BHC	16.65	15.1	ug/kg	91				47	139	
	gamma-BHC (Lindane)	16.65	14.9	ug/kg	89				49	135	
	Heptachlor	16.65	15.4	ug/kg	92				47	136	
	Aldrin	16.65	15.1	ug/kg	91				45	136	
	Heptachlor epoxide	16.65	15.1	ug/kg	91				52	136	
	Endosulfan I	16.65	14.9	ug/kg	89				53	132	
	Dieldrin	16.65	15.6	ug/kg	94				56	136	
	4,4'-DDE	16.65	15.5	ug/kg	93				56	134	
	Endrin	16.65	15.6	ug/kg	94				57	140	
	Endosulfan II	16.65	15.8	ug/kg	95				53	134	
	4,4'-DDD	16.65	15.9	ug/kg	95				56	139	
	Endosulfan sulfate	16.65	16.0	ug/kg	96				55	136	
	4,4'-DDT	16.65	16.5	ug/kg	99				50	141	
	Methoxychlor	16.65	16.5	ug/kg	99				52	143	
	Endrin ketone	16.65	16.8	ug/kg	101				55	136	
	Endrin aldehyde	16.65	16.2	ug/kg	97				35	137	
	alpha-Chlordane	16.65	15.3	ug/kg	92				54	133	
	gamma-Chlordane	16.65	15.3	ug/kg	92				53	135	

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB170205BL

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: PB170205BL

Lab File ID: PD090756.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 10/22/2025

Date Analyzed (1): 10/22/2025

Date Analyzed (2): 10/22/2025

Time Analyzed (1): 15:13

Time Analyzed (2): 15:13

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB170205BS	PB170205BS	PD090757.D	10/22/2025	10/22/2025
COMP-ROLLOFF	Q3399-01	PD090758.D	10/22/2025	10/22/2025
COMP-ROLLOFFMS	Q3399-01MS	PD090768.D	10/22/2025	10/22/2025
COMP-ROLLOFFMSD	Q3399-01MSD	PD090769.D	10/22/2025	10/22/2025

COMMENTS: _____



SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19
- 20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/20/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/20/25
Client Sample ID:	COMP-ROLLOFF		SDG No.:	Q3399
Lab Sample ID:	Q3399-01		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	89.9
Sample Wt/Vol:	30.06 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541	Prep Date	10/22/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
319-84-6	alpha-BHC	0.37	U	1	0.14	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
319-85-7	beta-BHC	0.92	U	1	0.20	0.92	1.90	ug/kg	10/22/25 15:40	PB170205
319-86-8	delta-BHC	0.92	U	1	0.43	0.92	1.90	ug/kg	10/22/25 15:40	PB170205
58-89-9	gamma-BHC (Lindane)	0.37	U	1	0.16	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
76-44-8	Heptachlor	0.37	U	1	0.13	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
309-00-2	Aldrin	0.37	U	1	0.13	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
1024-57-3	Heptachlor epoxide	0.92	U	1	0.21	0.92	1.90	ug/kg	10/22/25 15:40	PB170205
959-98-8	Endosulfan I	0.37	U	1	0.16	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
60-57-1	Dieldrin	0.25	J	1	0.16	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
72-55-9	4,4-DDE	0.21	J	1	0.16	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
72-20-8	Endrin	0.38	JP	1	0.16	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
33213-65-9	Endosulfan II	0.92	U	1	0.32	0.92	1.90	ug/kg	10/22/25 15:40	PB170205
72-54-8	4,4-DDD	0.37	U	1	0.17	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
1031-07-8	Endosulfan Sulfate	0.37	U	1	0.14	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
50-29-3	4,4-DDT	0.75	JP	1	0.16	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
72-43-5	Methoxychlor	0.92	U	1	0.41	0.92	1.90	ug/kg	10/22/25 15:40	PB170205
53494-70-5	Endrin ketone	0.92	U	1	0.21	0.92	1.90	ug/kg	10/22/25 15:40	PB170205
7421-93-4	Endrin aldehyde	0.92	U	1	0.41	0.92	1.90	ug/kg	10/22/25 15:40	PB170205
5103-71-9	alpha-Chlordane	0.37	J	1	0.13	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
5103-74-2	gamma-Chlordane	0.37	U	1	0.17	0.37	1.90	ug/kg	10/22/25 15:40	PB170205
8001-35-2	Toxaphene	18.9	U	1	6.00	18.9	36.6	ug/kg	10/22/25 15:40	PB170205
SURROGATES										
2051-24-3	Decachlorobiphenyl	16.0			55 - 130		80%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	19.4			42 - 129		97%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3399-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
 Data File : PD090758.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 15:40
 Operator : AR\AJ
 Sample : Q3399-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

COMP-ROLLOFF

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/23/2025

Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 12:25:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.873	56865197	576.8E6	18.260m	19.369
28) SA Decachlor...	9.069	8.063	69161925	485.5E6	14.785	16.028
Target Compounds						
10) B gamma-Chl...	5.942	5.117	1658575	13259181	0.280	0.295
11) B alpha-Chl...	6.021	5.182	6161942	31133470	0.997	0.722 #
12) B 4,4'-DDE	6.191	5.366	2962743	21426427	0.556	0.489
13) MA Dieldrin	6.341	5.501	1168188	29786418	0.200m	0.674m#
14) MA Endrin	6.570	5.783	2492609	41471793	0.503m	1.024m#
17) MA 4,4'-DDT	7.017	6.173	2467837	71467482	0.600	2.038 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090758.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 15:40
Operator : AR\AJ
Sample : Q3399-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

COMP-ROLLOFF

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 10/23/2025

Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 12:25:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: -0.001 min
Response: 56865197
Conc: 18.26 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: -0.001 min
Response: 576762882
Conc: 19.37 ng/ml




#10 gamma-Chlordane

R.T.: 5.942 min
Delta R.T.: 0.002 min
Response: 1658575
Conc: 0.28 ng/ml



#10 gamma-Chlordane

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 13259181
Conc: 0.29 ng/ml




#11 alpha-Chlordane

R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 6161942
Conc: 1.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025




#11 alpha-Chlordane

R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 31133470
Conc: 0.72 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.001 min
Response: 2962743
Conc: 0.56 ng/ml



#12 4,4'-DDE

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 21426427
Conc: 0.49 ng/ml




#13 Dieldrin

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 1168188
Conc: 0.20 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025




#13 Dieldrin

R.T.: 5.501 min
Delta R.T.: -0.003 min
Response: 29786418
Conc: 0.67 ng/ml m




#14 Endrin

R.T.: 6.570 min
Delta R.T.: 0.001 min
Response: 2492609
Conc: 0.50 ng/ml m



#14 Endrin

R.T.: 5.783 min
Delta R.T.: 0.002 min
Response: 41471793
Conc: 1.02 ng/ml m




#17 4,4'-DDT

R.T.: 7.017 min
Delta R.T.: 0.002 min
Response: 2467837
Conc: 0.60 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025




#17 4,4'-DDT

R.T.: 6.173 min
Delta R.T.: 0.000 min
Response: 71467482
Conc: 2.04 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.069 min
Delta R.T.: 0.005 min
Response: 69161925
Conc: 14.78 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.063 min
Delta R.T.: 0.000 min
Response: 485518279
Conc: 16.03 ng/ml



CALIBRATION SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3399</u>
Instrument ID:	<u>ECD_D</u>	Calibration Date(s):	<u>10/17/2025</u> <u>10/17/2025</u>
		Calibration Times:	<u>11:34</u> <u>12:28</u>

 GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PD090650.D</u>	RT 075 = <u>PD090651.D</u>
RT 050 = <u>PD090652.D</u>	RT 025 = <u>PD090653.D</u>	RT 005 = <u>PD090654.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
4,4'-DDD	6.70	6.70	6.70	6.70	6.70	6.70	6.60	6.80
4,4'-DDE	6.19	6.19	6.19	6.19	6.19	6.19	6.09	6.29
4,4'-DDT	7.02	7.02	7.02	7.02	7.02	7.02	6.92	7.12
Aldrin	5.26	5.27	5.26	5.26	5.26	5.26	5.16	5.36
alpha-BHC	3.99	3.99	3.99	3.99	3.99	3.99	3.89	4.09
alpha-Chlordane	6.02	6.02	6.02	6.02	6.02	6.02	5.92	6.12
beta-BHC	4.51	4.51	4.51	4.51	4.51	4.51	4.41	4.61
Decachlorobiphenyl	9.07	9.07	9.06	9.07	9.07	9.07	8.97	9.17
delta-BHC	4.76	4.76	4.76	4.76	4.76	4.76	4.66	4.86
Dieldrin	6.34	6.34	6.34	6.34	6.34	6.34	6.24	6.44
Endosulfan I	6.07	6.07	6.07	6.07	6.07	6.07	5.97	6.17
Endosulfan II	6.78	6.78	6.78	6.78	6.78	6.78	6.68	6.88
Endosulfan sulfate	7.15	7.15	7.15	7.14	7.14	7.14	7.04	7.24
Endrin	6.57	6.57	6.57	6.57	6.57	6.57	6.47	6.67
Endrin aldehyde	6.91	6.91	6.91	6.91	6.91	6.91	6.81	7.01
Endrin ketone	7.63	7.63	7.63	7.63	7.62	7.62	7.52	7.72
gamma-BHC (Lindane)	4.32	4.33	4.33	4.32	4.32	4.32	4.22	4.42
gamma-Chlordane	5.94	5.94	5.94	5.94	5.94	5.94	5.84	6.04
Heptachlor	4.92	4.92	4.92	4.92	4.92	4.92	4.82	5.02
Heptachlor epoxide	5.69	5.69	5.68	5.68	5.68	5.68	5.58	5.78
Methoxychlor	7.49	7.49	7.49	7.49	7.49	7.49	7.39	7.59
Tetrachloro-m-xylene	3.54	3.55	3.55	3.54	3.54	3.54	3.44	3.64

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3399</u>
Instrument ID:	<u>ECD_D</u>	Calibration Date(s):	<u>10/17/2025</u> <u>10/17/2025</u>
		Calibration Times:	<u>11:34</u> <u>12:28</u>

 GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PD090650.D</u>	RT 075 = <u>PD090651.D</u>
RT 050 = <u>PD090652.D</u>	RT 025 = <u>PD090653.D</u>	RT 005 = <u>PD090654.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
4,4'-DDD	5.92	5.92	5.92	5.92	5.92	5.92	5.82	6.02
4,4'-DDE	5.37	5.37	5.37	5.37	5.37	5.37	5.27	5.47
4,4'-DDT	6.17	6.18	6.17	6.17	6.18	6.17	6.07	6.27
Aldrin	4.36	4.36	4.36	4.36	4.36	4.36	4.26	4.46
alpha-BHC	3.39	3.39	3.39	3.39	3.39	3.39	3.29	3.49
alpha-Chlordane	5.18	5.18	5.18	5.18	5.18	5.18	5.08	5.28
beta-BHC	4.02	4.02	4.02	4.02	4.02	4.02	3.92	4.12
Decachlorobiphenyl	8.06	8.06	8.06	8.06	8.06	8.06	7.96	8.16
delta-BHC	4.26	4.26	4.26	4.26	4.26	4.26	4.16	4.36
Dieldrin	5.50	5.50	5.50	5.50	5.50	5.50	5.40	5.60
Endosulfan I	5.24	5.24	5.24	5.24	5.24	5.24	5.14	5.34
Endosulfan II	6.07	6.07	6.07	6.07	6.07	6.07	5.97	6.17
Endosulfan sulfate	6.47	6.47	6.47	6.47	6.47	6.47	6.37	6.57
Endrin	5.78	5.78	5.78	5.78	5.78	5.78	5.68	5.88
Endrin aldehyde	6.25	6.25	6.25	6.25	6.25	6.25	6.15	6.35
Endrin ketone	6.98	6.98	6.98	6.98	6.98	6.98	6.88	7.08
gamma-BHC (Lindane)	3.72	3.72	3.72	3.72	3.72	3.72	3.62	3.82
gamma-Chlordane	5.12	5.12	5.12	5.12	5.12	5.12	5.02	5.22
Heptachlor	4.08	4.08	4.08	4.07	4.08	4.07	3.97	4.17
Heptachlor epoxide	4.86	4.87	4.86	4.86	4.86	4.86	4.76	4.96
Methoxychlor	6.75	6.75	6.75	6.74	6.75	6.74	6.64	6.84
Tetrachloro-m-xylene	2.87	2.88	2.88	2.87	2.87	2.87	2.77	2.97

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: ICES02
SDG NO.: Q3399
Calibration Date(s): 10/17/2025 10/17/2025
Calibration Times: 11:34 12:28

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD090650.D</u>	CF 075 =	<u>PD090651.D</u>		
CF 050 =		<u>PD090652.D</u>	CF 025 =	<u>PD090653.D</u>	CF 005 =	<u>PD090654.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	4150980000	4075270000	4055470000	3840760000	4148570000	4054210000	3
4,4'-DDE	5435610000	5370750000	5317610000	5053060000	5456160000	5326640000	3
4,4'-DDT	4256810000	4152600000	4117930000	3927880000	4097130000	4110470000	3
Aldrin	6694230000	6562230000	6516940000	6238140000	6724820000	6547270000	3
alpha-BHC	7752530000	7402380000	7252210000	6743070000	6735520000	7177140000	6
alpha-Chlordane	5966070000	5905470000	5938650000	6170460000	6920630000	6180250000	7
beta-BHC	2515210000	2489310000	2540370000	2560620000	2992650000	2619630000	8
Decachlorobiphenyl	4361180000	4315270000	4506230000	4707060000	5500070000	4677960000	10
delta-BHC	7169260000	6905720000	6771510000	6342450000	6619680000	6761720000	5
Dieldrin	5929890000	5841940000	5818180000	5550330000	6016400000	5831350000	3
Endosulfan I	5452270000	5404490000	5449050000	5348660000	6028770000	5536650000	5
Endosulfan II	4873230000	4835950000	4885080000	4789750000	5615800000	4999960000	7
Endosulfan sulfate	4557060000	4504850000	4556170000	4479960000	5092920000	4638190000	6
Endrin	5035940000	4962910000	4938250000	4711180000	5120260000	4953710000	3
Endrin aldehyde	3505880000	3507960000	3571680000	3579710000	4148280000	3662700000	7
Endrin ketone	4819770000	4736510000	4782320000	4695250000	5260400000	4858850000	5
gamma-BHC (Lindane)	7185880000	6905860000	6799650000	6417980000	6684750000	6798820000	4
gamma-Chlordane	5972450000	5894040000	5855640000	5592670000	6341090000	5931180000	5
Heptachlor	5757910000	5597510000	5545540000	5288860000	5757970000	5589560000	3
Heptachlor epoxide	5795570000	5708170000	5718680000	5584980000	6137920000	5789060000	4
Methoxychlor	2050060000	2015340000	2078640000	2094570000	2389310000	2125590000	7
Tetrachloro-m-xylene	3010860000	2951180000	3030100000	3124300000	3454390000	3114160000	6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: ICES02
SDG NO.: Q3399

Calibration Date(s): 10/17/2025 10/17/2025
Calibration Times: 11:34 12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD090650.D</u>	CF 075 =	<u>PD090651.D</u>		
CF 050 =		<u>PD090652.D</u>	CF 025 =	<u>PD090653.D</u>	CF 005 =	<u>PD090654.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	33708900000	34309400000	35560900000	36237500000	42373300000	36438000000	10
4,4'-DDE	40500800000	41191000000	42548600000	43577200000	51373700000	43838300000	10
4,4'-DDT	33888000000	33982600000	34868200000	34959100000	37667600000	35073100000	4
Aldrin	43083800000	43574800000	45033800000	46172400000	53729700000	46318900000	9
alpha-BHC	49118000000	48392000000	49683700000	50500300000	57784900000	51095800000	7
alpha-Chlordane	40004000000	40550200000	41893000000	42708900000	50532500000	43137700000	10
beta-BHC	18296700000	18365500000	19080800000	19804100000	23600800000	19829500000	11
Decachlorobiphenyl	28856800000	28247200000	29308000000	30057200000	34987600000	30291400000	9
delta-BHC	45324900000	45220200000	46466400000	47155800000	54416000000	47716700000	8
Dieldrin	40601100000	41425500000	43096200000	44203200000	51593900000	44184000000	10
Endosulfan I	35728600000	36586400000	38185700000	39478600000	46972400000	39390300000	11
Endosulfan II	34269700000	34968100000	36432200000	37520200000	44587200000	37555500000	11
Endosulfan sulfate	32963400000	33538900000	34960800000	35861000000	41900900000	35845000000	10
Endrin	36902000000	37827400000	39400900000	40693000000	47646500000	40494000000	11
Endrin aldehyde	25110900000	25739400000	26893100000	27767400000	32953200000	27692800000	11
Endrin ketone	34460000000	35020500000	36550700000	37816700000	44740900000	37717800000	11
gamma-BHC (Lindane)	44964800000	44604900000	45972800000	46808300000	53893500000	47248900000	8
gamma-Chlordane	41920800000	42422300000	43748100000	44346600000	52306700000	44948900000	9
Heptachlor	41272300000	41810600000	43239300000	44502900000	52189000000	44602800000	10
Heptachlor epoxide	38203400000	38891500000	40672900000	42374400000	53052500000	42639000000	14
Methoxychlor	16116900000	16316400000	17198900000	17735100000	19771700000	17427800000	8
Tetrachloro-m-xylene	27903400000	27689100000	28617400000	29726500000	34949400000	29777100000	10



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ICES02

Lab Code: ACE SDG NO.: Q3399

Instrument ID: ECD_D Date(s) Analyzed: 10/17/2025 10/17/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.24	6.14	6.34	32564400
		2	6.44	6.34	6.54	48621800
		3	7.14	7.04	7.24	85710600
		4	7.56	7.46	7.66	95737400
		5	7.92	7.82	8.02	60480500



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ICES02

Lab Code: ACE SDG NO.: Q3399

Instrument ID: ECD_D Date(s) Analyzed: 10/17/2025 10/17/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.47	5.37	5.57	263702000
		2	5.64	5.54	5.74	176560000
		3	6.75	6.65	6.85	658883000
		4	7.19	7.09	7.29	495104000
		5	7.32	7.22	7.42	246033000

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090650.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:34
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:33:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.874	301.1E6	2790.3E6	99.682	98.737
28) SA Decachlor...	9.066	8.061	436.1E6	2885.7E6	98.364	99.224
Target Compounds						
2) A alpha-BHC	3.994	3.386	775.3E6	4911.8E6	103.334	99.427
3) MA gamma-BHC...	4.324	3.722	718.6E6	4496.5E6	102.762	98.892
4) MA Heptachlor	4.922	4.075	575.8E6	4127.2E6	101.879	97.673
5) MB Aldrin	5.264	4.360	669.4E6	4308.4E6	101.342	97.787
6) B beta-BHC	4.511	4.019	251.5E6	1829.7E6	99.502	97.902
7) B delta-BHC	4.760	4.255	716.9E6	4532.5E6	102.853	98.756
8) B Heptachlo...	5.685	4.864	579.6E6	3820.3E6	100.668	96.869
9) A Endosulfan I	6.068	5.238	545.2E6	3572.9E6	100.030	96.676
10) B gamma-Chl...	5.940	5.117	597.2E6	4192.1E6	100.988	97.867
11) B alpha-Chl...	6.020	5.181	596.6E6	4000.4E6	100.230	97.693
12) B 4,4'-DDE	6.190	5.366	543.6E6	4050.1E6	101.097	97.534
13) MA Dieldrin	6.341	5.504	593.0E6	4060.1E6	100.951	97.019
14) MA Endrin	6.568	5.780	503.6E6	3690.2E6	100.979	96.725
15) B Endosulfa...	6.781	6.072	487.3E6	3427.0E6	99.879	96.941
16) A 4,4'-DDD	6.700	5.920	415.1E6	3370.9E6	101.164	97.326
17) MA 4,4'-DDT	7.016	6.174	425.7E6	3388.8E6	101.658	98.574
18) B Endrin al...	6.910	6.250	350.6E6	2511.1E6	99.070	96.573
19) B Endosulfa...	7.145	6.474	455.7E6	3296.3E6	100.010	97.059
20) A Methoxychlor	7.489	6.745	205.0E6	1611.7E6	99.308	96.752
21) B Endrin ke...	7.625	6.982	482.0E6	3446.0E6	100.390	97.056
22) Mirex	8.107	7.175	295.5E6	2331.6E6	97.723	97.238



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090650.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:34
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:33:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090650.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 301086012
Conc: 99.68 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100


Time Response_ Signal: PD090650.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 2790335064
Conc: 98.74 ng/ml


Time Response_ Signal: PD090650.D\ECD1A.ch



#2 alpha-BHC


R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 775252906
Conc: 103.33 ng/ml

Time Response_ Signal: PD090650.D\ECD2B.ch



#2 alpha-BHC


R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 4911798337
Conc: 99.43 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 718588040
Conc: 102.76 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100




#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 4496483491
Conc: 98.89 ng/ml



#4 Heptachlor


R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 575791141
Conc: 101.88 ng/ml



#4 Heptachlor

R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 4127234386
Conc: 97.67 ng/ml

Response_ Signal: PD090650.D\ECD1A.ch




#5 Aldrin

R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 669423227
Conc: 101.34 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100


Time Response_ Signal: PD090650.D\ECD2B.ch



#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 4308380411
Conc: 97.79 ng/ml


Time Response_ Signal: PD090650.D\ECD1A.ch



#6 beta-BHC

R.T.: 4.511 min
Delta R.T.: 0.000 min
Response: 251520811
Conc: 99.50 ng/ml

Time Response_ Signal: PD090650.D\ECD2B.ch



#6 beta-BHC

R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 1829668615
Conc: 97.90 ng/ml

Response_

Signal: PD090650.D\ECD1A.ch

#7 delta-BHC

R.T.: 4.760 min

Delta R.T.: 0.000 min

Response: 716926396


Conc: 102.85 ng/ml

Instrument :

ECD_D

ClientSampleId :

PSTDICC100



Time

Response_

Signal: PD090650.D\ECD2B.ch


#7 delta-BHC

R.T.: 4.255 min

Delta R.T.: 0.000 min

Response: 4532491918

Conc: 98.76 ng/ml



Time

Response_

Signal: PD090650.D\ECD1A.ch


#8 Heptachlor epoxide

R.T.: 5.685 min

Delta R.T.: 0.000 min

Response: 579557032

Conc: 100.67 ng/ml



Time

Response_

Signal: PD090650.D\ECD2B.ch


#8 Heptachlor epoxide

R.T.: 4.864 min


Delta R.T.: 0.000 min

Response: 3820342410

Conc: 96.87 ng/ml




Time



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 545227033
Conc: 100.03 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 3572859410
Conc: 96.68 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 597245354
Conc: 100.99 ng/ml



#10 gamma-Chlordane


R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 4192081714
Conc: 97.87 ng/ml



#11 alpha-Chlordane


R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 596606616
Conc: 100.23 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 4000403429
Conc: 97.69 ng/ml



#12 4,4'-DDE


R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 543560575
Conc: 101.10 ng/ml



#12 4,4'-DDE

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 4050079316
Conc: 97.53 ng/ml

Response_ Signal: PD090650.D\ECD1A.ch




#13 Dieldrin

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 592988915
Conc: 100.95 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100


Time Response_ Signal: PD090650.D\ECD2B.ch



#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 4060107897
Conc: 97.02 ng/ml


Time Response_ Signal: PD090650.D\ECD1A.ch



#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 503593872
Conc: 100.98 ng/ml

Time Response_ Signal: PD090650.D\ECD2B.ch



#14 Endrin

R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 3690195471
Conc: 96.72 ng/ml

Response_ Signal: PD090650.D\ECD1A.ch

#15 Endosulfan II

R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 487323424
Conc: 99.88 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Time

Response_ Signal: PD090650.D\ECD2B.ch

#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 3426973062
Conc: 96.94 ng/ml

Time

Response_ Signal: PD090650.D\ECD1A.ch

#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 415098011
Conc: 101.16 ng/ml

Time

Response_ Signal: PD090650.D\ECD2B.ch

#16 4,4'-DDD

R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 3370894623
Conc: 97.33 ng/ml

Time

Response_ Signal: PD090650.D\ECD1A.ch

#17 4,4'-DDT

R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 425681261
Conc: 101.66 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Time 6.60 6.70 6.80 6.90 7.00 7.10 7.20 7.30 7.40

Response_ Signal: PD090650.D\ECD2B.ch

#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 3388802413
Conc: 98.57 ng/ml

Time 5.95 6.00 6.05 6.10 6.15 6.20 6.25 6.30 6.35

Response_ Signal: PD090650.D\ECD1A.ch

#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 350587578
Conc: 99.07 ng/ml

Time 6.70 6.80 6.90 7.00 7.10

Response_ Signal: PD090650.D\ECD2B.ch

#18 Endrin aldehyde

R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 2511089389
Conc: 96.57 ng/ml


Time 6.00 6.10 6.20 6.30 6.40 6.50

Response_ Signal: PD090650.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 455705527
Conc: 100.01 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC100



Time Response_ Signal: PD090650.D\ECD2B.ch

#19 Endosulfan Sulfate


R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 3296343899
Conc: 97.06 ng/ml



Time Response_ Signal: PD090650.D\ECD1A.ch

#20 Methoxychlor


R.T.: 7.489 min
Delta R.T.: 0.000 min
Response: 205006247
Conc: 99.31 ng/ml



Time Response_ Signal: PD090650.D\ECD2B.ch

#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 1611694442
Conc: 96.75 ng/ml



Response_ Signal: PD090650.D\ECD1A.ch

#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 481977282
Conc: 100.39 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Time

Response_ Signal: PD090650.D\ECD2B.ch

#21 Endrin ketone

R.T.: 6.982 min
Delta R.T.: 0.000 min
Response: 3446004756
Conc: 97.06 ng/ml

Time

Response_ Signal: PD090650.D\ECD1A.ch

#22 Mirex

R.T.: 8.107 min
Delta R.T.: 0.000 min
Response: 295471777
Conc: 97.72 ng/ml

Time

Response_ Signal: PD090650.D\ECD2B.ch

#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 2331618859
Conc: 97.24 ng/ml


Time

Response_

Signal: PD090650.D\ECD1A.ch

#28 Decachlorobiphenyl

ICAL Form



R.T.: 9.066 min
Delta R.T.: 0.000 min
Response: 436118427
Conc: 98.36 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Time

Response

Signal: PD090650.D\ECD2B.ch

#28 Decachlorobiphenyl



R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 2885681422
Conc: 99.22 ng/ml

Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090651.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 11:48
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 13:35:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 13:29:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.875	221.3E6	2076.7E6	73.844	73.982
28)	SA Decachlor...	9.065	8.062	323.6E6	2118.5E6	73.652	73.550
Target Compounds							
2)	A alpha-BHC	3.994	3.387	555.2E6	3629.4E6	74.331	73.972
3)	MA gamma-BHC...	4.325	3.723	517.9E6	3345.4E6	74.376	74.044
4)	MA Heptachlor	4.923	4.076	419.8E6	3135.8E6	74.519	74.471
5)	MB Aldrin	5.265	4.361	492.2E6	3268.1E6	74.671	74.449
6)	B beta-BHC	4.512	4.019	186.7E6	1377.4E6	74.235	74.130
7)	B delta-BHC	4.760	4.255	517.9E6	3391.5E6	74.535	74.261
8)	B Heptachlo...	5.685	4.865	428.1E6	2916.9E6	74.574	74.304
9)	A Endosulfan I	6.068	5.239	405.3E6	2744.0E6	74.575	74.497
10)	B gamma-Chl...	5.940	5.117	442.1E6	3181.7E6	74.831	74.517
11)	B alpha-Chl...	6.021	5.182	442.9E6	3041.3E6	74.605	74.512
12)	B 4,4'-DDE	6.191	5.366	402.8E6	3089.3E6	74.945	74.597
13)	MA Dieldrin	6.341	5.504	438.1E6	3106.9E6	74.726	74.493
14)	MA Endrin	6.568	5.781	372.2E6	2837.1E6	74.757	74.574
15)	B Endosulfa...	6.781	6.072	362.7E6	2622.6E6	74.556	74.457
16)	A 4,4'-DDD	6.700	5.921	305.6E6	2573.2E6	74.659	74.529
17)	MA 4,4'-DDT	7.016	6.175	311.4E6	2548.7E6	74.584	74.422
18)	B Endrin al...	6.910	6.251	263.1E6	1930.5E6	74.563	74.493
19)	B Endosulfa...	7.145	6.474	337.9E6	2515.4E6	74.430	74.374
20)	A Methoxychlor	7.488	6.745	151.2E6	1223.7E6	73.804	73.968
21)	B Endrin ke...	7.625	6.983	355.2E6	2626.5E6	74.325	74.314
22)	Mirex	8.108	7.176	222.5E6	1780.4E6	74.051	74.497



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090651.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:48
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:35:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm




Response_ Signal: PD090651.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 221338475
Conc: 73.84 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC075



Response_ Signal: PD090651.D\ECD2B.ch

#1 Tetrachloro-m-xylene


R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 2076681010
Conc: 73.98 ng/ml



Response_ Signal: PD090651.D\ECD1A.ch

#2 alpha-BHC


R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 555178575
Conc: 74.33 ng/ml



Response_ Signal: PD090651.D\ECD2B.ch

#2 alpha-BHC

R.T.: 3.387 min
Delta R.T.: 0.000 min
Response: 3629403435
Conc: 73.97 ng/ml



Response_

Signal: PD090651.D\ECD1A.ch

#3 gamma-BHC (Lindane)

R.T.: 4.325 min

Delta R.T.: 0.000 min

Response: 517939637


Conc: 74.38 ng/ml

Instrument :

ECD_D

ClientSampleId :

PSTDICC075



Time

Response_

Signal: PD090651.D\ECD2B.ch


#3 gamma-BHC (Lindane)

R.T.: 3.723 min

Delta R.T.: 0.000 min

Response: 3345366469

Conc: 74.04 ng/ml



Time

Response_

Signal: PD090651.D\ECD1A.ch


#4 Heptachlor

R.T.: 4.923 min

Delta R.T.: 0.000 min

Response: 419813050

Conc: 74.52 ng/ml



Time

Response_

Signal: PD090651.D\ECD2B.ch


#4 Heptachlor

R.T.: 4.076 min


Delta R.T.: 0.000 min

Response: 3135796942

Conc: 74.47 ng/ml




Time



#5 Aldrin


R.T.: 5.265 min
Delta R.T.: 0.000 min
Response: 492167044
Conc: 74.67 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075




#5 Aldrin

R.T.: 4.361 min
Delta R.T.: 0.000 min
Response: 3268108002
Conc: 74.45 ng/ml



#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 186698186
Conc: 74.24 ng/ml




#6 beta-BHC

R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 1377409505
Conc: 74.13 ng/ml

Response_

Signal: PD090651.D\ECD1A.ch

#7 delta-BHC



R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 517928984
Conc: 74.53 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Time

Response_

Signal: PD090651.D\ECD2B.ch

#7 delta-BHC




R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 3391516826
Conc: 74.26 ng/ml

Time

Response_

Signal: PD090651.D\ECD1A.ch

#8 Heptachlor epoxide




R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 428112423
Conc: 74.57 ng/ml

Time

Response_

Signal: PD090651.D\ECD2B.ch


#8 Heptachlor epoxide



R.T.: 4.865 min
Delta R.T.: 0.000 min
Response: 2916863262
Conc: 74.30 ng/ml

Time

Response_ Signal: PD090651.D\ECD1A.ch




#9 Endosulfan I

R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 405336624
Conc: 74.58 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075


Time Response_ Signal: PD090651.D\ECD2B.ch



#9 Endosulfan I

R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 2743979164
Conc: 74.50 ng/ml


Time Response_ Signal: PD090651.D\ECD1A.ch



#10 gamma-Chlordane


R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 442053358
Conc: 74.83 ng/ml

Time Response_ Signal: PD090651.D\ECD2B.ch




#10 gamma-Chlordane

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 3181670591
Conc: 74.52 ng/ml




R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 442910146
Conc: 74.61 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC075



R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 3041262707
Conc: 74.51 ng/ml



R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 402806011
Conc: 74.95 ng/ml




R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 3089327488
Conc: 74.60 ng/ml

Response_ Signal: PD090651.D\ECD1A.ch

#13 Dieldrin

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 438145830
Conc: 74.73 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC075



Time Response_ Signal: PD090651.D\ECD2B.ch

#13 Dieldrin


R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 3106916142
Conc: 74.49 ng/ml



Time Response_ Signal: PD090651.D\ECD1A.ch

#14 Endrin


R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 372217983
Conc: 74.76 ng/ml




Time Response_ Signal: PD090651.D\ECD2B.ch

#14 Endrin

R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 2837058459
Conc: 74.57 ng/ml






#15 Endosulfan II


R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 362696180
Conc: 74.56 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 2622605787
Conc: 74.46 ng/ml



#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 305645357
Conc: 74.66 ng/ml



#16 4,4'-DDD

R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 2573202277
Conc: 74.53 ng/ml

Response_ Signal: PD090651.D\ECD1A.ch

#17 4,4'-DDT

R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 311444736
Conc: 74.58 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Time

Response_ Signal: PD090651.D\ECD2B.ch

#17 4,4'-DDT

R.T.: 6.175 min
Delta R.T.: 0.000 min
Response: 2548691729
Conc: 74.42 ng/ml

Time

Response_ Signal: PD090651.D\ECD1A.ch

#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 263096696
Conc: 74.56 ng/ml

Time


Response_ Signal: PD090651.D\ECD2B.ch

#18 Endrin aldehyde

R.T.: 6.251 min
Delta R.T.: 0.000 min
Response: 1930452420
Conc: 74.49 ng/ml

Time

Response_ Signal: PD090651.D\ECD1A.ch




#19 Endosulfan Sulfate

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 337864118
Conc: 74.43 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075


Time Response_ Signal: PD090651.D\ECD2B.ch



#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 2515420876
Conc: 74.37 ng/ml


Time Response_ Signal: PD090651.D\ECD1A.ch



#20 Methoxychlor


R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 151150843
Conc: 73.80 ng/ml

Time Response_ Signal: PD090651.D\ECD2B.ch



#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 1223732210
Conc: 73.97 ng/ml



R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 355238454
Conc: 74.32 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Time

Response_

Signal: PD090651.D\ECD2B.ch

#21 Endrin ketone




R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 2626540426
Conc: 74.31 ng/ml

Time

Response_

Signal: PD090651.D\ECD1A.ch

#22 Mirex




R.T.: 8.108 min
Delta R.T.: 0.000 min
Response: 222493333
Conc: 74.05 ng/ml

Time

Response_


Signal: PD090651.D\ECD2B.ch

#22 Mirex



R.T.: 7.176 min
Delta R.T.: 0.000 min
Response: 1780357534
Conc: 74.50 ng/ml


Time



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.000 min
Response: 323645321
Conc: 73.65 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC075



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 2118537336
Conc: 73.55 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090652.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 12:01
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 13:30:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 13:29:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.875	151.5E6	1430.9E6	50.000	50.000
28)	SA Decachlor...	9.064	8.062	225.3E6	1465.4E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.994	3.387	362.6E6	2484.2E6	50.000	50.000
3)	MA gamma-BHC...	4.325	3.723	340.0E6	2298.6E6	50.000	50.000
4)	MA Heptachlor	4.923	4.075	277.3E6	2162.0E6	50.000	50.000
5)	MB Aldrin	5.264	4.361	325.8E6	2251.7E6	50.000	50.000
6)	B beta-BHC	4.512	4.019	127.0E6	954.0E6	50.000	50.000
7)	B delta-BHC	4.760	4.255	338.6E6	2323.3E6	50.000	50.000
8)	B Heptachlo...	5.684	4.864	285.9E6	2033.6E6	50.000	50.000
9)	A Endosulfan I	6.068	5.238	272.5E6	1909.3E6	50.000	50.000
10)	B gamma-Chl...	5.940	5.117	292.8E6	2187.4E6	50.000	50.000
11)	B alpha-Chl...	6.021	5.181	296.9E6	2094.7E6	50.000	50.000
12)	B 4,4'-DDE	6.190	5.366	265.9E6	2127.4E6	50.000	50.000
13)	MA Dieldrin	6.340	5.504	290.9E6	2154.8E6	50.000	50.000
14)	MA Endrin	6.568	5.781	246.9E6	1970.0E6	50.000	50.000
15)	B Endosulfa...	6.781	6.072	244.3E6	1821.6E6	50.000	50.000
16)	A 4,4'-DDD	6.700	5.921	202.8E6	1778.0E6	50.000	50.000
17)	MA 4,4'-DDT	7.015	6.174	205.9E6	1743.4E6	50.000	50.000
18)	B Endrin al...	6.910	6.250	178.6E6	1344.7E6	50.000	50.000
19)	B Endosulfa...	7.145	6.474	227.8E6	1748.0E6	50.000	50.000
20)	A Methoxychlor	7.488	6.745	103.9E6	859.9E6	50.000	50.000
21)	B Endrin ke...	7.625	6.983	239.1E6	1827.5E6	50.000	50.000
22)	Mirex	8.107	7.175	154.6E6	1232.0E6	50.000	50.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090652.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 12:01
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:30:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090652.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 151504775
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Time 3.20 3.30 3.40 3.50 3.60 3.70 3.80 3.90

Response_ Signal: PD090652.D\ECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 1430869153
Conc: 50.00 ng/ml

Time 2.70 2.75 2.80 2.85 2.90 2.95 3.00 3.05

Response_ Signal: PD090652.D\ECD1A.ch

#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 362610738
Conc: 50.00 ng/ml

Time 3.70 3.80 3.90 4.00 4.10 4.20 4.30

Response_ Signal: PD090652.D\ECD2B.ch

#2 alpha-BHC


R.T.: 3.387 min
Delta R.T.: 0.000 min
Response: 2484185326
Conc: 50.00 ng/ml

Time 3.20 3.30 3.40 3.50 3.60

Response_

Signal: PD090652.D\ECD1A.ch

#3 gamma-BHC (Lindane)



R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 339982682
Conc: 50.00 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Time

Response_

Signal: PD090652.D\ECD2B.ch

#3 gamma-BHC (Lindane)




R.T.: 3.723 min
Delta R.T.: 0.000 min
Response: 2298638612
Conc: 50.00 ng/ml

Time

Response_

Signal: PD090652.D\ECD1A.ch

#4 Heptachlor




R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 277277005
Conc: 50.00 ng/ml

Time

Response_


Signal: PD090652.D\ECD2B.ch

#4 Heptachlor



R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 2161964168
Conc: 50.00 ng/ml

Time



R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 325847234
Conc: 50.00 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Time

Response_

Signal: PD090652.D\ECD2B.ch

#5 Aldrin




R.T.: 4.361 min
Delta R.T.: 0.000 min
Response: 2251688760
Conc: 50.00 ng/ml

Time

Response_

Signal: PD090652.D\ECD1A.ch

#6 beta-BHC




R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 127018373
Conc: 50.00 ng/ml

Time

Response_


Signal: PD090652.D\ECD2B.ch

#6 beta-BHC



R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 954037739
Conc: 50.00 ng/ml


Time



#7 delta-BHC


R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 338575319
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050




#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 2323317526
Conc: 50.00 ng/ml




#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 285933842
Conc: 50.00 ng/ml



#8 Heptachlor epoxide


R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 2033647315
Conc: 50.00 ng/ml



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 272452330
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 1909283878
Conc: 50.00 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 292781895
Conc: 50.00 ng/ml



#10 gamma-Chlordane


R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 2187405429
Conc: 50.00 ng/ml



#11 alpha-Chlordane


R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 296932529
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2094650466
Conc: 50.00 ng/ml




#12 4,4'-DDE

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 265880490
Conc: 50.00 ng/ml



#12 4,4'-DDE


R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 2127431088
Conc: 50.00 ng/ml



#13 Dieldrin


R.T.: 6.340 min
Delta R.T.: 0.000 min
Response: 290909064
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050




#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 2154808119
Conc: 50.00 ng/ml



#14 Endrin


R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 246912451
Conc: 50.00 ng/ml



#14 Endrin

R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 1970044096
Conc: 50.00 ng/ml

Response_ Signal: PD090652.D\ECD1A.ch




#15 Endosulfan II

R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 244253888
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050


Time Response_ Signal: PD090652.D\ECD2B.ch



#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1821608734
Conc: 50.00 ng/ml


Time Response_ Signal: PD090652.D\ECD1A.ch



#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 202773551
Conc: 50.00 ng/ml


Time Response_ Signal: PD090652.D\ECD2B.ch



#16 4,4'-DDD

R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 1778046885
Conc: 50.00 ng/ml

Response_ Signal: PD090652.D\ECD1A.ch




#17 4,4'-DDT

R.T.: 7.015 min
Delta R.T.: 0.000 min
Response: 205896645
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050


Time Response_ Signal: PD090652.D\ECD2B.ch



#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1743410018
Conc: 50.00 ng/ml


Time Response_ Signal: PD090652.D\ECD1A.ch



#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 178583940
Conc: 50.00 ng/ml

Time Response_ Signal: PD090652.D\ECD2B.ch



#18 Endrin aldehyde

R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1344655981
Conc: 50.00 ng/ml

Response_ Signal: PD090652.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 227808376
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Time

Response_ Signal: PD090652.D\ECD2B.ch

#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1748039260
Conc: 50.00 ng/ml

Time

Response_ Signal: PD090652.D\ECD1A.ch

#20 Methoxychlor

R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 103932037
Conc: 50.00 ng/ml


Time

Response_ Signal: PD090652.D\ECD2B.ch

#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 859946427
Conc: 50.00 ng/ml


Time



#21 Endrin ketone


R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 239116023
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050




#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1827533277
Conc: 50.00 ng/ml




#22 Mirex

R.T.: 8.107 min
Delta R.T.: 0.000 min
Response: 154621594
Conc: 50.00 ng/ml



#22 Mirex


R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1232028398
Conc: 50.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.064 min
Delta R.T.: 0.000 min
Response: 225311337
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC050



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 1465399343
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090653.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 12:15
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:38:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.874	78107465	743.2E6	25.786	26.090
28) SA Decachlor...	9.065	8.061	117.7E6	751.4E6	26.311	25.807
Target Compounds						
2) A alpha-BHC	3.993	3.386	168.6E6	1262.5E6	23.132	25.545
3) MA gamma-BHC...	4.324	3.722	160.4E6	1170.2E6	23.501	25.669
4) MA Heptachlor	4.922	4.074	132.2E6	1112.6E6	23.835	26.052
5) MB Aldrin	5.263	4.360	156.0E6	1154.3E6	23.982	25.959
6) B beta-BHC	4.511	4.018	64015475	495.1E6	25.339	26.214
7) B delta-BHC	4.759	4.255	158.6E6	1178.9E6	23.327	25.605
8) B Heptachlo...	5.684	4.864	139.6E6	1059.4E6	24.488	26.460
9) A Endosulfan I	6.067	5.238	133.7E6	987.0E6	24.700	26.323
10) B gamma-Chl...	5.939	5.116	139.8E6	1108.7E6	23.988	25.717
11) B alpha-Chl...	6.020	5.181	154.3E6	1067.7E6	25.731	25.860
12) B 4,4'-DDE	6.190	5.365	126.3E6	1089.4E6	23.861	25.967
13) MA Dieldrin	6.340	5.503	138.8E6	1105.1E6	23.985	26.105
14) MA Endrin	6.567	5.780	117.8E6	1017.3E6	23.978	26.284
15) B Endosulfa...	6.781	6.071	119.7E6	938.0E6	24.710	26.203
16) A 4,4'-DDD	6.700	5.920	96019034	905.9E6	23.822	25.918
17) MA 4,4'-DDT	7.016	6.173	98197092	874.0E6	23.870	25.388
18) B Endrin al...	6.909	6.250	89492653	694.2E6	25.271	26.317
19) B Endosulfa...	7.144	6.473	112.0E6	896.5E6	24.754	26.114
20) A Methoxychlor	7.488	6.744	52364279	443.4E6	25.424	26.326
21) B Endrin ke...	7.625	6.982	117.4E6	945.4E6	24.668	26.289
22) Mirex	8.108	7.175	79715192	640.8E6	26.131	26.336



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090653.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 12:15
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:38:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm




Response_ Signal: PD090653.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 78107465
Conc: 25.79 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC025



Time Response_ Signal: PD090653.D\ECD2B.ch

#1 Tetrachloro-m-xylene


R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 743162855
Conc: 26.09 ng/ml



Time Response_ Signal: PD090653.D\ECD1A.ch

#2 alpha-BHC


R.T.: 3.993 min
Delta R.T.: 0.000 min
Response: 168576733
Conc: 23.13 ng/ml




Time Response_ Signal: PD090653.D\ECD2B.ch

#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 1262506307
Conc: 25.54 ng/ml






#3 gamma-BHC (Lindane)


R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 160449432
Conc: 23.50 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025




#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 1170208039
Conc: 25.67 ng/ml




#4 Heptachlor

R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 132221469
Conc: 23.83 ng/ml



#4 Heptachlor

R.T.: 4.074 min
Delta R.T.: 0.000 min
Response: 1112572417
Conc: 26.05 ng/ml



R.T.: 5.263 min
Delta R.T.: 0.000 min
Response: 155953472
Conc: 23.98 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Time

Response_

Signal: PD090653.D\ECD2B.ch

#5 Aldrin




R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 1154309416
Conc: 25.96 ng/ml

Time

Response_

Signal: PD090653.D\ECD1A.ch

#6 beta-BHC




R.T.: 4.511 min
Delta R.T.: 0.000 min
Response: 64015475
Conc: 25.34 ng/ml

Time

Response_


Signal: PD090653.D\ECD2B.ch

#6 beta-BHC



R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 495101315
Conc: 26.21 ng/ml


Time



#7 delta-BHC


R.T.: 4.759 min
Delta R.T.: 0.000 min
Response: 158561347
Conc: 23.33 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025




#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 1178894488
Conc: 25.60 ng/ml



#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 139624515
Conc: 24.49 ng/ml




#8 Heptachlor epoxide

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 1059359842
Conc: 26.46 ng/ml

Response_

Signal: PD090653.D\ECD1A.ch

#9 Endosulfan I



R.T.: 6.067 min
Delta R.T.: 0.000 min
Response: 133716398
Conc: 24.70 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Time

Response

Signal: PD090653.D\ECD2B.ch

#9 Endosulfan I




R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 986966005
Conc: 26.32 ng/ml

Time

Response_

Signal: PD090653.D\ECD1A.ch

#10 gamma-Chlordane




R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 139816717
Conc: 23.99 ng/ml

Time

Response


Signal: PD090653.D\ECD2B.ch

#10 gamma-Chlordane



R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 1108664102
Conc: 25.72 ng/ml

Time



R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 154261488
Conc: 25.73 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Time 5.80 5.85 5.90 5.95 6.00 6.05 6.10 6.15 6.20

Response

Signal: PD090653.D\ECD2B.ch

#11 alpha-Chlordane




R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 1067721561
Conc: 25.86 ng/ml

Time 5.05 5.10 5.15 5.20 5.25 5.30

Response_

Signal: PD090653.D\ECD1A.ch

#12 4,4'-DDE




R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 126326499
Conc: 23.86 ng/ml

Time 6.00 6.10 6.20 6.30 6.40

Response

Signal: PD090653.D\ECD2B.ch


#12 4,4'-DDE



R.T.: 5.365 min
Delta R.T.: 0.000 min
Response: 1089429298
Conc: 25.97 ng/ml

Time 5.20 5.30 5.40 5.50 5.60

Response_ Signal: PD090653.D\ECD1A.ch




#13 Dieldrin

R.T.: 6.340 min
Delta R.T.: 0.000 min
Response: 138758150
Conc: 23.99 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025


Time Response_ Signal: PD090653.D\ECD2B.ch



#13 Dieldrin

R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 1105079839
Conc: 26.11 ng/ml


Time Response_ Signal: PD090653.D\ECD1A.ch



#14 Endrin


R.T.: 6.567 min
Delta R.T.: 0.000 min
Response: 117779573
Conc: 23.98 ng/ml

Time Response_ Signal: PD090653.D\ECD2B.ch



#14 Endrin

R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 1017324913
Conc: 26.28 ng/ml



R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 119743666
Conc: 24.71 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Time

Response_

Signal: PD090653.D\ECD2B.ch

#15 Endosulfan II




R.T.: 6.071 min
Delta R.T.: 0.000 min
Response: 938003929
Conc: 26.20 ng/ml

Time

Response_

Signal: PD090653.D\ECD1A.ch

#16 4,4'-DDD




R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 96019034
Conc: 23.82 ng/ml

Time


Response_

Signal: PD090653.D\ECD2B.ch

#16 4,4'-DDD




R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 905938376
Conc: 25.92 ng/ml



#17 4,4'-DDT


R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 98197092
Conc: 23.87 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025




#17 4,4'-DDT

R.T.: 6.173 min
Delta R.T.: 0.000 min
Response: 873976959
Conc: 25.39 ng/ml




#18 Endrin aldehyde

R.T.: 6.909 min
Delta R.T.: 0.000 min
Response: 89492653
Conc: 25.27 ng/ml



#18 Endrin aldehyde

R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 694183960
Conc: 26.32 ng/ml



R.T.: 7.144 min
Delta R.T.: 0.000 min
Response: 111999004
Conc: 24.75 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC025

Time

Response_

Signal: PD090653.D\ECD2B.ch

#19 Endosulfan Sulfate




R.T.: 6.473 min
Delta R.T.: 0.000 min
Response: 896525775
Conc: 26.11 ng/ml

Time

Response_

Signal: PD090653.D\ECD1A.ch

#20 Methoxychlor




R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 52364279
Conc: 25.42 ng/ml

Time

Response_

Signal: PD090653.D\ECD2B.ch


#20 Methoxychlor



R.T.: 6.744 min
Delta R.T.: 0.000 min
Response: 443376884
Conc: 26.33 ng/ml

Time

Response_ Signal: PD090653.D\ECD1A.ch




#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 117381168
Conc: 24.67 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025


Time Response_ Signal: PD090653.D\ECD2B.ch



#21 Endrin ketone

R.T.: 6.982 min
Delta R.T.: 0.000 min
Response: 945417829
Conc: 26.29 ng/ml


Time Response_ Signal: PD090653.D\ECD1A.ch



#22 Mirex


R.T.: 8.108 min
Delta R.T.: 0.000 min
Response: 79715192
Conc: 26.13 ng/ml

Time Response_ Signal: PD090653.D\ECD2B.ch



#22 Mirex


R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 640790883
Conc: 26.34 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.000 min
Response: 117676418
Conc: 26.31 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC025



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 751430638
Conc: 25.81 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090654.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 12:28
 Operator : AR\AJ
 Sample : PSTDICC005
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDICC005

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/19/2025

Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 13:41:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 13:29:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.543	2.874	17271955	174.7E6	5.550m	5.868
28) SA Decachlor...	9.066	8.062	27500355	174.9E6	5.879	5.775
Target Compounds						
2) A alpha-BHC	3.993	3.386	33677596	288.9E6	4.692	5.655
3) MA gamma-BHC...	4.324	3.723	33423739	269.5E6	4.916	5.703
4) MA Heptachlor	4.922	4.075	28789856	260.9E6	5.151	5.850
5) MB Aldrin	5.264	4.360	33624118	268.6E6	5.136	5.800
6) B beta-BHC	4.512	4.019	14963256	118.0E6	5.712	5.951
7) B delta-BHC	4.760	4.255	33098385	272.1E6	4.895	5.702
8) B Heptachlo...	5.684	4.864	30689599	265.3E6	5.301	6.280
9) A Endosulfan I	6.067	5.238	30143871	234.9E6	5.444	5.962
10) B gamma-Chl...	5.940	5.117	31705442	261.5E6	5.346	5.818
11) B alpha-Chl...	6.019	5.181	34603140	252.7E6	5.576m	5.857
12) B 4,4'-DDE	6.190	5.366	27280775	256.9E6	5.122	5.859
13) MA Dieldrin	6.341	5.504	30082015	258.0E6	5.159	5.839
14) MA Endrin	6.568	5.781	25601278	238.2E6	5.168	5.883
15) B Endosulfa...	6.781	6.072	28078989	222.9E6	5.616	5.936
16) A 4,4'-DDD	6.700	5.921	20742842	211.9E6	5.116	5.814
17) MA 4,4'-DDT	7.016	6.175	20485670	188.3E6	4.984	5.370
18) B Endrin al...	6.910	6.250	20741409	164.8E6	5.663	5.950
19) B Endosulfa...	7.144	6.474	25464611	209.5E6	5.490	5.845
20) A Methoxychlor	7.489	6.745	11946535	98858490	5.620	5.672
21) B Endrin ke...	7.624	6.983	26302012	223.7E6	5.414m	5.931
22) Mirex	8.108	7.175	19009608	155.6E6	5.939	6.058

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090654.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 12:28
Operator : AR\AJ
Sample : PSTDICC005
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDICC005

Manual Integrations



APPROVED

Reviewed By :Abdul Mirza 10/19/2025


Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:41:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:29:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090654.D\ECD1A.ch



#1 Tetrachloro-m-xylene

R.T.: 3.543 min
Delta R.T.: 0.000 min
Response: 17271955
Conc: 5.55 ng/ml

Instrument :

ECD_D

ClientSampleId :

PSTDICC005

Manual Integrations


APPROVED

Reviewed By :Abdul Mirza 10/19/2025

Supervised By :mohammad ahmed 10/24/2025

Time 3.35 3.40 3.45 3.50 3.55 3.60 3.65 3.70

Response_ Signal: PD090654.D\ECD2B.ch




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 174747032
Conc: 5.87 ng/ml

Time 2.70 2.80 2.90 3.00

Response_ Signal: PD090654.D\ECD1A.ch




#2 alpha-BHC

R.T.: 3.993 min
Delta R.T.: 0.000 min
Response: 33677596
Conc: 4.69 ng/ml

Time 3.70 3.80 3.90 4.00 4.10 4.20


Response_ Signal: PD090654.D\ECD2B.ch



#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 288924308
Conc: 5.65 ng/ml

Time 3.20 3.25 3.30 3.35 3.40 3.45 3.50 3.55



#3 gamma-BHC (Lindane)

R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 33423739
Conc: 4.92 ng/ml

Instrument :

ECD_D

ClientSampleId :


PSTDICC005

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 10/19/2025

Supervised By :mohammad ahmed 10/24/2025




#3 gamma-BHC (Lindane)

R.T.: 3.723 min
Delta R.T.: 0.000 min
Response: 269467396
Conc: 5.70 ng/ml




#4 Heptachlor

R.T.: 4.922 min
Delta R.T.: 0.000 min
Response: 28789856
Conc: 5.15 ng/ml



#4 Heptachlor

R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 260945196
Conc: 5.85 ng/ml




#5 Aldrin

R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 33624118
Conc: 5.14 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 268648325
Conc: 5.80 ng/ml




#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 14963256
Conc: 5.71 ng/ml



#6 beta-BHC

R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 118003884
Conc: 5.95 ng/ml




R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 33098385
Conc: 4.89 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDICC005

Manual Integrations
APPROVED


Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 272079912
Conc: 5.70 ng/ml



R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 30689599
Conc: 5.30 ng/ml



R.T.: 4.864 min
Delta R.T.: 0.001 min
Response: 265262691
Conc: 6.28 ng/ml




#9 Endosulfan I

R.T.: 6.067 min
Delta R.T.: 0.000 min
Response: 30143871
Conc: 5.44 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 234861846
Conc: 5.96 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 31705442
Conc: 5.35 ng/ml



#10 gamma-Chlordane

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 261533283
Conc: 5.82 ng/ml




#11 alpha-Chlordane

R.T.: 6.019 min
Delta R.T.: 0.000 min
Response: 34603140
Conc: 5.58 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 252662629
Conc: 5.86 ng/ml




#12 4,4'-DDE

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 27280775
Conc: 5.12 ng/ml



#12 4,4'-DDE

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 256868291
Conc: 5.86 ng/ml




#13 Dieldrin

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 30082015
Conc: 5.16 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 257969409
Conc: 5.84 ng/ml




#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 25601278
Conc: 5.17 ng/ml



#14 Endrin

R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 238232371
Conc: 5.88 ng/ml




#15 Endosulfan II

R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 28078989
Conc: 5.62 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 222935892
Conc: 5.94 ng/ml




#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 20742842
Conc: 5.12 ng/ml



#16 4,4'-DDD

R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 211866590
Conc: 5.81 ng/ml




#17 4,4'-DDT

R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 20485670
Conc: 4.98 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#17 4,4'-DDT

R.T.: 6.175 min
Delta R.T.: 0.000 min
Response: 188338004
Conc: 5.37 ng/ml




#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 20741409
Conc: 5.66 ng/ml



#18 Endrin aldehyde

R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 164766045
Conc: 5.95 ng/ml



#19 Endosulfan Sulfate

R.T.: 7.144 min
Delta R.T.: 0.000 min
Response: 25464611
Conc: 5.49 ng/ml

Instrument :

ECD_D

ClientSampleId :


PSTDICC005

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 10/19/2025

Supervised By :mohammad ahmed 10/24/2025




#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 209504586
Conc: 5.84 ng/ml



#20 Methoxychlor


R.T.: 7.489 min
Delta R.T.: 0.000 min
Response: 11946535
Conc: 5.62 ng/ml



#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 98858490
Conc: 5.67 ng/ml

Response_ Signal: PD090654.D\ECD1A.ch



#21 Endrin ketone


R.T.: 7.624 min
Delta R.T.: 0.000 min
Response: 26302012
Conc: 5.41 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025


Time Response_ Signal: PD090654.D\ECD2B.ch



#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 223704577
Conc: 5.93 ng/ml


Response_ Signal: PD090654.D\ECD1A.ch



#22 Mirex


R.T.: 8.108 min
Delta R.T.: 0.000 min
Response: 19009608
Conc: 5.94 ng/ml

Time Response_ Signal: PD090654.D\ECD2B.ch



#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 155640388
Conc: 6.06 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.000 min
Response: 27500355
Conc: 5.88 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDICC005

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 174937832
Conc: 5.78 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090657.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 13:09
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:49:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:48:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.874	108.8E6	1215.0E6	50.000	50.000
28)	SA Decachlor...	9.066	8.062	167.8E6	1105.6E6	50.000	50.000
Target Compounds							
23)	Chlordane-1	4.708	3.898	92608674	605.2E6	500.000	500.000
24)	Chlordane-2	5.235	4.479	92865500	631.9E6	500.000	500.000
25)	Chlordane-3	5.941	5.117	379.1E6	1885.9E6	500.000	500.000
26)	Chlordane-4	6.026	5.181	458.8E6	1625.9E6	500.000	500.000
27)	Chlordane-5	6.865	6.081	74238996	730.3E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090657.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 13:09
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 13:49:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 13:48:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 108758097
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 1214996007
Conc: 50.00 ng/ml



#23 Chlordane-1


R.T.: 4.708 min
Delta R.T.: 0.000 min
Response: 92608674
Conc: 500.00 ng/ml



#23 Chlordane-1

R.T.: 3.898 min
Delta R.T.: 0.000 min
Response: 605185458
Conc: 500.00 ng/ml

Response_ Signal: PD090657.D\ECD1A.ch




#24 Chlordane-2

R.T.: 5.235 min
Delta R.T.: 0.000 min
Response: 92865500
Conc: 500.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500


Time Response_ Signal: PD090657.D\ECD2B.ch



#24 Chlordane-2

R.T.: 4.479 min
Delta R.T.: 0.000 min
Response: 631892083
Conc: 500.00 ng/ml


Time Response_ Signal: PD090657.D\ECD1A.ch



#25 Chlordane-3

R.T.: 5.941 min
Delta R.T.: 0.000 min
Response: 379095261
Conc: 500.00 ng/ml

Time Response_ Signal: PD090657.D\ECD2B.ch



#25 Chlordane-3

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 1885915535
Conc: 500.00 ng/ml

Response_ Signal: PD090657.D\ECD1A.ch

#26 Chlordane-4

R.T.: 6.026 min
Delta R.T.: 0.000 min
Response: 458791566
Conc: 500.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500

Time

Response_ Signal: PD090657.D\ECD2B.ch

#26 Chlordane-4

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 1625937453
Conc: 500.00 ng/ml

Time

Response_ Signal: PD090657.D\ECD1A.ch

#27 Chlordane-5

R.T.: 6.865 min
Delta R.T.: 0.000 min
Response: 74238996
Conc: 500.00 ng/ml


Time

Response_ Signal: PD090657.D\ECD2B.ch

#27 Chlordane-5

R.T.: 6.081 min
Delta R.T.: 0.000 min
Response: 730253582
Conc: 500.00 ng/ml


Time



#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.000 min
Response: 167755700
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORICC500



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 1105630741
Conc: 50.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090662.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 14:17
 Operator : AR\AJ
 Sample : PTOXICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 14:36:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 14:35:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.545	2.875	149.0E6	1451.5E6	50.000	50.000
7) SA Decachlor...	9.066	8.062	225.6E6	1489.1E6	50.000	50.000
Target Compounds						
2) Toxaphene-1	6.236	5.466	16282219	131.9E6	500.000	500.000
3) Toxaphene-2	6.436	5.638	24310877	88280081	500.000	500.000
4) Toxaphene-3	7.143	6.745	42855293	329.4E6	500.000	500.000
5) Toxaphene-4	7.558	7.189	47868684	247.6E6	500.000	500.000
6) Toxaphene-5	7.923	7.320	30240231	123.0E6	500.000	500.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090662.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 14:17
Operator : AR\AJ
Sample : PTOXICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
PTOXICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:36:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:35:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm



Response_ Signal: PD090662.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 149025126
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXICC500


Time Response_ Signal: PD090662.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 1451452961
Conc: 50.00 ng/ml


Time Response_ Signal: PD090662.D\ECD1A.ch



#2 Toxaphene-1


R.T.: 6.236 min
Delta R.T.: 0.000 min
Response: 16282219
Conc: 500.00 ng/ml

Time Response_ Signal: PD090662.D\ECD2B.ch



#2 Toxaphene-1


R.T.: 5.466 min
Delta R.T.: 0.000 min
Response: 131850816
Conc: 500.00 ng/ml



#3 Toxaphene-2


R.T.: 6.436 min
Delta R.T.: 0.000 min
Response: 24310877
Conc: 500.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXICC500




#3 Toxaphene-2

R.T.: 5.638 min
Delta R.T.: 0.000 min
Response: 88280081
Conc: 500.00 ng/ml




#4 Toxaphene-3

R.T.: 7.143 min
Delta R.T.: 0.000 min
Response: 42855293
Conc: 500.00 ng/ml



#4 Toxaphene-3


R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 329441315
Conc: 500.00 ng/ml



#5 Toxaphene-4


R.T.: 7.558 min
Delta R.T.: 0.000 min
Response: 47868684
Conc: 500.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXICC500




#5 Toxaphene-4

R.T.: 7.189 min
Delta R.T.: 0.000 min
Response: 247552187
Conc: 500.00 ng/ml



#6 Toxaphene-5

R.T.: 7.923 min
Delta R.T.: 0.000 min
Response: 30240231
Conc: 500.00 ng/ml



#6 Toxaphene-5


R.T.: 7.320 min
Delta R.T.: 0.000 min
Response: 123016477
Conc: 500.00 ng/ml

Response_

Signal: PD090662.D\ECD1A.ch

#7 Decachlorobiphenyl

ICAL Form



R.T.: 9.066 min
Delta R.T.: 0.000 min
Response: 225576698
Conc: 50.00 ng/ml


Instrument :
ECD_D
ClientSampleId :
PTOXICC500

Time

Response_

Signal: PD090662.D\ECD2B.ch

#7 Decachlorobiphenyl



R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 1489074983
Conc: 50.00 ng/ml

Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090665.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 16:46
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD101725

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 17:27:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 14:02:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	153.7E6	1464.9E6	49.347	49.195
28)	SA Decachlor...	9.065	8.062	232.3E6	1556.2E6	49.651	51.373
Target Compounds							
2)	A alpha-BHC	3.993	3.386	368.9E6	2547.7E6	51.395	49.860
3)	MA gamma-BHC...	4.325	3.722	346.1E6	2351.4E6	50.913	49.766
4)	MA Heptachlor	4.923	4.074	269.7E6	2210.7E6	48.243	49.564
5)	MB Aldrin	5.264	4.360	332.4E6	2302.8E6	50.770	49.716
6)	B beta-BHC	4.512	4.018	128.8E6	974.3E6	49.166	49.134
7)	B delta-BHC	4.760	4.255	343.7E6	2372.2E6	50.832	49.714
8)	B Heptachlo...	5.684	4.864	292.1E6	2073.7E6	50.455	48.633
9)	A Endosulfan I	6.068	5.238	276.7E6	1960.5E6	49.982	49.772
10)	B gamma-Chl...	5.940	5.117	297.4E6	2236.0E6	50.146	49.745
11)	B alpha-Chl...	6.020	5.181	299.6E6	2144.4E6	48.473	49.711
12)	B 4,4'-DDE	6.190	5.366	269.1E6	2179.5E6	50.526	49.716
13)	MA Dieldrin	6.341	5.503	295.7E6	2209.3E6	50.705	50.003
14)	MA Endrin	6.568	5.780	251.8E6	2041.8E6	50.824	50.421
15)	B Endosulfa...	6.781	6.072	251.4E6	1876.1E6	50.275	49.956
16)	A 4,4'-DDD	6.701	5.920	206.7E6	1821.0E6	50.990	49.977
17)	MA 4,4'-DDT	7.016	6.174	206.8E6	1798.7E6	50.320	51.285
18)	B Endrin al...	6.910	6.250	184.2E6	1401.3E6	50.283	50.600
19)	B Endosulfa...	7.144	6.474	231.2E6	1809.0E6	49.837	50.467
20)	A Methoxychlor	7.489	6.745	104.2E6	890.1E6	48.999	51.076
21)	B Endrin ke...	7.625	6.983	242.9E6	1911.1E6	49.982	50.668
22)	Mirex	8.108	7.176	158.1E6	1294.5E6	49.394	50.389



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090665.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 16:46
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
ICVPD101725

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 17:27:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090665.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: -0.001 min
Response: 153674114
Conc: 49.35 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725


Time Response_ Signal: PD090665.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 1464875316
Conc: 49.19 ng/ml


Time Response_ Signal: PD090665.D\ECD1A.ch



#2 alpha-BHC


R.T.: 3.993 min
Delta R.T.: 0.000 min
Response: 368866522
Conc: 51.39 ng/ml

Time Response_ Signal: PD090665.D\ECD2B.ch



#2 alpha-BHC


R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2547652740
Conc: 49.86 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 346148511
Conc: 50.91 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 2351401356
Conc: 49.77 ng/ml




#4 Heptachlor

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 269655012
Conc: 48.24 ng/ml



#4 Heptachlor

R.T.: 4.074 min
Delta R.T.: 0.000 min
Response: 2210699675
Conc: 49.56 ng/ml



R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 332404308
Conc: 50.77 ng/ml


Instrument :
ECD_D
ClientSampleId :
ICVPD101725

Time

Response

Signal: PD090665.D\ECD2B.ch

#5 Aldrin




R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 2302805919
Conc: 49.72 ng/ml

Time

Response

Signal: PD090665.D\ECD1A.ch

#6 beta-BHC




R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 128796459
Conc: 49.17 ng/ml

Time

Response

Signal: PD090665.D\ECD2B.ch

#6 beta-BHC



R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 974305644
Conc: 49.13 ng/ml

Time

Response_ Signal: PD090665.D\ECD1A.ch

#7 delta-BHC

R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 343712775
Conc: 50.83 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725

Time

Response_ Signal: PD090665.D\ECD2B.ch

#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 2372178334
Conc: 49.71 ng/ml

Time

Response_ Signal: PD090665.D\ECD1A.ch

#8 Heptachlor epoxide

R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 292086915
Conc: 50.45 ng/ml


Time

Response_ Signal: PD090665.D\ECD2B.ch

#8 Heptachlor epoxide

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 2073663307
Conc: 48.63 ng/ml


Time



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 276731064
Conc: 49.98 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 1960516174
Conc: 49.77 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 297423108
Conc: 50.15 ng/ml



#10 gamma-Chlordane


R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 2235964406
Conc: 49.74 ng/ml



#11 alpha-Chlordane


R.T.: 6.020 min
Delta R.T.: 0.000 min
Response: 299573543
Conc: 48.47 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2144431564
Conc: 49.71 ng/ml



#12 4,4'-DDE


R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 269131461
Conc: 50.53 ng/ml



#12 4,4'-DDE

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 2179474893
Conc: 49.72 ng/ml

Response_ Signal: PD090665.D\ECD1A.ch




#13 Dieldrin

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 295678653
Conc: 50.71 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725


Time Response_ Signal: PD090665.D\ECD2B.ch



#13 Dieldrin

R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 2209322407
Conc: 50.00 ng/ml


Time Response_ Signal: PD090665.D\ECD1A.ch



#14 Endrin


R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 251765100
Conc: 50.82 ng/ml

Time Response_ Signal: PD090665.D\ECD2B.ch



#14 Endrin


R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 2041761986
Conc: 50.42 ng/ml



#15 Endosulfan II


R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 251374054
Conc: 50.28 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1876116999
Conc: 49.96 ng/ml




#16 4,4'-DDD

R.T.: 6.701 min
Delta R.T.: 0.000 min
Response: 206725943
Conc: 50.99 ng/ml



#16 4,4'-DDD


R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 1821049286
Conc: 49.98 ng/ml



#17 4,4'-DDT


R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 206839113
Conc: 50.32 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1798730996
Conc: 51.29 ng/ml



#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 184171641
Conc: 50.28 ng/ml



#18 Endrin aldehyde

R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1401255725
Conc: 50.60 ng/ml

Response_ Signal: PD090665.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.144 min
Delta R.T.: 0.000 min
Response: 231154352
Conc: 49.84 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725

Time

Response_ Signal: PD090665.D\ECD2B.ch

#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1808994693
Conc: 50.47 ng/ml

Time

Response_ Signal: PD090665.D\ECD1A.ch

#20 Methoxychlor

R.T.: 7.489 min
Delta R.T.: 0.000 min
Response: 104150715
Conc: 49.00 ng/ml


Time

Response_ Signal: PD090665.D\ECD2B.ch

#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 890144630
Conc: 51.08 ng/ml


Time



#21 Endrin ketone


R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 242853020
Conc: 49.98 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725




#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1911076638
Conc: 50.67 ng/ml




#22 Mirex

R.T.: 8.108 min
Delta R.T.: 0.000 min
Response: 158102045
Conc: 49.39 ng/ml



#22 Mirex


R.T.: 7.176 min
Delta R.T.: 0.000 min
Response: 1294533618
Conc: 50.39 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.065 min
Delta R.T.: 0.000 min
Response: 232264787
Conc: 49.65 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 1556154030
Conc: 51.37 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090666.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 17:50
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD101725CHLOR

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 18:03:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.547	2.873	151.2E6	1640.5E6	48.564	55.094
28)	SA Decachlor...	9.069	8.063	245.8E6	1635.1E6	52.541	53.979
Target Compounds							
23)	Chlordane-1	4.711	3.897	81712455	547.0E6	432.404	438.923
24)	Chlordane-2	5.238	4.479	83577133	567.3E6	432.751	434.873
25)	Chlordane-3	5.944	5.117	339.4E6	1711.3E6	444.973	439.739
26)	Chlordane-4	6.029	5.182	443.9E6	1471.1E6	458.160	440.289
27)	Chlordane-5	6.868	6.082	67446272	668.4E6	448.868	452.832



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090666.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 17:50
Operator : AR\AJ
Sample : PCHLORICV500
Misc :
ALS Vial : 21 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
ICVPD101725CHLOR

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 18:03:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090666.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 3.547 min
Delta R.T.: 0.002 min
Response: 151237324
Conc: 48.56 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725CHLOR


Time Response_ Signal: PD090666.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: -0.001 min
Response: 1640546394
Conc: 55.09 ng/ml


Time Response_ Signal: PD090666.D\ECD1A.ch



#23 Chlordane-1


R.T.: 4.711 min
Delta R.T.: 0.004 min
Response: 81712455
Conc: 432.40 ng/ml

Time Response_ Signal: PD090666.D\ECD2B.ch



#23 Chlordane-1


R.T.: 3.897 min
Delta R.T.: 0.000 min
Response: 547017680
Conc: 438.92 ng/ml



#24 Chlordane-2


R.T.: 5.238 min
Delta R.T.: 0.003 min
Response: 83577133
Conc: 432.75 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725CHLOR




#24 Chlordane-2

R.T.: 4.479 min
Delta R.T.: 0.000 min
Response: 567301247
Conc: 434.87 ng/ml



#25 Chlordane-3

R.T.: 5.944 min
Delta R.T.: 0.003 min
Response: 339408061
Conc: 444.97 ng/ml



#25 Chlordane-3

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 1711276835
Conc: 439.74 ng/ml

Response_ Signal: PD090666.D\ECD1A.ch

#26 Chlordane-4

R.T.: 6.029 min
Delta R.T.: 0.004 min
Response: 443904717
Conc: 458.16 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725CHLOR

Time

Response_ Signal: PD090666.D\ECD2B.ch

#26 Chlordane-4

R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 1471108584
Conc: 440.29 ng/ml

Time

Response_ Signal: PD090666.D\ECD1A.ch

#27 Chlordane-5

R.T.: 6.868 min
Delta R.T.: 0.003 min
Response: 67446272
Conc: 448.87 ng/ml


Time

Response_ Signal: PD090666.D\ECD2B.ch

#27 Chlordane-5

R.T.: 6.082 min
Delta R.T.: 0.000 min
Response: 668394616
Conc: 452.83 ng/ml


Time



#28 Decachlorobiphenyl

R.T.: 9.069 min
Delta R.T.: 0.005 min
Response: 245784139
Conc: 52.54 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725CHLOR



#28 Decachlorobiphenyl

R.T.: 8.063 min
Delta R.T.: 0.001 min
Response: 1635082214
Conc: 53.98 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090667.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 18:42
 Operator : AR\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD101725TOX

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/21/2025
 Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 20 09:07:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 16:43:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.550	2.874	144.8E6	1414.7E6	47.850	47.865
7) SA Decachlor...	9.073	8.063	222.4E6	1485.6E6	48.104	48.572
Target Compounds						
2) Toxaphene-1	6.243	5.468	16199211	128.4E6	486.326	472.241
3) Toxaphene-2	6.442	5.639	23160349	87664602	469.047	488.530
4) Toxaphene-3	7.149	6.748	41739789	339.6E6	481.161	526.232
5) Toxaphene-4	7.565	7.190	46882437	247.3E6	496.038	505.430
6) Toxaphene-5	7.930	7.320	29219348	133.4E6	485.054	589.259m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090667.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 18:42
Operator : AR\AJ
Sample : PTOXICV500
Misc :
ALS Vial : 22 Sample Multiplier: 1


Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 20 09:07:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 16:43:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0. Signal #2 Info : 30M x 0.32mm x0.25µm



Manual Integrations
APPROVED
Reviewed By: Abdul Mirza 10/21/2025
Supervised By: mohammad ahmed 10/24/2025

Instrument :
ECD_D
ClientSample :
ICVPD101725TOX



#1 Tetrachloro-m-xylene

R.T.: 3.550 min
Delta R.T.: 0.000 min
Response: 144773410
Conc: 47.85 ng/ml

ICAL Form

Instrument :

ECD_D

ClientSampleId :


ICVPD101725TOX

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 10/21/2025

Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 1414733633
Conc: 47.87 ng/ml




#2 Toxaphene-1

R.T.: 6.243 min
Delta R.T.: 0.001 min
Response: 16199211
Conc: 486.33 ng/ml



#2 Toxaphene-1

R.T.: 5.468 min
Delta R.T.: 0.000 min
Response: 128400708
Conc: 472.24 ng/ml




#3 Toxaphene-2

R.T.: 6.442 min
Delta R.T.: 0.000 min
Response: 23160349
Conc: 469.05 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725TOX


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025




#3 Toxaphene-2

R.T.: 5.639 min
Delta R.T.: 0.000 min
Response: 87664602
Conc: 488.53 ng/ml



#4 Toxaphene-3


R.T.: 7.149 min
Delta R.T.: 0.001 min
Response: 41739789
Conc: 481.16 ng/ml



#4 Toxaphene-3

R.T.: 6.748 min
Delta R.T.: 0.005 min
Response: 339600612
Conc: 526.23 ng/ml

Response_ Signal: PD090667.D\ECD1A.ch



#5 Toxaphene-4


R.T.: 7.565 min
Delta R.T.: 0.000 min
Response: 46882437
Conc: 496.04 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725TOX

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025


Time Response_ Signal: PD090667.D\ECD2B.ch



#5 Toxaphene-4

R.T.: 7.190 min
Delta R.T.: 0.000 min
Response: 247294883
Conc: 505.43 ng/ml


Time Response_ Signal: PD090667.D\ECD1A.ch



#6 Toxaphene-5


R.T.: 7.930 min
Delta R.T.: 0.000 min
Response: 29219348
Conc: 485.05 ng/ml

Time Response_ Signal: PD090667.D\ECD2B.ch



#6 Toxaphene-5

R.T.: 7.320 min
Delta R.T.: -0.001 min
Response: 133404883
Conc: 589.26 ng/ml m




#7 Decachlorobiphenyl

R.T.: 9.073 min
Delta R.T.: 0.000 min
Response: 222440569
Conc: 48.10 ng/ml

Instrument :
ECD_D
ClientSampleId :
ICVPD101725TOX

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/21/2025
Supervised By :mohammad ahmed 10/24/2025



#7 Decachlorobiphenyl

R.T.: 8.063 min
Delta R.T.: 0.000 min
Response: 1485641961
Conc: 48.57 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/22/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 14:59

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.06	8.96	9.16	-0.01
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.27	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.69	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.15	7.15	7.05	7.25	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.63	7.63	7.53	7.73	0.00
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/22/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 14:59

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.88	2.78	2.98	0.01
alpha-BHC	3.39	3.39	3.29	3.49	0.00
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.25	4.26	4.16	4.36	0.01
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.08	3.98	4.18	0.01
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.24	5.24	5.14	5.34	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.37	5.37	5.27	5.47	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.18	6.17	6.07	6.27	-0.01
Methoxychlor	6.75	6.75	6.65	6.85	0.01
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.12	5.12	5.02	5.22	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025
Client Sample No.: CCAL01 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090755.D **Time Analyzed:** 14:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.702	6.600	6.800	50.540	50.000	1.1
4,4'-DDE	6.192	6.090	6.290	47.660	50.000	-4.7
4,4'-DDT	7.018	6.915	7.115	50.860	50.000	1.7
Aldrin	5.265	5.164	5.364	48.440	50.000	-3.1
alpha-BHC	3.994	3.894	4.094	48.140	50.000	-3.7
alpha-Chlordane	6.022	5.921	6.121	48.240	50.000	-3.5
beta-BHC	4.513	4.412	4.612	47.070	50.000	-5.9
Decachlorobiphenyl	9.068	8.964	9.164	46.320	50.000	-7.4
delta-BHC	4.761	4.660	4.860	48.380	50.000	-3.2
Dieldrin	6.342	6.240	6.440	52.600	50.000	5.2
Endosulfan I	6.069	5.968	6.168	48.290	50.000	-3.4
Endosulfan II	6.783	6.681	6.881	48.960	50.000	-2.1
Endosulfan sulfate	7.146	7.045	7.245	48.270	50.000	-3.5
Endrin	6.570	6.468	6.668	47.990	50.000	-4.0
Endrin aldehyde	6.912	6.810	7.010	49.240	50.000	-1.5
Endrin ketone	7.627	7.525	7.725	49.560	50.000	-0.9
gamma-BHC (Lindane)	4.326	4.225	4.425	47.990	50.000	-4.0
gamma-Chlordane	5.941	5.840	6.040	48.010	50.000	-4.0
Heptachlor	4.924	4.823	5.023	57.070	50.000	14.1
Heptachlor epoxide	5.685	5.584	5.784	49.030	50.000	-1.9
Methoxychlor	7.490	7.388	7.588	49.890	50.000	-0.2
Tetrachloro-m-xylene	3.545	3.445	3.645	46.510	50.000	-7.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090755.D **Time Analyzed:** 14:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.920	5.821	6.021	50.670	50.000	1.3
4,4'-DDE	5.366	5.266	5.466	49.060	50.000	-1.9
4,4'-DDT	6.175	6.074	6.274	52.370	50.000	4.7
Aldrin	4.360	4.261	4.461	48.160	50.000	-3.7
alpha-BHC	3.386	3.287	3.487	47.870	50.000	-4.3
alpha-Chlordane	5.181	5.081	5.281	48.660	50.000	-2.7
beta-BHC	4.018	3.919	4.119	47.530	50.000	-4.9
Decachlorobiphenyl	8.062	7.962	8.162	54.910	50.000	9.8
delta-BHC	4.254	4.155	4.355	48.500	50.000	-3.0
Dieldrin	5.504	5.404	5.604	48.850	50.000	-2.3
Endosulfan I	5.238	5.138	5.338	48.350	50.000	-3.3
Endosulfan II	6.072	5.972	6.172	50.380	50.000	0.8
Endosulfan sulfate	6.474	6.374	6.574	50.840	50.000	1.7
Endrin	5.781	5.681	5.881	48.490	50.000	-3.0
Endrin aldehyde	6.251	6.150	6.350	51.380	50.000	2.8
Endrin ketone	6.983	6.883	7.083	53.380	50.000	6.8
gamma-BHC (Lindane)	3.722	3.623	3.823	49.500	50.000	-1.0
gamma-Chlordane	5.117	5.017	5.217	48.880	50.000	-2.2
Heptachlor	4.074	3.975	4.175	49.470	50.000	-1.1
Heptachlor epoxide	4.864	4.764	4.964	47.790	50.000	-4.4
Methoxychlor	6.745	6.645	6.845	52.980	50.000	6.0
Tetrachloro-m-xylene	2.874	2.775	2.975	47.350	50.000	-5.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090755.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 14:59
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:21:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
System Monitoring Compounds						
1) SA Tetrachlo...	3.545	2.874	144.8E6	1409.9E6	46.513	47.348
28) SA Decachlor...	9.068	8.062	216.7E6	1663.4E6	46.324	54.912
Target Compounds						
2) A alpha-BHC	3.994	3.386	345.5E6	2446.0E6	48.136	47.871
3) MA gamma-BHC...	4.326	3.722	326.2E6	2338.9E6	47.986	49.501
4) MA Heptachlor	4.924	4.074	319.0E6	2206.6E6	57.071	49.471
5) MB Aldrin	5.265	4.360	317.1E6	2230.7E6	48.437	48.160
6) B beta-BHC	4.513	4.018	123.3E6	942.5E6	47.072	47.528
7) B delta-BHC	4.761	4.254	327.1E6	2314.0E6	48.379	48.495
8) B Heptachlo...	5.685	4.864	283.9E6	2037.7E6	49.034	47.791
9) A Endosulfan I	6.069	5.238	267.3E6	1904.6E6	48.287	48.352
10) B gamma-Chl...	5.941	5.117	284.7E6	2197.2E6	48.006	48.882
11) B alpha-Chl...	6.022	5.181	298.1E6	2099.1E6	48.241	48.659
12) B 4,4'-DDE	6.192	5.366	253.9E6	2150.6E6	47.659	49.058
13) MA Dieldrin	6.342	5.504	306.7E6	2158.4E6	52.602	48.850
14) MA Endrin	6.570	5.781	237.7E6	1963.5E6	47.994	48.488
15) B Endosulfa...	6.783	6.072	244.8E6	1892.2E6	48.956	50.384
16) A 4,4'-DDD	6.702	5.920	204.9E6	1846.3E6	50.537	50.669
17) MA 4,4'-DDT	7.018	6.175	209.1E6	1836.9E6	50.858	52.374
18) B Endrin al...	6.912	6.251	180.4E6	1422.9E6	49.244	51.383
19) B Endosulfa...	7.146	6.474	223.9E6	1822.2E6	48.268	50.837
20) A Methoxychlor	7.490	6.745	106.0E6	923.3E6	49.892	52.976
21) B Endrin ke...	7.627	6.983	240.8E6	2013.4E6	49.559	53.380
22) Mirex	8.110	7.175	156.4E6	1363.3E6	48.872	53.066


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090755.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 14:59
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:21:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 144847643
Conc: 46.51 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 1409902951
Conc: 47.35 ng/ml



#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 345475866
Conc: 48.14 ng/ml



#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2446012186
Conc: 47.87 ng/ml

Response_

Signal: PD090755.D\ECD1A.ch

#3 gamma-BHC (Lindane)

R.T.: 4.326 min

Delta R.T.: 0.000 min

Response: 326249917


Conc: 47.99 ng/ml

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050



Time

Response_

Signal: PD090755.D\ECD2B.ch


#3 gamma-BHC (Lindane)

R.T.: 3.722 min

Delta R.T.: 0.000 min

Response: 2338869799

Conc: 49.50 ng/ml



Time

Response_

Signal: PD090755.D\ECD1A.ch


#4 Heptachlor

R.T.: 4.924 min

Delta R.T.: 0.001 min

Response: 319001387

Conc: 57.07 ng/ml



Time

Response_

Signal: PD090755.D\ECD2B.ch


#4 Heptachlor

R.T.: 4.074 min


Delta R.T.: 0.000 min

Response: 2206550938

Conc: 49.47 ng/ml




Time



#5 Aldrin


R.T.: 5.265 min
Delta R.T.: 0.000 min
Response: 317130181
Conc: 48.44 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 2230723911
Conc: 48.16 ng/ml




#6 beta-BHC

R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 123310645
Conc: 47.07 ng/ml



#6 beta-BHC


R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 942466285
Conc: 47.53 ng/ml



#7 delta-BHC


R.T.: 4.761 min
Delta R.T.: 0.000 min
Response: 327128187
Conc: 48.38 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.254 min
Delta R.T.: 0.000 min
Response: 2314028327
Conc: 48.50 ng/ml




#8 Heptachlor epoxide

R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 283860408
Conc: 49.03 ng/ml



#8 Heptachlor epoxide


R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 2037746487
Conc: 47.79 ng/ml



#9 Endosulfan I


R.T.: 6.069 min
Delta R.T.: 0.002 min
Response: 267346476
Conc: 48.29 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 1904585701
Conc: 48.35 ng/ml




#10 gamma-Chlordane

R.T.: 5.941 min
Delta R.T.: 0.001 min
Response: 284730211
Conc: 48.01 ng/ml



#10 gamma-Chlordane


R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 2197178543
Conc: 48.88 ng/ml



#11 alpha-Chlordane


R.T.: 6.022 min
Delta R.T.: 0.001 min
Response: 298142828
Conc: 48.24 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2099054612
Conc: 48.66 ng/ml




#12 4,4'-DDE

R.T.: 6.192 min
Delta R.T.: 0.002 min
Response: 253861776
Conc: 47.66 ng/ml



#12 4,4'-DDE


R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 2150611512
Conc: 49.06 ng/ml



#13 Dieldrin


R.T.: 6.342 min
Delta R.T.: 0.002 min
Response: 306739347
Conc: 52.60 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 2158387471
Conc: 48.85 ng/ml




#14 Endrin

R.T.: 6.570 min
Delta R.T.: 0.002 min
Response: 237748554
Conc: 47.99 ng/ml



#14 Endrin


R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 1963466551
Conc: 48.49 ng/ml



#15 Endosulfan II


R.T.: 6.783 min
Delta R.T.: 0.002 min
Response: 244779005
Conc: 48.96 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1892195027
Conc: 50.38 ng/ml




#16 4,4'-DDD

R.T.: 6.702 min
Delta R.T.: 0.002 min
Response: 204889504
Conc: 50.54 ng/ml



#16 4,4'-DDD


R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 1846288621
Conc: 50.67 ng/ml



#17 4,4'-DDT


R.T.: 7.018 min
Delta R.T.: 0.002 min
Response: 209051518
Conc: 50.86 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.175 min
Delta R.T.: 0.000 min
Response: 1836911213
Conc: 52.37 ng/ml




#18 Endrin aldehyde

R.T.: 6.912 min
Delta R.T.: 0.002 min
Response: 180365864
Conc: 49.24 ng/ml



#18 Endrin aldehyde


R.T.: 6.251 min
Delta R.T.: 0.000 min
Response: 1422933932
Conc: 51.38 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.146 min
Delta R.T.: 0.002 min
Response: 223878368
Conc: 48.27 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1822248047
Conc: 50.84 ng/ml




#20 Methoxychlor

R.T.: 7.490 min
Delta R.T.: 0.002 min
Response: 106049041
Conc: 49.89 ng/ml



#20 Methoxychlor


R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 923250709
Conc: 52.98 ng/ml



#21 Endrin ketone


R.T.: 7.627 min
Delta R.T.: 0.002 min
Response: 240800953
Conc: 49.56 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 2013363970
Conc: 53.38 ng/ml



#22 Mirex

R.T.: 8.110 min
Delta R.T.: 0.003 min
Response: 156432817
Conc: 48.87 ng/ml




#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1363313824
Conc: 53.07 ng/ml

Signal: PD090755.D\ECD1A.ch

#28 Decachlorobiphenyl




R.T.: 9.068 min
Delta R.T.: 0.004 min
Response: 216701028
Conc: 46.32 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Signal: PD090755.D\ECD2B.ch

#28 Decachlorobiphenyl



R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 1663352556
Conc: 54.91 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/22/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 17:17

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.06	8.96	9.16	-0.01
Tetrachloro-m-xylene	3.54	3.55	3.45	3.65	0.01
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.32	4.33	4.23	4.43	0.01
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.27	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.69	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.15	7.15	7.05	7.25	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.63	7.63	7.53	7.73	0.00
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/22/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 17:17

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.88	2.78	2.98	0.01
alpha-BHC	3.39	3.39	3.29	3.49	0.00
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.25	4.26	4.16	4.36	0.01
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.08	3.98	4.18	0.01
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.24	5.24	5.14	5.34	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.37	5.37	5.27	5.47	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.75	6.75	6.65	6.85	0.01
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.12	5.12	5.02	5.22	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025
Client Sample No.: CCAL02 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090764.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.701	6.600	6.800	48.540	50.000	-2.9
4,4'-DDE	6.191	6.090	6.290	45.890	50.000	-8.2
4,4'-DDT	7.016	6.915	7.115	47.220	50.000	-5.6
Aldrin	5.265	5.164	5.364	47.430	50.000	-5.1
alpha-BHC	3.994	3.894	4.094	47.810	50.000	-4.4
alpha-Chlordane	6.022	5.921	6.121	45.360	50.000	-9.3
beta-BHC	4.513	4.412	4.612	46.500	50.000	-7.0
Decachlorobiphenyl	9.067	8.964	9.164	44.170	50.000	-11.7
delta-BHC	4.761	4.660	4.860	47.540	50.000	-4.9
Dieldrin	6.342	6.240	6.440	50.320	50.000	0.6
Endosulfan I	6.069	5.968	6.168	46.400	50.000	-7.2
Endosulfan II	6.782	6.681	6.881	46.850	50.000	-6.3
Endosulfan sulfate	7.146	7.045	7.245	45.950	50.000	-8.1
Endrin	6.569	6.468	6.668	46.250	50.000	-7.5
Endrin aldehyde	6.911	6.810	7.010	46.610	50.000	-6.8
Endrin ketone	7.627	7.525	7.725	46.940	50.000	-6.1
gamma-BHC (Lindane)	4.324	4.225	4.425	47.530	50.000	-4.9
gamma-Chlordane	5.941	5.840	6.040	46.490	50.000	-7.0
Heptachlor	4.923	4.823	5.023	55.620	50.000	11.2
Heptachlor epoxide	5.685	5.584	5.784	46.870	50.000	-6.3
Methoxychlor	7.489	7.388	7.588	46.180	50.000	-7.6
Tetrachloro-m-xylene	3.544	3.445	3.645	46.320	50.000	-7.4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090764.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.920	5.821	6.021	48.200	50.000	-3.6
4,4'-DDE	5.366	5.266	5.466	47.300	50.000	-5.4
4,4'-DDT	6.174	6.074	6.274	48.220	50.000	-3.6
Aldrin	4.360	4.261	4.461	47.330	50.000	-5.3
alpha-BHC	3.386	3.287	3.487	47.470	50.000	-5.1
alpha-Chlordane	5.181	5.081	5.281	47.290	50.000	-5.4
beta-BHC	4.018	3.919	4.119	47.020	50.000	-6.0
Decachlorobiphenyl	8.061	7.962	8.162	48.710	50.000	-2.6
delta-BHC	4.254	4.155	4.355	47.810	50.000	-4.4
Dieldrin	5.503	5.404	5.604	47.130	50.000	-5.7
Endosulfan I	5.238	5.138	5.338	46.640	50.000	-6.7
Endosulfan II	6.072	5.972	6.172	47.490	50.000	-5.0
Endosulfan sulfate	6.474	6.374	6.574	47.640	50.000	-4.7
Endrin	5.780	5.681	5.881	46.650	50.000	-6.7
Endrin aldehyde	6.250	6.150	6.350	47.580	50.000	-4.8
Endrin ketone	6.983	6.883	7.083	48.300	50.000	-3.4
gamma-BHC (Lindane)	3.722	3.623	3.823	48.880	50.000	-2.2
gamma-Chlordane	5.116	5.017	5.217	47.570	50.000	-4.9
Heptachlor	4.074	3.975	4.175	48.710	50.000	-2.6
Heptachlor epoxide	4.864	4.764	4.964	46.510	50.000	-7.0
Methoxychlor	6.745	6.645	6.845	48.070	50.000	-3.9
Tetrachloro-m-xylene	2.874	2.775	2.975	47.070	50.000	-5.9

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090764.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 17:17
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:22:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.874	144.2E6	1401.5E6	46.319	47.068
28) SA Decachlor...	9.067	8.061	206.6E6	1475.5E6	44.170	48.711
Target Compounds						
2) A alpha-BHC	3.994	3.386	343.1E6	2425.5E6	47.807	47.470
3) MA gamma-BHC...	4.324	3.722	323.2E6	2309.7E6	47.532	48.884
4) MA Heptachlor	4.923	4.074	310.9E6	2172.7E6	55.618	48.712
5) MB Aldrin	5.265	4.360	310.5E6	2192.4E6	47.432	47.333
6) B beta-BHC	4.513	4.018	121.8E6	932.3E6	46.501	47.017
7) B delta-BHC	4.761	4.254	321.4E6	2281.3E6	47.537	47.810
8) B Heptachlo...	5.685	4.864	271.3E6	1983.1E6	46.870	46.509
9) A Endosulfan I	6.069	5.238	256.9E6	1837.3E6	46.404	46.643
10) B gamma-Chl...	5.941	5.116	275.7E6	2138.4E6	46.485	47.574
11) B alpha-Chl...	6.022	5.181	280.3E6	2040.1E6	45.359	47.293
12) B 4,4'-DDE	6.191	5.366	244.4E6	2073.4E6	45.887	47.296
13) MA Dieldrin	6.342	5.503	293.5E6	2082.5E6	50.324	47.133
14) MA Endrin	6.569	5.780	229.1E6	1889.2E6	46.251	46.653
15) B Endosulfa...	6.782	6.072	234.2E6	1783.7E6	46.846	47.495
16) A 4,4'-DDD	6.701	5.920	196.8E6	1756.4E6	48.536	48.203
17) MA 4,4'-DDT	7.016	6.174	194.1E6	1691.1E6	47.222	48.217
18) B Endrin al...	6.911	6.250	170.7E6	1317.5E6	46.611	47.577
19) B Endosulfa...	7.146	6.474	213.1E6	1707.5E6	45.955	47.636
20) A Methoxychlor	7.489	6.745	98153519	837.8E6	46.177	48.074
21) B Endrin ke...	7.627	6.983	228.1E6	1821.7E6	46.940	48.298
22) Mirex	8.109	7.175	147.7E6	1224.1E6	46.154	47.647



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090764.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 17:17
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:22:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090764.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 144244640
Conc: 46.32 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Time

Response_ Signal: PD090764.D\ECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 1401542493
Conc: 47.07 ng/ml

Time

Response_ Signal: PD090764.D\ECD1A.ch

#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 343119547
Conc: 47.81 ng/ml

Time

Response_ Signal: PD090764.D\ECD2B.ch

#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2425516091
Conc: 47.47 ng/ml

Time


Response_

Signal: PD090764.D\ECD1A.ch

#3 gamma-BHC (Lindane)

R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 323159059
Conc: 47.53 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




Time

Response_

Signal: PD090764.D\ECD2B.ch

#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: -0.001 min
Response: 2309716498
Conc: 48.88 ng/ml




Time

Response_

Signal: PD090764.D\ECD1A.ch

#4 Heptachlor

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 310877514
Conc: 55.62 ng/ml




Time

Response_


Signal: PD090764.D\ECD2B.ch

#4 Heptachlor

R.T.: 4.074 min
Delta R.T.: 0.000 min
Response: 2172692086
Conc: 48.71 ng/ml




Time



#5 Aldrin


R.T.: 5.265 min
Delta R.T.: 0.000 min
Response: 310547737
Conc: 47.43 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.360 min
Delta R.T.: -0.001 min
Response: 2192416099
Conc: 47.33 ng/ml



#6 beta-BHC

R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 121816014
Conc: 46.50 ng/ml



#6 beta-BHC

R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 932327111
Conc: 47.02 ng/ml

Response_ Signal: PD090764.D\ECD1A.ch

#7 delta-BHC

R.T.: 4.761 min
Delta R.T.: 0.000 min
Response: 321429638
Conc: 47.54 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Time

Response_ Signal: PD090764.D\ECD2B.ch

#7 delta-BHC

R.T.: 4.254 min
Delta R.T.: 0.000 min
Response: 2281340622
Conc: 47.81 ng/ml

Time

Response_ Signal: PD090764.D\ECD1A.ch

#8 Heptachlor epoxide

R.T.: 5.685 min
Delta R.T.: 0.001 min
Response: 271335940
Conc: 46.87 ng/ml

Time

Response_ Signal: PD090764.D\ECD2B.ch

#8 Heptachlor epoxide

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 1983083749
Conc: 46.51 ng/ml

Time

Response_

Signal: PD090764.D\ECD1A.ch

#9 Endosulfan I

R.T.: 6.069 min

Delta R.T.: 0.000 min

Response: 256925211

Conc: 46.40 ng/ml

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

2e+07
1e+07
0

6.067

+

5.90 5.95 6.00 6.05 6.10 6.15 6.20 6.25

Time

Response_

Signal: PD090764.D\ECD2B.ch

#9 Endosulfan I

R.T.: 5.238 min

Delta R.T.: 0.000 min

Response: 1837287369

Conc: 46.64 ng/ml

2.5e+08
2e+08
1.5e+08
1e+08
5e+07
0

5.237

+

5.10 5.20 5.30 5.40

Time

Response_

Signal: PD090764.D\ECD1A.ch

#10 gamma-Chlordane

R.T.: 5.941 min

Delta R.T.: 0.000 min

Response: 275713646

Conc: 46.49 ng/ml

2e+07
1e+07
0

5.939

+

5.75 5.80 5.85 5.90 5.95 6.00 6.05 6.10

Time

Response_

Signal: PD090764.D\ECD2B.ch

#10 gamma-Chlordane

R.T.: 5.116 min

Delta R.T.: 0.000 min

Response: 2138379541

Conc: 47.57 ng/ml

2.5e+08
2e+08
1.5e+08
1e+08
5e+07
0

5.115

+

4.90 5.00 5.10 5.20 5.30

Time

Response_

Signal: PD090764.D\ECD1A.ch

#11 alpha-Chlordane

R.T.: 6.022 min

Delta R.T.: 0.001 min

Response: 280330427


Conc: 45.36 ng/ml

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050



Time

Response_

Signal: PD090764.D\ECD2B.ch


#11 alpha-Chlordane

R.T.: 5.181 min

Delta R.T.: 0.000 min

Response: 2040129063

Conc: 47.29 ng/ml



Time

Response_

Signal: PD090764.D\ECD1A.ch


#12 4,4'-DDE

R.T.: 6.191 min

Delta R.T.: 0.001 min

Response: 244423556

Conc: 45.89 ng/ml



Time

Response_

Signal: PD090764.D\ECD2B.ch


#12 4,4'-DDE

R.T.: 5.366 min

Delta R.T.: 0.000 min

Response: 2073368168


Conc: 47.30 ng/ml



Time

Response_ Signal: PD090764.D\ECD1A.ch

#13 Dieldrin




R.T.: 6.342 min
Delta R.T.: 0.001 min
Response: 293456933
Conc: 50.32 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Time Response_ Signal: PD090764.D\ECD2B.ch


#13 Dieldrin



R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 2082529046
Conc: 47.13 ng/ml

Time Response_ Signal: PD090764.D\ECD1A.ch


#14 Endrin




R.T.: 6.569 min
Delta R.T.: 0.000 min
Response: 229113454
Conc: 46.25 ng/ml

Time Response_ Signal: PD090764.D\ECD2B.ch

#14 Endrin




R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 1889175654
Conc: 46.65 ng/ml



#15 Endosulfan II


R.T.: 6.782 min
Delta R.T.: 0.001 min
Response: 234228888
Conc: 46.85 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1783678423
Conc: 47.49 ng/ml




#16 4,4'-DDD

R.T.: 6.701 min
Delta R.T.: 0.001 min
Response: 196776180
Conc: 48.54 ng/ml



#16 4,4'-DDD


R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 1756410429
Conc: 48.20 ng/ml



#17 4,4'-DDT


R.T.: 7.016 min
Delta R.T.: 0.001 min
Response: 194104242
Conc: 47.22 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1691126008
Conc: 48.22 ng/ml



#18 Endrin aldehyde

R.T.: 6.911 min
Delta R.T.: 0.000 min
Response: 170722731
Conc: 46.61 ng/ml



#18 Endrin aldehyde


R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1317548084
Conc: 47.58 ng/ml

Response_ Signal: PD090764.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.146 min
Delta R.T.: 0.000 min
Response: 213147227
Conc: 45.95 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDCCC050



Time Response_ Signal: PD090764.D\ECD2B.ch

#19 Endosulfan Sulfate


R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1707529758
Conc: 47.64 ng/ml



Time Response_ Signal: PD090764.D\ECD1A.ch

#20 Methoxychlor


R.T.: 7.489 min
Delta R.T.: 0.001 min
Response: 98153519
Conc: 46.18 ng/ml



Time Response_ Signal: PD090764.D\ECD2B.ch

#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 837824203
Conc: 48.07 ng/ml




Response_ Signal: PD090764.D\ECD1A.ch

#21 Endrin ketone

R.T.: 7.627 min
Delta R.T.: 0.002 min
Response: 228073439
Conc: 46.94 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDCCC050



Time Response_ Signal: PD090764.D\ECD2B.ch

#21 Endrin ketone


R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1821709703
Conc: 48.30 ng/ml



Time Response_ Signal: PD090764.D\ECD1A.ch

#22 Mirex


R.T.: 8.109 min
Delta R.T.: 0.001 min
Response: 147733151
Conc: 46.15 ng/ml




Time Response_ Signal: PD090764.D\ECD2B.ch

#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1224093737
Conc: 47.65 ng/ml






#28 Decachlorobiphenyl

R.T.: 9.067 min
Delta R.T.: 0.003 min
Response: 206625069
Conc: 44.17 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 1475527555
Conc: 48.71 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/22/2025

Initial Calibration Date(s): 10/17/2025 10/17/2025

Continuing Calib Time: 19:19

Initial Calibration Time(s): 11:34 12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.06	8.96	9.16	-0.01
Tetrachloro-m-xylene	3.54	3.55	3.45	3.65	0.01
alpha-BHC	3.99	3.99	3.89	4.09	0.00
beta-BHC	4.51	4.51	4.41	4.61	0.00
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.32	4.33	4.23	4.43	0.01
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.69	5.68	5.58	5.78	0.00
Endosulfan I	6.07	6.07	5.97	6.17	0.00
Dieldrin	6.34	6.34	6.24	6.44	0.00
4,4'-DDE	6.19	6.19	6.09	6.29	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Endosulfan II	6.78	6.78	6.68	6.88	0.00
4,4'-DDD	6.70	6.70	6.60	6.80	0.00
Endosulfan sulfate	7.15	7.15	7.05	7.25	0.01
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.63	7.63	7.53	7.73	0.00
Endrin aldehyde	6.91	6.91	6.81	7.01	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.94	5.84	6.04	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/22/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 19:19

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.88	2.78	2.98	0.01
alpha-BHC	3.39	3.39	3.29	3.49	0.01
beta-BHC	4.02	4.02	3.92	4.12	0.00
delta-BHC	4.25	4.26	4.16	4.36	0.01
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.08	3.98	4.18	0.01
Aldrin	4.36	4.36	4.26	4.46	0.00
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endosulfan I	5.24	5.24	5.14	5.34	0.00
Dieldrin	5.50	5.50	5.40	5.60	0.00
4,4'-DDE	5.37	5.37	5.27	5.47	0.01
Endrin	5.78	5.78	5.68	5.88	0.00
Endosulfan II	6.07	6.07	5.97	6.17	0.00
4,4'-DDD	5.92	5.92	5.82	6.02	0.00
Endosulfan sulfate	6.47	6.47	6.37	6.57	0.00
4,4'-DDT	6.17	6.17	6.07	6.27	0.00
Methoxychlor	6.74	6.75	6.65	6.85	0.01
Endrin ketone	6.98	6.98	6.88	7.08	0.00
Endrin aldehyde	6.25	6.25	6.15	6.35	0.00
alpha-Chlordane	5.18	5.18	5.08	5.28	0.00
gamma-Chlordane	5.12	5.12	5.02	5.22	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090773.D **Time Analyzed:** 19:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.700	6.600	6.800	51.820	50.000	3.6
4,4'-DDE	6.190	6.090	6.290	48.730	50.000	-2.5
4,4'-DDT	7.016	6.915	7.115	50.780	50.000	1.6
Aldrin	5.264	5.164	5.364	49.570	50.000	-0.9
alpha-BHC	3.994	3.894	4.094	49.430	50.000	-1.1
alpha-Chlordane	6.021	5.921	6.121	48.200	50.000	-3.6
beta-BHC	4.512	4.412	4.612	48.100	50.000	-3.8
Decachlorobiphenyl	9.066	8.964	9.164	48.090	50.000	-3.8
delta-BHC	4.760	4.660	4.860	49.750	50.000	-0.5
Dieldrin	6.341	6.240	6.440	53.220	50.000	6.4
Endosulfan I	6.068	5.968	6.168	49.150	50.000	-1.7
Endosulfan II	6.781	6.681	6.881	49.660	50.000	-0.7
Endosulfan sulfate	7.145	7.045	7.245	49.130	50.000	-1.7
Endrin	6.568	6.468	6.668	49.010	50.000	-2.0
Endrin aldehyde	6.910	6.810	7.010	49.600	50.000	-0.8
Endrin ketone	7.625	7.525	7.725	50.440	50.000	0.9
gamma-BHC (Lindane)	4.324	4.225	4.425	49.170	50.000	-1.7
gamma-Chlordane	5.939	5.840	6.040	49.270	50.000	-1.5
Heptachlor	4.923	4.823	5.023	57.890	50.000	15.8
Heptachlor epoxide	5.685	5.584	5.784	49.640	50.000	-0.7
Methoxychlor	7.489	7.388	7.588	49.960	50.000	-0.1
Tetrachloro-m-xylene	3.544	3.445	3.645	47.760	50.000	-4.5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090773.D **Time Analyzed:** 19:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.920	5.821	6.021	51.850	50.000	3.7
4,4'-DDE	5.365	5.266	5.466	49.630	50.000	-0.7
4,4'-DDT	6.174	6.074	6.274	51.950	50.000	3.9
Aldrin	4.360	4.261	4.461	48.750	50.000	-2.5
alpha-BHC	3.385	3.287	3.487	48.890	50.000	-2.2
alpha-Chlordane	5.181	5.081	5.281	49.370	50.000	-1.3
beta-BHC	4.018	3.919	4.119	48.510	50.000	-3.0
Decachlorobiphenyl	8.061	7.962	8.162	55.760	50.000	11.5
delta-BHC	4.254	4.155	4.355	49.470	50.000	-1.1
Dieldrin	5.503	5.404	5.604	50.610	50.000	1.2
Endosulfan I	5.238	5.138	5.338	48.830	50.000	-2.3
Endosulfan II	6.072	5.972	6.172	51.180	50.000	2.4
Endosulfan sulfate	6.473	6.374	6.574	51.930	50.000	3.9
Endrin	5.780	5.681	5.881	49.660	50.000	-0.7
Endrin aldehyde	6.250	6.150	6.350	51.550	50.000	3.1
Endrin ketone	6.982	6.883	7.083	53.780	50.000	7.6
gamma-BHC (Lindane)	3.722	3.623	3.823	50.230	50.000	0.5
gamma-Chlordane	5.116	5.017	5.217	49.480	50.000	-1.0
Heptachlor	4.073	3.975	4.175	49.520	50.000	-1.0
Heptachlor epoxide	4.863	4.764	4.964	48.340	50.000	-3.3
Methoxychlor	6.744	6.645	6.845	53.040	50.000	6.1
Tetrachloro-m-xylene	2.874	2.775	2.975	48.360	50.000	-3.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090773.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 19:19
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:24:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	148.7E6	1439.9E6	47.761	48.356
28)	SA Decachlor...	9.066	8.061	225.0E6	1689.1E6	48.088	55.761
Target Compounds							
2)	A alpha-BHC	3.994	3.385	354.8E6	2498.0E6	49.429	48.889
3)	MA gamma-BHC...	4.324	3.722	334.3E6	2373.4E6	49.172	50.233
4)	MA Heptachlor	4.923	4.073	323.6E6	2208.8E6	57.892	49.521
5)	MB Aldrin	5.264	4.360	324.5E6	2258.0E6	49.567	48.749
6)	B beta-BHC	4.512	4.018	126.0E6	961.8E6	48.100	48.505
7)	B delta-BHC	4.760	4.254	336.4E6	2360.4E6	49.752	49.467
8)	B Heptachlo...	5.685	4.863	287.4E6	2061.3E6	49.643	48.343
9)	A Endosulfan I	6.068	5.238	272.1E6	1923.6E6	49.153	48.834
10)	B gamma-Chl...	5.939	5.116	292.3E6	2224.2E6	49.275	49.483
11)	B alpha-Chl...	6.021	5.181	297.9E6	2129.6E6	48.203	49.367
12)	B 4,4'-DDE	6.190	5.365	259.6E6	2175.5E6	48.734	49.625
13)	MA Dieldrin	6.341	5.503	310.4E6	2236.2E6	53.222	50.612
14)	MA Endrin	6.568	5.780	242.8E6	2010.9E6	49.010	49.660
15)	B Endosulfa...	6.781	6.072	248.3E6	1922.0E6	49.662	51.179
16)	A 4,4'-DDD	6.700	5.920	210.1E6	1889.4E6	51.818	51.853
17)	MA 4,4'-DDT	7.016	6.174	208.7E6	1822.0E6	50.781	51.949
18)	B Endrin al...	6.910	6.250	181.7E6	1427.6E6	49.604	51.550
19)	B Endosulfa...	7.145	6.473	227.9E6	1861.6E6	49.133	51.934
20)	A Methoxychlor	7.489	6.744	106.2E6	924.3E6	49.956	53.035
21)	B Endrin ke...	7.625	6.982	245.1E6	2028.5E6	50.441	53.781
22)	Mirex	8.108	7.175	159.6E6	1367.1E6	49.856	53.215



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090773.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 19:19
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:24:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090773.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: -0.001 min
Response: 148734863
Conc: 47.76 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Time

Response_ Signal: PD090773.D\ECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 1439910437
Conc: 48.36 ng/ml

Time

Response_ Signal: PD090773.D\ECD1A.ch

#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 354760484
Conc: 49.43 ng/ml

Time

Response_ Signal: PD090773.D\ECD2B.ch

#2 alpha-BHC

R.T.: 3.385 min
Delta R.T.: -0.001 min
Response: 2498020591
Conc: 48.89 ng/ml

Time

Response_ Signal: PD090773.D\ECD1A.ch

#3 gamma-BHC (Lindane)

R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 334308516
Conc: 49.17 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Time

Response_ Signal: PD090773.D\ECD2B.ch

#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: -0.001 min
Response: 2373435161
Conc: 50.23 ng/ml

Time

Response_ Signal: PD090773.D\ECD1A.ch

#4 Heptachlor

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 323591660
Conc: 57.89 ng/ml


Time

Response_ Signal: PD090773.D\ECD2B.ch

#4 Heptachlor

R.T.: 4.073 min
Delta R.T.: -0.002 min
Response: 2208783581
Conc: 49.52 ng/ml


Time



#5 Aldrin


R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 324528594
Conc: 49.57 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 2257981502
Conc: 48.75 ng/ml



#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 126004403
Conc: 48.10 ng/ml




#6 beta-BHC

R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 961832177
Conc: 48.51 ng/ml

Response_ Signal: PD090773.D\ECD1A.ch

#7 delta-BHC




R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 336412047
Conc: 49.75 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Time Response_ Signal: PD090773.D\ECD2B.ch


#7 delta-BHC



R.T.: 4.254 min
Delta R.T.: -0.001 min
Response: 2360380737
Conc: 49.47 ng/ml

Time Response_ Signal: PD090773.D\ECD1A.ch


#8 Heptachlor epoxide




R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 287385858
Conc: 49.64 ng/ml

Time Response_ Signal: PD090773.D\ECD2B.ch

#8 Heptachlor epoxide




R.T.: 4.863 min
Delta R.T.: 0.000 min
Response: 2061286618
Conc: 48.34 ng/ml



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 272140478
Conc: 49.15 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 1923606058
Conc: 48.83 ng/ml




#10 gamma-Chlordane

R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 292257254
Conc: 49.27 ng/ml



#10 gamma-Chlordane


R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 2224190087
Conc: 49.48 ng/ml



#11 alpha-Chlordane


R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 297906594
Conc: 48.20 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2129579894
Conc: 49.37 ng/ml




#12 4,4'-DDE

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 259587501
Conc: 48.73 ng/ml



#12 4,4'-DDE


R.T.: 5.365 min
Delta R.T.: 0.000 min
Response: 2175479533
Conc: 49.63 ng/ml



#13 Dieldrin


R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 310353890
Conc: 53.22 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 2236248876
Conc: 50.61 ng/ml




#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 242779954
Conc: 49.01 ng/ml



#14 Endrin


R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 2010934915
Conc: 49.66 ng/ml



#15 Endosulfan II


R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 248306801
Conc: 49.66 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1922041967
Conc: 51.18 ng/ml




#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 210080583
Conc: 51.82 ng/ml



#16 4,4'-DDD


R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 1889412060
Conc: 51.85 ng/ml



#17 4,4'-DDT


R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 208735791
Conc: 50.78 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1822016426
Conc: 51.95 ng/ml



#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 181684301
Conc: 49.60 ng/ml



#18 Endrin aldehyde


R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1427554208
Conc: 51.55 ng/ml

Response_ Signal: PD090773.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 227886039
Conc: 49.13 ng/ml


Instrument :
ECD_D
ClientSampleId :
PSTDCCC050



Time Response_ Signal: PD090773.D\ECD2B.ch

#19 Endosulfan Sulfate


R.T.: 6.473 min
Delta R.T.: 0.000 min
Response: 1861576873
Conc: 51.93 ng/ml



Time Response_ Signal: PD090773.D\ECD1A.ch

#20 Methoxychlor


R.T.: 7.489 min
Delta R.T.: 0.000 min
Response: 106185108
Conc: 49.96 ng/ml




Time Response_ Signal: PD090773.D\ECD2B.ch

#20 Methoxychlor

R.T.: 6.744 min
Delta R.T.: 0.000 min
Response: 924292649
Conc: 53.04 ng/ml






#21 Endrin ketone


R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 245084800
Conc: 50.44 ng/ml

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.982 min
Delta R.T.: 0.000 min
Response: 2028515873
Conc: 53.78 ng/ml




#22 Mirex

R.T.: 8.108 min
Delta R.T.: 0.000 min
Response: 159581481
Conc: 49.86 ng/ml



#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1367146785
Conc: 53.22 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.066 min

Delta R.T.: 0.002 min

Response: 224953584


Conc: 48.09 ng/ml

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 8.061 min

Delta R.T.: 0.000 min

Response: 1689087164

Conc: 55.76 ng/ml

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090648.D Date Analyzed: 10/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:07

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.066	8.970	9.170	22.860	20.000	14.3
Tetrachloro-m-xylene	3.544	3.490	3.590	21.710	20.000	8.6
alpha-BHC	3.994	3.940	4.040	9.050	10.000	-9.5
beta-BHC	4.512	4.460	4.560	10.070	10.000	0.7
gamma-BHC (Lindane)	4.324	4.270	4.370	9.330	10.000	-6.7
Endrin	6.568	6.500	6.640	49.200	50.000	-1.6
4,4'-DDT	7.016	6.950	7.090	100.270	100.000	0.3
Methoxychlor	7.489	7.420	7.560	236.810	250.000	-5.3

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090648.D Date Analyzed: 10/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:07

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.062	7.960	8.160	21.400	20.000	7.0
Tetrachloro-m-xylene	2.875	2.820	2.930	21.220	20.000	6.1
alpha-BHC	3.386	3.340	3.440	10.300	10.000	3.0
beta-BHC	4.019	3.970	4.070	10.460	10.000	4.6
gamma-BHC (Lindane)	3.723	3.670	3.770	10.340	10.000	3.4
Endrin	5.781	5.710	5.850	47.940	50.000	-4.1
4,4'-DDT	6.174	6.100	6.240	95.670	100.000	-4.3
Methoxychlor	6.745	6.670	6.820	201.830	250.000	-19.3

PEM
Data File: PD090648.D Date Acquired 10/17/2025 11:07
Operator: AR\AJ

ENDRIN BREAK DOWN

Column #1

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin	6.57	243728295.7	249054181.9	5325886.17	2.14
Endrin aldehyde	6.91	1372426.584			
Endrin ketone	7.63	3953459.588			

Column #2

Name	RT	Response	Response [E+EA+EK]	Response [EA+EK]	% Break Down
Endrin #2	5.78	1941282940	1991060515	49777575.7	2.50
Endrin aldehyde #2	6.25	19940698.04			
Endrin ketone #2	6.98	29836877.65			

DDT BREAK DOWN

Column #1

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT	7.02	412144601.3	415215115.8	3070514.5	0.74
4,4'-DDE	0.00	0			
4,4'-DDD	6.70	3070514.503			

Column #2

Name	RT	Response	Response [DDT+DDE+DDD]	Response [DDE+DDD]	% Break Down
4,4'-DDT #2	6.17	3355351424	3394188860	38837436.1	1.14
4,4'-DDE #2	0.00	0			
4,4'-DDD #2	5.92	38837436.09			

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090648.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 11:07
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 10/19/2025
 Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 14:04:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 14:02:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.875	67593250	632.0E6	21.705	21.224
28)	SA Decachlor...	9.066	8.062	106.9E6	648.1E6	22.859	21.395
Target Compounds							
2)	A alpha-BHC	3.994	3.386	64971308	526.5E6	9.053	10.304
3)	MA gamma-BHC...	4.324	3.723	63399907	488.4E6	9.325	10.337
6)	B beta-BHC	4.512	4.019	26386693	207.4E6	10.073	10.459
14)	MA Endrin	6.568	5.781	243.7E6	1941.3E6	49.201	47.940
16)	A 4,4'-DDD	6.701	5.923	3070515	38837436	0.757	1.066 #
17)	MA 4,4'-DDT	7.016	6.174	412.1E6	3355.4E6	100.267	95.667
18)	B Endrin al...	6.913	6.248	1372427	19940698	0.375m	0.720 #
20)	A Methoxychlor	7.489	6.745	503.3E6	3517.5E6	236.805	201.834
21)	B Endrin ke...	7.625	6.981	3953460	29836878	0.814	0.791m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090648.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:07
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:04:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 67593250
Conc: 21.71 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 631992202
Conc: 21.22 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 64971308
Conc: 9.05 ng/ml



#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 526499291
Conc: 10.30 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 4.324 min
Delta R.T.: 0.000 min
Response: 63399907
Conc: 9.33 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#3 gamma-BHC (Lindane)

R.T.: 3.723 min
Delta R.T.: 0.000 min
Response: 488402734
Conc: 10.34 ng/ml




#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 26386693
Conc: 10.07 ng/ml



#6 beta-BHC

R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 207399327
Conc: 10.46 ng/ml




#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 243728296
Conc: 49.20 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#14 Endrin

R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 1941282940
Conc: 47.94 ng/ml




#16 4,4'-DDD

R.T.: 6.701 min
Delta R.T.: 0.000 min
Response: 3070515
Conc: 0.76 ng/ml



#16 4,4'-DDD

R.T.: 5.923 min
Delta R.T.: 0.003 min
Response: 38837436
Conc: 1.07 ng/ml




#17 4,4'-DDT

R.T.: 7.016 min
Delta R.T.: 0.000 min
Response: 412144601
Conc: 100.27 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 3355351424
Conc: 95.67 ng/ml




#18 Endrin aldehyde

R.T.: 6.913 min
Delta R.T.: 0.003 min
Response: 1372427
Conc: 0.37 ng/ml m



#18 Endrin aldehyde

R.T.: 6.248 min
Delta R.T.: -0.002 min
Response: 19940698
Conc: 0.72 ng/ml



#20 Methoxychlor

R.T.: 7.489 min
Delta R.T.: 0.000 min
Response: 503349948
Conc: 236.81 ng/ml

Instrument :

ECD_D

ClientSampleId :


PEM

Manual Integrations

APPROVED


Reviewed By :Abdul Mirza 10/19/2025

Supervised By :mohammad ahmed 10/24/2025




#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 3517519034
Conc: 201.83 ng/ml




#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 3953460
Conc: 0.81 ng/ml



#21 Endrin ketone

R.T.: 6.981 min
Delta R.T.: -0.002 min
Response: 29836878
Conc: 0.79 ng/ml m




#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.002 min
Response: 106933782
Conc: 22.86 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 648085268
Conc: 21.40 ng/ml

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090754.D Date Analyzed: 10/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 14:45

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.070	8.970	9.170	22.040	20.000	10.2
Tetrachloro-m-xylene	3.544	3.490	3.590	20.920	20.000	4.6
alpha-BHC	3.994	3.940	4.040	8.790	10.000	-12.1
beta-BHC	4.513	4.460	4.560	10.210	10.000	2.1
gamma-BHC (Lindane)	4.325	4.270	4.380	9.200	10.000	-8.0
Endrin	6.570	6.500	6.640	48.180	50.000	-3.6
4,4'-DDT	7.018	6.950	7.090	100.910	100.000	0.9
Methoxychlor	7.491	7.420	7.560	235.430	250.000	-5.8

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090754.D Date Analyzed: 10/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 14:45

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.062	7.960	8.160	25.380	20.000	26.9
Tetrachloro-m-xylene	2.873	2.820	2.920	21.110	20.000	5.6
alpha-BHC	3.385	3.330	3.440	10.320	10.000	3.2
beta-BHC	4.017	3.970	4.070	10.570	10.000	5.7
gamma-BHC (Lindane)	3.721	3.670	3.770	10.640	10.000	6.4
Endrin	5.779	5.710	5.850	49.190	50.000	-1.6
4,4'-DDT	6.173	6.100	6.240	103.470	100.000	3.5
Methoxychlor	6.745	6.670	6.820	217.530	250.000	-13.0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
 Data File : PD090754.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 14:45
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 14:21:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.873	65151970	628.7E6	20.921	21.112
28)	SA Decachlor...	9.070	8.062	103.1E6	768.9E6	22.036	25.383
Target Compounds							
2)	A alpha-BHC	3.994	3.385	63097425	527.4E6	8.791	10.322
3)	MA gamma-BHC...	4.325	3.721	62558349	502.6E6	9.201	10.636
6)	B beta-BHC	4.513	4.017	26755969	209.6E6	10.214	10.572
14)	MA Endrin	6.570	5.779	238.6E6	1992.0E6	48.175	49.193
16)	A 4,4'-DDD	6.702	5.920	11104415	112.6E6	2.739	3.090
17)	MA 4,4'-DDT	7.018	6.173	414.8E6	3628.9E6	100.907	103.468
18)	B Endrin al...	6.913	6.248	3608496	30772791	0.985	1.111
20)	A Methoxychlor	7.491	6.745	500.4E6	3791.1E6	235.432	217.534
21)	B Endrin ke...	7.627	6.983	8409255	82813597	1.731	2.196 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090754.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 14:45
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:21:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 65151970
Conc: 20.92 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM




#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: -0.002 min
Response: 628665474
Conc: 21.11 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 63097425
Conc: 8.79 ng/ml



#2 alpha-BHC


R.T.: 3.385 min
Delta R.T.: -0.002 min
Response: 527416985
Conc: 10.32 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 62558349
Conc: 9.20 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM




#3 gamma-BHC (Lindane)

R.T.: 3.721 min
Delta R.T.: -0.001 min
Response: 502562004
Conc: 10.64 ng/ml




#6 beta-BHC

R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 26755969
Conc: 10.21 ng/ml



#6 beta-BHC


R.T.: 4.017 min
Delta R.T.: -0.001 min
Response: 209646424
Conc: 10.57 ng/ml



#14 Endrin


R.T.: 6.570 min
Delta R.T.: 0.001 min
Response: 238646896
Conc: 48.18 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM




#14 Endrin

R.T.: 5.779 min
Delta R.T.: -0.001 min
Response: 1991998893
Conc: 49.19 ng/ml




#16 4,4'-DDD

R.T.: 6.702 min
Delta R.T.: 0.002 min
Response: 11104415
Conc: 2.74 ng/ml



#16 4,4'-DDD


R.T.: 5.920 min
Delta R.T.: 0.000 min
Response: 112577983
Conc: 3.09 ng/ml



#17 4,4'-DDT


R.T.: 7.018 min
Delta R.T.: 0.002 min
Response: 414774704
Conc: 100.91 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM




#17 4,4'-DDT

R.T.: 6.173 min
Delta R.T.: 0.000 min
Response: 3628931057
Conc: 103.47 ng/ml




#18 Endrin aldehyde

R.T.: 6.913 min
Delta R.T.: 0.003 min
Response: 3608496
Conc: 0.99 ng/ml



#18 Endrin aldehyde


R.T.: 6.248 min
Delta R.T.: -0.002 min
Response: 30772791
Conc: 1.11 ng/ml



#20 Methoxychlor


R.T.: 7.491 min
Delta R.T.: 0.003 min
Response: 500431728
Conc: 235.43 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM




#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 3791144088
Conc: 217.53 ng/ml




#21 Endrin ketone

R.T.: 7.627 min
Delta R.T.: 0.003 min
Response: 8409255
Conc: 1.73 ng/ml



#21 Endrin ketone


R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 82813597
Conc: 2.20 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.070 min
Delta R.T.: 0.005 min
Response: 103083892
Conc: 22.04 ng/ml

Instrument :
ECD_D
ClientSampleId :
PEM



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 768875869
Conc: 25.38 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090649.D
Acq On : 17 Oct 2025 11:20
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e

Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Title : GC Extractables
Last Update : Fri Oct 17 17:43:28 2025
Integrator: ChemStation

RT#1	RT#2	Resolution
3.544	5.940	100.00%
5.940	6.068	100.00%
6.068	6.190	100.00%
6.190	6.341	100.00%
6.341	7.144	100.00%
7.144	7.488	100.00%
7.488	7.625	100.00%
7.625	9.066	100.00%

Signal #2

2.874	5.116	100.00%
5.116	5.238	100.00%
5.238	5.366	100.00%
5.366	5.503	100.00%
5.503	6.474	100.00%
6.474	6.745	100.00%
6.745	6.983	100.00%
6.983	8.062	100.00%

PD101725.M Sat Oct 25 06:10:53 2025

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
 Data File : PD090649.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2025 11:20
 Operator : AR\AJ
 Sample : RESCHK
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 RESCHK

Manual Integrations APPROVED

Reviewed By :Abdul Mirza 10/19/2025
 Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 14:04:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 14:02:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.874	55125186	545.1E6	17.701m	18.307
28)	SA Decachlor...	9.066	8.062	88858238	547.9E6	18.995	18.088
Target Compounds							
9)	A Endosulfan I	6.068	5.238	43926319	330.7E6	7.934	8.397
10)	B gamma-Chl...	5.940	5.116	47696248	387.3E6	8.042	8.616
12)	B 4,4'-DDE	6.190	5.366	86477351	754.1E6	16.235	17.202
13)	MA Dieldrin	6.341	5.503	92552800	748.0E6	15.872	16.930
19)	B Endosulfa...	7.144	6.474	76754913	612.1E6	16.548	17.078
20)	A Methoxychlor	7.488	6.745	169.7E6	1331.6E6	79.844	76.409
21)	B Endrin ke...	7.625	6.983	80182457	638.1E6	16.502	16.917

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090649.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 11:20
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

RESCHK

Manual Integrations



APPROVED


Reviewed By :Abdul Mirza 10/19/2025

Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:04:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.543 min
Delta R.T.: -0.002 min
Response: 55125186
Conc: 17.70 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 545126372
Conc: 18.31 ng/ml




#9 Endosulfan I

R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 43926319
Conc: 7.93 ng/ml



#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 330745194
Conc: 8.40 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 47696248
Conc: 8.04 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#10 gamma-Chlordane

R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 387274886
Conc: 8.62 ng/ml




#12 4,4'-DDE

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 86477351
Conc: 16.23 ng/ml



#12 4,4'-DDE

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 754114700
Conc: 17.20 ng/ml




#13 Dieldrin

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 92552800
Conc: 15.87 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#13 Dieldrin

R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 748035470
Conc: 16.93 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.144 min
Delta R.T.: 0.000 min
Response: 76754913
Conc: 16.55 ng/ml



#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 612147653
Conc: 17.08 ng/ml



#20 Methoxychlor

R.T.: 7.488 min
Delta R.T.: 0.000 min
Response: 169715831
Conc: 79.84 ng/ml


Instrument :

ECD_D

ClientSampleId :
RESCHK


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025




#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 1331648566
Conc: 76.41 ng/ml




#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 80182457
Conc: 16.50 ng/ml



#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 638061752
Conc: 16.92 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.001 min
Response: 88858238
Conc: 19.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
RESCHK

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/19/2025
Supervised By :mohammad ahmed 10/24/2025



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 547910430
Conc: 18.09 ng/ml

Analytical Sequence

Client: ICE Service Group, Inc.	SDG No.: Q3399
Project: NWIRP Northrup Grumman Site – Bethpage	Instrument ID: ECD_D
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 10/17/2025 10/17/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	10/17/2025	10:53	PD090647.D	9.07	3.55
PEM	PEM	10/17/2025	11:07	PD090648.D	9.07	3.54
RESCHK	RESCHK	10/17/2025	11:20	PD090649.D	9.07	3.54
PSTDICCC100	PSTDICCC100	10/17/2025	11:34	PD090650.D	9.07	3.54
PSTDICCC075	PSTDICCC075	10/17/2025	11:48	PD090651.D	9.07	3.55
PSTDICCC050	PSTDICCC050	10/17/2025	12:01	PD090652.D	9.06	3.55
PSTDICCC025	PSTDICCC025	10/17/2025	12:15	PD090653.D	9.07	3.54
PSTDICCC005	PSTDICCC005	10/17/2025	12:28	PD090654.D	9.07	3.54
PCHLORICC500	PCHLORICC500	10/17/2025	13:09	PD090657.D	9.07	3.55
PTOXICC500	PTOXICC500	10/17/2025	14:17	PD090662.D	9.07	3.55
IBLK	IBLK	10/22/2025	14:32	PD090753.D	9.08	3.55
PEM	PEM	10/22/2025	14:45	PD090754.D	9.07	3.54
PSTDCCC050	PSTDCCC050	10/22/2025	14:59	PD090755.D	9.07	3.55
PB170205BL	PB170205BL	10/22/2025	15:13	PD090756.D	9.07	3.54
PB170205BS	PB170205BS	10/22/2025	15:26	PD090757.D	9.07	3.54
COMP-ROLLOFF	Q3399-01	10/22/2025	15:40	PD090758.D	9.07	3.54
IBLK	IBLK	10/22/2025	17:03	PD090763.D	9.07	3.54
PSTDCCC050	PSTDCCC050	10/22/2025	17:17	PD090764.D	9.07	3.54
COMP-ROLLOFFMS	Q3399-01MS	10/22/2025	18:11	PD090768.D	9.07	3.54
COMP-ROLLOFFMSD	Q3399-01MSD	10/22/2025	18:25	PD090769.D	9.07	3.54
IBLK	IBLK	10/22/2025	19:06	PD090772.D	9.07	3.54
PSTDCCC050	PSTDCCC050	10/22/2025	19:19	PD090773.D	9.07	3.54

Analytical Sequence

Client: ICE Service Group, Inc.	SDG No.: Q3399
Project: NWIRP Northrup Grumman Site – Bethpage	Instrument ID: ECD_D
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 10/17/2025 10/17/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	10/17/2025	10:53	PD090647.D	8.06	2.88
PEM	PEM	10/17/2025	11:07	PD090648.D	8.06	2.88
RESCHK	RESCHK	10/17/2025	11:20	PD090649.D	8.06	2.87
PSTDICC100	PSTDICC100	10/17/2025	11:34	PD090650.D	8.06	2.87
PSTDICC075	PSTDICC075	10/17/2025	11:48	PD090651.D	8.06	2.88
PSTDICC050	PSTDICC050	10/17/2025	12:01	PD090652.D	8.06	2.88
PSTDICC025	PSTDICC025	10/17/2025	12:15	PD090653.D	8.06	2.87
PSTDICC005	PSTDICC005	10/17/2025	12:28	PD090654.D	8.06	2.87
PCHLORICC500	PCHLORICC500	10/17/2025	13:09	PD090657.D	8.06	2.87
PTOXICC500	PTOXICC500	10/17/2025	14:17	PD090662.D	8.06	2.88
IBLK	IBLK	10/22/2025	14:32	PD090753.D	8.06	2.87
PEM	PEM	10/22/2025	14:45	PD090754.D	8.06	2.87
PSTDCCC050	PSTDCCC050	10/22/2025	14:59	PD090755.D	8.06	2.87
PB170205BL	PB170205BL	10/22/2025	15:13	PD090756.D	8.06	2.87
PB170205BS	PB170205BS	10/22/2025	15:26	PD090757.D	8.06	2.87
COMP-ROLLOFF	Q3399-01	10/22/2025	15:40	PD090758.D	8.06	2.87
IBLK	IBLK	10/22/2025	17:03	PD090763.D	8.06	2.87
PSTDCCC050	PSTDCCC050	10/22/2025	17:17	PD090764.D	8.06	2.87
COMP-ROLLOFFMS	Q3399-01MS	10/22/2025	18:11	PD090768.D	8.06	2.87
COMP-ROLLOFFMSD	Q3399-01MSD	10/22/2025	18:25	PD090769.D	8.06	2.87
IBLK	IBLK	10/22/2025	19:06	PD090772.D	8.06	2.87
PSTDCCC050	PSTDCCC050	10/22/2025	19:19	PD090773.D	8.06	2.87

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

COMP-ROLLOFF

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: Q3399-01

Date(s) Analyzed: 10/22/2025 10/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin	1	6.57	6.52	6.62	0.19	68.4
	2	5.78	5.73	5.83	0.38	
4,4'-DDT	1	7.02	6.97	7.07	0.22	109.1
	2	6.17	6.12	6.22	0.75	
alpha-Chlordane	1	6.02	5.97	6.07	0.37	32.6
	2	5.18	5.13	5.23	0.27	
4,4'-DDE	1	6.19	6.14	6.24	0.21	13.3
	2	5.37	5.32	5.42	0.18	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

COMP-ROLLOFFMS

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: Q3399-01MS

Date(s) Analyzed: 10/22/2025 10/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.70	6.65	6.75	18.4	5.6
	2	5.92	5.87	5.97	17.4	
4,4'-DDT	1	7.02	6.97	7.07	18.5	3.7
	2	6.17	6.12	6.22	19.2	
alpha-BHC	1	3.99	3.94	4.04	18.2	3.4
	2	3.39	3.34	3.44	17.6	
Aldrin	1	5.26	5.21	5.31	17.8	0
	2	4.36	4.31	4.41	17.8	
beta-BHC	1	4.51	4.46	4.56	17.5	0
	2	4.02	3.97	4.07	17.5	
delta-BHC	1	4.76	4.71	4.81	18.1	2.2
	2	4.26	4.21	4.31	17.7	
Endosulfan I	1	6.07	6.02	6.12	17.7	0
	2	5.24	5.19	5.29	17.7	
alpha-Chlordane	1	6.02	5.97	6.07	17.7	1.1
	2	5.18	5.13	5.23	17.9	
4,4'-DDE	1	6.19	6.14	6.24	17.6	2.2
	2	5.37	5.32	5.42	18.0	
Dieldrin	1	6.34	6.29	6.39	18.0	0.6
	2	5.50	5.45	5.55	17.9	
Endosulfan II	1	6.78	6.73	6.83	16.8	8
	2	6.07	6.02	6.12	18.2	
Endrin aldehyde	1	6.91	6.86	6.96	17.6	3.9
	2	6.25	6.20	6.30	18.3	
Endosulfan sulfate	1	7.15	7.10	7.20	17.4	6.1
	2	6.47	6.42	6.52	18.5	
Methoxychlor	1	7.49	7.44	7.54	18.0	4.9
	2	6.75	6.70	6.80	18.9	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

COMP-ROLLOFFMS

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: Q3399-01MS

Date(s) Analyzed: 10/22/2025 10/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin ketone	1	7.63	7.58	7.68	18.0	4.9
	2	6.98	6.93	7.03	18.9	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	17.7	0.6
	2	3.72	3.67	3.77	17.6	
Heptachlor	1	4.92	4.87	4.97	20.9	14.4
	2	4.08	4.03	4.13	18.1	
Heptachlor epoxide	1	5.69	5.64	5.74	18.1	2.8
	2	4.86	4.81	4.91	17.6	
gamma-Chlordane	1	5.94	5.89	5.99	17.6	1.7
	2	5.12	5.07	5.17	17.9	
Endrin	1	6.57	6.52	6.62	17.9	0
	2	5.78	5.73	5.83	17.9	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

COMP-ROLLOFFMSD

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: Q3399-01MSD

Date(s) Analyzed: 10/22/2025 10/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan II	1	6.78	6.73	6.83	17.1	8.4
	2	6.07	6.02	6.12	18.6	
4,4'-DDD	1	6.70	6.65	6.75	18.7	5.5
	2	5.92	5.87	5.97	17.7	
4,4'-DDT	1	7.02	6.97	7.07	19.0	3.6
	2	6.17	6.12	6.22	19.7	
Endrin aldehyde	1	6.91	6.86	6.96	18.0	3.3
	2	6.25	6.20	6.30	18.6	
Endosulfan sulfate	1	7.15	7.10	7.20	17.7	6
	2	6.47	6.42	6.52	18.8	
Methoxychlor	1	7.49	7.44	7.54	18.3	5.3
	2	6.75	6.70	6.80	19.3	
Endrin ketone	1	7.63	7.58	7.68	18.4	4.3
	2	6.98	6.93	7.03	19.2	
alpha-BHC	1	3.99	3.94	4.04	18.5	3.3
	2	3.39	3.34	3.44	17.9	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	18.0	0.6
	2	3.72	3.67	3.77	17.9	
Heptachlor	1	4.92	4.87	4.97	21.3	15.2
	2	4.08	4.03	4.13	18.3	
Aldrin	1	5.26	5.21	5.31	18.1	0.6
	2	4.36	4.31	4.41	18.0	
beta-BHC	1	4.51	4.46	4.56	17.7	0
	2	4.02	3.97	4.07	17.7	
delta-BHC	1	4.76	4.71	4.81	18.5	2.7
	2	4.26	4.21	4.31	18.0	
Heptachlor epoxide	1	5.69	5.64	5.74	18.3	2.8
	2	4.86	4.81	4.91	17.8	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

COMP-ROLLOFFMSD

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: Q3399-01MSD

Date(s) Analyzed: 10/22/2025 10/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endosulfan I	1	6.07	6.02	6.12	18.0	0.6
	2	5.24	5.19	5.29	17.9	
gamma-Chlordane	1	5.94	5.89	5.99	17.9	1.7
	2	5.12	5.07	5.17	18.2	
alpha-Chlordane	1	6.02	5.97	6.07	17.7	2.8
	2	5.18	5.13	5.23	18.2	
4,4'-DDE	1	6.19	6.14	6.24	18.1	1.1
	2	5.37	5.32	5.42	18.3	
Dieldrin	1	6.34	6.29	6.39	18.3	0.5
	2	5.50	5.45	5.55	18.2	
Endrin	1	6.57	6.52	6.62	18.3	1.7
	2	5.78	5.73	5.83	18.0	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB170205BS

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: PB170205BS

Date(s) Analyzed: 10/22/2025 10/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
4,4'-DDD	1	6.70	6.65	6.75	16.1	1.3
	2	5.92	5.87	5.97	15.9	
4,4'-DDE	1	6.19	6.14	6.24	15.0	3.3
	2	5.37	5.32	5.42	15.5	
4,4'-DDT	1	7.02	6.97	7.07	16.2	1.8
	2	6.17	6.12	6.22	16.5	
alpha-BHC	1	3.99	3.94	4.04	15.0	1.3
	2	3.39	3.34	3.44	14.8	
Aldrin	1	5.26	5.21	5.31	15.2	0.7
	2	4.36	4.31	4.41	15.1	
alpha-Chlordane	1	6.02	5.97	6.07	14.8	3.3
	2	5.18	5.13	5.23	15.3	
Endosulfan II	1	6.78	6.73	6.83	15.6	1.3
	2	6.07	6.02	6.12	15.8	
Endosulfan sulfate	1	7.15	7.10	7.20	15.4	3.8
	2	6.47	6.42	6.52	16.0	
beta-BHC	1	4.51	4.46	4.56	14.7	1.4
	2	4.02	3.97	4.07	14.9	
delta-BHC	1	4.76	4.71	4.81	15.2	0.7
	2	4.26	4.21	4.31	15.1	
Endosulfan I	1	6.07	6.02	6.12	15.2	2
	2	5.24	5.19	5.29	14.9	
Dieldrin	1	6.34	6.29	6.39	15.5	0.6
	2	5.50	5.45	5.55	15.6	
Endrin aldehyde	1	6.91	6.86	6.96	15.8	2.5
	2	6.25	6.20	6.30	16.2	
Methoxychlor	1	7.49	7.44	7.54	16.0	3.1
	2	6.75	6.70	6.80	16.5	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB170205BS

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: PB170205BS

Date(s) Analyzed: 10/22/2025 10/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Endrin ketone	1	7.63	7.58	7.68	15.8	6.1
	2	6.98	6.93	7.03	16.8	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	14.9	0
	2	3.72	3.67	3.77	14.9	
Heptachlor	1	4.92	4.87	4.97	17.9	15
	2	4.08	4.03	4.13	15.4	
Heptachlor epoxide	1	5.69	5.64	5.74	15.4	2
	2	4.86	4.81	4.91	15.1	
gamma-Chlordane	1	5.94	5.89	5.99	15.1	1.3
	2	5.12	5.07	5.17	15.3	
Endrin	1	6.57	6.52	6.62	15.4	1.3
	2	5.78	5.73	5.83	15.6	



QC SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19
- 20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: PB170205BL
Lab Sample ID: PB170205BL
Analytical Method: 8081B
Sample Wt/Vol: 30.01 g Final Vol: 10000 uL
Prep Method: SW3541 Prep Date 10/22/25

Date Collected:
Date Received:
SDG No.: Q3399
Matrix: SOIL
% Solid: 100
Test: Pesticide-TCL

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
319-84-6	alpha-BHC	0.33	U	1	0.13	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
319-85-7	beta-BHC	0.83	U	1	0.18	0.83	1.70	ug/kg	10/22/25 15:13	PB170205
319-86-8	delta-BHC	0.83	U	1	0.39	0.83	1.70	ug/kg	10/22/25 15:13	PB170205
58-89-9	gamma-BHC (Lindane)	0.33	U	1	0.14	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
76-44-8	Heptachlor	0.33	U	1	0.12	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
309-00-2	Aldrin	0.33	U	1	0.12	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
1024-57-3	Heptachlor epoxide	0.83	U	1	0.19	0.83	1.70	ug/kg	10/22/25 15:13	PB170205
959-98-8	Endosulfan I	0.33	U	1	0.14	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
60-57-1	Dieldrin	0.33	U	1	0.14	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
72-55-9	4,4-DDE	0.33	U	1	0.14	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
72-20-8	Endrin	0.33	U	1	0.14	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
33213-65-9	Endosulfan II	0.83	U	1	0.29	0.83	1.70	ug/kg	10/22/25 15:13	PB170205
72-54-8	4,4-DDD	0.33	U	1	0.15	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
1031-07-8	Endosulfan Sulfate	0.33	U	1	0.13	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
50-29-3	4,4-DDT	0.33	U	1	0.14	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
72-43-5	Methoxychlor	0.83	U	1	0.37	0.83	1.70	ug/kg	10/22/25 15:13	PB170205
53494-70-5	Endrin ketone	0.83	U	1	0.19	0.83	1.70	ug/kg	10/22/25 15:13	PB170205
7421-93-4	Endrin aldehyde	0.83	U	1	0.37	0.83	1.70	ug/kg	10/22/25 15:13	PB170205
5103-71-9	alpha-Chlordane	0.33	U	1	0.12	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
5103-74-2	gamma-Chlordane	0.33	U	1	0.15	0.33	1.70	ug/kg	10/22/25 15:13	PB170205
8001-35-2	Toxaphene	17.0	U	1	5.40	17.0	33.0	ug/kg	10/22/25 15:13	PB170205
SURROGATES										
2051-24-3	Decachlorobiphenyl	24.3			55 - 130		121%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.4			42 - 129		102%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090756.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 15:13
Operator : AR\AJ
Sample : PB170205BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB170205BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:22:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.874	61840766	608.5E6	19.858	20.434
2) SA Decachlor...	9.067	8.062	100.5E6	736.0E6	21.487	24.299

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090756.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 15:13
Operator : AR\AJ
Sample : PB170205BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB170205BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:22:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 61840766
Conc: 19.86 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170205BL




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: -0.001 min
Response: 608479551
Conc: 20.43 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.067 min
Delta R.T.: 0.003 min
Response: 100514162
Conc: 21.49 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 736049267
Conc: 24.30 ng/ml

Report of Analysis

Client: ICE Service Group, Inc.	Date Collected: 10/17/25	
Project: NWIRP Northrup Grumman Site – Bethpage, NY	Date Received: 10/17/25	
Client Sample ID: PIBLK-PD090647.D	SDG No.: Q3399	
Lab Sample ID: I.BLK-PD090647.D	Matrix: WATER	
Analytical Method: 8081B	% Solid: 0	
Sample Wt/Vol: 1000 mL	Final Vol: 10000 uL	Test: Pesticide-TCL
Prep Method:	Prep Date	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
319-84-6	alpha-BHC	0.010	U	1	0.0039	0.010	0.050	ug/L	10/17/25	pd101725
319-85-7	beta-BHC	0.010	U	1	0.0049	0.010	0.050	ug/L	10/17/25	pd101725
319-86-8	delta-BHC	0.025	U	1	0.011	0.025	0.050	ug/L	10/17/25	pd101725
58-89-9	gamma-BHC (Lindane)	0.010	U	1	0.0037	0.010	0.050	ug/L	10/17/25	pd101725
76-44-8	Heptachlor	0.010	U	1	0.0027	0.010	0.050	ug/L	10/17/25	pd101725
309-00-2	Aldrin	0.010	U	1	0.0036	0.010	0.050	ug/L	10/17/25	pd101725
1024-57-3	Heptachlor epoxide	0.025	U	1	0.0096	0.025	0.050	ug/L	10/17/25	pd101725
959-98-8	Endosulfan I	0.010	U	1	0.0031	0.010	0.050	ug/L	10/17/25	pd101725
60-57-1	Dieldrin	0.010	U	1	0.0036	0.010	0.050	ug/L	10/17/25	pd101725
72-55-9	4,4-DDE	0.010	U	1	0.0037	0.010	0.050	ug/L	10/17/25	pd101725
72-20-8	Endrin	0.010	U	1	0.0032	0.010	0.050	ug/L	10/17/25	pd101725
33213-65-9	Endosulfan II	0.025	U	1	0.0079	0.025	0.050	ug/L	10/17/25	pd101725
72-54-8	4,4-DDD	0.025	U	1	0.0071	0.025	0.050	ug/L	10/17/25	pd101725
1031-07-8	Endosulfan Sulfate	0.010	U	1	0.0037	0.010	0.050	ug/L	10/17/25	pd101725
50-29-3	4,4-DDT	0.010	U	1	0.0035	0.010	0.050	ug/L	10/17/25	pd101725
72-43-5	Methoxychlor	0.025	U	1	0.011	0.025	0.050	ug/L	10/17/25	pd101725
53494-70-5	Endrin ketone	0.025	U	1	0.0093	0.025	0.050	ug/L	10/17/25	pd101725
7421-93-4	Endrin aldehyde	0.025	U	1	0.011	0.025	0.050	ug/L	10/17/25	pd101725
5103-71-9	alpha-Chlordane	0.010	U	1	0.0035	0.010	0.050	ug/L	10/17/25	pd101725
5103-74-2	gamma-Chlordane	0.010	U	1	0.0039	0.010	0.050	ug/L	10/17/25	pd101725
8001-35-2	Toxaphene	0.50	U	1	0.17	0.50	1.00	ug/L	10/17/25	pd101725
SURROGATES										
2051-24-3	Decachlorobiphenyl	21.3			30 - 135		106%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.0			44 - 124		100%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3399-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090647.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 10:53
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:03:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.875	60526543	594.5E6	19.436	19.964
28)	SA Decachlor...	9.066	8.062	99514776	616.2E6	21.273	20.342

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101725\
Data File : PD090647.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2025 10:53
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 14:03:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 14:02:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 60526543
Conc: 19.44 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.875 min
Delta R.T.: 0.000 min
Response: 594475365
Conc: 19.96 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.066 min
Delta R.T.: 0.002 min
Response: 99514776
Conc: 21.27 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 616188932
Conc: 20.34 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/22/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/22/25
Client Sample ID:	PIBLK-PD090753.D		SDG No.:	Q3399
Lab Sample ID:	I.BLK-PD090753.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL	Test:
Prep Method:		Prep Date:		Pesticide-TCL

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
319-84-6	alpha-BHC	0.010	U	1	0.0039	0.010	0.050	ug/L	10/22/25	PD102225
319-85-7	beta-BHC	0.010	U	1	0.0049	0.010	0.050	ug/L	10/22/25	PD102225
319-86-8	delta-BHC	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25	PD102225
58-89-9	gamma-BHC (Lindane)	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25	PD102225
76-44-8	Heptachlor	0.010	U	1	0.0027	0.010	0.050	ug/L	10/22/25	PD102225
309-00-2	Aldrin	0.010	U	1	0.0036	0.010	0.050	ug/L	10/22/25	PD102225
1024-57-3	Heptachlor epoxide	0.025	U	1	0.0096	0.025	0.050	ug/L	10/22/25	PD102225
959-98-8	Endosulfan I	0.010	U	1	0.0031	0.010	0.050	ug/L	10/22/25	PD102225
60-57-1	Dieldrin	0.010	U	1	0.0036	0.010	0.050	ug/L	10/22/25	PD102225
72-55-9	4,4-DDE	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25	PD102225
72-20-8	Endrin	0.010	U	1	0.0032	0.010	0.050	ug/L	10/22/25	PD102225
33213-65-9	Endosulfan II	0.025	U	1	0.0079	0.025	0.050	ug/L	10/22/25	PD102225
72-54-8	4,4-DDD	0.025	U	1	0.0071	0.025	0.050	ug/L	10/22/25	PD102225
1031-07-8	Endosulfan Sulfate	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25	PD102225
50-29-3	4,4-DDT	0.010	U	1	0.0035	0.010	0.050	ug/L	10/22/25	PD102225
72-43-5	Methoxychlor	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25	PD102225
53494-70-5	Endrin ketone	0.025	U	1	0.0093	0.025	0.050	ug/L	10/22/25	PD102225
7421-93-4	Endrin aldehyde	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25	PD102225
5103-71-9	alpha-Chlordane	0.010	U	1	0.0035	0.010	0.050	ug/L	10/22/25	PD102225
5103-74-2	gamma-Chlordane	0.010	U	1	0.0039	0.010	0.050	ug/L	10/22/25	PD102225
8001-35-2	Toxaphene	0.50	U	1	0.17	0.50	1.00	ug/L	10/22/25	PD102225
SURROGATES										
2051-24-3	Decachlorobiphenyl	25.1			30 - 135		126%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.7			44 - 124		104%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3399-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090753.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 14:32
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:21:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.552	2.873	61839307	616.6E6	19.857	20.707
28)	SA Decachlor...	9.080	8.064	103.5E6	760.3E6	22.133	25.098

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090753.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 14:32
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:21:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.552 min
Delta R.T.: 0.007 min
Response: 61839307
Conc: 19.86 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: -0.002 min
Response: 616610152
Conc: 20.71 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.080 min
Delta R.T.: 0.016 min
Response: 103539116
Conc: 22.13 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.064 min
Delta R.T.: 0.002 min
Response: 760256842
Conc: 25.10 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/22/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/22/25
Client Sample ID:	PIBLK-PD090763.D		SDG No.:	Q3399
Lab Sample ID:	I.BLK-PD090763.D		Matrix:	WATER
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL	Test:
Prep Method:		Prep Date:		Pesticide-TCL

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
319-84-6	alpha-BHC	0.010	U	1	0.0039	0.010	0.050	ug/L	10/22/25	PD102225
319-85-7	beta-BHC	0.010	U	1	0.0049	0.010	0.050	ug/L	10/22/25	PD102225
319-86-8	delta-BHC	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25	PD102225
58-89-9	gamma-BHC (Lindane)	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25	PD102225
76-44-8	Heptachlor	0.010	U	1	0.0027	0.010	0.050	ug/L	10/22/25	PD102225
309-00-2	Aldrin	0.010	U	1	0.0036	0.010	0.050	ug/L	10/22/25	PD102225
1024-57-3	Heptachlor epoxide	0.025	U	1	0.0096	0.025	0.050	ug/L	10/22/25	PD102225
959-98-8	Endosulfan I	0.010	U	1	0.0031	0.010	0.050	ug/L	10/22/25	PD102225
60-57-1	Dieldrin	0.010	U	1	0.0036	0.010	0.050	ug/L	10/22/25	PD102225
72-55-9	4,4-DDE	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25	PD102225
72-20-8	Endrin	0.010	U	1	0.0032	0.010	0.050	ug/L	10/22/25	PD102225
33213-65-9	Endosulfan II	0.025	U	1	0.0079	0.025	0.050	ug/L	10/22/25	PD102225
72-54-8	4,4-DDD	0.025	U	1	0.0071	0.025	0.050	ug/L	10/22/25	PD102225
1031-07-8	Endosulfan Sulfate	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25	PD102225
50-29-3	4,4-DDT	0.010	U	1	0.0035	0.010	0.050	ug/L	10/22/25	PD102225
72-43-5	Methoxychlor	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25	PD102225
53494-70-5	Endrin ketone	0.025	U	1	0.0093	0.025	0.050	ug/L	10/22/25	PD102225
7421-93-4	Endrin aldehyde	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25	PD102225
5103-71-9	alpha-Chlordane	0.010	U	1	0.0035	0.010	0.050	ug/L	10/22/25	PD102225
5103-74-2	gamma-Chlordane	0.010	U	1	0.0039	0.010	0.050	ug/L	10/22/25	PD102225
8001-35-2	Toxaphene	0.50	U	1	0.17	0.50	1.00	ug/L	10/22/25	PD102225
SURROGATES										
2051-24-3	Decachlorobiphenyl	21.0			30 - 135		105%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.6			44 - 124		103%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3399-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090763.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 17:03
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:22:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	62142204	613.9E6	19.955	20.617
28)	SA Decachlor...	9.068	8.062	98339448	636.0E6	21.022	20.996

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090763.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 17:03
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:22:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090763.D\ECD1A.ch

#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 62142204
Conc: 19.95 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK

Time 3.20 3.30 3.40 3.50 3.60 3.70 3.80 3.90

Response_ Signal: PD090763.D\ECD2B.ch

#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 613909551
Conc: 20.62 ng/ml

Time 2.70 2.80 2.90 3.00 3.10

Response_ Signal: PD090763.D\ECD1A.ch

#28 Decachlorobiphenyl

R.T.: 9.068 min
Delta R.T.: 0.004 min
Response: 98339448
Conc: 21.02 ng/ml

Time 8.70 8.80 8.90 9.00 9.10 9.20 9.30 9.40

Response_ Signal: PD090763.D\ECD2B.ch

#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 635993503
Conc: 21.00 ng/ml

Time 7.80 7.90 8.00 8.10 8.20 8.30

Report of Analysis

Client: ICE Service Group, Inc.	Date Collected: 10/22/25	
Project: NWIRP Northrup Grumman Site – Bethpage, NY	Date Received: 10/22/25	
Client Sample ID: PIBLK-PD090772.D	SDG No.: Q3399	
Lab Sample ID: I.BLK-PD090772.D	Matrix: WATER	
Analytical Method: 8081B	% Solid: 0	
Sample Wt/Vol: 1000 mL	Final Vol: 10000 uL	Test: Pesticide-TCL
Prep Method:	Prep Date	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
319-84-6	alpha-BHC	0.010	U	1	0.0039	0.010	0.050	ug/L	10/22/25	PD102225
319-85-7	beta-BHC	0.010	U	1	0.0049	0.010	0.050	ug/L	10/22/25	PD102225
319-86-8	delta-BHC	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25	PD102225
58-89-9	gamma-BHC (Lindane)	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25	PD102225
76-44-8	Heptachlor	0.010	U	1	0.0027	0.010	0.050	ug/L	10/22/25	PD102225
309-00-2	Aldrin	0.010	U	1	0.0036	0.010	0.050	ug/L	10/22/25	PD102225
1024-57-3	Heptachlor epoxide	0.025	U	1	0.0096	0.025	0.050	ug/L	10/22/25	PD102225
959-98-8	Endosulfan I	0.010	U	1	0.0031	0.010	0.050	ug/L	10/22/25	PD102225
60-57-1	Dieldrin	0.010	U	1	0.0036	0.010	0.050	ug/L	10/22/25	PD102225
72-55-9	4,4-DDE	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25	PD102225
72-20-8	Endrin	0.010	U	1	0.0032	0.010	0.050	ug/L	10/22/25	PD102225
33213-65-9	Endosulfan II	0.025	U	1	0.0079	0.025	0.050	ug/L	10/22/25	PD102225
72-54-8	4,4-DDD	0.025	U	1	0.0071	0.025	0.050	ug/L	10/22/25	PD102225
1031-07-8	Endosulfan Sulfate	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25	PD102225
50-29-3	4,4-DDT	0.010	U	1	0.0035	0.010	0.050	ug/L	10/22/25	PD102225
72-43-5	Methoxychlor	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25	PD102225
53494-70-5	Endrin ketone	0.025	U	1	0.0093	0.025	0.050	ug/L	10/22/25	PD102225
7421-93-4	Endrin aldehyde	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25	PD102225
5103-71-9	alpha-Chlordane	0.010	U	1	0.0035	0.010	0.050	ug/L	10/22/25	PD102225
5103-74-2	gamma-Chlordane	0.010	U	1	0.0039	0.010	0.050	ug/L	10/22/25	PD102225
8001-35-2	Toxaphene	0.50	U	1	0.17	0.50	1.00	ug/L	10/22/25	PD102225
SURROGATES										
2051-24-3	Decachlorobiphenyl	24.0			30 - 135		120%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.5			44 - 124		103%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3399-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090772.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 19:06
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:24:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.874	62069474	610.6E6	19.931	20.504
2) SA Decachlor...	9.067	8.062	100.3E6	727.8E6	21.432	24.027

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090772.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 19:06
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:24:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090772.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 62069474
Conc: 19.93 ng/ml

Instrument :
ECD_D
ClientSampleId :
I.BLK


Response_ Signal: PD090772.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 610560919
Conc: 20.50 ng/ml


Response_ Signal: PD090772.D\ECD1A.ch



#28 Decachlorobiphenyl

R.T.: 9.067 min
Delta R.T.: 0.002 min
Response: 100258011
Conc: 21.43 ng/ml

Response_ Signal: PD090772.D\ECD2B.ch



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 727822735
Conc: 24.03 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: PB170205BS
Lab Sample ID: PB170205BS
Analytical Method: 8081B
Sample Wt/Vol: 30.03 g Final Vol: 10000 uL
Prep Method: SW3541 Prep Date 10/22/25

Date Collected:
Date Received:
SDG No.: Q3399
Matrix: SOIL
% Solid: 100
Test: Pesticide-TCL

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
319-84-6	alpha-BHC	15.0		1	0.13	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
319-85-7	beta-BHC	14.9		1	0.18	0.83	1.70	ug/kg	10/22/25 15:26	PB170205
319-86-8	delta-BHC	15.2		1	0.39	0.83	1.70	ug/kg	10/22/25 15:26	PB170205
58-89-9	gamma-BHC (Lindane)	14.9		1	0.14	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
76-44-8	Heptachlor	17.9		1	0.12	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
309-00-2	Aldrin	15.2		1	0.12	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
1024-57-3	Heptachlor epoxide	15.4		1	0.19	0.83	1.70	ug/kg	10/22/25 15:26	PB170205
959-98-8	Endosulfan I	15.2		1	0.14	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
60-57-1	Dieldrin	15.6		1	0.14	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
72-55-9	4,4-DDE	15.5		1	0.14	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
72-20-8	Endrin	15.6		1	0.14	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
33213-65-9	Endosulfan II	15.8		1	0.29	0.83	1.70	ug/kg	10/22/25 15:26	PB170205
72-54-8	4,4-DDD	16.1		1	0.15	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
1031-07-8	Endosulfan Sulfate	16.0		1	0.13	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
50-29-3	4,4-DDT	16.5		1	0.14	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
72-43-5	Methoxychlor	16.5		1	0.37	0.83	1.70	ug/kg	10/22/25 15:26	PB170205
53494-70-5	Endrin ketone	16.8		1	0.19	0.83	1.70	ug/kg	10/22/25 15:26	PB170205
7421-93-4	Endrin aldehyde	16.2		1	0.37	0.83	1.70	ug/kg	10/22/25 15:26	PB170205
5103-71-9	alpha-Chlordane	15.3		1	0.12	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
5103-74-2	gamma-Chlordane	15.3		1	0.15	0.33	1.70	ug/kg	10/22/25 15:26	PB170205
8001-35-2	Toxaphene	17.0	U	1	5.40	17.0	33.0	ug/kg	10/22/25 15:26	PB170205
SURROGATES										
2051-24-3	Decachlorobiphenyl	24.5			55 - 130		123%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	17.3			42 - 129		86%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3399-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
 Data File : PD090757.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 15:26
 Operator : AR\AJ
 Sample : PB170205BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB170205BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 14:22:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.873	52460372	514.1E6	16.846	17.266
28)	SA Decachlor...	9.067	8.062	101.0E6	743.2E6	21.596	24.536
Target Compounds							
2)	A alpha-BHC	3.994	3.386	324.2E6	2276.5E6	45.167	44.553
3)	MA gamma-BHC...	4.325	3.722	305.1E6	2117.3E6	44.880	44.811
4)	MA Heptachlor	4.923	4.075	300.8E6	2063.8E6	53.818	46.271
5)	MB Aldrin	5.264	4.360	297.9E6	2099.6E6	45.504	45.329
6)	B beta-BHC	4.512	4.018	116.0E6	889.1E6	44.284	44.835
7)	B delta-BHC	4.760	4.255	308.4E6	2160.5E6	45.606	45.278
8)	B Heptachlo...	5.685	4.864	267.0E6	1933.7E6	46.122	45.350
9)	A Endosulfan I	6.068	5.238	253.5E6	1765.2E6	45.788	44.814
10)	B gamma-Chl...	5.940	5.116	269.8E6	2068.4E6	45.490	46.017
11)	B alpha-Chl...	6.021	5.181	274.3E6	1983.4E6	44.377	45.979
12)	B 4,4'-DDE	6.190	5.365	240.2E6	2036.5E6	45.096	46.454
13)	MA Dieldrin	6.341	5.503	271.2E6	2072.0E6	46.509	46.895
14)	MA Endrin	6.568	5.780	228.9E6	1894.3E6	46.206	46.779
15)	B Endosulfa...	6.781	6.072	234.1E6	1779.6E6	46.814	47.386
16)	A 4,4'-DDD	6.700	5.921	196.3E6	1738.1E6	48.426	47.700
17)	MA 4,4'-DDT	7.016	6.174	199.6E6	1733.1E6	48.561	49.415
18)	B Endrin al...	6.910	6.250	174.1E6	1350.3E6	47.527	48.762
19)	B Endosulfa...	7.145	6.474	214.1E6	1725.0E6	46.168	48.125
20)	A Methoxychlor	7.490	6.745	101.9E6	862.7E6	47.917	49.504
21)	B Endrin ke...	7.625	6.983	229.8E6	1904.4E6	47.298	50.492
22)	Mirex	8.109	7.175	158.9E6	1361.8E6	49.646	53.006



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090757.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 15:26
Operator : AR\AJ
Sample : PB170205BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB170205BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:22:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 52460372
Conc: 16.85 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170205BS




#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: -0.001 min
Response: 514128562
Conc: 17.27 ng/ml



#2 alpha-BHC


R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 324173183
Conc: 45.17 ng/ml



#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2276486192
Conc: 44.55 ng/ml

Response_ Signal: PD090757.D\ECD1A.ch




#3 gamma-BHC (Lindane)

R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 305130987
Conc: 44.88 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170205BS


Time Response_ Signal: PD090757.D\ECD2B.ch



#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 2117279323
Conc: 44.81 ng/ml


Time Response_ Signal: PD090757.D\ECD1A.ch



#4 Heptachlor

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 300818590
Conc: 53.82 ng/ml

Time Response_ Signal: PD090757.D\ECD2B.ch



#4 Heptachlor


R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 2063826928
Conc: 46.27 ng/ml

Response_ Signal: PD090757.D\ECD1A.ch

#5 Aldrin

R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 297926320
Conc: 45.50 ng/ml


Instrument :
ECD_D
ClientSampleId :
PB170205BS



Time Response_ Signal: PD090757.D\ECD2B.ch

#5 Aldrin


R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 2099599075
Conc: 45.33 ng/ml



Time Response_ Signal: PD090757.D\ECD1A.ch

#6 beta-BHC


R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 116008207
Conc: 44.28 ng/ml



Time Response_ Signal: PD090757.D\ECD2B.ch

#6 beta-BHC

R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 889058142
Conc: 44.84 ng/ml



Response_ Signal: PD090757.D\ECD1A.ch

#7 delta-BHC

R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 308377761
Conc: 45.61 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170205BS

Time

Response_ Signal: PD090757.D\ECD2B.ch

#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 2160508855
Conc: 45.28 ng/ml

Time

Response_ Signal: PD090757.D\ECD1A.ch

#8 Heptachlor epoxide

R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 267004164
Conc: 46.12 ng/ml

Time

Response_ Signal: PD090757.D\ECD2B.ch

#8 Heptachlor epoxide

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 1933684395
Conc: 45.35 ng/ml


Time

Response_ Signal: PD090757.D\ECD1A.ch

#9 Endosulfan I

R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 253513363
Conc: 45.79 ng/ml


Instrument :
ECD_D
ClientSampleId :
PB170205BS



Time Response_ Signal: PD090757.D\ECD2B.ch

#9 Endosulfan I


R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 1765236984
Conc: 44.81 ng/ml



Time Response_ Signal: PD090757.D\ECD1A.ch

#10 gamma-Chlordane


R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 269809384
Conc: 45.49 ng/ml




Time Response_ Signal: PD090757.D\ECD2B.ch

#10 gamma-Chlordane

R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 2068434487
Conc: 46.02 ng/ml






#11 alpha-Chlordane


R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 274258448
Conc: 44.38 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170205BS




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 1983443695
Conc: 45.98 ng/ml



#12 4,4'-DDE


R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 240209618
Conc: 45.10 ng/ml



#12 4,4'-DDE

R.T.: 5.365 min
Delta R.T.: 0.000 min
Response: 2036450152
Conc: 46.45 ng/ml

Response_ Signal: PD090757.D\ECD1A.ch




#13 Dieldrin

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 271208242
Conc: 46.51 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170205BS


Time Response_ Signal: PD090757.D\ECD2B.ch



#13 Dieldrin

R.T.: 5.503 min
Delta R.T.: 0.000 min
Response: 2072023796
Conc: 46.90 ng/ml


Time Response_ Signal: PD090757.D\ECD1A.ch



#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 228891906
Conc: 46.21 ng/ml

Time Response_ Signal: PD090757.D\ECD2B.ch



#14 Endrin

R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 1894283295
Conc: 46.78 ng/ml

Response_ Signal: PD090757.D\ECD1A.ch

#15 Endosulfan II

R.T.: 6.781 min
Delta R.T.: 0.000 min
Response: 234069858
Conc: 46.81 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170205BS

Time

Response_ Signal: PD090757.D\ECD2B.ch

#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1779595213
Conc: 47.39 ng/ml

Time

Response_ Signal: PD090757.D\ECD1A.ch

#16 4,4'-DDD

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 196329837
Conc: 48.43 ng/ml


Time

Response_ Signal: PD090757.D\ECD2B.ch

#16 4,4'-DDD

R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 1738079654
Conc: 47.70 ng/ml


Time



#17 4,4'-DDT


R.T.: 7.016 min
Delta R.T.: 0.001 min
Response: 199607104
Conc: 48.56 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170205BS




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1733144260
Conc: 49.42 ng/ml



#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 174077411
Conc: 47.53 ng/ml



#18 Endrin aldehyde


R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1350344806
Conc: 48.76 ng/ml

Response_ Signal: PD090757.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 214134631
Conc: 46.17 ng/ml


Instrument :
ECD_D
ClientSampleId :
PB170205BS



Time Response_ Signal: PD090757.D\ECD2B.ch

#19 Endosulfan Sulfate


R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1725040998
Conc: 48.12 ng/ml



Time Response_ Signal: PD090757.D\ECD1A.ch

#20 Methoxychlor


R.T.: 7.490 min
Delta R.T.: 0.001 min
Response: 101851013
Conc: 47.92 ng/ml




Time Response_ Signal: PD090757.D\ECD2B.ch

#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 862741493
Conc: 49.50 ng/ml



Response_ Signal: PD090757.D\ECD1A.ch




#21 Endrin ketone

R.T.: 7.625 min
Delta R.T.: 0.000 min
Response: 229815542
Conc: 47.30 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170205BS


Time Response_ Signal: PD090757.D\ECD2B.ch



#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1904441465
Conc: 50.49 ng/ml


Time Response_ Signal: PD090757.D\ECD1A.ch



#22 Mirex


R.T.: 8.109 min
Delta R.T.: 0.001 min
Response: 158909880
Conc: 49.65 ng/ml

Time Response_ Signal: PD090757.D\ECD2B.ch



#22 Mirex


R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1361762368
Conc: 53.01 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.067 min
Delta R.T.: 0.003 min
Response: 101024427
Conc: 21.60 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB170205BS



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 743221359
Conc: 24.54 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/20/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/20/25
Client Sample ID:	COMP-ROLLOFFMS		SDG No.:	Q3399
Lab Sample ID:	Q3399-01MS		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	89.9
Sample Wt/Vol:	30.09 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541	Prep Date	10/22/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
319-84-6	alpha-BHC	18.2		1	0.14	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
319-85-7	beta-BHC	17.5		1	0.20	0.92	1.90	ug/kg	10/22/25 18:11	PB170205
319-86-8	delta-BHC	18.1		1	0.43	0.92	1.90	ug/kg	10/22/25 18:11	PB170205
58-89-9	gamma-BHC (Lindane)	17.7		1	0.16	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
76-44-8	Heptachlor	20.9		1	0.13	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
309-00-2	Aldrin	17.8		1	0.13	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
1024-57-3	Heptachlor epoxide	18.1		1	0.21	0.92	1.90	ug/kg	10/22/25 18:11	PB170205
959-98-8	Endosulfan I	17.7		1	0.16	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
60-57-1	Dieldrin	18.0		1	0.16	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
72-55-9	4,4-DDE	18.0		1	0.16	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
72-20-8	Endrin	17.9		1	0.16	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
33213-65-9	Endosulfan II	18.2		1	0.32	0.92	1.90	ug/kg	10/22/25 18:11	PB170205
72-54-8	4,4-DDD	18.4		1	0.17	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
1031-07-8	Endosulfan Sulfate	18.5		1	0.14	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
50-29-3	4,4-DDT	19.2		1	0.16	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
72-43-5	Methoxychlor	18.9		1	0.41	0.92	1.90	ug/kg	10/22/25 18:11	PB170205
53494-70-5	Endrin ketone	18.9		1	0.21	0.92	1.90	ug/kg	10/22/25 18:11	PB170205
7421-93-4	Endrin aldehyde	18.3		1	0.41	0.92	1.90	ug/kg	10/22/25 18:11	PB170205
5103-71-9	alpha-Chlordane	17.9		1	0.13	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
5103-74-2	gamma-Chlordane	17.9		1	0.17	0.37	1.90	ug/kg	10/22/25 18:11	PB170205
8001-35-2	Toxaphene	18.9	U	1	6.00	18.9	36.6	ug/kg	10/22/25 18:11	PB170205
SURROGATES										
2051-24-3	Decachlorobiphenyl	15.3			55 - 130		77%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	19.1			42 - 129		96%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3399-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
 Data File : PD090768.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 18:11
 Operator : AR\AJ
 Sample : Q3399-01MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COMP-ROLLOFFMS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/23/2025
 Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 14:23:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.874	55259130	570.1E6	17.744m	19.146
28)	SA Decachlor...	9.067	8.062	65859422	464.0E6	14.079	15.317
Target Compounds							
2)	A alpha-BHC	3.994	3.386	353.9E6	2433.0E6	49.315	47.616
3)	MA gamma-BHC...	4.325	3.722	325.3E6	2250.0E6	47.843	47.620
4)	MA Heptachlor	4.923	4.075	315.9E6	2185.3E6	56.513	48.994
5)	MB Aldrin	5.264	4.360	314.9E6	2225.7E6	48.102	48.052
6)	B beta-BHC	4.513	4.018	123.8E6	938.0E6	47.251	47.302
7)	B delta-BHC	4.760	4.255	331.8E6	2286.7E6	49.066	47.923
8)	B Heptachlo...	5.685	4.864	283.6E6	2032.2E6	48.993	47.661
9)	A Endosulfan I	6.068	5.238	265.3E6	1883.5E6	47.918	47.817
10)	B gamma-Chl...	5.940	5.116	283.0E6	2180.3E6	47.715	48.506
11)	B alpha-Chl...	6.021	5.181	296.7E6	2094.3E6	48.014	48.550
12)	B 4,4'-DDE	6.191	5.366	254.1E6	2138.5E6	47.709	48.782
13)	MA Dieldrin	6.341	5.504	283.6E6	2141.6E6	48.629	48.469
14)	MA Endrin	6.568	5.780	239.7E6	1957.8E6	48.393	48.349
15)	B Endosulfa...	6.782	6.072	226.8E6	1847.3E6	45.363	49.188
16)	A 4,4'-DDD	6.701	5.921	201.3E6	1717.5E6	49.653	47.134
17)	MA 4,4'-DDT	7.016	6.174	206.0E6	1822.3E6	50.113	51.957
18)	B Endrin al...	6.910	6.250	174.4E6	1373.5E6	47.625	49.597
19)	B Endosulfa...	7.145	6.474	218.3E6	1795.4E6	47.056	50.089
20)	A Methoxychlor	7.490	6.745	103.3E6	889.0E6	48.607	51.010
21)	B Endrin ke...	7.627	6.983	236.4E6	1933.5E6	48.654	51.263
22)	Mirex	8.109	7.175	150.2E6	1272.8E6	46.913	49.543

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090768.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 18:11
Operator : AR\AJ
Sample : Q3399-01MS
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS

Manual Integrations
APPROVED


Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:23:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090768.D\ECD1A.ch



#1 Tetrachloro-m-xylene


R.T.: 3.543 min
Delta R.T.: -0.002 min
Response: 55259130
Conc: 17.74 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025


Response_ Signal: PD090768.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 570118937
Conc: 19.15 ng/ml


Response_ Signal: PD090768.D\ECD1A.ch



#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 353943515
Conc: 49.32 ng/ml


Response_ Signal: PD090768.D\ECD2B.ch



#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2432979872
Conc: 47.62 ng/ml

Response_ Signal: PD090768.D\ECD1A.ch



#3 gamma-BHC (Lindane)


R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 325274369
Conc: 47.84 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025


Time Response_ Signal: PD090768.D\ECD2B.ch



#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 2249990116
Conc: 47.62 ng/ml


Time Response_ Signal: PD090768.D\ECD1A.ch



#4 Heptachlor


R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 315881480
Conc: 56.51 ng/ml

Time Response_ Signal: PD090768.D\ECD2B.ch



#4 Heptachlor

R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 2185284603
Conc: 48.99 ng/ml




#5 Aldrin

R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 314935367
Conc: 48.10 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025




#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 2225696544
Conc: 48.05 ng/ml




#6 beta-BHC

R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 123781257
Conc: 47.25 ng/ml



#6 beta-BHC

R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 937968947
Conc: 47.30 ng/ml




#7 delta-BHC

R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 331768828
Conc: 49.07 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025




#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 2286739044
Conc: 47.92 ng/ml




#8 Heptachlor epoxide

R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 283621153
Conc: 48.99 ng/ml



#8 Heptachlor epoxide

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 2032230337
Conc: 47.66 ng/ml




#9 Endosulfan I

R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 265306892
Conc: 47.92 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 1883530425
Conc: 47.82 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 283008513
Conc: 47.72 ng/ml



#10 gamma-Chlordane

R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 2180269957
Conc: 48.51 ng/ml




#11 alpha-Chlordane

R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 296739854
Conc: 48.01 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2094323635
Conc: 48.55 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 254128833
Conc: 47.71 ng/ml



#12 4,4'-DDE

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 2138528914
Conc: 48.78 ng/ml




#13 Dieldrin

R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 283574965
Conc: 48.63 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025




#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 2141551063
Conc: 48.47 ng/ml



#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 239724746
Conc: 48.39 ng/ml



#14 Endrin

R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 1957839756
Conc: 48.35 ng/ml

Response_ Signal: PD090768.D\ECD1A.ch

#15 Endosulfan II

R.T.: 6.782 min
Delta R.T.: 0.000 min
Response: 226813846
Conc: 45.36 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025

Time

Response_ Signal: PD090768.D\ECD2B.ch

#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1847281980
Conc: 49.19 ng/ml

Time

Response_ Signal: PD090768.D\ECD1A.ch

#16 4,4'-DDD

R.T.: 6.701 min
Delta R.T.: 0.000 min
Response: 201302989
Conc: 49.65 ng/ml


Time

Response_ Signal: PD090768.D\ECD2B.ch

#16 4,4'-DDD

R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 1717457499
Conc: 47.13 ng/ml

Time




#17 4,4'-DDT

R.T.: 7.016 min
Delta R.T.: 0.001 min
Response: 205987070
Conc: 50.11 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1822278697
Conc: 51.96 ng/ml



#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 174437307
Conc: 47.63 ng/ml



#18 Endrin aldehyde

R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1373492836
Conc: 49.60 ng/ml

Response_ Signal: PD090768.D\ECD1A.ch

#19 Endosulfan Sulfate

R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 218256060
Conc: 47.06 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS

Manual Integrations
APPROVED


Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025

Time

Response_ Signal: PD090768.D\ECD2B.ch

#19 Endosulfan Sulfate


R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1795448252
Conc: 50.09 ng/ml



Response_ Signal: PD090768.D\ECD1A.ch

#20 Methoxychlor


R.T.: 7.490 min
Delta R.T.: 0.001 min
Response: 103317287
Conc: 48.61 ng/ml




Response_ Signal: PD090768.D\ECD2B.ch

#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 888989148
Conc: 51.01 ng/ml






#21 Endrin ketone

R.T.: 7.627 min
Delta R.T.: 0.002 min
Response: 236402703
Conc: 48.65 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025




#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1933535947
Conc: 51.26 ng/ml




#22 Mirex

R.T.: 8.109 min
Delta R.T.: 0.001 min
Response: 150162319
Conc: 46.91 ng/ml



#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1272797690
Conc: 49.54 ng/ml




#28 Decachlorobiphenyl

R.T.: 9.067 min
Delta R.T.: 0.003 min
Response: 65859422
Conc: 14.08 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/23/2025
Supervised By :mohammad ahmed 10/24/2025



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 463986691
Conc: 15.32 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/20/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/20/25
Client Sample ID:	COMP-ROLLOFFMSD		SDG No.:	Q3399
Lab Sample ID:	Q3399-01MSD		Matrix:	SOIL
Analytical Method:	8081B		% Solid:	89.9
Sample Wt/Vol:	30.08 g	Final Vol:	10000 uL	Test:
Prep Method:	SW3541	Prep Date	10/22/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
319-84-6	alpha-BHC	18.5		1	0.14	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
319-85-7	beta-BHC	17.7		1	0.20	0.92	1.90	ug/kg	10/22/25 18:25	PB170205
319-86-8	delta-BHC	18.5		1	0.43	0.92	1.90	ug/kg	10/22/25 18:25	PB170205
58-89-9	gamma-BHC (Lindane)	18.0		1	0.16	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
76-44-8	Heptachlor	21.3		1	0.13	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
309-00-2	Aldrin	18.1		1	0.13	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
1024-57-3	Heptachlor epoxide	18.3		1	0.21	0.92	1.90	ug/kg	10/22/25 18:25	PB170205
959-98-8	Endosulfan I	18.0		1	0.16	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
60-57-1	Dieldrin	18.3		1	0.16	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
72-55-9	4,4-DDE	18.3		1	0.16	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
72-20-8	Endrin	18.3		1	0.16	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
33213-65-9	Endosulfan II	18.6		1	0.32	0.92	1.90	ug/kg	10/22/25 18:25	PB170205
72-54-8	4,4-DDD	18.7		1	0.17	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
1031-07-8	Endosulfan Sulfate	18.8		1	0.14	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
50-29-3	4,4-DDT	19.7		1	0.16	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
72-43-5	Methoxychlor	19.3		1	0.41	0.92	1.90	ug/kg	10/22/25 18:25	PB170205
53494-70-5	Endrin ketone	19.2		1	0.21	0.92	1.90	ug/kg	10/22/25 18:25	PB170205
7421-93-4	Endrin aldehyde	18.6		1	0.41	0.92	1.90	ug/kg	10/22/25 18:25	PB170205
5103-71-9	alpha-Chlordane	18.2		1	0.13	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
5103-74-2	gamma-Chlordane	18.2		1	0.17	0.37	1.90	ug/kg	10/22/25 18:25	PB170205
8001-35-2	Toxaphene	18.9	U	1	6.00	18.9	36.6	ug/kg	10/22/25 18:25	PB170205
SURROGATES										
2051-24-3	Decachlorobiphenyl	15.7			55 - 130		79%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	19.4			42 - 129		97%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

Q3399-Pesticide-TCL

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
 Data File : PD090769.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 18:25
 Operator : AR\AJ
 Sample : Q3399-01MSD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COMP-ROLLOFFMSD

Manual Integrations
 APPROVED

Reviewed By :Abdul
 Mirza

10/23/2025
 Supervised By :mohammad
 ahmed

10/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 14:23:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
System Monitoring Compounds						
1) SA Tetrachlo...	3.543	2.874	54579548	578.4E6	17.526m	19.425
28) SA Decachlor...	9.068	8.062	67109061	476.8E6	14.346	15.741
Target Compounds						
2) A alpha-BHC	3.994	3.386	358.4E6	2475.1E6	49.931	48.440
3) MA gamma-BHC...	4.325	3.722	331.9E6	2281.5E6	48.810	48.287
4) MA Heptachlor	4.923	4.075	322.2E6	2211.8E6	57.644	49.589
5) MB Aldrin	5.264	4.360	320.3E6	2251.0E6	48.919	48.597
6) B beta-BHC	4.512	4.019	125.7E6	951.4E6	47.986	47.977
7) B delta-BHC	4.760	4.255	337.7E6	2326.8E6	49.941	48.763
8) B Heptachlo...	5.685	4.864	286.7E6	2057.0E6	49.525	48.243
9) A Endosulfan I	6.069	5.239	270.0E6	1908.1E6	48.767	48.440
10) B gamma-Chl...	5.940	5.117	287.5E6	2211.2E6	48.476	49.194
11) B alpha-Chl...	6.022	5.181	296.3E6	2125.8E6	47.945	49.279
12) B 4,4'-DDE	6.191	5.365	260.3E6	2166.5E6	48.861	49.420
13) MA Dieldrin	6.342	5.504	289.1E6	2168.8E6	49.583	49.086
14) MA Endrin	6.569	5.781	244.9E6	1970.4E6	49.430	48.660
15) B Endosulfa...	6.782	6.072	231.9E6	1885.7E6	46.374	50.211
16) A 4,4'-DDD	6.701	5.921	204.9E6	1742.0E6	50.544	47.806
17) MA 4,4'-DDT	7.017	6.174	211.0E6	1863.8E6	51.335	53.142
18) B Endrin al...	6.910	6.251	178.4E6	1394.3E6	48.714	50.347
19) B Endosulfa...	7.146	6.474	222.2E6	1818.0E6	47.910	50.718
20) A Methoxychlor	7.490	6.745	105.1E6	908.4E6	49.463	52.124
21) B Endrin ke...	7.626	6.983	241.7E6	1962.1E6	49.740	52.020
22) Mirex	8.109	7.175	151.1E6	1297.4E6	47.218	50.499

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090769.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 18:25
Operator : AR\AJ
Sample : Q3399-01MSD
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

COMP-ROLLOFFMSD

Manual Integrations

APPROVED

Reviewed By :Abdul
Mirza



10/23/2025

Supervised By :mohammad
ahmed


10/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:23:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PD090769.D\ECD1A.ch



#1 Tetrachloro-m-xylene

R.T.: 3.543 min
Delta R.T.: -0.002 min
Response: 54579548
Conc: 17.53 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD


Manual Integrations
APPROVED

Reviewed By :Abdul
Mirza

10/23/2025
Supervised By :mohammad
ahmed

10/24/2025


Response_ Signal: PD090769.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 578415516
Conc: 19.42 ng/ml


Response_ Signal: PD090769.D\ECD1A.ch



#2 alpha-BHC


R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 358360933
Conc: 49.93 ng/ml

Response_ Signal: PD090769.D\ECD2B.ch



#2 alpha-BHC

R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2475094258
Conc: 48.44 ng/ml



#3 gamma-BHC (Lindane)

R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 331851325
Conc: 48.81 ng/ml


Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD

Manual Integrations
APPROVED

Reviewed By :Abdul
Mirza


10/23/2025
Supervised By :mohammad
ahmed

10/24/2025




#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 2281514650
Conc: 48.29 ng/ml




#4 Heptachlor

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 322201675
Conc: 57.64 ng/ml



#4 Heptachlor

R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 2211816269
Conc: 49.59 ng/ml



#5 Aldrin

R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 320284321
Conc: 48.92 ng/ml


Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD

Manual Integrations
APPROVED

Reviewed By :Abdul
Mirza


10/23/2025
Supervised By :mohammad
ahmed

10/24/2025




#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 2250954350
Conc: 48.60 ng/ml




#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 125705624
Conc: 47.99 ng/ml



#6 beta-BHC

R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 951364052
Conc: 47.98 ng/ml



#7 delta-BHC

R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 337690238
Conc: 49.94 ng/ml

Instrument :

ECD_D

ClientSampleId :

COMP-ROLLOFFMSD


Manual Integrations
APPROVED

Reviewed By :Abdul
Mirza

10/23/2025


Supervised By :mohammad
ahmed

10/24/2025




#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 2326791858
Conc: 48.76 ng/ml




#8 Heptachlor epoxide

R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 286703039
Conc: 49.52 ng/ml



#8 Heptachlor epoxide

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 2057020731
Conc: 48.24 ng/ml



#9 Endosulfan I

R.T.: 6.069 min
Delta R.T.: 0.000 min
Response: 270006575
Conc: 48.77 ng/ml


Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD

Manual Integrations
APPROVED

Reviewed By :Abdul
Mirza


10/23/2025
Supervised By :mohammad
ahmed

10/24/2025




#9 Endosulfan I

R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 1908057666
Conc: 48.44 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 287520328
Conc: 48.48 ng/ml



#10 gamma-Chlordane

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 2211234747
Conc: 49.19 ng/ml



#11 alpha-Chlordane

R.T.: 6.022 min
Delta R.T.: 0.001 min
Response: 296310307
Conc: 47.94 ng/ml


Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD

Manual Integrations
APPROVED

Reviewed By :Abdul
Mirza


10/23/2025
Supervised By :mohammad
ahmed

10/24/2025




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2125775960
Conc: 49.28 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 260264333
Conc: 48.86 ng/ml



#12 4,4'-DDE

R.T.: 5.365 min
Delta R.T.: 0.000 min
Response: 2166466646
Conc: 49.42 ng/ml



#13 Dieldrin

R.T.: 6.342 min
Delta R.T.: 0.001 min
Response: 289136583
Conc: 49.58 ng/ml


Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD

Manual Integrations
APPROVED

Reviewed By :Abdul
Mirza


10/23/2025
Supervised By :mohammad
ahmed

10/24/2025




#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 2168799726
Conc: 49.09 ng/ml




#14 Endrin

R.T.: 6.569 min
Delta R.T.: 0.000 min
Response: 244864046
Conc: 49.43 ng/ml



#14 Endrin

R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 1970443496
Conc: 48.66 ng/ml



#15 Endosulfan II

R.T.: 6.782 min
Delta R.T.: 0.001 min
Response: 231867559
Conc: 46.37 ng/ml


Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD

Manual Integrations
APPROVED

Reviewed By :Abdul
Mirza


10/23/2025
Supervised By :mohammad
ahmed

10/24/2025




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1885706637
Conc: 50.21 ng/ml




#16 4,4'-DDD

R.T.: 6.701 min
Delta R.T.: 0.001 min
Response: 204914718
Conc: 50.54 ng/ml



#16 4,4'-DDD

R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 1741968091
Conc: 47.81 ng/ml



#17 4,4'-DDT

R.T.: 7.017 min
Delta R.T.: 0.002 min
Response: 211011087
Conc: 51.34 ng/ml


Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD

Manual Integrations
APPROVED

Reviewed By :Abdul
Mirza


10/23/2025
Supervised By :mohammad
ahmed

10/24/2025




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1863845638
Conc: 53.14 ng/ml



#18 Endrin aldehyde


R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 178426338
Conc: 48.71 ng/ml



#18 Endrin aldehyde

R.T.: 6.251 min
Delta R.T.: 0.000 min
Response: 1394257530
Conc: 50.35 ng/ml

Response_ Signal: PD090769.D\ECD1A.ch



#19 Endosulfan Sulfate

R.T.: 7.146 min
Delta R.T.: 0.001 min
Response: 222214770
Conc: 47.91 ng/ml

Instrument :

ECD_D

ClientSampleId :

COMP-ROLLOFFMSD

Manual Integrations
APPROVED


Reviewed By :Abdul
Mirza

10/23/2025

Supervised By :mohammad
ahmed

10/24/2025


Time Response_ Signal: PD090769.D\ECD2B.ch



#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1817999761
Conc: 50.72 ng/ml


Time Response_ Signal: PD090769.D\ECD1A.ch



#20 Methoxychlor


R.T.: 7.490 min
Delta R.T.: 0.001 min
Response: 105138524
Conc: 49.46 ng/ml

Time Response_ Signal: PD090769.D\ECD2B.ch



#20 Methoxychlor

R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 908402672
Conc: 52.12 ng/ml



#21 Endrin ketone

R.T.: 7.626 min
Delta R.T.: 0.001 min
Response: 241679099
Conc: 49.74 ng/ml


Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD

Manual Integrations
APPROVED

Reviewed By :Abdul
Mirza


10/23/2025
Supervised By :mohammad
ahmed

10/24/2025




#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1962091852
Conc: 52.02 ng/ml




#22 Mirex

R.T.: 8.109 min
Delta R.T.: 0.002 min
Response: 151136710
Conc: 47.22 ng/ml



#22 Mirex

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1297356869
Conc: 50.50 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.068 min
Delta R.T.: 0.003 min
Response: 67109061
Conc: 14.35 ng/ml


Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD

Manual Integrations
APPROVED

Reviewed By :Abdul
Mirza

10/23/2025
Supervised By :mohammad
ahmed

10/24/2025



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 476816665
Conc: 15.74 ng/ml

Manual Integration Report

Sequence:	pd101725	Instrument	ECD_d

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
PEM	PD090648.D	Endrin aldehyde	Abdul	10/19/2025 5:20:46 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PEM	PD090648.D	Endrin ketone #2	Abdul	10/19/2025 5:20:46 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
RESCHK	PD090649.D	Tetrachloro-m-xylene	Abdul	10/19/2025 5:20:37 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PSTDICC005	PD090654.D	alpha-Chlordane	Abdul	10/19/2025 5:19:40 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PSTDICC005	PD090654.D	Endrin ketone	Abdul	10/19/2025 5:19:40 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PSTDICC005	PD090654.D	Tetrachloro-m-xylene	Abdul	10/19/2025 5:19:40 PM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PTOXICV500	PD090667.D	Toxaphene-5 #2	Abdul	10/21/2025 8:49:16 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PEM	PD090669.D	4,4"-DDD	Abdul	10/21/2025 8:49:12 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PEM	PD090669.D	4,4"-DDD #2	Abdul	10/21/2025 8:49:12 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software
PEM	PD090669.D	Endrin aldehyde	Abdul	10/21/2025 8:49:12 AM	mohammad	10/24/2025 1:04:58	Peak Integrated by Software

Manual Integration Report

Sequence:	PD102225	Instrument	ECD_d

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q3399-01	PD090758.D	Dieldrin	Abdul	10/23/2025 3:58:29 PM	mohammad	10/24/2025 1:06:26	Peak Integrated by Software
Q3399-01	PD090758.D	Dieldrin #2	Abdul	10/23/2025 3:58:29 PM	mohammad	10/24/2025 1:06:26	Peak Integrated by Software
Q3399-01	PD090758.D	Endrin	Abdul	10/23/2025 3:58:29 PM	mohammad	10/24/2025 1:06:26	Peak Integrated by Software
Q3399-01	PD090758.D	Endrin #2	Abdul	10/23/2025 3:58:29 PM	mohammad	10/24/2025 1:06:26	Peak Integrated by Software
Q3399-01	PD090758.D	Tetrachloro-m-xylene	Abdul	10/23/2025 3:58:29 PM	mohammad	10/24/2025 1:06:26	Peak Integrated by Software
Q3399-01MS	PD090768.D	Tetrachloro-m-xylene	Abdul	10/23/2025 3:57:49 PM	mohammad	10/24/2025 1:06:26	Peak Integrated by Software
Q3399-01MSD	PD090769.D	Tetrachloro-m-xylene	Abdul	10/23/2025 3:57:43 PM	mohammad	10/24/2025 1:06:26	Peak Integrated by Software

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD090646.D	17 Oct 2025 10:40	AR\AJ	Ok
2	I.BLK	PD090647.D	17 Oct 2025 10:53	AR\AJ	Ok
3	PEM	PD090648.D	17 Oct 2025 11:07	AR\AJ	Ok,M
4	RESCHK	PD090649.D	17 Oct 2025 11:20	AR\AJ	Ok,M
5	PSTDICC100	PD090650.D	17 Oct 2025 11:34	AR\AJ	Ok
6	PSTDICC075	PD090651.D	17 Oct 2025 11:48	AR\AJ	Ok
7	PSTDICC050	PD090652.D	17 Oct 2025 12:01	AR\AJ	Ok
8	PSTDICC025	PD090653.D	17 Oct 2025 12:15	AR\AJ	Ok
9	PSTDICC005	PD090654.D	17 Oct 2025 12:28	AR\AJ	Ok,M
10	PCHLORICC1000	PD090655.D	17 Oct 2025 12:42	AR\AJ	Ok
11	PCHLORICC750	PD090656.D	17 Oct 2025 12:56	AR\AJ	Ok
12	PCHLORICC500	PD090657.D	17 Oct 2025 13:09	AR\AJ	Ok
13	PCHLORICC250	PD090658.D	17 Oct 2025 13:23	AR\AJ	Ok
14	PCHLORICC050	PD090659.D	17 Oct 2025 13:36	AR\AJ	Ok,M
15	PTOXICC1000	PD090660.D	17 Oct 2025 13:50	AR\AJ	Ok
16	PTOXICC750	PD090661.D	17 Oct 2025 14:03	AR\AJ	Ok
17	PTOXICC500	PD090662.D	17 Oct 2025 14:17	AR\AJ	Ok
18	PTOXICC250	PD090663.D	17 Oct 2025 14:30	AR\AJ	Ok
19	PTOXICC100	PD090664.D	17 Oct 2025 16:18	AR\AJ	Ok
20	PSTDICV050	PD090665.D	17 Oct 2025 16:46	AR\AJ	Ok
21	PCHLORICV500	PD090666.D	17 Oct 2025 17:50	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PD090667.D	17 Oct 2025 18:42	AR\AJ	Ok,M
23	I.BLK	PD090668.D	17 Oct 2025 19:50	AR\AJ	Ok
24	PEM	PD090669.D	17 Oct 2025 20:04	AR\AJ	Ok,M
25	PSTDCCC050	PD090670.D	17 Oct 2025 20:18	AR\AJ	Ok
26	PB170142BL	PD090671.D	17 Oct 2025 20:31	AR\AJ	Ok
27	PB170142BS	PD090672.D	17 Oct 2025 20:45	AR\AJ	Ok
28	Q3366-01	PD090673.D	17 Oct 2025 20:59	AR\AJ	Ok,M
29	Q3366-03	PD090674.D	17 Oct 2025 21:12	AR\AJ	Ok,M
30	Q3367-01	PD090675.D	17 Oct 2025 21:26	AR\AJ	Ok,M
31	Q3369-01	PD090676.D	17 Oct 2025 21:40	AR\AJ	Ok,M
32	Q3369-05	PD090677.D	17 Oct 2025 21:53	AR\AJ	Ok,M
33	Q3363-01	PD090678.D	17 Oct 2025 22:48	AR\AJ	Ok,M
34	Q3363-02	PD090679.D	17 Oct 2025 23:02	AR\AJ	Ok,M
35	Q3363-03	PD090680.D	17 Oct 2025 23:15	AR\AJ	Ok,M
36	Q3363-04	PD090681.D	17 Oct 2025 23:29	AR\AJ	Ok,M
37	Q3363-04MS	PD090682.D	17 Oct 2025 23:43	AR\AJ	Ok,M
38	Q3363-04MSD	PD090683.D	17 Oct 2025 23:56	AR\AJ	Ok,M
39	Q3363-05	PD090684.D	18 Oct 2025 00:10	AR\AJ	Ok,M
40	Q3343-04MS	PD090685.D	18 Oct 2025 00:24	AR\AJ	Ok,M
41	Q3343-04MSD	PD090686.D	18 Oct 2025 00:37	AR\AJ	Ok
42	I.BLK	PD090687.D	18 Oct 2025 00:51	AR\AJ	Ok
43	PSTDCCC050	PD090688.D	18 Oct 2025 02:13	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD102225

Review By	Abdul	Review On	10/23/2025 3:59:00 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:06:26 AM
SubDirectory	PD102225	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	P24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PD090752.D	22 Oct 2025 08:22	AR\AJ	Ok
2	I.BLK	PD090753.D	22 Oct 2025 14:32	AR\AJ	Ok
3	PEM	PD090754.D	22 Oct 2025 14:45	AR\AJ	Ok
4	PSTDCCC050	PD090755.D	22 Oct 2025 14:59	AR\AJ	Ok
5	PB170205BL	PD090756.D	22 Oct 2025 15:13	AR\AJ	Ok
6	PB170205BS	PD090757.D	22 Oct 2025 15:26	AR\AJ	Ok
7	Q3399-01	PD090758.D	22 Oct 2025 15:40	AR\AJ	Ok,M
8	Q3399-02	PD090759.D	22 Oct 2025 15:54	AR\AJ	Ok
9	Q3404-01	PD090760.D	22 Oct 2025 16:08	AR\AJ	Ok,M
10	Q3404-04	PD090761.D	22 Oct 2025 16:21	AR\AJ	Ok,M
11	Q3416-01	PD090762.D	22 Oct 2025 16:49	AR\AJ	Ok,M
12	I.BLK	PD090763.D	22 Oct 2025 17:03	AR\AJ	Ok
13	PSTDCCC050	PD090764.D	22 Oct 2025 17:17	AR\AJ	Ok
14	PB170218BL	PD090765.D	22 Oct 2025 17:30	AR\AJ	Ok
15	PB170218BS	PD090766.D	22 Oct 2025 17:44	AR\AJ	Ok
16	PB170172TB	PD090767.D	22 Oct 2025 17:58	AR\AJ	Ok
17	Q3399-01MS	PD090768.D	22 Oct 2025 18:11	AR\AJ	Ok,M
18	Q3399-01MSD	PD090769.D	22 Oct 2025 18:25	AR\AJ	Ok,M
19	Q3399-02MS	PD090770.D	22 Oct 2025 18:38	AR\AJ	Ok
20	Q3399-02MSD	PD090771.D	22 Oct 2025 18:52	AR\AJ	Ok,M
21	I.BLK	PD090772.D	22 Oct 2025 19:06	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD102225

Review By	Abdul	Review On	10/23/2025 3:59:00 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:06:26 AM
SubDirectory	PD102225	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	P24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PSTDCCC050	PD090773.D	22 Oct 2025 19:19	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24942,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD090646.D	17 Oct 2025 10:40		AR\AJ	Ok
2	I.BLK	I.BLK	PD090647.D	17 Oct 2025 10:53		AR\AJ	Ok
3	PEM	PEM	PD090648.D	17 Oct 2025 11:07		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD090649.D	17 Oct 2025 11:20		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PD090650.D	17 Oct 2025 11:34		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD090651.D	17 Oct 2025 11:48		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD090652.D	17 Oct 2025 12:01		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD090653.D	17 Oct 2025 12:15		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD090654.D	17 Oct 2025 12:28		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD090655.D	17 Oct 2025 12:42		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD090656.D	17 Oct 2025 12:56		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD090657.D	17 Oct 2025 13:09		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD090658.D	17 Oct 2025 13:23		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD090659.D	17 Oct 2025 13:36		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PD090660.D	17 Oct 2025 13:50		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD090661.D	17 Oct 2025 14:03		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD090662.D	17 Oct 2025 14:17		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD090663.D	17 Oct 2025 14:30		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD090664.D	17 Oct 2025 16:18		AR\AJ	Ok
20	PSTDICV050	ICVPD101725	PD090665.D	17 Oct 2025 16:46		AR\AJ	Ok
21	PCHLORICV500	ICVPD101725CHLOR	PD090666.D	17 Oct 2025 17:50		AR\AJ	Ok
22	PTOXICV500	ICVPD101725TOX	PD090667.D	17 Oct 2025 18:42		AR\AJ	Ok,M
23	I.BLK	I.BLK	PD090668.D	17 Oct 2025 19:50		AR\AJ	Ok
24	PEM	PEM	PD090669.D	17 Oct 2025 20:04		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PD090670.D	17 Oct 2025 20:18		AR\AJ	Ok
26	PB170142BL	PB170142BL	PD090671.D	17 Oct 2025 20:31		AR\AJ	Ok
27	PB170142BS	PB170142BS	PD090672.D	17 Oct 2025 20:45		AR\AJ	Ok
28	Q3366-01	TR-06-101625	PD090673.D	17 Oct 2025 20:59		AR\AJ	Ok,M
29	Q3366-03	TR-1-101625	PD090674.D	17 Oct 2025 21:12		AR\AJ	Ok,M
30	Q3367-01	BU-03-101625	PD090675.D	17 Oct 2025 21:26		AR\AJ	Ok,M
31	Q3369-01	MH-22	PD090676.D	17 Oct 2025 21:40		AR\AJ	Ok,M
32	Q3369-05	MH-21	PD090677.D	17 Oct 2025 21:53		AR\AJ	Ok,M
33	Q3363-01	PR132-SO1-085011-20	PD090678.D	17 Oct 2025 22:48		AR\AJ	Ok,M
34	Q3363-02	PR132-SO1-039085-20	PD090679.D	17 Oct 2025 23:02		AR\AJ	Ok,M
35	Q3363-03	PR132-SO3-028010-20	PD090680.D	17 Oct 2025 23:15		AR\AJ	Ok,M
36	Q3363-04	PR132-SO3-010013-20	PD090681.D	17 Oct 2025 23:29		AR\AJ	Ok,M
37	Q3363-04MS	PR132-SO3-010013-20	PD090682.D	17 Oct 2025 23:43		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD101725

Review By	Abdul	Review On	10/19/2025 5:21:31 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:04:58 AM
SubDirectory	PD101725	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q3363-04MSD	PR132-SO3-010013-20	PD090683.D	17 Oct 2025 23:56		AR\AJ	Ok,M
39	Q3363-05	COMP01-20251015	PD090684.D	18 Oct 2025 00:10		AR\AJ	Ok,M
40	Q3343-04MS	SOIL-ROLLOFF-WC-1	PD090685.D	18 Oct 2025 00:24		AR\AJ	Ok,M
41	Q3343-04MSD	SOIL-ROLLOFF-WC-1	PD090686.D	18 Oct 2025 00:37		AR\AJ	Ok
42	I.BLK	I.BLK	PD090687.D	18 Oct 2025 00:51		AR\AJ	Ok
43	PSTDCCC050	PSTDCCC050	PD090688.D	18 Oct 2025 02:13		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD102225

Review By	Abdul	Review On	10/23/2025 3:59:00 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:06:26 AM
SubDirectory	PD102225	HP Acquire Method	HP Processing Method pd101725 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24942,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	P24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD090752.D	22 Oct 2025 08:22		AR\AJ	Ok
2	I.BLK	I.BLK	PD090753.D	22 Oct 2025 14:32		AR\AJ	Ok
3	PEM	PEM	PD090754.D	22 Oct 2025 14:45		AR\AJ	Ok
4	PSTDCCC050	PSTDCCC050	PD090755.D	22 Oct 2025 14:59		AR\AJ	Ok
5	PB170205BL	PB170205BL	PD090756.D	22 Oct 2025 15:13		AR\AJ	Ok
6	PB170205BS	PB170205BS	PD090757.D	22 Oct 2025 15:26		AR\AJ	Ok
7	Q3399-01	COMP-ROLLOFF	PD090758.D	22 Oct 2025 15:40		AR\AJ	Ok,M
8	Q3399-02	COMP-ROLLOFF	PD090759.D	22 Oct 2025 15:54		AR\AJ	Ok
9	Q3404-01	MOO-25-0285-0288	PD090760.D	22 Oct 2025 16:08		AR\AJ	Ok,M
10	Q3404-04	MOO-25-0292	PD090761.D	22 Oct 2025 16:21		AR\AJ	Ok,M
11	Q3416-01	NB-08-10212025	PD090762.D	22 Oct 2025 16:49		AR\AJ	Ok,M
12	I.BLK	I.BLK	PD090763.D	22 Oct 2025 17:03		AR\AJ	Ok
13	PSTDCCC050	PSTDCCC050	PD090764.D	22 Oct 2025 17:17		AR\AJ	Ok
14	PB170218BL	PB170218BL	PD090765.D	22 Oct 2025 17:30		AR\AJ	Ok
15	PB170218BS	PB170218BS	PD090766.D	22 Oct 2025 17:44		AR\AJ	Ok
16	PB170172TB	PB170172TB	PD090767.D	22 Oct 2025 17:58		AR\AJ	Ok
17	Q3399-01MS	COMP-ROLLOFFMS	PD090768.D	22 Oct 2025 18:11		AR\AJ	Ok,M
18	Q3399-01MSD	COMP-ROLLOFFMSD	PD090769.D	22 Oct 2025 18:25		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD102225

Review By	Abdul	Review On	10/23/2025 3:59:00 PM
Supervise By	mohammad	Supervise On	10/24/2025 1:06:26 AM
SubDirectory	PD102225	HP Acquire Method	HP Processing Method pd101725 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24942,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	P24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q3399-02MS	COMP-ROLLOFFMS	PD090770.D	22 Oct 2025 18:38		AR\AJ	Ok
20	Q3399-02MSD	COMP-ROLLOFFMSD	PD090771.D	22 Oct 2025 18:52		AR\AJ	Ok,M
21	I.BLK	I.BLK	PD090772.D	22 Oct 2025 19:06		AR\AJ	Ok
22	PSTDCCC050	PSTDCCC050	PD090773.D	22 Oct 2025 19:19		AR\AJ	Ok

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/22/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 10/21/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:15
Out Date: 10/22/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137592

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q3399-01	COMP-ROLLOFF	13	1.15	10.96	12.11	11.00	89.9	
Q3401-01	41-CONCRETE	1	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3402-01	41-LA	2	1.15	10.90	12.05	8.69	69.2	
Q3402-02	41-LA-E2	3	1.13	10.64	11.77	10.2	85.2	
Q3402-04	41-RA	4	1.16	10.73	11.89	7.61	60.1	
Q3402-05	41-RA-E2	5	1.16	10.98	12.14	7.39	56.7	
Q3402-07	41-LB	6	1.16	10.77	11.93	8.26	65.9	
Q3402-08	41-LB-E2	7	1.16	10.95	12.11	6.22	46.2	
Q3402-10	41-RB	8	1.11	10.87	11.98	7.01	54.3	
Q3402-11	41-RB-E2	9	1.15	10.68	11.83	9.78	80.8	
Q3403-01	MH-18	10	1.19	10.49	11.68	10.75	91.1	
Q3403-02	MH-18-EPH	11	1.16	10.66	11.82	10.8	90.4	
Q3403-03	MH-18-VOC	12	1.19	11.37	12.56	11.59	91.5	
Q3404-01	MOO-25-0285-0288	14	1.15	10.74	11.89	10.22	84.5	
Q3404-03	MOO-25-0285-0288-E2	15	1.19	10.42	11.61	9.77	82.3	
Q3404-04	MOO-25-0292	16	1.18	10.22	11.4	10.5	91.2	
Q3404-07	MOO-25-0292-EPH-2	17	1.14	10.27	11.41	10.44	90.6	
Q3405-01	PAD-10-20-25-A	18	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3405-02	PAD-10-20-25-B	19	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3405-03	PAD-10-20-25-C	20	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3405-04	SW-10-20-25-A	21	1.00	1.00	2.00	2.00	100.0	Concrete sample
Q3406-01	ARS20-0037	22	1.18	11.19	12.37	11.24	89.9	
Q3406-03	ARS20-0037-E2	23	1.18	11.48	12.66	11.4	89.0	
Q3407-01	L35-A	24	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-02	L35-B	25	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-03	L36-A	26	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-04	L36-B	27	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-05	L37-A	28	1.00	1.00	2.00	2.00	100.0	PILC

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/22/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 10/21/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:15
Out Date: 10/22/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137592

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3407-06	L37-B	29	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-07	L38-A	30	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-08	L38-B	31	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-09	L39-A	32	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-10	L39-B	33	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-11	L40-A	34	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-12	L40-B	35	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-13	L41-A	36	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-14	L41-B	37	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-15	L42-A	38	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-16	L42-B	39	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-17	L43-A	40	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-18	L43-B	41	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-19	L44-A	42	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-20	L44-B	43	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-21	L45-A	44	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-22	L45-B	45	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-23	L46-A	46	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-24	L46-B	47	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-25	L47-A	48	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-26	L47-B	49	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-27	L48-A	50	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-28	L48-B	51	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-29	L49-A	52	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-30	L49-B	53	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-31	L50-A	54	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-32	L50-B	55	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-33	L51-A	56	1.00	1.00	2.00	2.00	100.0	PILC

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/22/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 10/21/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:15
Out Date: 10/22/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137592

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q3407-34	L51-B	57	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-35	L52-A	58	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-36	L52-B	59	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-37	L53-A	60	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-38	L53-B	61	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-39	245F69-A	62	1.00	1.00	2.00	2.00	100.0	PILC
Q3407-40	245F69-B	63	1.00	1.00	2.00	2.00	100.0	PILC
Q3408-01	ASPHALT-A	64	1.00	1.00	2.00	2.00	100.0	ASPHALT
Q3408-02	ASPHALT-B	65	1.00	1.00	2.00	2.00	100.0	ASPHALT
Q3408-03	ASPHALT-C	66	1.00	1.00	2.00	2.00	100.0	ASPHALT
Q3408-04	ASPHALT-D	67	1.00	1.00	2.00	2.00	100.0	ASPHALT
Q3409-01	EXCAVATION-AREA-A	68	1.00	1.00	2.00	2.00	100.0	Concreate sample
Q3409-02	EXCAVATION-AREA-B	69	1.00	1.00	2.00	2.00	100.0	Concreate sample
Q3410-01	WASTE	70	1.14	10.45	11.59	9.59	80.9	
Q3410-02	VOC	71	1.15	10.55	11.7	9.64	80.5	
Q3410-03	1	72	1.15	10.97	12.12	9.89	79.7	
Q3410-04	2	73	1.13	11.27	12.4	11.24	89.7	
Q3410-05	3	74	1.16	10.62	11.78	9.6	79.5	
Q3410-06	4	75	1.12	10.50	11.62	9.76	82.3	
Q3410-07	5	76	1.16	10.50	11.66	9.43	78.8	
Q3412-01	TP-6	77	1.14	10.81	11.95	10.84	89.7	
Q3412-02	TP-6-EPH	78	1.18	10.73	11.91	10.7	88.7	
Q3412-03	TP-6-VOC	79	1.15	10.52	11.67	10.38	87.7	
Q3413-01	BP-VPB-184-GW-980-982	80	1.14	11.55	12.69	2.00	7.4	sludge sample
Q3414-01	SB-1	81	1.13	10.12	11.25	11.17	99.2	
Q3414-02	SB-2	82	1.15	10.62	11.77	11.6	98.4	
Q3414-03	SB-3	83	1.13	10.53	11.66	11.4	97.5	
Q3414-04	SB-4	84	1.14	10.28	11.42	11.3	98.8	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/22/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 10/21/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:15
Out Date: 10/22/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137592

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q3414-05	SB-5	85	1.16	10.25	11.41	11.34	99.3	
Q3414-06	SB-6	86	1.18	10.36	11.54	11.49	99.5	
Q3414-07	SB-7	87	1.15	10.25	11.4	11.26	98.6	
Q3414-08	SB-8	88	1.19	10.25	11.44	11.14	97.1	
Q3414-09	SB-9	89	1.16	10.31	11.47	11.18	97.2	
Q3414-10	SB-10	90	1.14	10.55	11.69	11.6	99.1	
Q3414-11	SB-11	91	1.19	10.20	11.39	11.04	96.6	
Q3414-12	SB-12	92	1.18	10.10	11.28	10.9	96.2	
Q3414-13	SB-13	93	1.14	10.06	11.2	10.85	96.5	
Q3415-01	50841	94	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3415-02	50842	95	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3415-03	50843	96	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
Q3416-01	NB-08-10212025	97	1.15	10.52	11.67	11.02	93.8	
Q3416-02	NB-08-10212025-E2	98	1.18	10.92	12.1	11.43	93.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

127592

WorkList Name : %1-102125 WorkList ID : 192565 Department : Wet-Chemistry Date : 10-21-2025 08:08:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3399-01	COMP-ROLLOFF	Solid	Percent Solids	Cool 4 deg C	ICES02	D41	10/20/2025	Chemtech -SO
Q3401-01	41-CONCRETE	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3402-01	41-LA	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3402-02	41-LA-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3402-04	41-RA	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3402-05	41-RA-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3402-07	41-LB	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3402-08	41-LB-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3402-10	41-RB	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3402-11	41-RB-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3403-01	MH-18	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3403-02	MH-18-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3403-03	MH-18-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3404-01	MOO-25-0285-0288	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3404-03	MOO-25-0285-0288-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3404-04	MOO-25-0292	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3404-07	MOO-25-0292-EPH-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3405-01	PAD-10-20-25-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3405-02	PAD-10-20-25-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3405-03	PAD-10-20-25-C	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3405-04	SW-10-20-25-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO

Date/Time 10/21/25 15:00 Date/Time 10/21/25 17:13

Raw Sample Received by: [Signature] Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature] Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

137592

WorkList Name : %1-102125 WorkList ID : 192565 Department : Wet-Chemistry Date : 10-21-2025 08:08:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3406-01	ARS20-0037	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3406-03	ARS20-0037-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3407-01	L35-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-02	L35-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-03	L36-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-04	L36-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-05	L37-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-06	L37-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-07	L38-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-08	L38-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-09	L39-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-10	L39-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-11	L40-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-12	L40-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-13	L41-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-14	L41-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-15	L42-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-16	L42-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-17	L43-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-18	L43-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-19	L44-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO

Date/Time 10/21/25 15:00
 Raw Sample Received by: SS WOC
 Raw Sample Relinquished by: QFS

Date/Time 10/21/25 17:30
 Raw Sample Received by: QFS
 Raw Sample Relinquished by: SS WOC

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19
- 20

WORKLIST(Hardcopy Internal Chain)

137592

WorkList Name : %1-102125 WorkList ID : 192565 Department : Wet-Chemistry Date : 10-21-2025 08:08:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3407-20	L44-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-21	L45-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-22	L45-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-23	L46-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-24	L46-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-25	L47-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-26	L47-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-27	L48-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-28	L48-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-29	L49-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-30	L49-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-31	L50-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-32	L50-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-33	L51-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-34	L51-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-35	L52-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-36	L52-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-37	L53-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-38	L53-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-39	245F69-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3407-40	245F69-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO

Date/Time 10/21/25 151.00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 10/21/25
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

WORKLIST(Hardcopy Internal Chain)

137592

WorkList Name : %1-102125 WorkList ID : 192565 Department : Wet-Chemistry Date : 10-21-2025 08:08:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3408-01	ASPHALT-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3408-02	ASPHALT-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3408-03	ASPHALT-C	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3408-04	ASPHALT-D	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3409-01	EXCAVATION-AREA-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3409-02	EXCAVATION-AREA-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/20/2025	Chemtech -SO
Q3410-01	WASTE	Solid	Percent Solids	Cool 4 deg C	SCIA01	J11	10/20/2025	Chemtech -SO
Q3410-02	VOC	Solid	Percent Solids	Cool 4 deg C	SCIA01	J11	10/20/2025	Chemtech -SO
Q3410-03	1	Solid	Percent Solids	Cool 4 deg C	SCIA01	J11	10/20/2025	Chemtech -SO
Q3410-04	2	Solid	Percent Solids	Cool 4 deg C	SCIA01	J11	10/20/2025	Chemtech -SO
Q3410-05	3	Solid	Percent Solids	Cool 4 deg C	SCIA01	J11	10/20/2025	Chemtech -SO
Q3410-06	4	Solid	Percent Solids	Cool 4 deg C	SCIA01	J11	10/20/2025	Chemtech -SO
Q3410-07	5	Solid	Percent Solids	Cool 4 deg C	SCIA01	J11	10/20/2025	Chemtech -SO
Q3412-01	TP-6	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3412-02	TP-6-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3412-03	TP-6-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/20/2025	Chemtech -SO
Q3413-01	BP-VPB-184-GW-980-982	Solid	Percent Solids	Cool 4 deg C	TETR06	D31	10/20/2025	Chemtech -SO
Q3414-01	SB-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO
Q3414-02	SB-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO
Q3414-03	SB-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO
Q3414-04	SB-4	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO

Date/Time 10/21/25 15:00 Date/Time 10/21/25 17:20
 Raw Sample Received by: FB woc Raw Sample Received by: CP SN
 Raw Sample Relinquished by: OP SN Raw Sample Relinquished by: FB woc

WORKLIST(Hardcopy Internal Chain)

N-137592

WorkList Name : %1-102125 WorkList ID : 192565 Department : Wet-Chemistry Date : 10-21-2025 08:08:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3414-05	SB-5	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO
Q3414-06	SB-6	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO
Q3414-07	SB-7	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO
Q3414-08	SB-8	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO
Q3414-09	SB-9	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO
Q3414-10	SB-10	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO
Q3414-11	SB-11	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO
Q3414-12	SB-12	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO
Q3414-13	SB-13	Solid	Percent Solids	Cool 4 deg C	PSEG03	D32	10/21/2025	Chemtech -SO
Q3415-01	50841	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/21/2025	Chemtech -SO
Q3415-02	50842	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/21/2025	Chemtech -SO
Q3415-03	50843	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/21/2025	Chemtech -SO
Q3416-01	NB-08-10212025	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	10/21/2025	Chemtech -SO
Q3416-02	NB-08-10212025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	10/21/2025	Chemtech -SO
Q3418-01	101725	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/21/2025	Chemtech -SO
Q3418-02	101725-A	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/21/2025	Chemtech -SO
Q3418-03	101725-B	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/21/2025	Chemtech -SO

Date/Time

10/21/25 15:00

Raw Sample Received by:

SB WCC

Raw Sample Relinquished by:

cd sm

Date/Time

10/21/25

Raw Sample Received by:

cd sm

Raw Sample Relinquished by:

SB WCC

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/24/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/23/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:15
Out Date: 10/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137624

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q3399-03	COMP-ROLLOFF	15	1.15	10.96	12.11	11.00	89.9	
Q3437-03	SS-15R(2.0-2.5)	1	1.18	10.46	11.64	10.32	87.4	
Q3437-05	SS-26(0-0.5)	2	1.16	10.19	11.35	10.17	88.4	
Q3437-11	SS-27(0-0.5)	3	1.17	10.34	11.51	10.85	93.6	
Q3437-14	SS-25(0-0.5)	4	1.11	10.33	11.44	10.34	89.4	
Q3437-16	SS-23(0-0.5)	5	1.13	10.84	11.97	11.26	93.5	
Q3439-01	TP-9	6	1.18	10.29	11.47	10.55	91.1	
Q3439-02	TP-9-EPH	7	1.19	10.61	11.8	10.87	91.2	
Q3439-03	TP-9-VOC	8	1.18	10.14	11.32	10.41	91.0	
Q3439-05	TP-10	9	1.15	10.60	11.75	10.82	91.2	
Q3439-06	TP-10-EPH	10	1.19	10.17	11.36	10.44	91.0	
Q3439-07	TP-10-VOC	11	1.18	10.31	11.49	10.37	89.1	
Q3440-01	JB-2	12	1.13	11.32	12.45	10.38	81.7	
Q3440-02	JB-2-EPH	13	1.17	10.42	11.59	9.79	82.7	
Q3440-03	JB-2-VOC	14	1.13	8.36	9.49	8.14	83.9	
Q3445-01	EO-01-102325	16	1.14	10.46	11.6	10.74	91.8	
Q3445-02	EO-01-102325-E2	17	1.16	10.40	11.56	10.57	90.5	
Q3446-01	SB-14	18	1.11	10.00	11.11	10.82	97.1	
Q3446-02	SB-15	19	1.14	10.54	11.68	11.64	99.6	
Q3446-03	SB-16	20	1.19	10.31	11.5	11.44	99.4	
Q3446-04	SB-17	21	1.19	10.13	11.32	11.27	99.5	
Q3446-05	SB-18	22	1.19	10.53	11.72	11.67	99.5	
Q3446-06	SB-19	23	1.15	10.44	11.59	11.32	97.4	
Q3446-07	SB-20	24	1.16	10.50	11.66	11.6	99.4	
Q3446-08	SB-21	25	1.14	10.08	11.22	10.9	96.8	
Q3447-01	P001-TW1-251023-01	26	1.00	1.00	2.00	2.00	100.0	TAR SAMPLE
Q3447-03	P001-TW2-251023-01	27	1.19	10.43	11.62	10.75	91.7	
Q3447-05	P001-TW3-251023-01	28	1.18	10.70	11.88	11.64	97.8	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/24/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 10/23/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:15
Out Date: 10/24/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC:LB137624

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q3447-07	P001-TW4-251023-01	29	1.00	1.00	2.00	2.00	100.0	TAR SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

10/23/2025

WorkList Name : %1-102325

WorkList ID : 192629

Department : Wet-Chemistry

Date : 10-23-2025 08:26:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3399-03	COMP-ROLLOFF	Solid	Percent Solids	Cool 4 deg C	ICES02	D41	10/20/2025	Chemtech -SO
Q3437-03	SS-15R(2.0-2.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D31	10/22/2025	Chemtech -SO
Q3437-05	SS-26(0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D31	10/22/2025	Chemtech -SO
Q3437-11	SS-27(0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D31	10/22/2025	Chemtech -SO
Q3437-14	SS-25(0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D31	10/22/2025	Chemtech -SO
Q3437-16	SS-23(0-0.5)	Solid	Percent Solids	Cool 4 deg C	MATR02	D31	10/22/2025	Chemtech -SO
Q3439-01	TP-9	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/22/2025	Chemtech -SO
Q3439-02	TP-9-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/22/2025	Chemtech -SO
Q3439-03	TP-9-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/22/2025	Chemtech -SO
Q3439-05	TP-10	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/22/2025	Chemtech -SO
Q3439-06	TP-10-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/23/2025	Chemtech -SO
Q3439-07	TP-10-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/23/2025	Chemtech -SO
Q3440-01	JB-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/23/2025	Chemtech -SO
Q3440-02	JB-2-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/22/2025	Chemtech -SO
Q3440-03	JB-2-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D41	10/22/2025	Chemtech -SO
Q3445-01	EO-01-102325	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	10/23/2025	Chemtech -SO
Q3445-02	EO-01-102325-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	10/23/2025	Chemtech -SO
Q3446-01	SB-14	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3446-02	SB-15	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3446-03	SB-16	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3446-04	SB-17	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO

Date/Time 10/23/25 15:00
 Raw Sample Received by: JH (wpc)
 Raw Sample Relinquished by: JH (wpc)

WORKLIST(Hardcopy Internal Chain)

10/23/25

WorkList Name : %1-102325

WorkList ID : 192629

Department : Wet-Chemistry

Date : 10-23-2025 08:26:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3446-05	SB-18	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3446-06	SB-19	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3446-07	SB-20	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3446-08	SB-21	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	10/23/2025	Chemtech -SO
Q3447-01	P001-TW1-251023-01	Solid	Percent Solids	Cool 4 deg C	ROYF02	D41	10/23/2025	Chemtech -SO
Q3447-03	P001-TW2-251023-01	Solid	Percent Solids	Cool 4 deg C	ROYF02	D41	10/23/2025	Chemtech -SO
Q3447-05	P001-TW3-251023-01	Solid	Percent Solids	Cool 4 deg C	ROYF02	D41	10/23/2025	Chemtech -SO
Q3447-07	P001-TW4-251023-01	Solid	Percent Solids	Cool 4 deg C	ROYF02	D41	10/23/2025	Chemtech -SO

Date/Time 10/23/25 15:00
 Raw Sample Received by: JG WWC
 Raw Sample Relinquished by: JG WWC

Date/Time 10/23/25 17:30
 Raw Sample Received by: JG WWC
 Raw Sample Relinquished by: JG WWC

SOP ID: M3541-ASE Extraction-15

Clean Up SOP #: Florisil

Matrix : Solid

Welgh By: EH **Extraction By:** EH

Balance check: EH **Filter By:** RS

Balance ID: EX-SC-2 **pH Meter ID:** N/A

pH Strip Lot#: N/A **Hood ID:** 3,7

Extraction Start Date : 10/22/2025

Extraction Start Time : 09:20

Extraction End Date : 10/22/2025

Extraction End Time : 12:45

Concentration By: EH

Supervisor By : RUPESH

Extraction Method: ☐ Seperatory Funnel ☐ Continous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standared Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	500 PPB	PP24766
Surrogate	1.0ML	200 PPB	PP24663
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2643
Baked Na2SO4	N/A	EP2652
Sand	N/A	E3951
Hexane	N/A	E3978
Florisil	N/A	E3976
9:1 Hexane:Acetone Mixture	N/A	EP2644
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40ML Vial Lot # 03-40BTS721.

KD Bath ID: N/A **Envap ID:** NE VAP-02


KD Bath Temperature: N/A **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/22/25	RS (Ext Lab)	R. Post/PCB Lab
12:50	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-15

Concentration Date: 10/22/2025

Sample ID	Client Sample ID	Test	g/ mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170205BL	PBLK205	Pesticide-TCL	30.01	N/A	Evelyn	RUPESH	10			U1-1
PB170205BS	PLCS205	Pesticide-TCL	30.03	N/A	Evelyn	RUPESH	10			2
Q3399-01	COMP-ROLLOFF	Pesticide-TCL	30.06	N/A	Evelyn	RUPESH	10	C		3
Q3399-01MS	COMP-ROLLOFFMS	Pesticide-TCL	30.09	N/A	Evelyn	RUPESH	10	C		4
Q3399-01MS D	COMP-ROLLOFFMSD	Pesticide-TCL	30.08	N/A	Evelyn	RUPESH	10	C		5
Q3404-01	MOO-25-0285-0288	Pesticide-TCL	30.04	N/A	Evelyn	RUPESH	10	E		6
Q3404-04	MOO-25-0292	Pesticide-TCL	30.07	N/A	Evelyn	RUPESH	10	E	Small Particles	U2-1
Q3416-01	NB-08-10212025	Pesticide-TCL	30.10	N/A	Evelyn	RUPESH	10	E		2



* Extracts relinquished on the same date as received.

5022020

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3416 WorkList ID : 192608 Department : Extraction Date : 10-22-2025 09:03:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3399-01	COMP-ROLLOFF	Solid	Pesticide-TCL	Cool 4 deg C	ICES02	D41	10/20/2025	8081B
Q3404-01	MOO-25-0285-0288	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	10/20/2025	8081B
Q3404-04	MOO-25-0292	Solid	Pesticide-TCL	Cool 4 deg C	PSEG03	D41	10/20/2025	8081B
Q3416-01	NB-08-10212025	Solid	Pesticide-TCL	Cool 4 deg C	PSEG05	D31	10/21/2025	8081B

Date/Time 10/22/25 9:15
Raw Sample Received by: RS (Ext-Lab)
Raw Sample Relinquished by: CSW

Date/Time 10/22/25 9:50
Raw Sample Received by: CSW
Raw Sample Relinquished by: RS (Ext-Lab)

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19
- 20

Prep Standard - Chemical Standard Summary

Order ID : Q3399
Test : Pesticide-TCL
Prepbatch ID : PB170205,
Sequence ID/Qc Batch ID: PD102225,

Standard ID :
EP2643,EP2652,PP24651,PP24663,PP24738,PP24739,PP24740,PP24741,PP24742,PP24744,PP24746,PP24747,PP24748,PP24749,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764,PP24766,PP24942,

Chemical ID :
E3875,E3941,E3944,E3951,E3955,E3956,E3971,E3972,E3976,E3978,P12604,P12610,P13038,P13041,P13196,P13246,P13774,P13780,P13786,P13788,P13862,P24754,P9053,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
230	1:1ACETONE/HEXANE	EP2643	09/16/2025	03/16/2026	Evelyn Huang	None	None	Riteshkumar Patel
09/16/2025								

FROM 4000.00000ml of E3971 + 4000.00000ml of E3972 = Final Quantity: 8000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2652	10/10/2025	01/28/2026	Evelyn Huang	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel
10/10/2025								

FROM 4000.00000gram of E3875 = Final Quantity: 4000.000 gram

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
4027	Pesticide resolution Check Mixture 8081	PP24651	06/16/2025	12/11/2025	Abdul Mirza	None	None	Yogesh Patel
								07/22/2025

FROM 1.00000ml of P13246 + 99.00000ml of E3941 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24663	06/24/2025	12/24/2025	Abdul Mirza	None	None	Yogesh Patel
								07/21/2025

FROM 1.00000ml of P13786 + 999.00000ml of E3944 = Final Quantity: 1000.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3629	20 PPM PEST stock Solution 1st source(RESTEK)	PP24738	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 1.00000ml of P13038 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1472	20 PPM Pest Stock Solution 2nd Source	PP24739	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 1.00000ml of P13041 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1273	20 PPM Mirex Stock (Primary Source)	PP24740	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P9053 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3663	20 PPM MIREX Stock STD (Secondary source)	PP24741	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13196 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24742	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 1.00000ml of P13788 + 9.00000ml of E3956 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3630	100/100 PPB PEST Working std.1st Source(RESTEK)	PP24744	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 98.50000ml of E3956 + 0.50000ml of PP24738 + 0.50000ml of PP24740 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
386	1000/100 PPB Chlordane STD (Restek)	PP24746	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P12604 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3746	1000/100 ppb Chlordane STD-RESTEK 2ND SOURCE	PP24747	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P12610 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
383	1000/100 PPB Toxaphene STD (Restek)	PP24748	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P13774 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3669	1000/100 PPB TOXAPHENE STD 2nd source (RESTEK)	PP24749	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.10000ml of P13862 + 99.40000ml of E3956 + 0.50000ml of PP24742 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3631	75 PPB ICAL PEST STD(RESTEK)	PP24750	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.25000ml of E3956 + 0.75000ml of PP24744 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3632	50 PPB ICAL PEST STD(RESTEK)	PP24751	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of E3956 + 0.50000ml of PP24744 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3633	25 PPB ICAL PEST STD(RESTEK)	PP24752	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.75000ml of E3956 + 0.25000ml of PP24744 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3634	5 PPB ICAL PEST STD(RESTEK)	PP24753	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.90000ml of E3956 + 0.10000ml of PP24751 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
528	CHLOR 750 PPB STD	PP24755	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.25000ml of E3956 + 0.75000ml of PP24746 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
529	CHLOR 500 PPB STD	PP24756	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of E3956 + 0.50000ml of PP24746 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
530	CHLOR 250 PPB STD	PP24757	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.75000ml of E3956 + 0.25000ml of PP24746 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3408	CHLOR 50 PPB STD	PP24758	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.90000ml of E3956 + 0.10000ml of PP24756 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
532	CHLOR 500 PPB ICV STD	PP24759	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.50000ml of E3956 + 0.50000ml of PP24747 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
533	TOX 750 PPB STD	PP24760	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								07/24/2025

FROM 0.25000ml of E3956 + 0.75000ml of PP24748 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
534	TOX 500 PPB STD	PP24761	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of E3956 + 0.50000ml of PP24748 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
535	TOX 250 PPB STD	PP24762	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.75000ml of E3956 + 0.25000ml of PP24748 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2217	TOX 100 PPB STD	PP24763	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.90000ml of E3956 + 0.10000ml of PP24748 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3670	TOX 500 PPB ICV std (RESTEK)	PP24764	07/21/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
07/24/2025								

FROM 0.50000ml of PP24749 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
79	500 PPB Pesticide Spike Solution	PP24766	07/28/2025	01/21/2026	Abdul Mirza	None	None	Yogesh Patel
								08/19/2025

FROM 95.00000ml of E3955 + 2.50000ml of PP24739 + 2.50000ml of PP24741 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
758	PEM Mix w/Surr	PP24942	09/24/2025	03/24/2026	Abdul Mirza	None	None	Yogesh Patel
								10/10/2025

FROM 1.00000ml of P13780 + 99.00000ml of E3971 = Final Quantity: 100.000 ml

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	07/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	12/11/2025	06/11/2025 / Rajesh	06/04/2025 / Rajesh	E3941

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	06/20/2025 / RUPESH	05/14/2025 / RUPESH	E3944

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	25A2756718	12/31/2028	07/09/2025 / RUPESH	04/28/2020 / RUPESH	E3951

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	04/18/2027	07/16/2025 / RUPESH	07/16/2025 / RUPESH	E3955

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	07/16/2025 / RUPESH	07/16/2025 / RUPESH	E3956

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	09/15/2025 / Evelyn	09/04/2025 / Riteshkumar	E3971

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	09/16/2025 / Evelyn	09/04/2025 / Riteshkumar	E3972

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agela Technologies Inc.	FS0006 / Cleanert Florisil cartridge	M07456	04/03/2026	10/03/2025 / Evelyn	10/03/2025 / Evelyn	E3976

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362006	04/13/2026	10/13/2025 / RUPESH	10/08/2025 / Riteshkumar	E3978

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0197993	01/21/2026	07/21/2025 / Abdul	07/03/2023 / Abdul	P12604

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32021 / Chlordane Std.	A0197993	01/21/2026	07/21/2025 / Abdul	07/03/2023 / Abdul	P12610

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0200423	01/21/2026	07/21/2025 / Abdul	12/26/2023 / Abdul	P13038

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32291 / Pesticide Mix, CLP method, organochlorine Std AB#1, 200ug/mL, hexane/toluene, 1mL/ampul	A0199099	07/21/2026	07/21/2025 / Abdul	12/26/2023 / Abdul	P13041

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	042022	01/21/2026	07/21/2025 / Abdul	01/17/2024 / Abdul	P13196

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	19161 / 8081 pesticide resolution check mixture	013124	12/17/2025	06/17/2025 / Abdul	02/09/2024 / Abdul	P13246

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0203038	01/21/2026	07/21/2025 / Abdul	05/03/2024 / Ankita	P13774

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32074 / Pesticide Performance Evaluation Mix w/Surrogate	A0213999	03/24/2026	09/24/2025 / Abdul	11/19/2024 / Ankita	P13780

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	12/24/2025	06/24/2025 / Abdul	11/19/2024 / Ankita	P13786

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0214495	01/21/2026	07/21/2025 / Abdul	11/19/2024 / Ankita	P13788

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32005 / Toxaphene Standard	A0210240	01/21/2026	07/21/2025 / Abdul	12/09/2024 / Abdul	P13862

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	79136 / Mirex, 1000 ug/ml	112018	01/21/2026	07/21/2025 / Abdul	11/01/2019 / Stephen	P9053



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT - 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Harout Sahagian

Recd. by RI on 6/11/25

E3941

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
 If there are any questions with this certificate, please call at (800) 227-6701.
 *Based on suggested storage condition.

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date: 2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3944

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20

Material	BDH9274-2.5KG
Material Description	BDH SAND STDD OTTAWA W+I 2.5KG
Grade	NOT APPLICABLE
Batch	25A2756718
Reassay Date	12/31/2028
CAS Number	14808-60-7
Molecular Formula	SiO ₂
Molecular Mass	60.09
Date of Manufacture	12/05/2024
Storage	Room Temperature

Characteristics	Specifications	Measured Values
Appearance	Beige granules.	Beige granules.
Moisture	<= 0.1 %	0.1 %
Particle Size 30-40 mesh	>= 80 %	99 %
CUSTOMER PART # BDH9274-2.5KG		

Received on 1/12/25.

E3951

Internal ID #: 793

Signature	Additional Information
We certify that this batch conforms to the specifications listed above. This document has been electronically produced and is valid without a signature. Leona Edwardson, Quality Control Sr. Manager - Solon VWR Chemicals, LLC. 28600 Fountain Parkway, Solon OH 44139 USA	Analysis may have been rounded to significant digits in specification limits Product meets analytical specifications of the grades listed.

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis



Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date: 2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/16/25

E 3955

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Received on 7/16/25

E3956

Jamie Croak
Director Quality Operations, Bioscience Production

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 25C0362005
Manufactured Date: 2025-01-29
Expiration Date: 2026-04-30
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3971

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 24H1462005
Manufactured Date: 2024-05-24
Expiration Date: 2027-05-24
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3972

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700



CAT# FS0006



MFG#: G02147

LOT#: M07456

1g/6ml 30/pkg

Cleanert Florisil



Made in China

固相萃取产品

Agela Technologies

E 3976



n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantorsTM**



Material No.: 9262-03

Batch No.: 25C0362006

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	4
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5\%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95\%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E 3978



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 Lot No.: A0199099

Description : Organochlorine Pesticide Mix AB #1

Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL Pkg Amt: > 1 mL

Expiration Date : June 30, 2027 Storage: 10°C or colder

Ship: Ambient

P130397 5
↓
P13043 1
✓
12-26-2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.0 µg/mL	+/- 8.9732
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	200.1 µg/mL	+/- 8.9762
3	beta-BHC	319-85-7	BCCC6425	99%	200.3 µg/mL	+/- 8.9844
4	delta-BHC	319-86-8	14450800	98%	200.0 µg/mL	+/- 8.9740
5	Heptachlor	76-44-8	813251	99%	200.1 µg/mL	+/- 8.9754
6	Aldrin	309-00-2	14389400	98%	200.0 µg/mL	+/- 8.9718
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.1 µg/mL	+/- 8.9754
8	trans-Chlordane	5103-74-2	32943	98%	199.9 µg/mL	+/- 8.9696
9	cis-Chlordane	5103-71-9	31766	98%	200.1 µg/mL	+/- 8.9762
10	Endosulfan I	959-98-8	BCCF4060	99%	200.1 µg/mL	+/- 8.9754
11	4,4'-DDE	72-55-9	GHYQG	99%	200.1 µg/mL	+/- 8.9777
12	Dieldrin	60-57-1	11129900	98%	200.0 µg/mL	+/- 8.9718
13	Endrin	72-20-8	14123200	98%	199.9 µg/mL	+/- 8.9696
14	4,4'-DDD	72-54-8	HAN02	99%	200.1 µg/mL	+/- 8.9777
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	200.0 µg/mL	+/- 8.9718

17	Endrin aldehyde	7421-93-4	30720	98%	200.1 µg/mL	+/- 8.9784
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.0 µg/mL	+/- 8.9732
19	Methoxychlor	72-43-5	13668200	99%	200.1 µg/mL	+/- 8.9777
20	Endrin ketone	53494-70-5	1-ABS-16-7	98%	200.0 µg/mL	+/- 8.9740

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P 13039
 ↓
 P 13043
 1
 5
 12/26/23

Quality Confirmation Test

Column:
 30m x .25mm x .2µm
 Rtx-CLP II (cat.# 11323)

Carrier Gas:
 helium-constant pressure 20 psi.

Temp. Program:
 150°C to 300°C
 @ 4°C/min. (hold 5 min.)


Inj. Temp:
 200°C

Det. Temp:
 300°C

Det. Type:
 ECD

Split Vent:
 Split ratio 50:1

Inj. Vol
 1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Josh McCloskey
 Josh McCloskey - Operations Technician I

Date Mixed: 19-Jun-2023 **Balance Serial #** 1128360905

Jennifer Pollino
 Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 23-Jun-2023

Manufactured under Restek's ISO 9001:2015
 Registered Quality System
 Certificate #FM 80397

**CERTIFIED WEIGHT REPORT**

Part Number:	<u>79136</u>
Lot Number:	<u>042022</u>
Description:	<u>Mirex</u>

Solvent(s):	Lot#
Acetone	81025

Formulated By: Prashant Chaughan DATE 042022

Expiration Date: 04/20/27

Recommended Storage: Refrigerate (4 °C)

Nominal Concentration ($\mu\text{g/mL}$):

NIST Test ID#:

Weight(s) shown below were combined and diluted to (mL):

5E-05 **Balance Uncertainty**
0.006 **Flask Uncertainty**

Formulated By:	 Prashant Chauhan	042022
Reviewed By:	 Pedro L. Ramias	042022

Reviewed By:

Pedro L. Rentas


042022
DATE

SDS Information

Compound	Lot Number	Nominal Conc (ug/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(ug/mL) (+/-) (ug/mL)	(Solvent Safety Info. On Attached pg.)	
								CAS#	LDSO
	RN#:							OSHA PEL (TWA)	

[illegible]

Method GC/MSD-1.M: Column: SPB-608 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 150°C (4min.), Temp 2 = 290°C (13.5 min.), Rate = 8°C/min., Injector B = 200°C, Detector B = 290°C. Split Ratio = 100:1, Scan Rate = 2. Analysis performed by Candice Warren.



- The certified values are concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified ($\pm$) 0.5% of the stated value, unless otherwise stated.

• All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.

- **Uncertainty Reference:** Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

$P_{13,195}$
 $\downarrow$
 $P_{13,199}$
 $\downarrow$
 (5)
 $\downarrow$

01/17/2024
HARRIS
D
3/3/19
5/5/19
5



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0197993
Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Ship: Ambient

P 12603
↓
P 12605
[3]
Kauf
7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	----%	1,005.0 µg/mL	+/- 55.7700

Solvent: Hexane
CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Morgan Craighead - Mix Technician


Date Mixed: 11-May-2023

Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-May-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 12603 } (3)
↓
P 12605

7/3/2023



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *chromatographic plus*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32021 **Lot No.:** A0197993
Description : Chlordane Standard
Chlordane Standard 1000µg/mL, Hexane, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Ship: Ambient

Handwritten: P 12606, P 12610, 5 Five, 7/3/2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Chlordane 10% trans-Chlordane; 9% cis-Chlordane; 81% other isomers	57-74-9	978545	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Tech Tips:

CAS #57-74-9 nomenclature is based on EPA method 8081B.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. { hold 10 min.}

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 11-May-2023

Balance Serial # 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-May-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32291 **Lot No.:** A0200423

Description : Organochlorine Pesticide Mix AB #1

Organochlorine Pesticide Mix AB #1 200µg/mL, Hexane/Toluene(50:50), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 10°C or colder

Ship: Ambient

P 13034
↓
P 13038
12.26.2023

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	alpha-BHC	319-84-6	14434500	99%	200.5 µg/mL	+/- 8.9956
2	gamma-BHC (Lindane)	58-89-9	14184400	98%	199.9 µg/mL	+/- 8.9696
3	beta-BHC	319-85-7	BCCC6425	99%	200.0 µg/mL	+/- 8.9732
4	delta-BHC	319-86-8	14450800	98%	199.9 µg/mL	+/- 8.9696
5	Heptachlor	76-44-8	813251	99%	202.0 µg/mL	+/- 9.0629
6	Aldrin	309-00-2	14389400	98%	200.9 µg/mL	+/- 9.0136
7	Heptachlor epoxide (isomer B)	1024-57-3	14448800	99%	200.0 µg/mL	+/- 8.9732
8	trans-Chlordane	5103-74-2	34616	99%	200.5 µg/mL	+/- 8.9956
9	cis-Chlordane	5103-71-9	31766	98%	201.4 µg/mL	+/- 9.0356
10	Endosulfan I	959-98-8	BCCF4060	99%	200.0 µg/mL	+/- 8.9732
11	4,4'-DDE	72-55-9	GHYQG	99%	201.5 µg/mL	+/- 9.0405
12	Dieldrin	60-57-1	14515000	98%	199.9 µg/mL	+/- 8.9696
13	Endrin	72-20-8	14485300	98%	200.4 µg/mL	+/- 8.9916
14	4,4'-DDD	72-54-8	HAN02	99%	200.5 µg/mL	+/- 8.9956
15	Endosulfan II	33213-65-9	14374700	99%	200.0 µg/mL	+/- 8.9732
16	4,4'-DDT	50-29-3	230410JLMA	98%	201.9 µg/mL	+/- 9.0575

17	Endrin aldehyde	7421-93-4	30720	98%	201.4 µg/mL	+/- 9.0356
18	Endosulfan sulfate	1031-07-8	BCCH9010	99%	200.5 µg/mL	+/- 8.9956
19	Methoxychlor	72-43-5	14563200	98%	200.9 µg/mL	+/- 9.0136
20	Endrin ketone	53494-70-5	14537700	98%	199.9 µg/mL	+/- 8.9696

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane/Toluene (50:50)
CAS # 110-54-3/108-88-3
Purity 99%

P13034
P13038
1
5
12/26/2023

Quality Confirmation Test

Column:
30m x .25mm x .2µm
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 50:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 31-Jul-2023

Balance Serial # B442140311

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 03-Aug-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT

Part Number:
Lot Number:
Description:

19161
013124
CLP Pesticides & PCBs Resolution Check Standard

Expiration Date:
Recommended Storage:
Nominal Concentration (µg/mL):

9 components
013129
Refrigerate (4 °C)
Varied
6UTB

Volume(s) shown below were combined and diluted to (mL):

5E-05
Balance Uncertainty
Pipet Uncertainty

<i>Lawrence Barry</i>	
Formulated By:	Lawrence Barry
013124	DATE
<i>Pedro L. Rentes</i>	
Reviewed By:	Pedro L. Rentes
013124	DATE

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty (mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty (±) µg/mL	CAS#	SDS Information (Solvent Safety Info. On Attached pg.)	LD50

1. trans-Chlordane	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	5103-74-2	0.5mg/m3 (skin)	or-rat 500mg/kg
2. Endosulfan I	19361	013124	0.010	1.00	0.004	101.3	1.0	0.02	959-98-8	0.1mg/m3 (skin)	or-rat 18mg/kg
3. 4,4'-DDE	19361	013124	0.010	1.00	0.004	201.6	2.0	0.03	72-55-9	N/A	or-rat 880mg/kg
4. Dieldrin	19361	013124	0.010	1.00	0.004	202.8	2.0	0.03	60-57-1	0.25mg/m3 (skin)	or-rat 38300µg/kg
5. Endosulfan sulfate	19361	013124	0.010	1.00	0.004	204.2	2.0	0.03	1031-07-8	N/A	or-rat 18mg/kg
6. Endrin ketone	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	53494-70-5	N/A	N/A
7. 4,4-Methoxychlor	19361	013124	0.010	1.00	0.004	1000.7	10.0	0.09	72-43-5	10mg/m3	or-rat 6000mg/kg
8. 2,4,5,6-Tetrachloro-m-xylene	19361	013124	0.010	1.00	0.004	202.6	2.0	0.03	877-09-8	N/A	N/A
9. Decachlorobiphenyl (209)	19361	013124	0.010	1.00	0.004	202.0	2.0	0.03	2051-24-3	N/A	N/A

1913243
1913244
1913245
1913246
1913247
1913248
1913249
1913250
1913251
1913252
1913253
1913254
1913255
1913256
1913257
1913258
1913259
1913260
1913261
1913262
1913263
1913264
1913265
1913266
1913267
1913268
1913269
1913270
1913271
1913272
1913273
1913274
1913275
1913276
1913277
1913278
1913279
1913280
1913281
1913282
1913283
1913284
1913285
1913286
1913287
1913288
1913289
1913290
1913291
1913292
1913293
1913294
1913295
1913296
1913297
1913298
1913299
1913300
1913301
1913302
1913303
1913304
1913305
1913306
1913307
1913308
1913309
1913310
1913311
1913312
1913313
1913314
1913315
1913316
1913317
1913318
1913319
1913320
1913321
1913322
1913323
1913324
1913325
1913326
1913327
1913328
1913329
1913330
1913331
1913332
1913333
1913334
1913335
1913336
1913337
1913338
1913339
1913340
1913341
1913342
1913343
1913344
1913345
1913346
1913347
1913348
1913349
1913350
1913351
1913352
1913353
1913354
1913355
1913356
1913357
1913358
1913359
1913360
1913361
1913362
1913363
1913364
1913365
1913366
1913367
1913368
1913369
1913370
1913371
1913372
1913373
1913374
1913375
1913376
1913377
1913378
1913379
1913380
1913381
1913382
1913383
1913384
1913385
1913386
1913387
1913388
1913389
1913390
1913391
1913392
1913393
1913394
1913395
1913396
1913397
1913398
1913399
1913400
1913401
1913402
1913403
1913404
1913405
1913406
1913407
1913408
1913409
1913410
1913411
1913412
1913413
1913414
1913415
1913416
1913417
1913418
1913419
1913420
1913421
1913422
1913423
1913424
1913425
1913426
1913427
1913428
1913429
1913430
1913431
1913432
1913433
1913434
1913435
1913436
1913437
1913438
1913439
1913440
1913441
1913442
1913443
1913444
1913445
1913446
1913447
1913448
1913449
1913450
1913451
1913452
1913453
1913454
1913455
1913456
1913457
1913458
1913459
1913460
1913461
1913462
1913463
1913464
1913465
1913466
1913467
1913468
1913469
1913470
1913471
1913472
1913473
1913474
1913475
1913476
1913477
1913478
1913479
1913480
1913481
1913482
1913483
1913484
1913485
1913486
1913487
1913488
1913489
1913490
1913491
1913492
1913493
1913494
1913495
1913496
1913497
1913498
1913499
1913500
1913501
1913502
1913503
1913504
1913505
1913506
1913507
1913508
1913509
1913510
1913511
1913512
1913513
1913514
1913515
1913516
1913517
1913518
1913519
1913520
1913521
1913522
1913523
1913524
1913525
1913526
1913527
1913528
1913529
1913530
1913531
1913532
1913533
1913534
1913535
1913536
1913537
1913538
1913539
1913540
1913541
1913542
1913543
1913544
1913545
1913546
1913547
1913548
1913549
1913550
1913551
1913552
1913553
1913554
1913555
1913556
1913557
1913558
1913559
1913560
1913561
1913562
1913563
1913564
1913565
1913566
1913567
1913568
1913569
1913570
1913571
1913572
1913573
1913574
1913575
1913576
1913577
1913578
1913579
1913580
1913581
1913582
1913583
1913584
1913585
1913586
1913587
1913588
1913589
1913590
1913591
1913592
1913593
1913594
1913595
1913596
1913597
1913598
1913599
1913600
1913601
1913602
1913603
1913604
1913605
1913606
1913607
1913608
1913609
1913610
1913611
1913612
1913613
1913614
1913615
1913616
1913617
1913618
1913619
1913620
1913621
1913622
1913623
1913624
1913625
1913626
1913627
1913628
1913629
1913630
1913631
1913632
1913633
1913634
1913635
1913636
1913637
1913638
1913639
1913640
1913641
1913642
1913643
1913644
1913645
1913646
1913647
1913648
1913649
1913650
1913651
1913652
1913653
1913654
1913655
1913656
1913657
1913658
1913659
1913660
1913661
1913662
1913663
1913664
1913665
1913666
1913667
1913668
1913669
1913670
1913671
1913672
1913673
1913674
1913675
1913676
1913677
1913678
1913679
1913680
1913681
1913682
1913683
1913684
1913685
1913686
1913687
1913688
1913689
1913690
1913691
1913692
1913693
1913694
1913695
1913696
1913697
1913698
1913699
1913700
1913701
1913702
1913703
1913704
1913705
1913706
1913707
1913708
1913709
1913710
1913711
1913712
1913713
1913714
1913715
1913716
1913717
1913718
1913719
1913720
1913721
1913722
1913723
1913724
1913725
1913726
1913727
1913728
1913729
1913730
1913731
1913732
1913733
1913734
1913735
1913736
1913737
1913738
1913739
1913740
1913741
1913742
1913743
1913744
1913745
1913746
1913747
1913748
1913749
1913750
1913751
1913752
1913753
1913754
1913755
1913756
1913757
1913758
1913759
1913760
1913761
1913762
1913763
1913764
1913765
1913766
1913767
1913768
1913769
1913770
1913771
1913772
1913773
1913774
1913775
1913776
1913777
1913778
1913779
1913780
1913781
1913782
1913783
1913784
1913785
1913786
1913787
1913788
1913789
1913790
1913791
1913792
1913793
1913794
1913795
1913796
1913797
1913798
1913799
1913800
1913801
1913802
1913803
1913804
1913805
1913806
1913807
1913808
1913809
1913810
1913811
1913812
1913813
1913814
1913815
1913816
1913817
1913818
1913819
1913820
1913821
1913822
1913823
1913824
1913825
1913826
1913827
1913828
1913829
1913830
1913831
1913832
1913833
1913834
1913835
1913836
1913837
1913838
1913839
1913840
1913841
1913842
1913843
1913844
1913845
1913846
1913847
1913848
1913849
1913850
1913851
1913852
1913853
1913854
1913855
1913856
1913857
1913858
1913859
1913860
1913861
1913862
1913863
1913864
1913865
1913866
1913867
1913868
1913869
1913870
1913871
1913872
1913873
1913874
1913875
1913876
1913877
1913878
1913879
1913880
1913881
1913882
1913883
1913884
1913885
1913886
1913887
1913888
1913889
1913890
1913891
1913892
1913893
1913894
1913895
1913896
1913897
1913898
1913899
1913900
1913901
1913902
1913903
1913904
1913905
1913906
1913907
1913908
1913909
1913910
1913911
1913912
1913913
1913914
1913915
1913916
1913917
1913918
1913919
1913920
1913921
1913922
1913923
1913924
1913925
1913926
1913927
1913928
1913929
1913930
1913931
1913932
1913933
1913934
1913935
1913936
1913937
1913938
1913939
1913940
1913941
1913942
1913943
1913944
1913945
1913946
1913947
1913948
1913949
1913950
1913951
1913952
1913953
1913954
1913955
1913956
1913957
1913958
1913959
1913960
1913961
1913962
1913963
1913964
1913965
1913966
1913967
1913968
1913969
1913970
1913971
1913972
1913973
1913974
1913975
1913976
1913977
1913978
1913979
1913980
1913981
1913982
1913983
1913984
1913985
1913986
1913987
1913988
1913989
1913990
1913991
1913992
1913993
1913994
1913995
1913996
1913997
1913998
1913999
1914000
1914001
1914002
1914003
1914004
1914005
1914006
1914007
1914008
1914009
1914010
1914011
1914012
1914013
1914014
1914015
1914016
1914017
1914018
1914019
1914020
1914021
1914022
1914023
1914024
1914025
1914026
1914027
1914028
1914029
1914030
1914031
1914032
1914033
1914034
1914035
1914036
1914037
1914038
1914039
1914040
1914041
1914042
1914043
1914044
1914045
1914046
1914047
1914048
1914049
1914050
1914051
1914052
1914053
1914054
1914055
1914056
1914057
1914058
1914059
1914060
1914061
1914062
1914063
1914064
1914065
1914066
1914067
1914068
1914069
1914070
1914071
1914072
1914073
1914074
1914075
1914076
1914077
1914078
1914079
1914080
1914081
1914082
1914083
1914084
1914085
1914086
1914087
1914088
1914089
1914090
1914091
1914092
1914093
1914094
1914095
1914096
1914097
1914098
1914099
1914100
1914101
1914102
1914103
1914104
1914105
1914106
1914107
1914108
1914109
1914110
1914111
1914112
1914113
1914114
1914115
1914116
1914117
1914118
1914119
1914120
1914121
1914122
1914123
1914124
1914125
1914126
1914127
1914128
1914129
1914130
1914131
1914132
1914133
1914134
1914135
1914136
1914137
1914138
1914139
1914140
1914141
1914142
1914143
1914144
1914145
1914146
1914147
1914148
1914149
1914150
1914151
1914152
1914153
1914154
1914155
1914156
1914157
1914158
1914159
1914160
1914161
1914162
1914163
1914164
1914165
1914166
1914167
1914168
1914169
1914170
1914171
1914172
1914173
1914174
1914175
1914176
1914177
1914178
1914179
1914180
1914181
1914182
1914183
1914184
1914185
1914186
1914187
1914188
1914189
1914190
1914191
1914192
1914193
1914194
1914195
1914196
1914197
1914198
1914199
1914200
1914201
1914202
1914203
1914204
1914205
1914206
1914207
1914208
1914209
1914210
1914211
1914212
1914213
1914214
1914215
1914216
1914217
1914218
1914219
1914220
1914221
1914222
1914223
1914224
1914225
1914226
1914227
1914228
1914229
1914230
1914231
1914232
1914233
1914234
1914235
1914236
1914237
1914238
1914239
1914240
1914241
1914242
1914243
1914244
1914245
1914246
1914247
1914248
1914249
1914250
1914251
1914252
1914253
1914254
1914255
1914256
1914257
1914258
1914259
1914260
1914261
1914262
1914263
1914264
1914265
1914266
1914267
1914268
1914269
1914270
1914271
1914272
1914273
1914274
1914275
1914276
1914277
1914278
1914279
1914280
1914281
1914282
1914283
1914284
1914285
1914286
1914287
1914288
1914289
1914290
1914291
1914292
1914293
1914294
1914295
1914296
1914297
1914298
1914299
1914300
1914301
1914302
1914303
1914304
1914305
1914306
1914307
1914308
1914309
1914310
1914311
1914312
1914313
1914314
1914315
1914316
1914317
1914318
1914319
1914320
1914321
1914322
1914323
1914324
1914325
1914326
1914327
1914328
1914329
1914330
1914331
1914332
1914333
1914334
1914335
1914336
1914337
1914338
1914339
1914340
1914341
1914342
1914343
1914344
1914345
1914346
1914347
1914348
1914349
1914350
1914351
1914352
1914353
1914354
1914355
1914356
1914357
1914358
1914359
191436





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis *chromatographic plus*



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.
This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0203038

Description : Toxaphene Standard

Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL


Expiration Date : January 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES						
Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	----%	1,009.0 µg/mL	+/- 55.9920


* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%



Quality Confirmation Test

Column:
30m x .25mm x .2um
Flex-CLP II (cat.# 11323)
Carrier Gas:
Helium-constant pressure 20 psi.
Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)
Inj. Temp:
250°C
Det. Temp:
300°C
Det. Type:
ECD
Split Vent:
300 ml/min.
Inj. Vol
0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

J.P.

Dakota Parson - Operations Technician I

Date Mixed: 10-Oct-2023 Balance Serial # 1128353505

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397

Handwritten notes:
A bracket groups the first two peaks, with an arrow pointing to the peak at ~4.5 minutes.
P13773
P19779
A7
11/19/24
A signature is written over the date.





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32074

Lot No.: A0213999

Description:

Pesticide Performance Eval Mix w/Surrogate

Performance Evaluation Std. 3/90 SOW w/surrogates 1-25µg/mL,
Hexane, 1mL/ampul

Container Size:

2 mL

Pkg Amt: > 1 mL

Expiration Date:

July 31, 2028

Storage: 10°C or colder

Handling:

Contains PCBs - sonicate prior to
use.

Ship: Ambient

PI8780 AJ
11/19/24

PI8784

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	2.0 µg/mL	+/- 0.1506
2	alpha-BHC	319-84-6	15279000	99%	1.0 µg/mL	+/- 0.0749
3	gamma-BHC (Lindane)	58-89-9	14931300	97%	1.0 µg/mL	+/- 0.0756
4	beta-BHC	319-85-7	BCCC6425	99%	1.0 µg/mL	+/- 0.0749
5	Endrin	72-20-8	15076100	98%	5.0 µg/mL	+/- 0.3769
6	4,4'-DDT	50-29-3	231103JLM	97%	10.0 µg/mL	+/- 0.7512
7	Methoxychlor	72-43-5	14979500	98%	24.8 µg/mL	+/- 1.8543
8	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	2.0 µg/mL	+/- 0.1513

Solvent:

Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x .25mm x .2µm
Rtx-CLP II (cat.# 111323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
150°C to 300°C
@ 4°C/min. (hold 5 min.)


Inj. Temp:
200°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
Split ratio 50:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Malina Homan

Malina Homan - Operations Technician I

Date Mixed: 17-Jul-2024

Balance Serial # 1128353505

Jennifer Pollino

Jennifer Pollino - Operations Tech III - ARPM QC

Date Passed: 31-Jul-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000

Lot No.: A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : October 31, 2030

Storage: 10°C or colder

Handling: Contains PCBs - sonicate prior to use.

Ship: Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone

CAS # 67-64-1

Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Aaron Enyart
Aaron Enyart - Operations Tech I

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0214495

Description : Pesticide Surrogate Mix

Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : October 31, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P19785
AJ
11/19/24
P19789

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.2 µg/mL	+/- 11.1087
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30679	99%	201.4 µg/mL	+/- 11.1753

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent:

Acetone

CAS # 67-64-1

Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:
30m x 25mm x 2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:
helium-constant pressure 20 psi.

Temp. Program:
200°C to 300°C
@ 25°C/min. (hold 10 min.)


Inj. Temp:
250°C

Det. Temp:
300°C

Det. Type:
ECD

Split Vent:
10 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Aaron Enyart
Aaron Enyart - Operations Tech I

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Mixed: 29-Jul-2024 **Balance Serial #** B345965662

Date Passed: 01-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32005 **Lot No.:** A0210240

Description : Toxaphene Standard
Toxaphene Standard 1000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2028 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Toxaphene	8001-35-2	1051817	----%	1,009.3 µg/mL	+/- 56.0105

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13861
P13862

12/9/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Amanda Miller - Operations Tech III - ARM QC

Date Mixed: 11-Apr-2024

Balance Serial # B442140311



Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 26-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13861 } ②
P13862 }
↑

12/9/2024



Certified Reference Material CRM



CERTIFIED WEIGHT REPORT


Part Number: 72072
Lot Number: 112018
Description: n-Tetracosane-d50
Expiration Date: 112028
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 1000
NIST Test ID#: 2684186
Weight(s) shown below were combined and diluted to (mL): 200.0

Solvent(s): Methylene chloride
Lot#: 102669
Received by
SG on 11/11/19
p9044-p9053
5E-05 Balance Uncertainty
0.058 Flask Uncertainty

Formulated By: Prashant Chauhan 112018 DATE
Reviewed By: Pedro Rentas 112018 DATE

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.)
1. n-Tetracosane-d50	2072	PR-17753/09216TC1	1000	98	0.2	0.20411	0.20415	1000.2	4.2	18416-32-3 N/A
Method GC8MSD-3.M: Column:SPB-5 (30m X 0.25mm ID X 0.25µm film thickness) Temp 1 = 50°C (1min.), Temp 2 = 300°C (9min.), Rate = 10°C/min., Injector B= 250°C, Detector B = 275°C, Split Ratio = 100:1, Scan Rate = 2. Analysis performed by: Candice Warren.										

TC 720720



• The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
• Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
• All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
• Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).




Run 40, "P72072 L112018 [1000µg/mL in MeCl2]"

Run Length: 35.00 min, 20999 points at 10 points/second.
Created: Thu, Nov 22, 2018 at 7:23:18 AM.
Sampled: Sequence "112018-GC4M1", Method "GC4-M1".
Analyzed using Method "GC4-M1".

Comments

GC4-M1 Analysis by Melissa Stonier
Column ID SPB5 L#60062-01A : 30 meter x 0.53mm x 1.5µm Film Thickness
Flow rates: Total Flow = 300 mL/min, Helium (carrier) = 6.5 mL, Hydrogen (detector) = 30 mL,
Air (detector) = 360 mL
Oven Temp 1 = 50°C (1 min), Rate = 10°C/min, Oven Temp 2 = 300°C (9 min), Total Run Time = 35 Minutes.
Injector Temp = 200°C, FID Temp = 300°C, FID Signal = eDAQ Channel 1.
Gas Chromatograph = HP 5890, Auto Sampler = HP 7673, Standard Injection = 0.5 µL, Range = 3





SHIPPING DOCUMENTS

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: TCE Service Group, Inc.
ADDRESS: 125 Bloomingdale Rd
CITY Levittown STATE: N.Y. ZIP:
ATTENTION: Dennis d. Morgan, II
PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: ICES 02
PROJECT NO.: LOCATION:
PROJECT MANAGER:
e-mail:
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#:
ADDRESS:
CITY Stun STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS*
HARDCOPY (DATA PACKAGE): 72 HR TAT DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

1. TCL 2. TCLP 3. Herb 4. DF 5. PC 6. 7. 8. 9.

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9		
1.	<u>Comp-Rolloff</u>	<u>50L</u>	<u>X</u>		<u>10-20-25</u>	<u>12:00</u>	<u>6</u>			<u>X</u>	<u>X</u>	<u>X</u>					<u>PiD = 1.3</u>	
2.		<u>I</u>		<u>X</u>	<u>I</u>	<u>12:04</u>	<u>3</u>	<u>X</u>	<u>X</u>								<u>PiD = 1.3</u>	
3.																		
4.																		
5.																		
6.																		
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>[Signature]</u>	DATE/TIME: <u>1300</u> <u>10-20-2025</u>	RECEIVED BY: 1. <u>[Signature]</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☒ COOLER TEMP <u>35</u> °C Comments: <u>*PiD Meter Calibrated 10-20-2025 *</u> <u>Equal Volume of soil used,</u>
RELINQUISHED BY SAMPLER: 2. <u>[Signature]</u>	DATE/TIME: <u></u>	RECEIVED BY: 2. <u>[Signature]</u>	
RELINQUISHED BY SAMPLER: 3. <u>[Signature]</u>	DATE/TIME: <u>1739</u> <u>10-20-2025</u>	RECEIVED BY: 3. <u>[Signature]</u>	

Page of

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3399 ICES02
Client Name : ICE Service Group, Inc.
Client Contact : Dennis D. Morgan, II
Invoice Name : ICE Service Group, Inc.
Invoice Contact : Dennis D. Morgan, II

Order Date : 10/20/2025 1:23:08 PM
Project Name : NWIRP Northrup Grumman
Receive Date Time : 10/20/2025 12:00:00 AM
Purchase Order : 5:54 PM

Project Mgr :
Report Type : NYS ASP A
EDD Type : NYSDEC EDD V-4
Hard Copy Date :
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3399-03	COMP-ROLLOFF	Solid	10/20/2025	12:04	VOC-TCLVOA-10		8260D		3 Bus. Days

Relinquished By :

Date / Time : 10/21/25 0740

Received By :

Date / Time : 10/21/25 08:00

Storage Area : VOA Refridgerator Room

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
 Data File : PD090758.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 15:40
 Operator : AR\AJ
 Sample : Q3399-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COMP-ROLLOFF

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 12:25:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.873	96546017	576.8E6	31.002	19.369 #
28)	SA Decachlor...	9.069	8.063	69161925	485.5E6	14.785	16.028
Target Compounds							
2)	A alpha-BHC	4.004	3.376	1722281	2935427	0.240	0.057 #
3)	MA gamma-BHC...	4.302f	3.726	631783	394343	0.093	0.008 #
4)	MA Heptachlor	4.916	4.079	3356691	7038248	0.601	0.158 #
5)	MB Aldrin	0.000	4.358	0	804746	N.D.	0.017 #
6)	B beta-BHC	4.492f	4.029	622256	10013947	0.238	0.505 #
7)	B delta-BHC	4.772	4.255	7229478	5195354	1.069	0.109 #
8)	B Heptachlo...	0.000	4.861	0	2237036	N.D.	0.052 #
9)	A Endosulfan I	0.000	5.262f	0	7850792	N.D.	0.199 #
10)	B gamma-Chl...	5.942	5.117	1658575	13259181	0.280	0.295
11)	B alpha-Chl...	6.021	5.182	6161942	31133470	0.997	0.722 #
12)	B 4,4'-DDE	6.191	5.366	2962743	21426427	0.556	0.489
13)	MA Dieldrin	6.342	5.502	955836	26552632	0.164	0.601 #
14)	MA Endrin	6.570	5.783	1270304	31881731	0.256	0.787 #
15)	B Endosulfa...	0.000	6.086	0	4915412	N.D.	0.131 #
16)	A 4,4'-DDD	6.683f	0.000	163225	0	0.040	N.D. #
17)	MA 4,4'-DDT	7.017	6.173	2467837	71467482	0.600	2.038 #
18)	B Endrin al...	6.937f	6.265	1114729	3737389	0.304	0.135 #
19)	B Endosulfa...	0.000	6.472	0	1539032	N.D.	0.043 #
20)	A Methoxychlor	0.000	6.729f	0	9010587	N.D.	0.517 #
21)	B Endrin ke...	7.614	6.963f	1008355	6678302	0.208	0.177
22)	Mirex	8.123f	7.168	632320	15591748	0.198	0.607 #
23)	Chlordane-1	0.000	3.893	0	4790617	N.D.	3.844 #
24)	Chlordane-2	5.214f	4.494	1396941	13448346	7.233	10.309 #
25)	Chlordane-3	5.942	5.117	1658575	13259181	2.174	3.407 #
26)	Chlordane-4	6.021	5.182	6161942	31133470	6.360	9.318 #
27)	Chlordane-5	6.838f	6.086	2530802	4915412	16.843	3.330 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090758.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 15:40
Operator : AR\AJ
Sample : Q3399-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 12:25:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 96546017
Conc: 31.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: -0.001 min
Response: 576762882
Conc: 19.37 ng/ml




#2 alpha-BHC

R.T.: 4.004 min
Delta R.T.: 0.010 min
Response: 1722281
Conc: 0.24 ng/ml



#2 alpha-BHC


R.T.: 3.376 min
Delta R.T.: -0.010 min
Response: 2935427
Conc: 0.06 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.302 min
Delta R.T.: -0.023 min
Response: 631783
Conc: 0.09 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#3 gamma-BHC (Lindane)

R.T.: 3.726 min
Delta R.T.: 0.003 min
Response: 394343
Conc: 0.01 ng/ml




#4 Heptachlor

R.T.: 4.916 min
Delta R.T.: -0.007 min
Response: 3356691
Conc: 0.60 ng/ml



#4 Heptachlor


R.T.: 4.079 min
Delta R.T.: 0.005 min
Response: 7038248
Conc: 0.16 ng/ml



#5 Aldrin


R.T.: 0.000 min
Exp R.T.: 5.264 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#5 Aldrin

R.T.: 4.358 min
Delta R.T.: -0.003 min
Response: 804746
Conc: 0.02 ng/ml




#6 beta-BHC

R.T.: 4.492 min
Delta R.T.: -0.020 min
Response: 622256
Conc: 0.24 ng/ml



#6 beta-BHC


R.T.: 4.029 min
Delta R.T.: 0.011 min
Response: 10013947
Conc: 0.51 ng/ml



#7 delta-BHC


R.T.: 4.772 min
Delta R.T.: 0.012 min
Response: 7229478
Conc: 1.07 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 5195354
Conc: 0.11 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 5.684 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 4.861 min
Delta R.T.: -0.003 min
Response: 2237036
Conc: 0.05 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 6.068 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#9 Endosulfan I

R.T.: 5.262 min
Delta R.T.: 0.024 min
Response: 7850792
Conc: 0.20 ng/ml




#10 gamma-Chlordane

R.T.: 5.942 min
Delta R.T.: 0.002 min
Response: 1658575
Conc: 0.28 ng/ml



#10 gamma-Chlordane


R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 13259181
Conc: 0.29 ng/ml



#11 alpha-Chlordane


R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 6161942
Conc: 1.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#11 alpha-Chlordane

R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 31133470
Conc: 0.72 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.001 min
Response: 2962743
Conc: 0.56 ng/ml



#12 4,4'-DDE


R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 21426427
Conc: 0.49 ng/ml



#13 Dieldrin


R.T.: 6.342 min
Delta R.T.: 0.002 min
Response: 955836
Conc: 0.16 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#13 Dieldrin

R.T.: 5.502 min
Delta R.T.: -0.002 min
Response: 26552632
Conc: 0.60 ng/ml




#14 Endrin

R.T.: 6.570 min
Delta R.T.: 0.002 min
Response: 1270304
Conc: 0.26 ng/ml



#14 Endrin


R.T.: 5.783 min
Delta R.T.: 0.003 min
Response: 31881731
Conc: 0.79 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T.: 6.781 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#15 Endosulfan II

R.T.: 6.086 min
Delta R.T.: 0.014 min
Response: 4915412
Conc: 0.13 ng/ml




#16 4,4'-DDD

R.T.: 6.683 min
Delta R.T.: -0.018 min
Response: 163225
Conc: 0.04 ng/ml



#16 4,4'-DDD


R.T.: 0.000 min
Exp R.T.: 5.921 min
Response: 0
Conc: N.D.



#17 4,4'-DDT


R.T.: 7.017 min
Delta R.T.: 0.002 min
Response: 2467837
Conc: 0.60 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#17 4,4'-DDT

R.T.: 6.173 min
Delta R.T.: 0.000 min
Response: 71467482
Conc: 2.04 ng/ml




#18 Endrin aldehyde

R.T.: 6.937 min
Delta R.T.: 0.028 min
Response: 1114729
Conc: 0.30 ng/ml



#18 Endrin aldehyde


R.T.: 6.265 min
Delta R.T.: 0.015 min
Response: 3737389
Conc: 0.13 ng/ml



#19 Endosulfan Sulfate


R.T.: 0.000 min
Exp R.T.: 7.145 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#19 Endosulfan Sulfate

R.T.: 6.472 min
Delta R.T.: -0.002 min
Response: 1539032
Conc: 0.04 ng/ml




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T.: 7.488 min
Response: 0
Conc: N.D.



#20 Methoxychlor


R.T.: 6.729 min
Delta R.T.: -0.016 min
Response: 9010587
Conc: 0.52 ng/ml



#21 Endrin ketone


R.T.: 7.614 min
Delta R.T.: -0.011 min
Response: 1008355
Conc: 0.21 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#21 Endrin ketone

R.T.: 6.963 min
Delta R.T.: -0.020 min
Response: 6678302
Conc: 0.18 ng/ml




#22 Mirex

R.T.: 8.123 min
Delta R.T.: 0.016 min
Response: 632320
Conc: 0.20 ng/ml



#22 Mirex


R.T.: 7.168 min
Delta R.T.: -0.008 min
Response: 15591748
Conc: 0.61 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T.: 4.709 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#23 Chlordane-1

R.T.: 3.893 min
Delta R.T.: -0.005 min
Response: 4790617
Conc: 3.84 ng/ml




#24 Chlordane-2

R.T.: 5.214 min
Delta R.T.: -0.021 min
Response: 1396941
Conc: 7.23 ng/ml



#24 Chlordane-2


R.T.: 4.494 min
Delta R.T.: 0.015 min
Response: 13448346
Conc: 10.31 ng/ml



#25 Chlordane-3


R.T.: 5.942 min
Delta R.T.: 0.000 min
Response: 1658575
Conc: 2.17 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#25 Chlordane-3

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 13259181
Conc: 3.41 ng/ml




#26 Chlordane-4

R.T.: 6.021 min
Delta R.T.: -0.004 min
Response: 6161942
Conc: 6.36 ng/ml



#26 Chlordane-4


R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 31133470
Conc: 9.32 ng/ml



#27 Chlordane-5


R.T.: 6.838 min
Delta R.T.: -0.028 min
Response: 2530802
Conc: 16.84 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFF




#27 Chlordane-5

R.T.: 6.086 min
Delta R.T.: 0.005 min
Response: 4915412
Conc: 3.33 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.069 min
Delta R.T.: 0.005 min
Response: 69161925
Conc: 14.78 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.063 min
Delta R.T.: 0.000 min
Response: 485518279
Conc: 16.03 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
 Data File : PD090768.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 18:11
 Operator : AR\AJ
 Sample : Q3399-01MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COMP-ROLLOFFMS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 14:23:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.874	87949601	570.1E6	28.242	19.146 #
28)	SA Decachlor...	9.067	8.062	65859422	464.0E6	14.079	15.317
Target Compounds							
2)	A alpha-BHC	3.994	3.386	353.9E6	2433.0E6	49.315	47.616
3)	MA gamma-BHC...	4.325	3.722	325.3E6	2250.0E6	47.843	47.620
4)	MA Heptachlor	4.923	4.075	315.9E6	2185.3E6	56.513	48.994
5)	MB Aldrin	5.264	4.360	314.9E6	2225.7E6	48.102	48.052
6)	B beta-BHC	4.513	4.018	123.8E6	938.0E6	47.251	47.302
7)	B delta-BHC	4.760	4.255	331.8E6	2286.7E6	49.066	47.923
8)	B Heptachlo...	5.685	4.864	283.6E6	2032.2E6	48.993	47.661
9)	A Endosulfan I	6.068	5.238	265.3E6	1883.5E6	47.918	47.817
10)	B gamma-Chl...	5.940	5.116	283.0E6	2180.3E6	47.715	48.506
11)	B alpha-Chl...	6.021	5.181	296.7E6	2094.3E6	48.014	48.550
12)	B 4,4'-DDE	6.191	5.366	254.1E6	2138.5E6	47.709	48.782
13)	MA Dieldrin	6.341	5.504	283.6E6	2141.6E6	48.629	48.469
14)	MA Endrin	6.568	5.780	239.7E6	1957.8E6	48.393	48.349
15)	B Endosulfa...	6.782	6.072	226.8E6	1847.3E6	45.363	49.188
16)	A 4,4'-DDD	6.701	5.921	201.3E6	1717.5E6	49.653	47.134
17)	MA 4,4'-DDT	7.016	6.174	206.0E6	1822.3E6	50.113	51.957
18)	B Endrin al...	6.910	6.250	174.4E6	1373.5E6	47.625	49.597
19)	B Endosulfa...	7.145	6.474	218.3E6	1795.4E6	47.056	50.089
20)	A Methoxychlor	7.490	6.745	103.3E6	889.0E6	48.607	51.010
21)	B Endrin ke...	7.627	6.983	236.4E6	1933.5E6	48.654	51.263
22)	Mirex	8.109	7.175	150.2E6	1272.8E6	46.913	49.543
23)	Chlordane-1	0.000	3.894	0	5287673	N.D.	4.243 #
24)	Chlordane-2	5.214f	4.497	1572436	28151014	8.142	21.580 #
25)	Chlordane-3	5.940	5.116	283.0E6	2180.3E6	371.032	560.255 #
26)	Chlordane-4	6.021	5.181	296.7E6	2094.3E6	306.269	626.811 #
27)	Chlordane-5	0.000	6.072	0	1847.3E6	N.D.	1251.517 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090768.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 18:11
Operator : AR\AJ
Sample : Q3399-01MS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:23:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.545 min
Delta R.T.: 0.000 min
Response: 87949601
Conc: 28.24 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 570118937
Conc: 19.15 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 353943515
Conc: 49.32 ng/ml



#2 alpha-BHC


R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2432979872
Conc: 47.62 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 325274369
Conc: 47.84 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 2249990116
Conc: 47.62 ng/ml




#4 Heptachlor

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 315881480
Conc: 56.51 ng/ml



#4 Heptachlor


R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 2185284603
Conc: 48.99 ng/ml



#5 Aldrin


R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 314935367
Conc: 48.10 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 2225696544
Conc: 48.05 ng/ml




#6 beta-BHC

R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 123781257
Conc: 47.25 ng/ml



#6 beta-BHC


R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 937968947
Conc: 47.30 ng/ml



#7 delta-BHC


R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 331768828
Conc: 49.07 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 2286739044
Conc: 47.92 ng/ml




#8 Heptachlor epoxide

R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 283621153
Conc: 48.99 ng/ml



#8 Heptachlor epoxide


R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 2032230337
Conc: 47.66 ng/ml



#9 Endosulfan I


R.T.: 6.068 min
Delta R.T.: 0.000 min
Response: 265306892
Conc: 47.92 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#9 Endosulfan I

R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 1883530425
Conc: 47.82 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 283008513
Conc: 47.72 ng/ml



#10 gamma-Chlordane


R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 2180269957
Conc: 48.51 ng/ml



#11 alpha-Chlordane


R.T.: 6.021 min
Delta R.T.: 0.000 min
Response: 296739854
Conc: 48.01 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2094323635
Conc: 48.55 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 254128833
Conc: 47.71 ng/ml



#12 4,4'-DDE


R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 2138528914
Conc: 48.78 ng/ml



#13 Dieldrin


R.T.: 6.341 min
Delta R.T.: 0.000 min
Response: 283574965
Conc: 48.63 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 2141551063
Conc: 48.47 ng/ml




#14 Endrin

R.T.: 6.568 min
Delta R.T.: 0.000 min
Response: 239724746
Conc: 48.39 ng/ml



#14 Endrin


R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 1957839756
Conc: 48.35 ng/ml



#15 Endosulfan II


R.T.: 6.782 min
Delta R.T.: 0.000 min
Response: 226813846
Conc: 45.36 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1847281980
Conc: 49.19 ng/ml




#16 4,4'-DDD

R.T.: 6.701 min
Delta R.T.: 0.000 min
Response: 201302989
Conc: 49.65 ng/ml



#16 4,4'-DDD


R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 1717457499
Conc: 47.13 ng/ml



#17 4,4'-DDT


R.T.: 7.016 min
Delta R.T.: 0.001 min
Response: 205987070
Conc: 50.11 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1822278697
Conc: 51.96 ng/ml




#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 174437307
Conc: 47.63 ng/ml



#18 Endrin aldehyde


R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 1373492836
Conc: 49.60 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.145 min
Delta R.T.: 0.000 min
Response: 218256060
Conc: 47.06 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1795448252
Conc: 50.09 ng/ml




#20 Methoxychlor

R.T.: 7.490 min
Delta R.T.: 0.001 min
Response: 103317287
Conc: 48.61 ng/ml



#20 Methoxychlor


R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 888989148
Conc: 51.01 ng/ml



#21 Endrin ketone


R.T.: 7.627 min
Delta R.T.: 0.002 min
Response: 236402703
Conc: 48.65 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1933535947
Conc: 51.26 ng/ml




#22 Mirex

R.T.: 8.109 min
Delta R.T.: 0.001 min
Response: 150162319
Conc: 46.91 ng/ml



#22 Mirex


R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1272797690
Conc: 49.54 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T.: 4.709 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#23 Chlordane-1

R.T.: 3.894 min
Delta R.T.: -0.003 min
Response: 5287673
Conc: 4.24 ng/ml




#24 Chlordane-2

R.T.: 5.214 min
Delta R.T.: -0.022 min
Response: 1572436
Conc: 8.14 ng/ml



#24 Chlordane-2


R.T.: 4.497 min
Delta R.T.: 0.018 min
Response: 28151014
Conc: 21.58 ng/ml



#25 Chlordane-3


R.T.: 5.940 min
Delta R.T.: -0.001 min
Response: 283008513
Conc: 371.03 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#25 Chlordane-3

R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 2180269957
Conc: 560.25 ng/ml




#26 Chlordane-4

R.T.: 6.021 min
Delta R.T.: -0.004 min
Response: 296739854
Conc: 306.27 ng/ml



#26 Chlordane-4


R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2094323635
Conc: 626.81 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T.: 6.866 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMS




#27 Chlordane-5

R.T.: 6.072 min
Delta R.T.: -0.009 min
Response: 1847281980
Conc: 1251.52 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.067 min
Delta R.T.: 0.003 min
Response: 65859422
Conc: 14.08 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 463986691
Conc: 15.32 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
 Data File : PD090769.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 18:25
 Operator : AR\AJ
 Sample : Q3399-01MSD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COMP-ROLLOFFMSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 14:23:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	89753448	578.4E6	28.821	19.425 #
28)	SA Decachlor...	9.068	8.062	67109061	476.8E6	14.346	15.741
Target Compounds							
2)	A alpha-BHC	3.994	3.386	358.4E6	2475.1E6	49.931	48.440
3)	MA gamma-BHC...	4.325	3.722	331.9E6	2281.5E6	48.810	48.287
4)	MA Heptachlor	4.923	4.075	322.2E6	2211.8E6	57.644	49.589
5)	MB Aldrin	5.264	4.360	320.3E6	2251.0E6	48.919	48.597
6)	B beta-BHC	4.512	4.019	125.7E6	951.4E6	47.986	47.977
7)	B delta-BHC	4.760	4.255	337.7E6	2326.8E6	49.941	48.763
8)	B Heptachlo...	5.685	4.864	286.7E6	2057.0E6	49.525	48.243
9)	A Endosulfan I	6.069	5.239	270.0E6	1908.1E6	48.767	48.440
10)	B gamma-Chl...	5.940	5.117	287.5E6	2211.2E6	48.476	49.194
11)	B alpha-Chl...	6.022	5.181	296.3E6	2125.8E6	47.945	49.279
12)	B 4,4'-DDE	6.191	5.365	260.3E6	2166.5E6	48.861	49.420
13)	MA Dieldrin	6.342	5.504	289.1E6	2168.8E6	49.583	49.086
14)	MA Endrin	6.569	5.781	244.9E6	1970.4E6	49.430	48.660
15)	B Endosulfa...	6.782	6.072	231.9E6	1885.7E6	46.374	50.211
16)	A 4,4'-DDD	6.701	5.921	204.9E6	1742.0E6	50.544	47.806
17)	MA 4,4'-DDT	7.017	6.174	211.0E6	1863.8E6	51.335	53.142
18)	B Endrin al...	6.910	6.251	178.4E6	1394.3E6	48.714	50.347
19)	B Endosulfa...	7.146	6.474	222.2E6	1818.0E6	47.910	50.718
20)	A Methoxychlor	7.490	6.745	105.1E6	908.4E6	49.463	52.124
21)	B Endrin ke...	7.626	6.983	241.7E6	1962.1E6	49.740	52.020
22)	Mirex	8.109	7.175	151.1E6	1297.4E6	47.218	50.499
23)	Chlordane-1	0.000	3.893	0	4684974	N.D.	3.759 #
24)	Chlordane-2	5.213f	4.496	1562593	28726309	8.091	22.021 #
25)	Chlordane-3	5.940	5.117	287.5E6	2211.2E6	376.947	568.211 #
26)	Chlordane-4	6.022	5.181	296.3E6	2125.8E6	305.826	636.224 #
27)	Chlordane-5	0.000	6.072	0	1885.7E6	N.D.	1277.550 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090769.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 18:25
Operator : AR\AJ
Sample : Q3399-01MSD
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:23:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.544 min
Delta R.T.: 0.000 min
Response: 89753448
Conc: 28.82 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#1 Tetrachloro-m-xylene

R.T.: 2.874 min
Delta R.T.: 0.000 min
Response: 578415516
Conc: 19.42 ng/ml




#2 alpha-BHC

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 358360933
Conc: 49.93 ng/ml



#2 alpha-BHC


R.T.: 3.386 min
Delta R.T.: 0.000 min
Response: 2475094258
Conc: 48.44 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.325 min
Delta R.T.: 0.000 min
Response: 331851325
Conc: 48.81 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#3 gamma-BHC (Lindane)

R.T.: 3.722 min
Delta R.T.: 0.000 min
Response: 2281514650
Conc: 48.29 ng/ml




#4 Heptachlor

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 322201675
Conc: 57.64 ng/ml



#4 Heptachlor


R.T.: 4.075 min
Delta R.T.: 0.000 min
Response: 2211816269
Conc: 49.59 ng/ml



#5 Aldrin


R.T.: 5.264 min
Delta R.T.: 0.000 min
Response: 320284321
Conc: 48.92 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#5 Aldrin

R.T.: 4.360 min
Delta R.T.: 0.000 min
Response: 2250954350
Conc: 48.60 ng/ml




#6 beta-BHC

R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 125705624
Conc: 47.99 ng/ml



#6 beta-BHC


R.T.: 4.019 min
Delta R.T.: 0.000 min
Response: 951364052
Conc: 47.98 ng/ml



#7 delta-BHC


R.T.: 4.760 min
Delta R.T.: 0.000 min
Response: 337690238
Conc: 49.94 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#7 delta-BHC

R.T.: 4.255 min
Delta R.T.: 0.000 min
Response: 2326791858
Conc: 48.76 ng/ml




#8 Heptachlor epoxide

R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 286703039
Conc: 49.52 ng/ml



#8 Heptachlor epoxide


R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 2057020731
Conc: 48.24 ng/ml



#9 Endosulfan I


R.T.: 6.069 min
Delta R.T.: 0.000 min
Response: 270006575
Conc: 48.77 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#9 Endosulfan I

R.T.: 5.239 min
Delta R.T.: 0.000 min
Response: 1908057666
Conc: 48.44 ng/ml




#10 gamma-Chlordane

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 287520328
Conc: 48.48 ng/ml



#10 gamma-Chlordane


R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 2211234747
Conc: 49.19 ng/ml



#11 alpha-Chlordane


R.T.: 6.022 min
Delta R.T.: 0.001 min
Response: 296310307
Conc: 47.94 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#11 alpha-Chlordane

R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2125775960
Conc: 49.28 ng/ml




#12 4,4'-DDE

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 260264333
Conc: 48.86 ng/ml



#12 4,4'-DDE


R.T.: 5.365 min
Delta R.T.: 0.000 min
Response: 2166466646
Conc: 49.42 ng/ml



#13 Dieldrin


R.T.: 6.342 min
Delta R.T.: 0.001 min
Response: 289136583
Conc: 49.58 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#13 Dieldrin

R.T.: 5.504 min
Delta R.T.: 0.000 min
Response: 2168799726
Conc: 49.09 ng/ml




#14 Endrin

R.T.: 6.569 min
Delta R.T.: 0.000 min
Response: 244864046
Conc: 49.43 ng/ml



#14 Endrin


R.T.: 5.781 min
Delta R.T.: 0.000 min
Response: 1970443496
Conc: 48.66 ng/ml



#15 Endosulfan II


R.T.: 6.782 min
Delta R.T.: 0.001 min
Response: 231867559
Conc: 46.37 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#15 Endosulfan II

R.T.: 6.072 min
Delta R.T.: 0.000 min
Response: 1885706637
Conc: 50.21 ng/ml




#16 4,4'-DDD

R.T.: 6.701 min
Delta R.T.: 0.001 min
Response: 204914718
Conc: 50.54 ng/ml



#16 4,4'-DDD


R.T.: 5.921 min
Delta R.T.: 0.000 min
Response: 1741968091
Conc: 47.81 ng/ml



#17 4,4'-DDT


R.T.: 7.017 min
Delta R.T.: 0.002 min
Response: 211011087
Conc: 51.34 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#17 4,4'-DDT

R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 1863845638
Conc: 53.14 ng/ml




#18 Endrin aldehyde

R.T.: 6.910 min
Delta R.T.: 0.000 min
Response: 178426338
Conc: 48.71 ng/ml



#18 Endrin aldehyde


R.T.: 6.251 min
Delta R.T.: 0.000 min
Response: 1394257530
Conc: 50.35 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.146 min
Delta R.T.: 0.001 min
Response: 222214770
Conc: 47.91 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 1817999761
Conc: 50.72 ng/ml




#20 Methoxychlor

R.T.: 7.490 min
Delta R.T.: 0.001 min
Response: 105138524
Conc: 49.46 ng/ml



#20 Methoxychlor


R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 908402672
Conc: 52.12 ng/ml



#21 Endrin ketone


R.T.: 7.626 min
Delta R.T.: 0.001 min
Response: 241679099
Conc: 49.74 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#21 Endrin ketone

R.T.: 6.983 min
Delta R.T.: 0.000 min
Response: 1962091852
Conc: 52.02 ng/ml




#22 Mirex

R.T.: 8.109 min
Delta R.T.: 0.002 min
Response: 151136710
Conc: 47.22 ng/ml



#22 Mirex


R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1297356869
Conc: 50.50 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T.: 4.709 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#23 Chlordane-1

R.T.: 3.893 min
Delta R.T.: -0.005 min
Response: 4684974
Conc: 3.76 ng/ml




#24 Chlordane-2

R.T.: 5.213 min
Delta R.T.: -0.022 min
Response: 1562593
Conc: 8.09 ng/ml



#24 Chlordane-2


R.T.: 4.496 min
Delta R.T.: 0.018 min
Response: 28726309
Conc: 22.02 ng/ml



#25 Chlordane-3


R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 287520328
Conc: 376.95 ng/ml

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#25 Chlordane-3

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 2211234747
Conc: 568.21 ng/ml




#26 Chlordane-4

R.T.: 6.022 min
Delta R.T.: -0.004 min
Response: 296310307
Conc: 305.83 ng/ml



#26 Chlordane-4


R.T.: 5.181 min
Delta R.T.: 0.000 min
Response: 2125775960
Conc: 636.22 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T.: 6.866 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
COMP-ROLLOFFMSD




#27 Chlordane-5

R.T.: 6.072 min
Delta R.T.: -0.009 min
Response: 1885706637
Conc: 1277.55 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.068 min
Delta R.T.: 0.003 min
Response: 67109061
Conc: 14.35 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.062 min
Delta R.T.: 0.000 min
Response: 476816665
Conc: 15.74 ng/ml