

Surrogate Summary

SDG No.: Q3399

Client: ICE Service Group, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		30	135
		Tetrachloro-m-xyl	1	20	19.4	97		44	124
		Decachlorobiphen	2	20	20.3	102		30	135
		Tetrachloro-m-xyl	2	20	20.0	100		44	124
I.BLK-PD090753.D	PIBLK-PD090753.D	Decachlorobiphen	1	20	22.1	111		30	135
		Tetrachloro-m-xyl	1	20	19.9	99		44	124
		Decachlorobiphen	2	20	25.1	126		30	135
		Tetrachloro-m-xyl	2	20	20.7	104		44	124
PB170205BL	PB170205BL	Decachlorobiphen	1	20	21.5	107		55	130
		Tetrachloro-m-xyl	1	20	19.9	99		42	129
		Decachlorobiphen	2	20	24.3	121		55	130
		Tetrachloro-m-xyl	2	20	20.4	102		42	129
PB170205BS	PB170205BS	Decachlorobiphen	1	20	21.6	108		55	130
		Tetrachloro-m-xyl	1	20	16.9	84		42	129
		Decachlorobiphen	2	20	24.5	123		55	130
		Tetrachloro-m-xyl	2	20	17.3	86		42	129
Q3399-01	COMP-ROLLOFF	Decachlorobiphen	1	20	14.8	74		55	130
		Tetrachloro-m-xyl	1	20	18.3	91		42	129
		Decachlorobiphen	2	20	16.0	80		55	130
		Tetrachloro-m-xyl	2	20	19.4	97		42	129
I.BLK-PD090763.D	PIBLK-PD090763.D	Decachlorobiphen	1	20	21.0	105		30	135
		Tetrachloro-m-xyl	1	20	19.9	100		44	124
		Decachlorobiphen	2	20	21.0	105		30	135
		Tetrachloro-m-xyl	2	20	20.6	103		44	124
Q3399-01MS	COMP-ROLLOFFMS	Decachlorobiphen	1	20	14.1	70		55	130
		Tetrachloro-m-xyl	1	20	17.7	89		42	129
		Decachlorobiphen	2	20	15.3	77		55	130
		Tetrachloro-m-xyl	2	20	19.1	96		42	129
Q3399-01MSD	COMP-ROLLOFFMSD	Decachlorobiphen	1	20	14.3	72		55	130
		Tetrachloro-m-xyl	1	20	17.5	88		42	129
		Decachlorobiphen	2	20	15.7	79		55	130
		Tetrachloro-m-xyl	2	20	19.4	97		42	129
I.BLK-PD090772.D	PIBLK-PD090772.D	Decachlorobiphen	1	20	21.4	107		30	135
		Tetrachloro-m-xyl	1	20	19.9	100		44	124
		Decachlorobiphen	2	20	24.0	120		30	135
		Tetrachloro-m-xyl	2	20	20.5	103		44	124