

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ICE Service Group, Inc.</u>	DataFile :	<u>PD090768.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
			Result								High	RPD
Lab Sample ID:	Q3399-01MS (Column 1)	Client Sample ID:		COMP-ROLLOFFMS								
	alpha-BHC	18.48	0	18.2	ug/kg	98				45	137	
	beta-BHC	18.48	0	17.5	ug/kg	95				50	136	
	delta-BHC	18.48	0	18.1	ug/kg	98				47	139	
	gamma-BHC (Lindane)	18.48	0	17.7	ug/kg	96				49	135	
	Heptachlor	18.48	0	20.9	ug/kg	113				47	136	
	Aldrin	18.48	0	17.8	ug/kg	96				45	136	
	Heptachlor epoxide	18.48	0	18.1	ug/kg	98				52	136	
	Endosulfan I	18.48	0	17.7	ug/kg	96				53	132	
	Dieldrin	18.48	0	18.0	ug/kg	97				56	136	
	4,4'-DDE	18.48	0.21	17.6	ug/kg	94				56	134	
	Endrin	18.48	0.19	17.9	ug/kg	96				57	140	
	Endosulfan II	18.48	0	16.8	ug/kg	91				53	134	
	4,4'-DDD	18.48	0	18.4	ug/kg	100				56	139	
	Endosulfan sulfate	18.48	0	17.4	ug/kg	94				55	136	
	4,4'-DDT	18.48	0.22	18.5	ug/kg	99				50	141	
	Methoxychlor	18.48	0	18.0	ug/kg	97				52	143	
	Endrin ketone	18.48	0	18.0	ug/kg	97				55	136	
	Endrin aldehyde	18.48	0	17.6	ug/kg	95				35	137	
	alpha-Chlordane	18.48	0.37	17.7	ug/kg	94				54	133	
	gamma-Chlordane	18.48	0	17.6	ug/kg	95				53	135	
Lab Sample ID:	Q3399-01MS (Column 2)	Client Sample ID:		COMP-ROLLOFFMS								
	alpha-BHC	18.48	0	17.6	ug/kg	95				45	137	
	beta-BHC	18.48	0	17.5	ug/kg	95				50	136	
	delta-BHC	18.48	0	17.7	ug/kg	96				47	139	
	gamma-BHC (Lindane)	18.48	0	17.6	ug/kg	95				49	135	
	Heptachlor	18.48	0	18.1	ug/kg	98				47	136	
	Aldrin	18.48	0	17.8	ug/kg	96				45	136	
	Heptachlor epoxide	18.48	0	17.6	ug/kg	95				52	136	
	Endosulfan I	18.48	0	17.7	ug/kg	96				53	132	
	Dieldrin	18.48	0.25	17.9	ug/kg	96				56	136	
	4,4'-DDE	18.48	0.18	18.0	ug/kg	96				56	134	
	Endrin	18.48	0.38	17.9	ug/kg	95				57	140	
	Endosulfan II	18.48	0	18.2	ug/kg	98				53	134	
	4,4'-DDD	18.48	0	17.4	ug/kg	94				56	139	
	Endosulfan sulfate	18.48	0	18.5	ug/kg	100				55	136	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ICE Service Group, Inc.</u>	DataFile :	<u>PD090768.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
		Result	Result							High	RPD
4,4'-DDT	18.48	0.75	19.2	ug/kg	100				50	141	
Methoxychlor	18.48	0	18.9	ug/kg	102				52	143	
Endrin ketone	18.48	0	18.9	ug/kg	102				55	136	
Endrin aldehyde	18.48	0	18.3	ug/kg	99				35	137	
alpha-Chlordane	18.48	0.27	17.9	ug/kg	95				54	133	
gamma-Chlordane	18.48	0	17.9	ug/kg	97				53	135	