

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ICE Service Group, Inc.</u>	DataFile :	<u>PD090769.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD		Low	Limits	
			Result					RPD	Qual		High	RPD
Lab Sample ID:	Q3399-01MSD (Column 1)	Client Sample ID:		COMP-ROLLOFFMSD								
	alpha-BHC	18.49	0	18.5	ug/kg	100		2		45	137	20
	beta-BHC	18.49	0	17.7	ug/kg	96		1		50	136	20
	delta-BHC	18.49	0	18.5	ug/kg	100		2		47	139	20
	gamma-BHC (Lindane)	18.49	0	18.0	ug/kg	97		1		49	135	20
	Heptachlor	18.49	0	21.3	ug/kg	115		2		47	136	20
	Aldrin	18.49	0	18.1	ug/kg	98		2		45	136	20
	Heptachlor epoxide	18.49	0	18.3	ug/kg	99		1		52	136	20
	Endosulfan I	18.49	0	18.0	ug/kg	97		1		53	132	20
	Dieldrin	18.49	0	18.3	ug/kg	99		2		56	136	20
	4,4'-DDE	18.49	0.21	18.1	ug/kg	97		3		56	134	20
	Endrin	18.49	0.19	18.3	ug/kg	98		2		57	140	20
	Endosulfan II	18.49	0	17.1	ug/kg	92		1		53	134	20
	4,4'-DDD	18.49	0	18.7	ug/kg	101		1		56	139	20
	Endosulfan sulfate	18.49	0	17.7	ug/kg	96		2		55	136	20
	4,4'-DDT	18.49	0.22	19.0	ug/kg	102		3		50	141	20
	Methoxychlor	18.49	0	18.3	ug/kg	99		2		52	143	20
	Endrin ketone	18.49	0	18.4	ug/kg	100		3		55	136	20
	Endrin aldehyde	18.49	0	18.0	ug/kg	97		2		35	137	20
	alpha-Chlordane	18.49	0.37	17.7	ug/kg	94		0		54	133	20
	gamma-Chlordane	18.49	0	17.9	ug/kg	97		2		53	135	20
Lab Sample ID:	Q3399-01MSD (Column 2)	Client Sample ID:		COMP-ROLLOFFMSD								
	alpha-BHC	18.49	0	17.9	ug/kg	97		2		45	137	20
	beta-BHC	18.49	0	17.7	ug/kg	96		1		50	136	20
	delta-BHC	18.49	0	18.0	ug/kg	97		1		47	139	20
	gamma-BHC (Lindane)	18.49	0	17.9	ug/kg	97		2		49	135	20
	Heptachlor	18.49	0	18.3	ug/kg	99		1		47	136	20
	Aldrin	18.49	0	18.0	ug/kg	97		1		45	136	20
	Heptachlor epoxide	18.49	0	17.8	ug/kg	96		1		52	136	20
	Endosulfan I	18.49	0	17.9	ug/kg	97		1		53	132	20
	Dieldrin	18.49	0.25	18.2	ug/kg	97		1		56	136	20
	4,4'-DDE	18.49	0.18	18.3	ug/kg	98		2		56	134	20
	Endrin	18.49	0.38	18.0	ug/kg	95		0		57	140	20
	Endosulfan II	18.49	0	18.6	ug/kg	101		3		53	134	20
	4,4'-DDD	18.49	0	17.7	ug/kg	96		2		56	139	20
	Endosulfan sulfate	18.49	0	18.8	ug/kg	102		2		55	136	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ICE Service Group, Inc.</u>	DataFile :	<u>PD090769.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD		Limits High	RPD
		Result	Result				RPD	Qual		
4,4'-DDT	18.49	0.75	19.7	ug/kg	102		2		50	20
Methoxychlor	18.49	0	19.3	ug/kg	104		2		52	20
Endrin ketone	18.49	0	19.2	ug/kg	104		2		55	20
Endrin aldehyde	18.49	0	18.6	ug/kg	101		2		35	20
alpha-Chlordane	18.49	0.27	18.2	ug/kg	97		2		54	20
gamma-Chlordane	18.49	0	18.2	ug/kg	98		1		53	20