



Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>	
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3399</u>	
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):	<u>10/13/2025</u>	<u>10/13/2025</u>
		Calibration Times:	<u>10:53</u>	<u>12:29</u>

LAB FILE ID:	RT 200 =	<u>PS031983.D</u>	RT 500 =	<u>PS031984.D</u>
RT 750 =	<u>PS031985.D</u>	RT 1000 =	<u>PS031986.D</u>	RT 1500 = <u>PS031987.D</u>

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3399</u>
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):	<u>10/13/2025</u> <u>10/13/2025</u>
		Calibration Times:	<u>10:53</u> <u>12:29</u>

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS031983.D</u>	RT 500 = <u>PS031984.D</u>
RT 750 = <u>PS031985.D</u>	RT 1000 = <u>PS031986.D</u>	RT 1500 = <u>PS031987.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW FROM TO	
2,4,5-TP(Silvex)	9.90	9.89	9.89	9.89	9.89	9.89	9.79	9.99
2,4-D	9.02	9.01	9.00	9.00	9.00	9.01	8.91	9.11
2,4-DCAA	7.76	7.75	7.74	7.74	7.74	7.75	7.65	7.85